

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032512.D **Time Analyzed:** 14:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.569	9.478	9.678	701.760	712.500	-1.5
2,4,5-TP(Silvex)	9.269	9.180	9.380	719.350	712.500	1.0
2,4-D	8.405	8.314	8.514	677.370	705.000	-3.9
2,4-DB	10.146	10.055	10.255	683.670	712.500	-4.0
2,4-DCAA	7.277	7.183	7.383	736.690	750.000	-1.8
DICAMBA	7.458	7.366	7.566	715.660	705.000	1.5
DICHLORPROP	8.165	8.074	8.274	689.620	705.000	-2.2
Dinoseb	11.333	11.245	11.445	704.240	705.000	-0.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032512.D **Time Analyzed:** 14:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.169	10.072	10.272	708.010	712.500	-0.6
2,4,5-TP(Silvex)	9.741	9.644	9.844	712.800	712.500	0.0
2,4-D	8.862	8.764	8.964	692.110	705.000	-1.8
2,4-DB	10.735	10.637	10.837	671.710	712.500	-5.7
2,4-DCAA	7.615	7.517	7.717	746.580	750.000	-0.5
DICAMBA	7.805	7.708	7.908	717.400	705.000	1.8
DICHLORPROP	8.519	8.422	8.622	702.650	705.000	-0.3
Dinoseb	11.102	11.005	11.205	691.480	705.000	-1.9