

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032524.D **Time Analyzed:** 20:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.567	9.478	9.678	700.150	712.500	-1.7
2,4,5-TP(Silvex)	9.268	9.180	9.380	714.900	712.500	0.3
2,4-D	8.403	8.314	8.514	692.990	705.000	-1.7
2,4-DB	10.143	10.055	10.255	707.280	712.500	-0.7
2,4-DCAA	7.276	7.183	7.383	733.280	750.000	-2.2
DICAMBA	7.457	7.366	7.566	706.980	705.000	0.3
DICHLORPROP	8.164	8.074	8.274	685.060	705.000	-2.8
Dinoseb	11.332	11.245	11.445	694.260	705.000	-1.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032524.D **Time Analyzed:** 20:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.168	10.072	10.272	721.370	712.500	1.2
2,4,5-TP(Silvex)	9.740	9.644	9.844	706.900	712.500	-0.8
2,4-D	8.861	8.764	8.964	706.870	705.000	0.3
2,4-DB	10.733	10.637	10.837	699.420	712.500	-1.8
2,4-DCAA	7.614	7.517	7.717	756.340	750.000	0.8
DICAMBA	7.805	7.708	7.908	715.060	705.000	1.4
DICHLORPROP	8.518	8.422	8.622	706.190	705.000	0.2
Dinoseb	11.101	11.005	11.205	689.070	705.000	-2.3