

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111325\
 Data File : PS032399.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 15:31
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/14/2025
 Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 03:37:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 04 04:59:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.282	7.615	3711.9E6	3016.5E6	819.177	827.473m
Target Compounds							
1) T	Dalapon	2.658	2.602	4609.1E6	4537.6E6	642.041	652.512
2) T	3,5-DICHL...	6.448	6.584	4970.8E6	4688.9E6	768.733	780.314
3) T	4-Nitroph...	7.106	7.186	995.3E6	1789.8E6	741.378	788.488
5) T	DICAMBA	7.463	7.806	14751.7E6	13599.3E6	781.913	787.338
6) T	MCPD	7.640	7.904	837.8E6	1100.3E6	76.931	82.723
7) T	MCPA	7.789	8.148	1073.9E6	1329.5E6	79.985	81.540m
8) T	DICHLORPROP	8.171	8.521	3342.1E6	3702.5E6	751.233	776.345
9) T	2,4-D	8.411	8.864	3459.1E6	4079.8E6	766.668	777.753
10) T	Pentachlo...	8.704	9.356	49371.3E6	46223.0E6	786.930	787.409
11) T	2,4,5-TP ...	9.275	9.743	20310.6E6	21047.1E6	780.829	783.765
12) T	2,4,5-T	9.575	10.171	18982.8E6	18647.8E6	796.320	786.230
13) T	2,4-DB	10.153	10.737	2038.5E6	2931.6E6	773.602	777.339
14) T	DINOSEB	11.339	11.104	15679.6E6	13379.4E6	753.550	756.436
15) T	Picloram	11.185	12.229	14021.6E6	22342.0E6	660.599m	600.880
16) T	DCPA	11.640	12.138	26471.6E6	24694.3E6	788.972	775.074m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111325\
Data File : PS032399.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 15:31
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 03:37:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
Quant Title : 8080.M
QLast Update : Tue Nov 04 04:59:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

