

Surrogate Summary

SDG No.: Q3618

Client: ENTACT

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PS032225.D	PIBLK-PS032225.D	2,4-DCAA	1	500	359	72		61	136
		2,4-DCAA	2	500	508	102		61	136
I.BLK-PS032398.D	PIBLK-PS032398.D	2,4-DCAA	1	500	439	88		24	167
		2,4-DCAA	2	500	517	103		24	167
PB170525BL	PB170525BL	2,4-DCAA	1	500	407	81		10	141
		2,4-DCAA	2	500	275	55		10	141
PB170525BS	PB170525BS	2,4-DCAA	1	500	533	107		10	141
		2,4-DCAA	2	500	573	115		10	141
Q3618-01	EME-TOP-SOIL	2,4-DCAA	1	500	262	52		10	141
		2,4-DCAA	2	500	290	58		10	141
Q3618-02	EME-GENERAL-FILL	2,4-DCAA	1	500	324	65		10	141
		2,4-DCAA	2	500	369	74		10	141
I.BLK-PS032412.D	PIBLK-PS032412.D	2,4-DCAA	1	500	419	84		24	167
		2,4-DCAA	2	500	536	107		24	167
I.BLK-PS032511.D	PIBLK-PS032511.D	2,4-DCAA	1	500	331	66		24	167
		2,4-DCAA	2	500	477	95		24	167
Q3609-01MS	AU-713-COMP-01MS	2,4-DCAA	1	500	54.8	11		10	141
		2,4-DCAA	2	500	154	31		10	141
Q3609-01MSD	AU-713-COMP-01MSD	2,4-DCAA	1	500	57.7	12		10	141
		2,4-DCAA	2	500	159	32		10	141
I.BLK-PS032523.D	PIBLK-PS032523.D	2,4-DCAA	1	500	356	71		24	167
		2,4-DCAA	2	500	477	95		24	167