

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3618</u>	Analytical Method:	<u>8151A</u>
Client:	<u>ENTACT</u>	DataFile :	<u>PS032514.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Low	Limits		RPD
			Result	Result			Qual	Qual	High		High		
Lab Sample ID:	Q3609-01MSD (Column 1)	Client Sample ID:		AU-713-COMP-01MSD									
	DICAMBA	203.7	0	98.7	ug/Kg	48		0		10	112		20
	DICHLORPROP	203.7	0	98.5	ug/Kg	48		2		10	113		20
	2,4-D	203.7	0	97.9	ug/Kg	48		2		10	144		20
	2,4,5-TP(Silvex)	203.7	0	95.6	ug/Kg	47		0		17	99		20
	2,4,5-T	203.7	0	95.7	ug/Kg	47		0		10	127		20
	2,4-DB	203.7	0	97.1	ug/Kg	48		0		10	91		20
	Dinoseb	203.7	0	87.0	ug/Kg	43		0		10	118		20
Lab Sample ID:	Q3609-01MSD (Column 2)	Client Sample ID:		AU-713-COMP-01MSD									
	DICAMBA	203.7	0	103	ug/Kg	51		0		10	112		20
	DICHLORPROP	203.7	0	93.7	ug/Kg	46		2		10	113		20
	2,4-D	203.7	0	104	ug/Kg	51		2		10	144		20
	2,4,5-TP(Silvex)	203.7	0	107	ug/Kg	53		0		17	99		20
	2,4,5-T	203.7	0	102	ug/Kg	50		0		10	127		20
	2,4-DB	203.7	0	78.7	ug/Kg	39		3		10	91		20
	Dinoseb	203.7	0	82.6	ug/Kg	41		0		10	118		20