

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 17:06

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.81	5.80	5.70	5.90	-0.01
Aroclor-1248-2 (2)	6.06	6.07	5.97	6.17	0.01
Aroclor-1248-3 (3)	6.26	6.27	6.17	6.37	0.01
Aroclor-1248-4 (4)	6.66	6.67	6.57	6.77	0.01
Aroclor-1248-5 (5)	6.70	6.71	6.61	6.81	0.01
Tetrachloro-m-xylene	4.63	4.65	4.55	4.75	0.02
Decachlorobiphenyl	10.40	10.42	10.32	10.52	0.02

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 17:06

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1248-2	(2)	5.09	5.10	5.00	5.20	0.01
Aroclor-1248-3	(3)	5.13	5.14	5.04	5.24	0.01
Aroclor-1248-4	(4)	5.30	5.31	5.21	5.41	0.01
Aroclor-1248-5	(5)	5.69	5.70	5.60	5.80	0.01
Tetrachloro-m-xylene		3.77	3.78	3.68	3.88	0.01
Decachlorobiphenyl		8.77	8.78	8.68	8.88	0.01