

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115067.D **Time Analyzed:** 09:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.486	5.387	5.587	497.130	500.000	-0.6
Aroclor-1016-2	5.508	5.408	5.608	494.520	500.000	-1.1
Aroclor-1016-3	5.570	5.470	5.670	511.420	500.000	2.3
Aroclor-1016-4	5.666	5.566	5.766	498.630	500.000	-0.3
Aroclor-1016-5	5.957	5.858	6.058	488.530	500.000	-2.3
Aroclor-1260-1	7.070	6.971	7.171	466.870	500.000	-6.6
Aroclor-1260-2	7.324	7.225	7.425	505.050	500.000	1.0
Aroclor-1260-3	7.681	7.582	7.782	501.150	500.000	0.2
Aroclor-1260-4	7.903	7.803	8.003	495.530	500.000	-0.9
Aroclor-1260-5	8.210	8.110	8.310	487.850	500.000	-2.4
Decachlorobiphenyl	9.929	9.830	10.030	49.920	50.000	-0.2
Tetrachloro-m-xylene	4.347	4.248	4.448	51.210	50.000	2.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115067.D **Time Analyzed:** 09:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.628	4.828	486.870	500.000	-2.6
Aroclor-1016-2	4.745	4.647	4.847	490.870	500.000	-1.8
Aroclor-1016-3	4.919	4.821	5.021	496.780	500.000	-0.6
Aroclor-1016-4	4.961	4.864	5.064	496.970	500.000	-0.6
Aroclor-1016-5	5.173	5.076	5.276	493.030	500.000	-1.4
Aroclor-1260-1	6.204	6.106	6.306	467.570	500.000	-6.5
Aroclor-1260-2	6.391	6.294	6.494	462.450	500.000	-7.5
Aroclor-1260-3	6.543	6.446	6.646	467.390	500.000	-6.5
Aroclor-1260-4	7.015	6.917	7.117	475.890	500.000	-4.8
Aroclor-1260-5	7.255	7.158	7.358	467.640	500.000	-6.5
Decachlorobiphenyl	8.638	8.541	8.741	51.050	50.000	2.1
Tetrachloro-m-xylene	3.648	3.548	3.748	49.180	50.000	-1.6