

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076483.D **Time Analyzed:** 10:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.791	5.695	5.895	555.270	500.000	11.1
Aroclor-1016-2	5.813	5.717	5.917	558.660	500.000	11.7
Aroclor-1016-3	5.876	5.780	5.980	566.670	500.000	13.3
Aroclor-1016-4	5.973	5.877	6.077	589.070	500.000	17.8
Aroclor-1016-5	6.265	6.169	6.369	581.940	500.000	16.4
Aroclor-1260-1	7.384	7.287	7.487	597.430	500.000	19.5
Aroclor-1260-2	7.637	7.540	7.740	588.990	500.000	17.8
Aroclor-1260-3	7.993	7.899	8.099	565.600	500.000	13.1
Aroclor-1260-4	8.222	8.127	8.327	587.500	500.000	17.5
Aroclor-1260-5	8.548	8.452	8.652	557.700	500.000	11.5
Decachlorobiphenyl	10.414	10.319	10.519	51.470	50.000	2.9
Tetrachloro-m-xylene	4.638	4.542	4.742	55.020	50.000	10.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076483.D **Time Analyzed:** 10:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.869	4.778	4.978	534.270	500.000	6.9
Aroclor-1016-2	4.927	4.836	5.036	547.220	500.000	9.4
Aroclor-1016-3	5.047	4.954	5.154	471.120	500.000	-5.8
Aroclor-1016-4	5.087	4.996	5.196	553.910	500.000	10.8
Aroclor-1016-5	5.301	5.210	5.410	504.950	500.000	1.0
Aroclor-1260-1	6.329	6.238	6.438	547.410	500.000	9.5
Aroclor-1260-2	6.516	6.425	6.625	570.640	500.000	14.1
Aroclor-1260-3	6.670	6.578	6.778	541.850	500.000	8.4
Aroclor-1260-4	7.138	7.047	7.247	579.430	500.000	15.9
Aroclor-1260-5	7.378	7.287	7.487	549.210	500.000	9.8
Decachlorobiphenyl	8.770	8.682	8.882	46.560	50.000	-6.9
Tetrachloro-m-xylene	3.774	3.681	3.881	54.780	50.000	9.6