

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114869.D	04 Nov 2025 09:03		YP/AJ	Ok
2	I.BLK	I.BLK	PO114870.D	04 Nov 2025 09:21		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO114871.D	04 Nov 2025 09:39		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO114872.D	04 Nov 2025 09:58		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO114873.D	04 Nov 2025 10:16		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO114874.D	04 Nov 2025 10:34		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO114875.D	04 Nov 2025 10:53		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO114876.D	04 Nov 2025 11:11		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO114877.D	04 Nov 2025 11:30		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO114878.D	04 Nov 2025 11:48		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO114879.D	04 Nov 2025 12:06		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO114880.D	04 Nov 2025 12:25		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO114881.D	04 Nov 2025 12:42		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO114882.D	04 Nov 2025 13:00		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO114883.D	04 Nov 2025 13:19		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO114884.D	04 Nov 2025 13:37		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO114885.D	04 Nov 2025 13:56		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO114886.D	04 Nov 2025 14:14		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PO114887.D	04 Nov 2025 14:33		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO114888.D	04 Nov 2025 14:51		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO114889.D	04 Nov 2025 15:09		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO114890.D	04 Nov 2025 15:28		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO114891.D	04 Nov 2025 16:04		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO114892.D	04 Nov 2025 16:23		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO114893.D	04 Nov 2025 17:00		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO114894.D	04 Nov 2025 17:18		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO114895.D	04 Nov 2025 17:36		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO114896.D	04 Nov 2025 17:55		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO114897.D	04 Nov 2025 18:13		YP/AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PO114898.D	04 Nov 2025 18:31		YP/AJ	Ok
31	PO110425ICV500	ICVPO110425	PO114899.D	04 Nov 2025 19:08		YP/AJ	Ok
32	AR1242ICV500	ICVPO110425AR1242	PO114900.D	04 Nov 2025 20:21		YP/AJ	Ok,M
33	AR1248ICV500	ICVPO110425AR1248	PO114901.D	04 Nov 2025 20:58		YP/AJ	Ok
34	AR1254ICV500	ICVPO110425AR1254	PO114902.D	04 Nov 2025 21:35		YP/AJ	Ok
35	AR1268ICV500	ICVPO110425AR1268	PO114903.D	04 Nov 2025 22:12		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO114904.D	04 Nov 2025 22:48		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO111325

Review By	rahul	Review On	11/14/2025 1:01:17 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:23:00 PM
SubDirectory	PO111325	HP Acquire Method	HP Processing Method PO110425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO115066.D	13 Nov 2025 09:24		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO115067.D	13 Nov 2025 09:43		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO115068.D	13 Nov 2025 10:11		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO115069.D	13 Nov 2025 10:29		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO115070.D	13 Nov 2025 10:47		YP/AJ	Ok,M
6	I.BLK	I.BLK	PO115071.D	13 Nov 2025 11:05		YP/AJ	Ok
7	PB170526BL	PB170526BL	PO115072.D	13 Nov 2025 12:11		YP/AJ	Ok
8	PB170526BS	PB170526BS	PO115073.D	13 Nov 2025 12:29		YP/AJ	Ok
9	Q3609-01	AU-713-COMP-01	PO115074.D	13 Nov 2025 12:47		YP/AJ	Ok
10	Q3609-04	AU-713-COMP-02	PO115075.D	13 Nov 2025 13:06		YP/AJ	Ok
11	Q3609-07	AU-713-COMP-03	PO115076.D	13 Nov 2025 13:24		YP/AJ	Ok,M
12	Q3609-07MS	AU-713-COMP-03MS	PO115077.D	13 Nov 2025 13:42	AR1016 recovery fail	YP/AJ	Ok,M
13	Q3609-07MSD	AU-713-COMP-03MSD	PO115078.D	13 Nov 2025 14:00	AR1016 recovery fail	YP/AJ	Ok,M
14	Q3609-10	AU-713-COMP-04	PO115079.D	13 Nov 2025 14:18		YP/AJ	Ok
15	Q3609-13	AU-713-COMP-05	PO115080.D	13 Nov 2025 14:37		YP/AJ	Ok
16	Q3614-01	COMP-1	PO115081.D	13 Nov 2025 14:55		YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO115082.D	13 Nov 2025 16:09		YP/AJ	Ok
18	AR1242CCC500	AR1242CCC500	PO115083.D	13 Nov 2025 17:04		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO111325

Review By	rahul	Review On	11/14/2025 1:01:17 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:23:00 PM
SubDirectory	PO111325	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO115084.D	13 Nov 2025 17:22		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO115085.D	13 Nov 2025 17:41		YP/AJ	Ok
21	I.BLK	I.BLK	PO115086.D	13 Nov 2025 17:59		YP/AJ	Ok
22	Q3623-02	RBR200007-SILICA	PO115087.D	13 Nov 2025 18:18		YP/AJ	Ok
23	Q3624-01	OK-02-11122025	PO115088.D	13 Nov 2025 18:36	DCB high in both column	YP/AJ	ReRun
24	AR1660CCC500	AR1660CCC500	PO115089.D	13 Nov 2025 19:48		YP/AJ	Ok,M
25	AR1242CCC500	AR1242CCC500	PO115090.D	13 Nov 2025 20:44		YP/AJ	Ok
26	AR1248CCC500	AR1248CCC500	PO115091.D	13 Nov 2025 21:01		YP/AJ	Ok
27	AR1254CCC500	AR1254CCC500	PO115092.D	13 Nov 2025 21:19		YP/AJ	Ok
28	I.BLK	I.BLK	PO115093.D	13 Nov 2025 21:38		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075886.D	21 Oct 2025 08:33		YP\AJ	Ok
2	I.BLK	I.BLK	PP075887.D	21 Oct 2025 08:49		YP\AJ	Ok,M
3	AR1660ICC1000	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP075889.D	21 Oct 2025 09:22		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP075890.D	21 Oct 2025 09:38		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP075891.D	21 Oct 2025 09:54		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP075892.D	21 Oct 2025 10:11		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP075893.D	21 Oct 2025 10:27		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP075894.D	21 Oct 2025 10:43		YP\AJ	Ok,M
10	AR1242ICC1000	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59		YP\AJ	Ok,M
11	AR1242ICC750	AR1242ICC750	PP075896.D	21 Oct 2025 11:16		YP\AJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP075897.D	21 Oct 2025 11:48		YP\AJ	Ok,M
13	AR1242ICC250	AR1242ICC250	PP075898.D	21 Oct 2025 12:04		YP\AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP075899.D	21 Oct 2025 12:21		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41		YP\AJ	Ok,M
16	AR1248ICC750	AR1248ICC750	PP075901.D	21 Oct 2025 12:57		YP\AJ	Ok,M
17	AR1248ICC500	AR1248ICC500	PP075902.D	21 Oct 2025 13:13		YP\AJ	Ok,M
18	AR1248ICC250	AR1248ICC250	PP075903.D	21 Oct 2025 13:30		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PP075904.D	21 Oct 2025 13:46		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP075906.D	21 Oct 2025 14:18		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP075907.D	21 Oct 2025 14:35		YPIAJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP075908.D	21 Oct 2025 14:51		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP075909.D	21 Oct 2025 15:07		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP075910.D	21 Oct 2025 15:40		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP075912.D	21 Oct 2025 16:12		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP075913.D	21 Oct 2025 16:29		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP075914.D	21 Oct 2025 16:45		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PP075915.D	21 Oct 2025 17:01		YPIAJ	Ok
31	PP102125ICV500	ICVPP102125	PP075916.D	21 Oct 2025 17:34		YPIAJ	Ok
32	AR1242ICV500	ICVPP102125AR1242	PP075917.D	21 Oct 2025 18:06		YPIAJ	Ok,M
33	AR1248ICV500	ICVPP102125AR1248	PP075918.D	21 Oct 2025 18:39		YPIAJ	Ok,M
34	AR1254ICV500	ICVPP102125AR1254	PP075919.D	22 Oct 2025 08:49		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP102125AR1268	PP075920.D	22 Oct 2025 09:07		YPIAJ	Ok
36	DDT ANALOG	DDT ANALOG	PP075921.D	22 Oct 2025 09:25		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111325

Review By	rahul	Review On	11/14/2025 1:00:32 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:22:02 PM
SubDirectory	PP111325	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076482.D	13 Nov 2025 09:20		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076483.D	13 Nov 2025 10:22		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP076484.D	13 Nov 2025 10:43		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076485.D	13 Nov 2025 10:59		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076486.D	13 Nov 2025 11:16		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP076487.D	13 Nov 2025 11:32		YP\AJ	Ok
7	PB170501BS	PB170501BS	PP076488.D	13 Nov 2025 11:48		YP\AJ	Ok,M
8	Q3601-13	TP-4	PP076489.D	13 Nov 2025 12:04	TCMX high in 2nd column	YP\AJ	Ok
9	Q3615-01	East Bathhouse- Concr	PP076490.D	13 Nov 2025 12:21		YP\AJ	Ok
10	Q3614-02	COMP-2	PP076491.D	13 Nov 2025 12:37		YP\AJ	Ok,M
11	Q3614-03	COMP-3	PP076492.D	13 Nov 2025 12:53		YP\AJ	Ok
12	Q3618-01	EME-TOP-SOIL	PP076493.D	13 Nov 2025 13:09		YP\AJ	Ok
13	Q3618-02	EME-GENERAL-FILL	PP076494.D	13 Nov 2025 13:26		YP\AJ	Ok
14	Q3621-02	VNJ-231	PP076495.D	13 Nov 2025 13:42	TCMX high in both column	YP\AJ	ReRun
15	Q3622-01	RT3071-SOLID	PP076496.D	13 Nov 2025 13:58	AR1260 Hit	YP\AJ	Ok,M
16	Q3623-01	RBR200007-SOIL	PP076497.D	13 Nov 2025 14:14	DCB high in 1st column , AR1260 Hit	YP\AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PP076498.D	13 Nov 2025 16:01		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111325

Review By	rahul	Review On	11/14/2025 1:00:32 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:22:02 PM
SubDirectory	PP111325	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP076499.D	13 Nov 2025 16:50		YPIAJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PP076500.D	13 Nov 2025 17:06		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP076501.D	13 Nov 2025 17:23		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP076502.D	13 Nov 2025 17:39		YPIAJ	Ok

M : Manual Integration