

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q3618	OrderDate:	11/12/2025 2:35:43 PM
Client:	ENTACT	Project:	Whittaker Coatings Site – E9125B
Contact:	Austin Farmerie	Location:	D41,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3618-01	EME-TOP-SOIL	SOIL	PCB	8082A	11/12/25	11/13/25	11/13/25	11/12/25
Q3618-02	EME-GENERAL-FILL	SOIL	PCB	8082A	11/12/25	11/13/25	11/13/25	11/12/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3618

Order ID: Q3618

Client: ENTACT

Project ID: Whittaker Coatings Site – E9125B

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000



QC SUMMARY

Surrogate Summary

SDG No.: Q3618

Client: ENTACT

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO114870.D	PIBLK-PO114870.D	Tetrachloro-m-xyl	1	20	15.5	77		60	140
		Decachlorobiphen	1	20	16.8	84		60	140
		Tetrachloro-m-xyl	2	20	16.2	81		60	140
		Decachlorobiphen	2	20	17.3	86		60	140
I.BLK-PO115071.D	PIBLK-PO115071.D	Tetrachloro-m-xyl	1	20	17.3	87		60	140
		Decachlorobiphen	1	20	17.4	87		60	140
		Tetrachloro-m-xyl	2	20	17.3	86		60	140
		Decachlorobiphen	2	20	18.5	93		60	140
PB170526BL	PB170526BL	Tetrachloro-m-xyl	1	20	21.9	109		21	165
		Decachlorobiphen	1	20	22.6	113		10	170
		Tetrachloro-m-xyl	2	20	20.9	105		21	165
		Decachlorobiphen	2	20	23.6	118		10	170
PB170526BS	PB170526BS	Tetrachloro-m-xyl	1	20	23.6	118		21	165
		Decachlorobiphen	1	20	23.7	118		10	170
		Tetrachloro-m-xyl	2	20	21.8	109		21	165
		Decachlorobiphen	2	20	25.0	125		10	170
Q3609-07MS	AU-713-COMP-03MS	Tetrachloro-m-xyl	1	20	21.8	109		21	165
		Decachlorobiphen	1	20	18.5	92		10	170
		Tetrachloro-m-xyl	2	20	21.1	106		21	165
		Decachlorobiphen	2	20	19.5	97		10	170
Q3609-07MSD	AU-713-COMP-03MSD	Tetrachloro-m-xyl	1	20	21.9	109		21	165
		Decachlorobiphen	1	20	18.4	92		10	170
		Tetrachloro-m-xyl	2	20	21.1	105		21	165
		Decachlorobiphen	2	20	19.4	97		10	170
I.BLK-PO115086.D	PIBLK-PO115086.D	Tetrachloro-m-xyl	1	20	16.8	84		60	140
		Decachlorobiphen	1	20	17.0	85		60	140
		Tetrachloro-m-xyl	2	20	14.6	73		60	140
		Decachlorobiphen	2	20	17.8	89		60	140
I.BLK-PP075887.D	PIBLK-PP075887.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	20.4	102		60	140
		Tetrachloro-m-xyl	2	20	18.8	94		60	140
		Decachlorobiphen	2	20	19.3	96		60	140
I.BLK-PP076487.D	PIBLK-PP076487.D	Tetrachloro-m-xyl	1	20	22.6	113		60	140
		Decachlorobiphen	1	20	22.8	114		60	140
		Tetrachloro-m-xyl	2	20	20.7	103		60	140
		Decachlorobiphen	2	20	18.7	94		60	140
Q3618-01	EME-TOP-SOIL	Tetrachloro-m-xyl	1	20	23.4	117		21	165
		Decachlorobiphen	1	20	11.5	58		10	170
		Tetrachloro-m-xyl	2	20	20.4	102		21	165
		Decachlorobiphen	2	20	10.3	52		10	170
Q3618-02	EME-GENERAL-FILL	Tetrachloro-m-xyl	1	20	28.6	143		21	165

Surrogate Summary

SDG No.: Q3618

Client: ENTACT

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3618-02	EME-GENERAL-FILL	Decachlorobiphen	1	20	22.5	112		10	170
		Tetrachloro-m-xyl	2	20	25.6	128		21	165
		Decachlorobiphen	2	20	19.6	98		10	170
I.BLK-PP076502.D	PIBLK-PP076502.D	Tetrachloro-m-xyl	1	20	22.1	111		60	140
		Decachlorobiphen	1	20	21.6	108		60	140
		Tetrachloro-m-xyl	2	20	19.1	96		60	140
		Decachlorobiphen	2	20	18.2	91		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3618</u>	Analytical Method:	<u>8082A</u>
Client:	<u>ENTACT</u>	DataFile :	<u>PO115077.D</u>

		Sample				Rec		RPD		Limits	
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3609-07MS										
	(Column 1)										
AR1016	195.1	0	858	ug/kg	440	*			39	159	
AR1260	195.1	0	244	ug/kg	125				19	160	
Lab Sample ID:	Q3609-07MS										
	(Column 2)										
AR1016	195.1	0	425	ug/kg	218	*			39	159	
AR1260	195.1	0	320	ug/kg	164				19	160	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3618</u>	Analytical Method:	<u>8082A</u>
Client:	<u>ENTACT</u>	DataFile :	<u>PO115078.D</u>

	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q3609-07MSD (Column 1)		Client Sample ID:	AU-713-COMP-03MSD								
	AR1016	194.9	0	889	ug/kg	456	*	4		39	159	21
	AR1260	194.9	0	247	ug/kg	127		2		19	160	15
Lab Sample ID:	Q3609-07MSD (Column 2)		Client Sample ID:	AU-713-COMP-03MSD								
	AR1016	194.9	0	415	ug/kg	213	*	2		39	159	21
	AR1260	194.9	0	332	ug/kg	170		4		19	160	15



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3618

Analytical Method: 8082A

Client: ENTACT

Datafile : PO115073.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB170526BS (Column 1)	AR1016	166.6	150	ug/kg	90				71	120	
	AR1260	166.6	141	ug/kg	85				65	130	
PB170526BS (Column 2)	AR1016	166.6	143	ug/kg	86				71	120	
	AR1260	166.6	136	ug/kg	82				65	130	

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB170526BL

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Lab Sample ID: PB170526BL

Lab File ID: PO115072.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 11/13/2025

Date Analyzed (1): 11/13/2025

Date Analyzed (2): 11/13/2025

Time Analyzed (1): 12:11

Time Analyzed (2): 12:11

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB170526BS	PB170526BS	PO115073.D	11/13/2025	11/13/2025
AU-713-COMP-03MS	Q3609-07MS	PO115077.D	11/13/2025	11/13/2025
AU-713-COMP-03MSD	Q3609-07MSD	PO115078.D	11/13/2025	11/13/2025
EME-TOP-SOIL	Q3618-01	PP076493.D	11/13/2025	11/13/2025
EME-GENERAL-FILL	Q3618-02	PP076494.D	11/13/2025	11/13/2025

COMMENTS: _____



SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT		Date Collected:	11/12/25
Project:	Whittaker Coatings Site – E9125B		Date Received:	11/12/25
Client Sample ID:	EME-TOP-SOIL		SDG No.:	Q3618
Lab Sample ID:	Q3618-01		Matrix:	SOIL
Analytical Method:	8082A		% Solid:	70.3
Sample Wt/Vol:	30.08 g	Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B	Prep Date: 11/13/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	5.60	U	1	5.60	24.1	ug/kg	11/13/25 13:09	PB170526
11104-28-2	Aroclor-1221	5.70	U	1	5.70	24.1	ug/kg	11/13/25 13:09	PB170526
11141-16-5	Aroclor-1232	5.30	U	1	5.30	24.1	ug/kg	11/13/25 13:09	PB170526
53469-21-9	Aroclor-1242	5.70	U	1	5.70	24.1	ug/kg	11/13/25 13:09	PB170526
12672-29-6	Aroclor-1248	8.40	U	1	8.40	24.1	ug/kg	11/13/25 13:09	PB170526
11097-69-1	Aroclor-1254	4.60	U	1	4.60	24.1	ug/kg	11/13/25 13:09	PB170526
37324-23-5	Aroclor-1262	7.10	U	1	7.10	24.1	ug/kg	11/13/25 13:09	PB170526
11100-14-4	Aroclor-1268	5.10	U	1	5.10	24.1	ug/kg	11/13/25 13:09	PB170526
11096-82-5	Aroclor-1260	4.60	U	1	4.60	24.1	ug/kg	11/13/25 13:09	PB170526
SURROGATES									
877-09-8	Tetrachloro-m-xylene	23.4			21 - 165	117%	SPK: 20		
2051-24-3	Decachlorobiphenyl	11.5			10 - 170	58%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111325\
Data File : PP076493.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 13:09
Operator : YP\AJ
Sample : Q3618-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
EME-TOP-SOIL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 00:10:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.640	3.778	33487797	135.0E6	23.433	20.402
2) SA Decachlor...	10.411	8.772	12425283	85116849	11.542	10.351

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111325\
Data File : PP076493.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 13:09
Operator : YP\AJ
Sample : Q3618-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
EME-TOP-SOIL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 00:10:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.640 min
Delta R.T.: -0.002 min
Response: 33487797
Conc: 23.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TOP-SOIL




#1 Tetrachloro-m-xylene

R.T.: 3.778 min
Delta R.T.: -0.004 min
Response: 135024815
Conc: 20.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.411 min
Delta R.T.: -0.008 min
Response: 12425283
Conc: 11.54 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.772 min
Delta R.T.: -0.010 min
Response: 85116849
Conc: 10.35 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT		Date Collected:	11/12/25
Project:	Whittaker Coatings Site – E9125B		Date Received:	11/12/25
Client Sample ID:	EME-GENERAL-FILL		SDG No.:	Q3618
Lab Sample ID:	Q3618-02		Matrix:	SOIL
Analytical Method:	8082A		% Solid:	88.2
Sample Wt/Vol:	30.02 g	Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B	Prep Date: 11/13/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.50	U	1	4.50	19.3	ug/kg	11/13/25 13:26	PB170526
11104-28-2	Aroclor-1221	4.60	U	1	4.60	19.3	ug/kg	11/13/25 13:26	PB170526
11141-16-5	Aroclor-1232	4.20	U	1	4.20	19.3	ug/kg	11/13/25 13:26	PB170526
53469-21-9	Aroclor-1242	4.50	U	1	4.50	19.3	ug/kg	11/13/25 13:26	PB170526
12672-29-6	Aroclor-1248	6.70	U	1	6.70	19.3	ug/kg	11/13/25 13:26	PB170526
11097-69-1	Aroclor-1254	3.60	U	1	3.60	19.3	ug/kg	11/13/25 13:26	PB170526
37324-23-5	Aroclor-1262	5.70	U	1	5.70	19.3	ug/kg	11/13/25 13:26	PB170526
11100-14-4	Aroclor-1268	4.10	U	1	4.10	19.3	ug/kg	11/13/25 13:26	PB170526
11096-82-5	Aroclor-1260	3.70	U	1	3.70	19.3	ug/kg	11/13/25 13:26	PB170526
SURROGATES									
877-09-8	Tetrachloro-m-xylene	28.6			21 - 165	143%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.5			10 - 170	112%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111325\
Data File : PP076494.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 13:26
Operator : YP\AJ
Sample : Q3618-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
EME-GENERAL-FILL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 00:10:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.637	3.775	40801934	169.4E6	28.551	25.594
2) SA Decachlor...	10.406	8.767	24200686	161.6E6	22.480	19.653

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111325\
Data File : PP076494.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 13:26
Operator : YP\AJ
Sample : Q3618-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
EME-GENERAL-FILL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 00:10:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.637 min
Delta R.T.: -0.005 min
Response: 40801934
Conc: 28.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-GENERAL-FILL




#1 Tetrachloro-m-xylene

R.T.: 3.775 min
Delta R.T.: -0.007 min
Response: 169386711
Conc: 25.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.406 min
Delta R.T.: -0.012 min
Response: 24200686
Conc: 22.48 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.767 min
Delta R.T.: -0.015 min
Response: 161607488
Conc: 19.65 ng/ml



CALIBRATION SUMMARY

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: **ENTA05**

Lab Code: ACE

SDG NO.: Q3618

Instrument ID: ECD_O

Calibration Date(s): 11/04/2025 11/04/2025

Calibration Times: 09:39 18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 =	<u>PO114871.D</u>	RT 750 =	<u>PO114872.D</u>	
RT 500 =	PO114873.D	RT 250 =	PO114874.D	RT 050 =	PO114875.D

[illegible]

[illegible]

Calibration Times: 09:39 18:31

LAB FILE ID: RT 1000 = PO114871.D RT 750 = PO114872.D
RT 500 = PO114873.D RT 250 = PO114874.D RT 050 = PO114875.D

[illegible]

[illegible]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: ENTA05
SDG NO.: Q3618

Calibration Date(s): 11/04/2025 11/04/2025
Calibration Times: 09:39 18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO114871.D</u>		CF 750 = <u>PO114872.D</u>				
CF 500 = <u>PO114873.D</u>		CF 250 = <u>PO114874.D</u>		CF 050 = <u>PO114875.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	235183609	242405119	252300612	252330724	247880040	246020021	3
Aroclor-1016-2	(2)	380671765	391174169	403435360	408228592	401844700	397070917	3
Aroclor-1016-3	(3)	214892716	220811199	230510330	231508016	217282560	223000964	3
Aroclor-1016-4	(4)	180735079	184273587	191842228	192742164	194782360	188875084	3
Aroclor-1016-5	(5)	166479276	173610481	179587174	180772736	190536000	178197133	5
Aroclor-1260-1	(1)	291080181	301659089	311914366	329562032	293951540	305633442	5
Aroclor-1260-2	(2)	393658847	402595176	414352186	418536840	350242960	395877202	7
Aroclor-1260-3	(3)	331096941	340266576	347857958	348147780	291720640	331817979	7
Aroclor-1260-4	(4)	303823322	310642864	321126472	321507392	268000240	305020058	7
Aroclor-1260-5	(5)	714393776	724690244	737226162	719415648	670367440	713218654	3
Decachlorobiphenyl		4451734610	4516548787	4612561720	4524869360	3995527600	4420248415	6
Tetrachloro-m-xylene		7020403280	7162783533	7307727940	7156075200	6543618600	7038121711	4
Aroclor-1242-1	(1)	180395803	184480117	189796832	192180524	184350480	186240751	2
Aroclor-1242-2	(2)	290607708	295647931	305830174	302003732	299791360	298776181	2
Aroclor-1242-3	(3)	164449629	167684461	173929682	174410736	176369820	171368866	3
Aroclor-1242-4	(4)	136844790	140633719	146402846	139660276	152376540	143183634	4
Aroclor-1242-5	(5)	138362947	141839337	145002188	138385972	141196180	140957325	2
Decachlorobiphenyl		4729850460	4804993427	4858375920	4746069120	4160138600	4659885505	6
Tetrachloro-m-xylene		7348004600	7445204373	7532777260	7262019280	6694579200	7256516943	5
Aroclor-1248-1	(1)	130126160	135244476	137909000	141742196	154206600	139845686	6
Aroclor-1248-2	(2)	182893435	187680161	194064692	196392780	211353600	194476934	5
Aroclor-1248-3	(3)	203368877	206306241	213983954	212927108	217587500	210834736	3
Aroclor-1248-4	(4)	240181487	246210361	255848644	255957264	264773260	252594203	4
Aroclor-1248-5	(5)	225312398	230738423	239469078	240115800	243213880	235769916	3
Decachlorobiphenyl		4590720290	4647354867	4744288880	4661202760	4211361200	4570985599	5
Tetrachloro-m-xylene		7316541360	7442934853	7528054520	7402608560	7218400400	7381707939	2
Aroclor-1254-1	(1)	240591852	247416493	258765478	257075152	263299240	253429643	3
Aroclor-1254-2	(2)	346021204	353529297	371589652	368897016	412184820	370444398	7
Aroclor-1254-3	(3)	372893964	380446979	397009938	384293176	389678940	384864599	2
Aroclor-1254-4	(4)	315539209	321424513	336840998	326904952	342871860	328716306	3
Aroclor-1254-5	(5)	295373361	300690525	312155076	300474784	300573260	301853401	2
Decachlorobiphenyl		4791769470	4896862187	5061968640	4831244600	4661831200	4848735219	3
Tetrachloro-m-xylene		7410585440	7526809333	7802184660	7481754040	7408041400	7525874975	2
Aroclor-1268-1	(1)	876279188	885717404	893041022	887529864	813508780	871215252	4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	748334854	756337347	765525536	759915712	702804360	746583562	3
Aroclor-1268-3	(3)	663299706	672085827	679866912	671661004	617886840	660960058	4
Aroclor-1268-4	(4)	274155738	277531193	282310020	282161716	286808980	280593529	2
Aroclor-1268-5	(5)	1966581606	1979440064	1985729368	1956798376	1807416340	1939193151	4
Decachlorobiphenyl		8288385350	8396434587	8504946880	8419357920	7722111200	8266247187	4
Tetrachloro-m-xylene		7475792880	7636683133	7655291620	7597134240	7121440400	7497268455	3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: ENTA05
SDG NO.: Q3618

Calibration Date(s): 11/04/2025 11/04/2025
Calibration Times: 09:39 18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO114871.D</u>		CF 750 = <u>PO114872.D</u>			
CF 500 = <u>PO114873.D</u>		CF 250 = <u>PO114874.D</u>		CF 050 = <u>PO114875.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	178795127	185535456	193106802	200932132	192907960	190255495
Aroclor-1016-2	(2)	260787154	267435697	277593168	286495324	262010880	270864445
Aroclor-1016-3	(3)	140310175	144533207	150201526	154962192	149858780	147973176
Aroclor-1016-4	(4)	108607533	112021955	117721682	122827936	125909940	117417809
Aroclor-1016-5	(5)	142235561	147493437	153611756	157990124	162236540	152713484
Aroclor-1260-1	(1)	272865489	283696256	302549666	306994580	314431380	296107474
Aroclor-1260-2	(2)	410849028	420744668	440173148	441650472	436738740	430031211
Aroclor-1260-3	(3)	328919579	344471297	362206406	366201912	353957800	351151399
Aroclor-1260-4	(4)	295043629	302500340	313130718	316509244	287959420	303028670
Aroclor-1260-5	(5)	819932115	830411719	851082708	846532168	760181660	821628074
Decachlorobiphenyl		3768213910	3862452533	4051635320	3997632280	3839185200	3903823849
Tetrachloro-m-xylene		4444199150	4532189360	4610968540	4573862320	4172678600	4466779594
Aroclor-1242-1	(1)	138347633	144438088	149108388	148877560	150336280	146221590
Aroclor-1242-2	(2)	199908386	204961789	213926082	211880332	205062860	207147890
Aroclor-1242-3	(3)	108216818	112229521	116418372	113485296	109537600	111977521
Aroclor-1242-4	(4)	101919263	105301932	110004430	109236748	109193360	107131147
Aroclor-1242-5	(5)	135980897	140103160	145653410	150301700	142549620	142917757
Decachlorobiphenyl		4093300380	4301738307	4403735460	4338775160	4380898600	4303689581
Tetrachloro-m-xylene		4800115840	4683347560	4961464680	4527345840	4437059000	4681866584
Aroclor-1248-1	(1)	101027399	105625799	111262386	113093188	115539800	109309714
Aroclor-1248-2	(2)	136540550	142594189	148530518	152692968	164144180	148900481
Aroclor-1248-3	(3)	143887712	149887199	155474250	158349236	169440400	155407759
Aroclor-1248-4	(4)	171266913	177666692	184171754	187513908	195642380	183252329
Aroclor-1248-5	(5)	174649655	180220384	187038042	189642400	190092660	184328628
Decachlorobiphenyl		4076042330	4166088587	4290124020	4403270520	4144788800	4216062851
Tetrachloro-m-xylene		4758509080	4829545453	4892334280	4811209560	4727034000	4803726475
Aroclor-1254-1	(1)	275439068	284238351	300243386	299923620	341055960	300180077
Aroclor-1254-2	(2)	245333366	253279232	267984098	268318040	313122940	269607535
Aroclor-1254-3	(3)	398238109	410614839	430545692	426530308	445294740	422244738
Aroclor-1254-4	(4)	284785979	294895399	305027804	301955520	300306220	297394184
Aroclor-1254-5	(5)	385708423	392326712	409498118	398089984	402006780	397526003
Decachlorobiphenyl		4273373360	4367226933	4604811080	4533103920	4272134600	4410129979
Tetrachloro-m-xylene		4791085320	4887004773	5044010700	4876428000	4843036000	4888312959
Aroclor-1268-1	(1)	948830555	965398547	981732904	993582148	991082540	976125339



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	784312077	796694211	811227528	817006596	768929160	795633914	2
Aroclor-1268-3	(3)	672635507	681509284	695723330	696225136	632763280	675771307	4
Aroclor-1268-4	(4)	270862616	275402580	280086838	282372280	267609400	275266743	2
Aroclor-1268-5	(5)	1768544208	1786014605	1818457210	1831952004	1717683400	1784530285	2
Decachlorobiphenyl		7349980390	7437822467	7706201340	7861559120	8321979800	7735508623	5
Tetrachloro-m-xylene		4791521570	4902547960	4915920660	4883273320	4724156000	4843483902	2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
Instrument ID: ECD_O **Date(s) Analyzed:** 11/04/2025 11/04/2025
GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.55	4.45	4.65	84124000
		2	4.63	4.53	4.73	60183600
		3	4.71	4.61	4.81	193335000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.71	4.61	4.81	150015000
		2	5.22	5.12	5.32	72567400
		3	5.51	5.41	5.61	158675000
		4	5.67	5.57	5.77	73795600
		5	5.76	5.66	5.86	51580600
Aroclor-1262	500	1	7.70	7.60	7.80	453404000
		2	8.23	8.13	8.33	707198000
		3	8.53	8.43	8.63	451420000
		4	8.62	8.52	8.72	336186000
		5	9.24	9.14	9.34	241816000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
Instrument ID: ECD_O **Date(s) Analyzed:** 11/04/2025 11/04/2025
GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.86	3.76	3.96	61799200
		2	3.95	3.85	4.05	45637000
		3	4.02	3.92	4.12	140257000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.02	3.92	4.12	111493000
		2	4.75	4.65	4.85	112095000
		3	4.92	4.82	5.02	59654200
		4	5.01	4.91	5.11	50889000
		5	5.18	5.08	5.28	56587600
Aroclor-1262	500	1	6.76	6.66	6.86	494206000
		2	7.26	7.16	7.36	795090000
		3	7.54	7.44	7.64	329028000
		4	7.61	7.51	7.71	538192000
		5	8.10	8.00	8.20	238020000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114871.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 09:39
 Operator : YP/AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 10:39:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 10:34:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	702.0E6	444.4E6	97.995	98.158
2) SA Decachlor...	9.931	8.642	445.2E6	376.8E6	98.226	96.376
Target Compounds						
3) L1 AR-1016-1	5.487	4.728	235.2E6	178.8E6	964.887	961.518
4) L1 AR-1016-2	5.509	4.747	380.7E6	260.8E6	970.969	968.784
5) L1 AR-1016-3	5.571	4.922	214.9E6	140.3E6	964.936	965.952
6) L1 AR-1016-4	5.666	4.964	180.7E6	108.6E6	970.188	959.731
7) L1 AR-1016-5	5.958	5.176	166.5E6	142.2E6	962.123	961.547
31) L7 AR-1260-1	7.072	6.207	291.1E6	272.9E6	965.449	948.413
32) L7 AR-1260-2	7.325	6.394	393.7E6	410.8E6	974.390	965.542
33) L7 AR-1260-3	7.683	6.547	331.1E6	328.9E6	975.314	951.837
34) L7 AR-1260-4	7.904	7.018	303.8E6	295.0E6	972.313	970.260
35) L7 AR-1260-5	8.212	7.259	714.4E6	819.9E6	984.271	981.358



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114871.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:39
Operator : YP/AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:39:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:34:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114872.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 09:58
 Operator : YP/AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 10:42:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 10:34:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	537.2E6	339.9E6	74.991	75.051
2) SA Decachlor...	9.931	8.642	338.7E6	289.7E6	74.828	74.390
Target Compounds						
3) L1 AR-1016-1	5.486	4.728	181.8E6	139.2E6	747.252	748.882
4) L1 AR-1016-2	5.508	4.747	293.4E6	200.6E6	748.878	746.734
5) L1 AR-1016-3	5.571	4.921	165.6E6	108.4E6	745.744	747.508
6) L1 AR-1016-4	5.666	4.964	138.2E6	84016466	744.572	744.934
7) L1 AR-1016-5	5.958	5.176	130.2E6	110.6E6	751.666	748.544
31) L7 AR-1260-1	7.071	6.206	226.2E6	212.8E6	750.268	742.996
32) L7 AR-1260-2	7.326	6.394	301.9E6	315.6E6	748.253	744.378
33) L7 AR-1260-3	7.683	6.546	255.2E6	258.4E6	751.161	748.419
34) L7 AR-1260-4	7.904	7.017	233.0E6	226.9E6	747.063	747.386
35) L7 AR-1260-5	8.211	7.258	543.5E6	622.8E6	749.228	746.944



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:58
Operator : YP/AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:42:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:34:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114873.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 10:16
 Operator : YP/AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 10:36:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 10:34:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	365.4E6	230.5E6	50.000	50.000
2) SA Decachlor...	9.930	8.641	230.6E6	202.6E6	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.487	4.728	126.2E6	96553401	500.000	500.000
4) L1 AR-1016-2	5.508	4.747	201.7E6	138.8E6	500.000	500.000
5) L1 AR-1016-3	5.570	4.921	115.3E6	75100763	500.000	500.000
6) L1 AR-1016-4	5.666	4.964	95921114	58860841	500.000	500.000
7) L1 AR-1016-5	5.958	5.176	89793587	76805878	500.000	500.000
31) L7 AR-1260-1	7.071	6.206	156.0E6	151.3E6	500.000	500.000
32) L7 AR-1260-2	7.325	6.394	207.2E6	220.1E6	500.000	500.000
33) L7 AR-1260-3	7.682	6.546	173.9E6	181.1E6	500.000	500.000
34) L7 AR-1260-4	7.903	7.017	160.6E6	156.6E6	500.000	500.000
35) L7 AR-1260-5	8.210	7.258	368.6E6	425.5E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114873.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 10:16
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:36:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:34:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114874.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 10:34
 Operator : YP/AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC250

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/05/2025
 Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 10:51:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 10:51:10 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	178.9E6	114.3E6	24.980	25.185
2) SA Decachlor...	9.931	8.641	113.1E6	99940807	24.991	25.495
Target Compounds						
3) L1 AR-1016-1	5.487	4.728	63082681	50233033	256.898	264.953
4) L1 AR-1016-2	5.508	4.747	102.1E6	71623831	257.800	262.284
5) L1 AR-1016-3	5.571	4.922	57877004	38740548	257.884	262.645
6) L1 AR-1016-4	5.666	4.964	48185541	30706984	257.129	266.335
7) L1 AR-1016-5	5.958	5.176	45193184	39497531	258.081	262.734
31) L7 AR-1260-1	7.071	6.206	82390508	76748645	267.021	260.958m
32) L7 AR-1260-2	7.325	6.393	104.6E6	110.4E6	256.906	256.862m
33) L7 AR-1260-3	7.682	6.546	87036945	91550478	254.611	261.237
34) L7 AR-1260-4	7.904	7.017	80376848	79127311	255.753	257.915
35) L7 AR-1260-5	8.211	7.258	179.9E6	211.6E6	248.441	252.850

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114874.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 10:34
Operator : YP/AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660ICC250

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:51:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:51:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114875.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 10:53
 Operator : YP/AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/05/2025
 Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 11:09:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 11:09:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.348	3.649	32718093	20863393	4.649	4.671
2)	SA Decachlor...	9.933	8.641	19977638	19195926	4.520	4.917
Target Compounds							
3)	L1 AR-1016-1	5.487	4.727	12394002	9645398	50.378	50.697
4)	L1 AR-1016-2	5.509	4.747	20092235	13100544	50.601	48.366
5)	L1 AR-1016-3	5.572	4.921	10864128	7492939	48.718	51.057m
6)	L1 AR-1016-4	5.666	4.964	9739118	6295497	51.564	53.616
7)	L1 AR-1016-5	5.959	5.175	9526800	8111827	53.462	53.507m
31)	L7 AR-1260-1	7.071	6.206	14697577	15721569	48.664m	53.302m
32)	L7 AR-1260-2	7.325	6.392	17512148	21836937	45.034m	51.096m
33)	L7 AR-1260-3	7.682	6.546	14586032	17697890	44.584m	50.400
34)	L7 AR-1260-4	7.904	7.017	13400012	14397971	44.028m	47.514
35)	L7 AR-1260-5	8.211	7.258	33518372	38009083	46.996	46.261

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114875.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 10:53
Operator : YP/AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 11:09:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 11:09:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114876.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 11:11
 Operator : YP/AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 11:57:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 11:57:11 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	367.5E6	224.3E6	50.000	50.000
2) SA Decachlor...	9.932	8.641	223.4E6	199.5E6	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.549	3.860	42062015	30899570	500.000	500.000
9) L2 AR-1221-2	4.632	3.945	30091751	22818522	500.000	500.000
10) L2 AR-1221-3	4.706	4.021	96667420	70128648	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114876.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 11:11
Operator : YP/AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 11:57:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 11:57:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114877.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 11:30
 Operator : YP/AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 12:15:50 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 11:57:11 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.649	373.4E6	232.5E6	50.000	50.000
2) SA Decachlor...	9.931	8.641	232.4E6	203.6E6	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.706	4.020	75007315	55746703	500.000	500.000
12) L3 AR-1232-2	5.224	4.747	36283683	56047490	500.000	500.000
13) L3 AR-1232-3	5.508	4.921	79337557	29827092	500.000	500.000
14) L3 AR-1232-4	5.666	5.005	36897763	25444495	500.000	500.000
15) L3 AR-1232-5	5.757	5.176	25790349	28293807	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114877.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 11:30
Operator : YP/AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:15:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 11:57:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114878.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 11:48
 Operator : YP/AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 12:52:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 12:48:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	734.8E6	480.0E6	98.758	98.347
2) SA Decachlor...	9.932	8.642	473.0E6	409.3E6	98.660	96.347
Target Compounds						
16) L4 AR-1242-1	5.487	4.728	180.4E6	138.3E6	974.605	962.566
17) L4 AR-1242-2	5.509	4.747	290.6E6	199.9E6	974.478	966.127
18) L4 AR-1242-3	5.570	4.921	164.4E6	108.2E6	971.984	963.489
19) L4 AR-1242-4	5.667	5.005	136.8E6	101.9E6	966.255	961.849
20) L4 AR-1242-5	6.392	5.526	138.4E6	136.0E6	976.570	965.656



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114878.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 11:48
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:52:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114879.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 12:06
 Operator : YP/AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 12:55:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 12:48:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	558.4E6	351.3E6	75.032	72.950
2) SA Decachlor...	9.930	8.641	360.4E6	322.6E6	75.113	75.624
Target Compounds						
16) L4 AR-1242-1	5.487	4.728	138.4E6	108.3E6	748.334	752.466
17) L4 AR-1242-2	5.508	4.747	221.7E6	153.7E6	745.677	745.260
18) L4 AR-1242-3	5.570	4.921	125.8E6	84172141	745.539	749.608
19) L4 AR-1242-4	5.666	5.006	105.5E6	78976449	746.496	746.880
20) L4 AR-1242-5	6.391	5.525	106.4E6	105.1E6	750.553	747.461



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114879.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 12:06
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:55:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114880.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 12:25
 Operator : YP/AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 12:49:50 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 12:48:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	376.6E6	248.1E6	50.000	50.000
2) SA Decachlor...	9.932	8.642	242.9E6	220.2E6	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.488	4.728	94898416	74554194	500.000	500.000
17) L4 AR-1242-2	5.509	4.747	152.9E6	107.0E6	500.000	500.000
18) L4 AR-1242-3	5.571	4.922	86964841	58209186	500.000	500.000
19) L4 AR-1242-4	5.667	5.006	73201423	55002215	500.000	500.000
20) L4 AR-1242-5	6.393	5.526	72501094	72826705	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114880.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 12:25
Operator : YP/AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:49:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114881.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 12:42
 Operator : YP/AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 12:59:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 12:48:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.649	181.6E6	113.2E6	24.544	23.863
2) SA Decachlor...	9.930	8.642	118.7E6	108.5E6	24.798	25.317
Target Compounds						
16) L4 AR-1242-1	5.485	4.728	48045131	37219390	257.320	256.344
17) L4 AR-1242-2	5.507	4.747	75500933	52970083	252.915	255.070
18) L4 AR-1242-3	5.569	4.921	43602684	28371324	256.308	251.994
19) L4 AR-1242-4	5.665	5.006	34915069	27309187	247.826	256.146
20) L4 AR-1242-5	6.391	5.526	34596493	37575425	245.544	262.747



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114881.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 12:42
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:59:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114882.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 13:00
 Operator : YP/AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC050

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/05/2025
 Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 13:15:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 13:15:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.649	33472896	22185295	4.613	4.739
2) SA Decachlor...	9.932	8.641	20800693	21904493	4.464	5.090
Target Compounds						
16) L4 AR-1242-1	5.487	4.729	9217524	7516814	49.493	51.407
17) L4 AR-1242-2	5.509	4.747	14989568	10253143	50.170	49.497
18) L4 AR-1242-3	5.571	4.922	8818491	5476880	51.459	48.911
19) L4 AR-1242-4	5.667	5.007	7618827	5459668	53.210	50.962
20) L4 AR-1242-5	6.392	5.526	7059809	7127481	50.085	49.823m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114882.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:00
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 13:15:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 13:15:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114883.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 13:19
 Operator : YP/AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 14:15:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 14:11:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	731.7E6	475.9E6	98.575	98.613
2) SA Decachlor...	9.931	8.641	459.1E6	407.6E6	98.355	97.441
Target Compounds						
21) L5 AR-1248-1	5.486	4.727	130.1E6	101.0E6	970.963	951.788
22) L5 AR-1248-2	5.757	4.963	182.9E6	136.5E6	970.365	957.940
23) L5 AR-1248-3	5.957	5.005	203.4E6	143.9E6	974.566	961.296
24) L5 AR-1248-4	6.354	5.175	240.2E6	171.3E6	968.415	963.693
25) L5 AR-1248-5	6.391	5.567	225.3E6	174.6E6	969.541	965.748



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114883.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:19
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:15:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114884.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 13:37
 Operator : YP/AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 14:18:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 14:11:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.650	558.2E6	362.2E6	75.139	75.043
2) SA Decachlor...	9.934	8.642	348.6E6	312.5E6	74.784	74.797
Target Compounds						
21) L5 AR-1248-1	5.487	4.728	101.4E6	79219349	754.563	747.551
22) L5 AR-1248-2	5.759	4.965	140.8E6	106.9E6	747.878	750.206
23) L5 AR-1248-3	5.958	5.007	154.7E6	112.4E6	744.299	750.689
24) L5 AR-1248-4	6.355	5.177	184.7E6	133.3E6	746.353	749.852
25) L5 AR-1248-5	6.393	5.568	173.1E6	135.2E6	746.437	748.274



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114884.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:37
Operator : YP/AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:18:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114885.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 13:56
 Operator : YP/AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 14:12:50 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 14:11:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.650	376.4E6	244.6E6	50.000	50.000
2) SA Decachlor...	9.932	8.642	237.2E6	214.5E6	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.487	4.728	68954500	55631193	500.000	500.000
22) L5 AR-1248-2	5.759	4.964	97032346	74265259	500.000	500.000
23) L5 AR-1248-3	5.959	5.006	107.0E6	77737125	500.000	500.000
24) L5 AR-1248-4	6.354	5.177	127.9E6	92085877	500.000	500.000
25) L5 AR-1248-5	6.393	5.568	119.7E6	93519021	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114885.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:56
Operator : YP/AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:12:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114886.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 14:14
 Operator : YP/AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 14:29:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 14:11:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.649	185.1E6	120.3E6	24.933	24.939
2) SA Decachlor...	9.931	8.642	116.5E6	110.1E6	25.002	26.000
Target Compounds						
21) L5 AR-1248-1	5.487	4.728	35435549	28273297	260.067	262.392
22) L5 AR-1248-2	5.758	4.964	49098195	38173242	258.061	263.101
23) L5 AR-1248-3	5.958	5.006	53231777	39587309	254.519	260.615
24) L5 AR-1248-4	6.354	5.176	63989316	46878477	256.419	260.212
25) L5 AR-1248-5	6.393	5.567	60028950	47410600	256.634	259.234



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114886.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 14:14
Operator : YP/AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:29:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114887.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 14:33
 Operator : YP/AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 14:49:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 14:48:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	36092002	23635170	4.889	4.920
2) SA Decachlor...	9.931	8.642	21056806	20723944	4.607	4.915
Target Compounds						
21) L5 AR-1248-1	5.487	4.727	7710330	5776990	55.135	52.850
22) L5 AR-1248-2	5.758	4.963	10567680	8207209	54.339	55.119
23) L5 AR-1248-3	5.958	5.005	10879375	8472020	51.601	54.515
24) L5 AR-1248-4	6.355	5.176	13238663	9782119	52.411	53.381
25) L5 AR-1248-5	6.393	5.567	12160694	9504633	51.579	51.564



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114887.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 14:33
Operator : YP/AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:49:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:48:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114888.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 14:51
 Operator : YP/AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:36:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 05:33:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.650	741.1E6	479.1E6	94.981	94.986
2) SA Decachlor...	9.932	8.642	479.2E6	427.3E6	94.662	92.802
Target Compounds						
26) L6 AR-1254-1	6.331	5.527	240.6E6	275.4E6	929.768	917.386
27) L6 AR-1254-2	6.547	5.675	346.0E6	245.3E6	931.192	915.477
28) L6 AR-1254-3	6.907	6.076	372.9E6	398.2E6	939.256	924.961
29) L6 AR-1254-4	7.188	6.304	315.5E6	284.8E6	936.760	933.639
30) L6 AR-1254-5	7.604	6.721	295.4E6	385.7E6	946.239	941.905



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114888.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 14:51
Operator : YP/AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114889.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 15:09
 Operator : YP/AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:36:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 05:33:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.349	3.650	564.5E6	366.5E6	72.353	72.665
2)	SA Decachlor...	9.933	8.642	367.3E6	327.5E6	72.554	71.130
Target Compounds							
26)	L6 AR-1254-1	6.331	5.527	185.6E6	213.2E6	717.106	710.020
27)	L6 AR-1254-2	6.547	5.675	265.1E6	190.0E6	713.548	708.846
28)	L6 AR-1254-3	6.907	6.077	285.3E6	308.0E6	718.711	715.281
29)	L6 AR-1254-4	7.189	6.304	241.1E6	221.2E6	715.674	725.087
30)	L6 AR-1254-5	7.604	6.721	225.5E6	294.2E6	722.455	718.550



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114889.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 15:09
Operator : YP/AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114890.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 15:28
 Operator : YP/AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:36:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 05:33:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.649	390.1E6	252.2E6	50.000	50.000
2) SA Decachlor...	9.932	8.642	253.1E6	230.2E6	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.331	5.527	129.4E6	150.1E6	500.000	500.000
27) L6 AR-1254-2	6.547	5.674	185.8E6	134.0E6	500.000	500.000
28) L6 AR-1254-3	6.907	6.076	198.5E6	215.3E6	500.000	500.000
29) L6 AR-1254-4	7.189	6.304	168.4E6	152.5E6	500.000	500.000
30) L6 AR-1254-5	7.603	6.720	156.1E6	204.7E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114890.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 15:28
Operator : YP/AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114891.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 16:04
 Operator : YP/AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:36:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 05:33:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.649	187.0E6	121.9E6	23.973	24.169
2) SA Decachlor...	9.932	8.642	120.8E6	113.3E6	23.861	24.611
Target Compounds						
26) L6 AR-1254-1	6.331	5.527	64268788	74980905	248.367	249.734
27) L6 AR-1254-2	6.548	5.674	92224254	67079510	248.188	250.312
28) L6 AR-1254-3	6.907	6.076	96073294	106.6E6	241.992	247.668
29) L6 AR-1254-4	7.189	6.304	81726238	75488880	242.626	247.482
30) L6 AR-1254-5	7.604	6.720	75118696	99522496	240.645	243.035


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 16:04
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114892.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 16:23
 Operator : YP/AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/05/2025
 Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:37:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 05:33:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	37040207	24215180	4.747	4.801
2) SA Decachlor...	9.931	8.642	23309156	21360673	4.605m	4.639
Target Compounds						
26) L6 AR-1254-1	6.330	5.526	13164962	17052798	50.876m	56.797
27) L6 AR-1254-2	6.547	5.674	20609241	15656147	55.462m	58.422
28) L6 AR-1254-3	6.907	6.076	19483947	22264737	49.077	51.713
29) L6 AR-1254-4	7.188	6.303	17143593	15015311	50.895m	49.226
30) L6 AR-1254-5	7.604	6.720	15028663	20100339	48.145	49.085

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114892.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 16:23
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_0
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:37:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114893.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 17:00
 Operator : YP/AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:08:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:07:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.368	3.649	362.5E6	234.2E6	50.000	50.000
2) SA Decachlor...	9.952	8.641	223.0E6	207.7E6	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	7.703	6.759	226.7E6	247.1E6	500.000	500.000
37) L8 AR-1262-2	8.231	7.258	353.6E6	397.5E6	500.000	500.000
38) L8 AR-1262-3	8.534	7.542	225.7E6	164.5E6	500.000	500.000
39) L8 AR-1262-4	8.616	7.605	168.1E6	269.1E6	500.000	500.000
40) L8 AR-1262-5	9.236	8.097	120.9E6	119.0E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114893.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:00
Operator : YP/AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:08:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:07:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 17:18
 Operator : YP/AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:23:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:21:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.649	747.6E6	479.2E6	97.655	97.469
2) SA Decachlor...	9.931	8.642	828.8E6	735.0E6	97.454	95.377
Target Compounds						
41) L9 AR-1268-1	8.511	7.542	876.3E6	948.8E6	981.231	966.485
42) L9 AR-1268-2	8.599	7.606	748.3E6	784.3E6	977.544	966.821
43) L9 AR-1268-3	8.820	7.814	663.3E6	672.6E6	975.632	966.815
44) L9 AR-1268-4	9.216	8.097	274.2E6	270.9E6	971.116	967.067
45) L9 AR-1268-5	9.609	8.387	1966.6E6	1768.5E6	990.357	972.552



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:18
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114895.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 17:36
 Operator : YP/AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:23:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:21:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	572.8E6	367.7E6	74.818	74.796
2) SA Decachlor...	9.932	8.641	629.7E6	557.8E6	74.043	72.388
Target Compounds						
41) L9 AR-1268-1	8.510	7.542	664.3E6	724.0E6	743.849	737.521
42) L9 AR-1268-2	8.599	7.606	567.3E6	597.5E6	740.998	736.564
43) L9 AR-1268-3	8.819	7.814	504.1E6	511.1E6	741.416	734.677
44) L9 AR-1268-4	9.215	8.096	208.1E6	206.6E6	737.304	737.457
45) L9 AR-1268-5	9.609	8.387	1484.6E6	1339.5E6	747.625	736.619



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114895.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:36
Operator : YP/AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114896.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 17:55
 Operator : YP/AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:23:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:21:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	382.8E6	245.8E6	50.000	50.000
2) SA Decachlor...	9.931	8.641	425.2E6	385.3E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.510	7.542	446.5E6	490.9E6	500.000	500.000
42) L9 AR-1268-2	8.598	7.606	382.8E6	405.6E6	500.000	500.000
43) L9 AR-1268-3	8.820	7.814	339.9E6	347.9E6	500.000	500.000
44) L9 AR-1268-4	9.215	8.096	141.2E6	140.0E6	500.000	500.000
45) L9 AR-1268-5	9.609	8.387	992.9E6	909.2E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:55
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114897.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 18:13
 Operator : YP/AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC250

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/05/2025
 Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:23:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:21:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	189.9E6	122.1E6	24.810	24.834
2) SA Decachlor...	9.932	8.641	210.5E6	196.5E6	24.748	25.504
Target Compounds						
41) L9 AR-1268-1	8.511	7.541	221.9E6	248.4E6	248.457	253.017m
42) L9 AR-1268-2	8.599	7.605	190.0E6	204.3E6	248.168	251.781
43) L9 AR-1268-3	8.820	7.813	167.9E6	174.1E6	246.983	250.180
44) L9 AR-1268-4	9.216	8.096	70540429	70593070	249.869	252.040
45) L9 AR-1268-5	9.610	8.387	489.2E6	458.0E6	246.358	251.855

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114897.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 18:13
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114898.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 18:31
 Operator : YP/AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:23:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:21:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	35607202	23620780	4.651	4.805
2) SA Decachlor...	9.931	8.640	38610556	41609899	4.540	5.400
Target Compounds						
41) L9 AR-1268-1	8.511	7.542	40675439	49554127	45.547	50.476
42) L9 AR-1268-2	8.599	7.606	35140218	38446458	45.903	47.393
43) L9 AR-1268-3	8.820	7.814	30894342	31638164	45.442	45.475
44) L9 AR-1268-4	9.214	8.096	14340449	13380470	50.797	47.773
45) L9 AR-1268-5	9.609	8.387	90370817	85884170	45.510	47.229



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114898.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 18:31
Operator : YP/AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114899.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 19:08
 Operator : YP/AJ
 Sample : PO110425ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO110425

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 02:03:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 02:02:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.348	3.649	350.6E6	216.0E6	49.817	48.358
2)	SA Decachlor...	9.932	8.640	217.2E6	202.0E6	49.145	51.742
Target Compounds							
3)	L1 AR-1016-1	5.487	4.728	119.9E6	92647170	487.281	486.962
4)	L1 AR-1016-2	5.508	4.747	193.0E6	132.4E6	485.938	488.627
5)	L1 AR-1016-3	5.570	4.921	110.7E6	72980433	496.242	493.200
6)	L1 AR-1016-4	5.666	4.963	92089690	57444659	487.569	489.233
7)	L1 AR-1016-5	5.957	5.176	84866211	74982175	476.249	490.999
31)	L7 AR-1260-1	7.071	6.205	144.7E6	141.9E6	473.335	479.180
32)	L7 AR-1260-2	7.325	6.393	200.2E6	209.0E6	505.658	486.103
33)	L7 AR-1260-3	7.682	6.546	167.5E6	170.3E6	504.732	484.991
34)	L7 AR-1260-4	7.904	7.017	152.7E6	150.0E6	500.629	495.013
35)	L7 AR-1260-5	8.210	7.257	350.1E6	404.4E6	490.903	492.224



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 19:08
Operator : YP/AJ
Sample : PO110425ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
ICVPO110425

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 02:03:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 02:02:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114900.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 20:21
 Operator : YP/AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO110425AR1242

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/05/2025
 Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:57:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:57:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	369.9E6	227.0E6	50.980	48.485
2) SA Decachlor...	9.932	8.639	237.1E6	216.0E6	50.883	50.180
Target Compounds						
16) L4 AR-1242-1	5.486	4.727	91685645	72678150	492.296m	497.041
17) L4 AR-1242-2	5.508	4.747	146.5E6	102.4E6	490.193m	494.324
18) L4 AR-1242-3	5.570	4.921	84401159	56402697	492.512m	503.697
19) L4 AR-1242-4	5.666	5.005	70483519	53471210	492.260m	499.119
20) L4 AR-1242-5	6.393	5.525	75170815	70192007	533.288	491.136

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 20:21
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1242

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:57:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:57:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114901.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 20:58
 Operator : YP/AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO110425AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:49:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:49:02 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	364.4E6	233.5E6	49.370	48.616
2) SA Decachlor...	9.931	8.640	231.0E6	213.4E6	50.545	50.607
Target Compounds						
21) L5 AR-1248-1	5.487	4.727	67941811	53066852	485.834	485.472
22) L5 AR-1248-2	5.758	4.963	96189175	72201389	494.605	484.897
23) L5 AR-1248-3	5.958	5.005	104.8E6	75367544	497.247	484.966
24) L5 AR-1248-4	6.354	5.175	125.7E6	89630325	497.795	489.109
25) L5 AR-1248-5	6.393	5.567	118.0E6	89347145	500.480	484.717



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 20:58
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:49:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:49:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114902.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 21:35
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO110425AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:47:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 05:47:04 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	365.7E6	233.3E6	48.598	47.717
2) SA Decachlor...	9.931	8.640	237.7E6	218.5E6	49.015	49.554
Target Compounds						
26) L6 AR-1254-1	6.331	5.526	122.7E6	141.2E6	484.213	470.374
27) L6 AR-1254-2	6.546	5.673	176.1E6	124.6E6	475.284	462.124
28) L6 AR-1254-3	6.906	6.075	193.7E6	200.1E6	503.340	473.848
29) L6 AR-1254-4	7.188	6.302	159.9E6	143.9E6	486.428	484.000
30) L6 AR-1254-5	7.603	6.719	147.9E6	190.6E6	490.023	479.436


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114902.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 21:35
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:47:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:47:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114903.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 22:12
 Operator : YP/AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO110425AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:39:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:38:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.647	376.8E6	240.7E6	50.252	49.688
2) SA Decachlor...	9.929	8.640	416.0E6	381.0E6	50.328	49.259
Target Compounds						
41) L9 AR-1268-1	8.509	7.541	437.5E6	476.9E6	502.163	488.581
42) L9 AR-1268-2	8.599	7.605	374.7E6	393.6E6	501.949	494.742
43) L9 AR-1268-3	8.819	7.814	331.8E6	336.4E6	502.049	497.748
44) L9 AR-1268-4	9.214	8.096	138.2E6	136.4E6	492.596	495.687
45) L9 AR-1268-5	9.608	8.385	968.0E6	889.7E6	499.171	498.578



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114903.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 22:12
Operator : YP/AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:39:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:38:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Contract: **ENTA05**

SDG NO.: Q3618

Calibration Date(s): **10/21/2025** **10/21/2025**

Calibration Times: 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP075888.D RT 750 = PP075889.D
RT 500 = PP075890.D RT 250 = PP075891.D RT 050 = PP075892.D

[illegible]

Calibration Times: 09:06 17:01

LAB FILE ID: RT 1000 = PP075888.D RT 750 = PP075889.D
RT 500 = PP075890.D RT 250 = PP075891.D RT 050 = PP075892.D

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.79	8.78	8.79	8.79	8.79	8.79	8.69	8.89
Tetrachloro-m-xylene	3.78	3.78	3.79	3.79	3.79	3.78	3.68	3.88



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: ENTA05
SDG NO.: Q3618

Calibration Date(s): 10/21/2025 10/21/2025
Calibration Times: 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP075888.D</u>		CF 750 = <u>PP075889.D</u>			
CF 500 = <u>PP075890.D</u>		CF 250 = <u>PP075891.D</u>		CF 050 = <u>PP075892.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	53893845	57848439	60228768	63056856	60602640	59126110
Aroclor-1016-2	(2)	81387326	86593457	91038396	94931280	88270120	88444116
Aroclor-1016-3	(3)	50025937	53493744	56764402	59544472	56926600	55351031
Aroclor-1016-4	(4)	41815325	44617539	46982858	48544784	45240520	45440205
Aroclor-1016-5	(5)	39075321	41377448	43150568	44782124	48642160	43405524
Aroclor-1260-1	(1)	64655691	67908073	70553690	74100180	66565440	68756615
Aroclor-1260-2	(2)	80744063	85234851	88594752	93987136	99325680	89577296
Aroclor-1260-3	(3)	64964687	67552583	69003280	72715708	72196540	69286560
Aroclor-1260-4	(4)	65448766	70711301	72542976	74553116	59600740	68571380
Aroclor-1260-5	(5)	139384380	146230631	152452738	155039368	146650940	147951611
Decachlorobiphenyl		1012132950	1071302787	1124785220	1137889120	1036523400	1076526695
Tetrachloro-m-xylene		1370219590	1453116893	1490571040	1483020960	1348537200	1429093137
Aroclor-1242-1	(1)	44019933	46315467	47674416	50255084	51268840	47906748
Aroclor-1242-2	(2)	64942800	67715737	70979556	73767248	76553360	70791740
Aroclor-1242-3	(3)	40140584	42044764	44589646	47135032	54298900	45641785
Aroclor-1242-4	(4)	33543618	35162207	36420600	36967860	39702420	36359341
Aroclor-1242-5	(5)	36405140	38662289	41888622	39296052	48467620	40943945
Decachlorobiphenyl		1136280900	1209279307	1231278960	1289202560	1357807000	1244769745
Tetrachloro-m-xylene		1466601570	1512648573	1545825660	1595207720	1573043800	1538665465
Aroclor-1248-1	(1)	33471805	35785521	36963520	39121904	34010240	35870598
Aroclor-1248-2	(2)	44816839	47584043	49402910	52441868	49476340	48744400
Aroclor-1248-3	(3)	49477903	52446567	61105936	57335280	50958020	54264741
Aroclor-1248-4	(4)	60435029	64167197	67505182	72174088	67539280	66364155
Aroclor-1248-5	(5)	58134534	61713151	65347636	71297512	67810300	64860627
Decachlorobiphenyl		1140637530	1190681093	1243623480	1306886720	1250686400	1226503045
Tetrachloro-m-xylene		1468701040	1529744213	1558324940	1626496320	1450854600	1526824223
Aroclor-1254-1	(1)	66261960	71898345	78270020	82464920	100572060	79893461
Aroclor-1254-2	(2)	87851253	92570580	96669370	102809560	120020480	99984249
Aroclor-1254-3	(3)	92035771	96900691	99787992	105988040	125061300	103954759
Aroclor-1254-4	(4)	73136868	76689867	79390286	83967352	104884540	83613783
Aroclor-1254-5	(5)	85440842	89720104	92252192	95013656	95427040	91570767
Decachlorobiphenyl		1194970230	1245235800	1296058720	1332579960	1514859800	1316740902
Tetrachloro-m-xylene		1499217980	1559170693	1583264260	1614158480	1720822800	1595326843
Aroclor-1268-1	(1)	186733090	195066952	205606148	210022232	235253460	206536376



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	172486766	179135464	188184104	190078376	211424800	188261902	8
Aroclor-1268-3	(3)	145245892	150281380	156894280	158090072	165408840	155184093	5
Aroclor-1268-4	(4)	71036137	74805685	76392484	76627212	94373740	78647052	12
Aroclor-1268-5	(5)	446420076	458917541	475352492	479939416	541304600	480386825	8
Decachlorobiphenyl		1988394870	2075434813	2173906840	2211711440	2356033600	2161096313	6
Tetrachloro-m-xylene		1515083040	1549265533	1603737900	1591085800	1659449000	1583724255	3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: ENTA05
SDG NO.: Q3618

Calibration Date(s): 10/21/2025 10/21/2025
Calibration Times: 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP075888.D</u>		CF 750 = <u>PP075889.D</u>			
CF 500 = <u>PP075890.D</u>		CF 250 = <u>PP075891.D</u>		CF 050 = <u>PP075892.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	708597432	755130533	750168220	768814916	718149060	740172032 3
Aroclor-1016-2	(2)	344688013	362255933	349033076	368003724	285681000	341932349 10
Aroclor-1016-3	(3)	184756021	198805969	201689770	204443556	237017160	205342495 9
Aroclor-1016-4	(4)	185429480	195612065	193130110	203643812	156612400	186885573 10
Aroclor-1016-5	(5)	202087874	223186380	221264342	249936892	243439240	227982946 8
Aroclor-1260-1	(1)	519115007	547143451	551523748	550034516	570658900	547695124 3
Aroclor-1260-2	(2)	816933203	847057133	836099344	841169992	727408040	813733542 6
Aroclor-1260-3	(3)	571186815	591919587	603061842	595667512	451948620	562756875 11
Aroclor-1260-4	(4)	562542796	557220415	556309062	533352684	387232680	519331527 14
Aroclor-1260-5	(5)	1443277342	1495374137	1448146362	1379476524	1011499060	1355554685 15
Decachlorobiphenyl		7970697090	8448130240	8416417320	8375691800	7903993200	8222985930 3
Tetrachloro-m-xylene		6687145430	7018651333	6804923780	6758234840	5822487600	6618288597 7
Aroclor-1242-1	(1)	571147464	588341980	597723718	569760388	546790280	574752766 3
Aroclor-1242-2	(2)	271622570	270496133	289123856	291348184	264965740	277511297 4
Aroclor-1242-3	(3)	148417684	150685240	162169734	157714788	168071120	157411713 5
Aroclor-1242-4	(4)	195275059	191450107	210824584	191270784	136740280	185112163 15
Aroclor-1242-5	(5)	229588702	236959084	244202752	234962408	205474700	230237529 6
Decachlorobiphenyl		8905080650	8993413907	9094156860	9113318400	9309338200	9083061603 2
Tetrachloro-m-xylene		7556133750	7417192413	7366207320	7201170040	6123421000	7132824905 8
Aroclor-1248-1	(1)	362937087	375551145	389160858	390611340	359825360	375617158 4
Aroclor-1248-2	(2)	226906631	241863020	243838414	257801160	259130280	245907901 5
Aroclor-1248-3	(3)	258884180	267447340	286041186	300772772	321318420	286892780 9
Aroclor-1248-4	(4)	264944699	288582168	281543548	284810560	293675780	282711351 4
Aroclor-1248-5	(5)	531974990	530976477	538260328	580030476	417051840	519658822 12
Decachlorobiphenyl		8995204870	9175718960	9269974040	9050625160	7344080200	8767120646 9
Tetrachloro-m-xylene		7343926730	7496283907	7432157100	7125361600	5416154400	6962776747 13
Aroclor-1254-1	(1)	570889099	597066183	605466528	586480320	460097860	563999998 11
Aroclor-1254-2	(2)	484331632	496825463	519240264	512723972	637282480	530080762 12
Aroclor-1254-3	(3)	866771846	893592493	892235862	873540968	729313040	851090842 8
Aroclor-1254-4	(4)	656603428	687178148	673132272	636725912	495267720	629781496 12
Aroclor-1254-5	(5)	714919063	793859644	718111150	679323440	693198580	719882375 6
Decachlorobiphenyl		9294247170	9652211800	9737198260	9430246600	10010120400	9624804846 3
Tetrachloro-m-xylene		7467434200	7759548133	7675903540	7263970760	7005463400	7434464007 4
Aroclor-1268-1	(1)	1831610079	1813456776	1858863308	1732000476	1817331660	1810652460 3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	1800363506	1772458256	1795415328	1662166904	1645167940	1735114387	4
Aroclor-1268-3	(3)	1369797429	1358701644	1402746042	1358715304	1345500120	1367092108	2
Aroclor-1268-4	(4)	542764588	536784571	540251440	501746924	465672060	517443917	6
Aroclor-1268-5	(5)	3855476951	3900359425	4013489614	3707995788	3381331260	3771730608	6
Decachlorobiphenyl		16718332370	16659738347	17037476180	16197346520	15850998600	16492778403	3
Tetrachloro-m-xylene		7499400880	7531972947	7553431460	7073074800	6855756800	7302727377	4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
Instrument ID: ECD_P **Date(s) Analyzed:** 10/21/2025 10/21/2025
GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.84	4.74	4.94	20176600
		2	4.93	4.83	5.03	15061400
		3	5.01	4.91	5.11	46915600
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	5.01	4.91	5.11	39314800
		2	5.53	5.43	5.63	20627600
		3	5.82	5.72	5.92	38639600
		4	5.98	5.88	6.08	20172600
		5	6.07	5.97	6.17	15662900
Aroclor-1262	500	1	8.23	8.13	8.33	84026000
		2	8.56	8.46	8.66	163669000
		3	8.88	8.78	8.98	121787000
		4	8.97	8.87	9.07	92111600
		5	9.64	9.54	9.74	67824600



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
Instrument ID: ECD_P **Date(s) Analyzed:** 10/21/2025 10/21/2025
GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.99	3.89	4.09	90009000
		2	4.08	3.98	4.18	63177800
		3	4.15	4.05	4.25	212036000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.15	4.05	4.25	168044000
		2	4.88	4.78	4.98	297516000
		3	5.06	4.96	5.16	81766800
		4	5.14	5.04	5.24	84943400
		5	5.31	5.21	5.41	100651000
Aroclor-1262	500	1	6.89	6.79	6.99	986952000
		2	7.15	7.05	7.25	703208000
		3	7.67	7.57	7.77	606960000
		4	7.74	7.64	7.84	1101340000
		5	8.23	8.13	8.33	461492000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075888.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 09:06
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:20:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.784	137.0E6	668.7E6	95.793	99.127
2)	SA Decachlor...	10.422	8.784	101.2E6	797.1E6	94.728	97.280
Target Compounds							
3)	L1 AR-1016-1	5.799	4.880	53893845	708.6E6	944.490	971.503
4)	L1 AR-1016-2	5.820	4.938	81387326	344.7E6	944.028	993.737
5)	L1 AR-1016-3	5.883	5.058	50025937	184.8E6	936.900	956.181
6)	L1 AR-1016-4	5.980	5.099	41815325	185.4E6	941.806	979.658
7)	L1 AR-1016-5	6.272	5.313	39075321	202.1E6	950.438	954.703
31)	L7 AR-1260-1	7.391	6.341	64655691	519.1E6	956.379	969.730
32)	L7 AR-1260-2	7.643	6.527	80744063	816.9E6	953.639	988.405
33)	L7 AR-1260-3	8.001	6.681	64964687	571.2E6	969.854	972.855
34)	L7 AR-1260-4	8.229	7.150	65448766	562.5E6	948.590	1005.572
35)	L7 AR-1260-5	8.555	7.390	139.4E6	1443.3E6	955.220	998.316



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075888.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:06
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:20:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075889.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 09:22
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:23:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.642	3.781	109.0E6	526.4E6	75.790	76.994
2) SA Decachlor...	10.419	8.782	80347709	633.6E6	75.133	76.538
Target Compounds						
3) L1 AR-1016-1	5.795	4.877	43386329	566.3E6	756.866	767.445
4) L1 AR-1016-2	5.817	4.935	64945093	271.7E6	752.204	771.869
5) L1 AR-1016-3	5.879	5.055	40120308	149.1E6	750.922	764.309
6) L1 AR-1016-4	5.977	5.096	33463154	146.7E6	752.456	766.543
7) L1 AR-1016-5	6.270	5.309	31033086	167.4E6	753.210	776.704
31) L7 AR-1260-1	7.387	6.338	50931055	410.4E6	752.240	760.963
32) L7 AR-1260-2	7.640	6.525	63926138	635.3E6	753.332	762.324
33) L7 AR-1260-3	7.999	6.678	50664437	443.9E6	754.232	754.073
34) L7 AR-1260-4	8.226	7.147	53033476	417.9E6	762.329	748.026
35) L7 AR-1260-5	8.552	7.387	109.7E6	1121.5E6	751.069	766.981



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075889.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:22
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:23:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075890.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 09:38
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:17:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.642	3.781	74528552	340.2E6	50.000	50.000
2) SA Decachlor...	10.419	8.782	56239261	420.8E6	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.795	4.878	30114384	375.1E6	500.000	500.000
4) L1 AR-1016-2	5.817	4.936	45519198	174.5E6	500.000	500.000
5) L1 AR-1016-3	5.880	5.054	28382201	100.8E6	500.000	500.000
6) L1 AR-1016-4	5.977	5.096	23491429	96565055	500.000	500.000
7) L1 AR-1016-5	6.269	5.310	21575284	110.6E6	500.000	500.000
31) L7 AR-1260-1	7.387	6.338	35276845	275.8E6	500.000	500.000
32) L7 AR-1260-2	7.640	6.525	44297376	418.0E6	500.000	500.000
33) L7 AR-1260-3	7.999	6.678	34501640	301.5E6	500.000	500.000
34) L7 AR-1260-4	8.227	7.147	36271488	278.2E6	500.000	500.000
35) L7 AR-1260-5	8.552	7.387	76226369	724.1E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075890.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:38
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:17:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075891.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 09:54
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:27:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.645	3.784	37075524	169.0E6	25.583	24.784
2) SA Decachlor...	10.422	8.784	28447228	209.4E6	26.182	25.220
Target Compounds						
3) L1 AR-1016-1	5.798	4.879	15764214	192.2E6	268.295	257.757
4) L1 AR-1016-2	5.819	4.938	23732820	92000931	268.205	258.433
5) L1 AR-1016-3	5.882	5.057	14886118	51110889	270.868	258.889
6) L1 AR-1016-4	5.979	5.098	12136196	50910953	266.787	261.815
7) L1 AR-1016-5	6.272	5.311	11195531	62484223	265.950	278.799
31) L7 AR-1260-1	7.390	6.340	18525045	137.5E6	267.300	253.727
32) L7 AR-1260-2	7.643	6.526	23496784	210.3E6	269.643	251.752
33) L7 AR-1260-3	8.001	6.680	18178927	148.9E6	265.157	252.205
34) L7 AR-1260-4	8.229	7.150	18638279	133.3E6	263.200	241.399
35) L7 AR-1260-5	8.554	7.389	38759842	344.9E6	261.402	239.232



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:54
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:27:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075892.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 10:11
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:33:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.645	3.781	6742686	29112438	4.718	4.278m
2) SA Decachlor...	10.422	8.783	5182617	39519966	4.814	4.806
Target Compounds						
3) L1 AR-1016-1	5.798	4.880	3030132	35907453	51.249	48.512
4) L1 AR-1016-2	5.819	4.943	4413506	14284050	49.902	41.774
5) L1 AR-1016-3	5.880	5.052	2846330	11850858	50.609m	58.316m
6) L1 AR-1016-4	5.978	5.094	2262026	7830620	48.653m	41.996m
7) L1 AR-1016-5	6.272	5.308	2432108	12171962	56.032	53.433m
31) L7 AR-1260-1	7.390	6.339	3328272	28532945	48.407	52.096
32) L7 AR-1260-2	7.643	6.528	4966284	36370402	55.441	44.696
33) L7 AR-1260-3	8.001	6.681	3609827	22597431	52.100	40.155
34) L7 AR-1260-4	8.229	7.150	2980037	19361634	43.459	37.282
35) L7 AR-1260-5	8.555	7.388	7332547	50574953	49.560	37.309

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075892.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:11
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:33:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075893.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 10:27
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:37:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:37:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.643	3.782	75268417	347.9E6	50.000	50.000
2) SA Decachlor...	10.419	8.781	55765598	413.6E6	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.843	3.993	10088314	45004530	500.000	500.000
9) L2 AR-1221-2	4.929	4.076	7530715	31588869	500.000	500.000
10) L2 AR-1221-3	5.005	4.154	23457783	106.0E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075893.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:27
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:37:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:37:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 10:43
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:45:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:45:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.644	3.783	76762692	367.2E6	50.000	50.000
2) SA Decachlor...	10.418	8.783	57573853	421.9E6	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	5.006	4.154	19657396	84022049	500.000	500.000
12) L3 AR-1232-2	5.532	4.877	10313847	148.8E6	500.000	500.000
13) L3 AR-1232-3	5.818	5.056	19319843	40883369	500.000	500.000
14) L3 AR-1232-4	5.978	5.136	10086255	42471740	500.000	480.688m
15) L3 AR-1232-5	6.068	5.311	7831441	50325540	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:43
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:45:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:45:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075895.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 10:59
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:57:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.784	146.7E6	755.6E6	97.370	101.273
2)	SA Decachlor...	10.422	8.784	113.6E6	890.5E6	95.988	98.950
Target Compounds							
16)	L4 AR-1242-1	5.799	4.880	44019933	571.1E6	960.145	977.263
17)	L4 AR-1242-2	5.820	4.938	64942800	271.6E6	955.587	968.789
18)	L4 AR-1242-3	5.882	5.058	40140584	148.4E6	947.491	955.722
19)	L4 AR-1242-4	5.980	5.140	33543618	195.3E6	958.879	946.968m
20)	L4 AR-1242-5	6.710	5.660	36405140	229.6E6	929.963	969.155

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075895.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:59
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:57:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075896.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 11:16
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:03:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.644	3.782	113.4E6	556.3E6	75.213	74.705
2)	SA Decachlor...	10.420	8.782	90695948	674.5E6	76.069	74.966
Target Compounds							
16)	L4 AR-1242-1	5.797	4.877	34736600	441.3E6	755.090	753.335
17)	L4 AR-1242-2	5.818	4.936	50786803	202.9E6	748.192	732.177
18)	L4 AR-1242-3	5.881	5.056	31533573	113.0E6	746.210	735.014
19)	L4 AR-1242-4	5.978	5.138	26371655	143.6E6	752.570	715.813m
20)	L4 AR-1242-5	6.708	5.658	28996717	177.7E6	743.785	750.134

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 11:16
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:03:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075897.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 11:48
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:54:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.647	3.784	77291283	368.3E6	50.000	50.000
2)	SA Decachlor...	10.422	8.785	61563948	454.7E6	50.000	50.000
Target Compounds							
16)	L4 AR-1242-1	5.799	4.880	23837208	298.9E6	498.805	500.000
17)	L4 AR-1242-2	5.821	4.938	35489778	144.6E6	500.000	500.000
18)	L4 AR-1242-3	5.884	5.059	22294823	81084867	500.000	500.000
19)	L4 AR-1242-4	5.981	5.141	18210300	105.4E6	500.000	515.181m
20)	L4 AR-1242-5	6.710	5.660	20944311	122.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075897.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 11:48
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:54:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075898.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 12:04
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:06:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.645	3.783	39880193	180.0E6	26.064	24.377
2)	SA Decachlor...	10.420	8.782	32230064	227.8E6	26.494	25.240
Target Compounds							
16)	L4 AR-1242-1	5.797	4.880	12563771	142.4E6	266.938	244.850
17)	L4 AR-1242-2	5.819	4.936	18441812	72837046	265.919	259.532
18)	L4 AR-1242-3	5.881	5.057	11783758	39428697	271.031	254.795
19)	L4 AR-1242-4	5.979	5.139	9241965	47817696	260.164	239.194m
20)	L4 AR-1242-5	6.707	5.659	9824013	58740602	248.605m	248.450

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075898.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:04
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:06:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075899.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 12:21
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:10:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.783	7865219	30617105	5.112	4.301m
2)	SA Decachlor...	10.422	8.783	6789035	46546691	5.454	5.125
Target Compounds							
16)	L4 AR-1242-1	5.798	4.882	2563442	27339514	53.509	47.567
17)	L4 AR-1242-2	5.820	4.939	3827668	13248287	54.069	47.740
18)	L4 AR-1242-3	5.883	5.054	2714945	8403556	59.484	53.386
19)	L4 AR-1242-4	5.979	5.138	1985121	6837014	54.597	36.934 #
20)	L4 AR-1242-5	6.708	5.660	2423381	10273735	59.725m	44.622 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:21
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:10:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075900.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 12:41
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:17:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.780	146.9E6	734.4E6	97.039	99.403
2) SA Decachlor...	10.426	8.783	114.1E6	899.5E6	95.681	98.496
Target Compounds						
21) L5 AR-1248-1	5.799	4.876	33471805	362.9E6	950.427	965.133
22) L5 AR-1248-2	6.071	5.095	44816839	226.9E6	951.326	964.032
23) L5 AR-1248-3	6.274	5.136	49477903	258.9E6	894.849	947.454m
24) L5 AR-1248-4	6.672	5.309	60435029	264.9E6	944.739	969.626
25) L5 AR-1248-5	6.711	5.698	58134534	532.0E6	941.586	994.127

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:41
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:17:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075901.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 12:57
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:20:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.644	3.784	114.7E6	562.2E6	75.534	75.729
2) SA Decachlor...	10.422	8.784	89301082	688.2E6	74.939	75.236
Target Compounds						
21) L5 AR-1248-1	5.797	4.879	26839141	281.7E6	758.019	749.338
22) L5 AR-1248-2	6.068	5.099	35688032	181.4E6	755.016	763.662
23) L5 AR-1248-3	6.271	5.139	39334925	200.6E6	723.821	728.506m
24) L5 AR-1248-4	6.670	5.311	48125398	216.4E6	751.539	777.551
25) L5 AR-1248-5	6.709	5.701	46284863	398.2E6	749.774	746.121

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:57
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:20:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075902.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 13:13
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:13:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.645	3.784	77916247	371.6E6	50.000	50.000
2) SA Decachlor...	10.424	8.784	62181174	463.5E6	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.798	4.877	18481760	194.6E6	500.000	500.000
22) L5 AR-1248-2	6.069	5.098	24701455	121.9E6	500.000	500.000
23) L5 AR-1248-3	6.273	5.138	30552968	143.0E6	500.000	491.112m
24) L5 AR-1248-4	6.671	5.312	33752591	140.8E6	500.000	500.000
25) L5 AR-1248-5	6.710	5.701	32673818	269.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075902.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:13
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:13:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075903.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 13:30
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:24:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.783	40662408	178.1E6	26.305	24.238
2)	SA Decachlor...	10.423	8.784	32672168	226.3E6	26.770	24.802
Target Compounds							
21)	L5 AR-1248-1	5.799	4.878	9780476	97652835	269.170	257.276
22)	L5 AR-1248-2	6.070	5.099	13110467	64450290	269.977	265.662
23)	L5 AR-1248-3	6.273	5.138	14333820	75193193	260.182	273.023m
24)	L5 AR-1248-4	6.671	5.312	18043522	71202640	273.096	254.322
25)	L5 AR-1248-5	6.710	5.701	17824378	145.0E6	277.971	265.917

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075903.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:30
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:24:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075904.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 13:46
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:29:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.645	3.780	7254273	27080772	4.751	3.914m
2) SA Decachlor...	10.422	8.783	6253432	36720401	5.099	4.188
Target Compounds						
21) L5 AR-1248-1	5.798	4.873	1700512	17991268	47.407	47.898
22) L5 AR-1248-2	6.069	5.101	2473817	12956514	50.751	52.688
23) L5 AR-1248-3	6.273	5.138	2547901	16065921	46.953	56.000
24) L5 AR-1248-4	6.669	5.313	3376964	14683789	49.218m	51.704m
25) L5 AR-1248-5	6.708	5.730	3390515	20852592	50.206m	38.783m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075904.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:46
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:29:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075905.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 14:02
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 15:57:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.783	149.9E6	746.7E6	97.273	98.623
2)	SA Decachlor...	10.424	8.784	119.5E6	929.4E6	95.942	97.673
Target Compounds							
26)	L6 AR-1254-1	6.646	5.662	66261960	570.9E6	872.670m	970.606
27)	L6 AR-1254-2	6.864	5.809	87851253	484.3E6	952.211	965.216
28)	L6 AR-1254-3	7.226	6.211	92035771	866.8E6	959.587	985.524
29)	L6 AR-1254-4	7.509	6.437	73136868	656.6E6	959.001	987.570
30)	L6 AR-1254-5	7.925	6.854	85440842	714.9E6	961.668	997.772

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075905.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:02
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 15:57:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075906.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 14:18
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:03:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.644	3.783	116.9E6	582.0E6	75.579	76.230
2)	SA Decachlor...	10.423	8.784	93392685	723.9E6	74.989	75.714
Target Compounds							
26)	L6 AR-1254-1	6.645	5.661	53923759	447.8E6	731.053m	757.518
27)	L6 AR-1254-2	6.863	5.808	69427935	372.6E6	751.680	745.041
28)	L6 AR-1254-3	7.225	6.210	72675518	670.2E6	755.137	757.967
29)	L6 AR-1254-4	7.507	6.436	57517400	515.4E6	752.790	766.592
30)	L6 AR-1254-5	7.923	6.851	67290078	595.4E6	754.900	785.342m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075906.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:18
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:03:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075907.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 14:35
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 15:24:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.785	79163213	383.8E6	50.000	50.000
2) SA Decachlor...	10.424	8.785	64802936	486.9E6	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.648	5.663	39135010	302.7E6	480.703m	500.000
27) L6 AR-1254-2	6.866	5.810	48334685	259.6E6	500.000	500.000
28) L6 AR-1254-3	7.228	6.212	49893996	446.1E6	500.000	500.000
29) L6 AR-1254-4	7.510	6.438	39695143	336.6E6	500.000	500.000
30) L6 AR-1254-5	7.926	6.854	46126096	359.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075907.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:35
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 15:24:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075908.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 14:51
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:07:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.784	40353962	181.6E6	25.803	24.079
2)	SA Decachlor...	10.422	8.784	33314499	235.8E6	26.290	24.742
Target Compounds							
26)	L6 AR-1254-1	6.647	5.662	20616230	146.6E6	278.580m	248.519
27)	L6 AR-1254-2	6.864	5.809	25702390	128.2E6	270.622	254.691
28)	L6 AR-1254-3	7.226	6.210	26497010	218.4E6	268.520	247.733
29)	L6 AR-1254-4	7.508	6.436	20991838	159.2E6	268.108	239.944
30)	L6 AR-1254-5	7.925	6.853	23753414	169.8E6	262.160	233.749

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075908.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:51
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:07:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075909.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 15:07
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:12:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.642	3.780	8604114	35027317	5.393	4.744m
2) SA Decachlor...	10.419	8.782	7574299	50050602	5.752	5.200
Target Compounds						
26) L6 AR-1254-1	6.644	5.659	5028603	23004893	64.906	40.789 #
27) L6 AR-1254-2	6.860	5.807	6001024	31864124	60.020	60.390
28) L6 AR-1254-3	7.222	6.206	6253065	36465652	60.152	42.846 #
29) L6 AR-1254-4	7.504	6.433	5244227	24763386	62.720	39.321 #
30) L6 AR-1254-5	7.921	6.850	4771352	34659929	52.106	48.147

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075909.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:07
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:12:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075910.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 15:40
 Operator : YP\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:21:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:20:51 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.785	76058633	352.0E6	50.000	50.000
2) SA Decachlor...	10.425	8.786	61011705	448.1E6	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.232	6.893	42013007	493.5E6	500.000	500.000
37) L8 AR-1262-2	8.557	7.150	81834582	351.6E6	500.000	500.000
38) L8 AR-1262-3	8.883	7.672	60893486	303.5E6	500.000	500.000
39) L8 AR-1262-4	8.971	7.736	46055772	550.7E6	500.000	500.000
40) L8 AR-1262-5	9.642	8.231	33912301	230.7E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075910.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:40
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:21:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:20:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075911.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 15:56
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:02:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.784	151.5E6	749.9E6	97.157	99.641
2) SA Decachlor...	10.424	8.785	198.8E6	1671.8E6	95.543	99.055
Target Compounds						
41) L9 AR-1268-1	8.878	7.673	186.7E6	1831.6E6	951.896	992.615
42) L9 AR-1268-2	8.973	7.737	172.5E6	1800.4E6	956.477	1001.376
43) L9 AR-1268-3	9.213	7.944	145.2E6	1369.8E6	961.447	988.116
44) L9 AR-1268-4	9.640	8.230	71036137	542.8E6	963.668	1002.321
45) L9 AR-1268-5	10.071	8.526	446.4E6	3855.5E6	968.612	979.920



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075911.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:56
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:02:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075912.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 16:12
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:06:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.783	116.2E6	564.9E6	74.674	75.037
2) SA Decachlor...	10.423	8.784	155.7E6	1249.5E6	74.863	74.351
Target Compounds						
41) L9 AR-1268-1	8.877	7.671	146.3E6	1360.1E6	747.184	741.339
42) L9 AR-1268-2	8.973	7.736	134.4E6	1329.3E6	746.666	742.894
43) L9 AR-1268-3	9.211	7.943	112.7E6	1019.0E6	747.385	739.990
44) L9 AR-1268-4	9.640	8.229	56104264	402.6E6	757.366	745.626
45) L9 AR-1268-5	10.070	8.525	344.2E6	2925.3E6	747.861	745.651



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075912.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:12
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:06:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075913.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 16:29
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:59:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.785	80186895	377.7E6	50.000	50.000
2) SA Decachlor...	10.424	8.785	108.7E6	851.9E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.878	7.673	102.8E6	929.4E6	500.000	500.000
42) L9 AR-1268-2	8.973	7.737	94092052	897.7E6	500.000	500.000
43) L9 AR-1268-3	9.213	7.943	78447140	701.4E6	500.000	500.000
44) L9 AR-1268-4	9.641	8.229	38196242	270.1E6	500.000	500.000
45) L9 AR-1268-5	10.070	8.526	237.7E6	2006.7E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075913.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:29
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:59:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075914.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 16:45
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:11:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.648	3.785	39777145	176.8E6	25.420	23.849
2) SA Decachlor...	10.427	8.786	55292786	404.9E6	26.176	24.316
Target Compounds						
41) L9 AR-1268-1	8.880	7.673	52505558	433.0E6	263.374	239.361
42) L9 AR-1268-2	8.975	7.737	47519594	415.5E6	260.422	236.426
43) L9 AR-1268-3	9.214	7.943	39522518	339.7E6	258.947	247.491
44) L9 AR-1268-4	9.642	8.231	19156803	125.4E6	256.397	236.500
45) L9 AR-1268-5	10.073	8.527	120.0E6	927.0E6	257.945	239.576



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075914.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:45
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:11:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075915.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 17:01
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:15:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.785	8297245	34278784	5.239	4.694
2) SA Decachlor...	10.424	8.786	11780168	79254993	5.451	4.805
Target Compounds						
41) L9 AR-1268-1	8.878	7.671	11762673	90866583	56.952	50.184
42) L9 AR-1268-2	8.973	7.736	10571240	82258397	56.152	47.408
43) L9 AR-1268-3	9.212	7.944	8270442	67275006	53.294	49.210
44) L9 AR-1268-4	9.641	8.228	4718687	23283603	59.998	44.997 #
45) L9 AR-1268-5	10.071	8.525	27065230	169.1E6	56.340	44.825



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075915.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 17:01
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:15:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075916.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 17:34
 Operator : YP\AJ
 Sample : PP102125ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:26:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.784	73895870	340.9E6	51.708	51.507
2)	SA Decachlor...	10.423	8.784	58947888	432.2E6	54.757	52.559
Target Compounds							
3)	L1 AR-1016-1	5.799	4.879	30579076	390.9E6	517.184	528.086
4)	L1 AR-1016-2	5.820	4.937	46283521	183.5E6	523.308	536.724
5)	L1 AR-1016-3	5.883	5.057	29014100	102.8E6	524.184	500.470
6)	L1 AR-1016-4	5.980	5.097	23843232	102.7E6	524.717	549.750
7)	L1 AR-1016-5	6.273	5.312	22384300	107.6E6	515.702	472.171
31)	L7 AR-1260-1	7.391	6.340	38313477	286.2E6	557.233	522.496
32)	L7 AR-1260-2	7.644	6.526	47189702	452.3E6	526.804	555.781
33)	L7 AR-1260-3	8.002	6.680	37186118	328.9E6	536.700	584.366
34)	L7 AR-1260-4	8.230	7.149	40498377	308.8E6	590.602	594.624
35)	L7 AR-1260-5	8.556	7.389	79566551	796.8E6	537.788	587.781



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075916.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 17:34
Operator : YP\AJ
Sample : PP102125ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP102125

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:26:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075917.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 18:06
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125AR1242

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/24/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 08:16:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.784	79486082	365.2E6	55.620	55.173
2) SA Decachlor...	10.423	8.785	64567530	484.9E6	59.978m	58.969
Target Compounds						
16) L4 AR-1242-1	5.799	4.879	24532478	301.7E6	512.088	524.891
17) L4 AR-1242-2	5.820	4.937	37086619	138.8E6	523.883	499.981
18) L4 AR-1242-3	5.883	5.057	22990427	78773054	503.714	500.427
19) L4 AR-1242-4	5.980	5.139	18909181	99357590	520.064	536.743m
20) L4 AR-1242-5	6.710	5.660	21676788	123.9E6	529.426	538.030

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 18:06
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1242

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 08:16:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075918.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 18:39
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125AR1248

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/24/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 08:17:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.784	77354608	375.4E6	54.128	56.717
2) SA Decachlor...	10.425	8.785	62090346	474.3E6	57.677	57.676
Target Compounds						
21) L5 AR-1248-1	5.799	4.878	18454423	193.8E6	514.472	515.973
22) L5 AR-1248-2	6.070	5.099	24718037	124.2E6	507.095	505.124
23) L5 AR-1248-3	6.273	5.138	27349933	132.7E6	504.009	462.616m
24) L5 AR-1248-4	6.671	5.312	33802442	144.9E6	509.348	512.423
25) L5 AR-1248-5	6.711	5.701	33065578	279.8E6	509.794	538.456

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075918.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 18:39
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1248

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 08:17:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075919.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 08:49
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125AR1254

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/24/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 09:02:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.654	3.784	73513288	361.3E6	51.441	54.591
2) SA Decachlor...	10.439	8.789	61994754	468.2E6	57.588	56.937
Target Compounds						
26) L6 AR-1254-1	6.657	5.662	33305659	337.3E6	416.876	597.997m#
27) L6 AR-1254-2	6.873	5.811	45460472	248.5E6	454.676	468.721
28) L6 AR-1254-3	7.235	6.214	47667437	420.5E6	458.540	494.093
29) L6 AR-1254-4	7.517	6.439	37597676	337.3E6	449.659	535.605
30) L6 AR-1254-5	7.934	6.855	44185805	348.6E6	482.532	484.302

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075919.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 08:49
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1254

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 09:02:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075920.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 09:07
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 15:40:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:38:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.648	3.782	80645176	382.7E6	50.921	52.402
2) SA Decachlor...	10.430	8.784	112.8E6	913.8E6	52.199	55.404
Target Compounds						
41) L9 AR-1268-1	8.881	7.671	106.5E6	989.5E6	515.825	546.466
42) L9 AR-1268-2	8.976	7.736	98184801	960.8E6	521.533	553.714
43) L9 AR-1268-3	9.216	7.942	81566200	749.4E6	525.609	548.157
44) L9 AR-1268-4	9.645	8.228	39672536	280.8E6	504.438	542.576
45) L9 AR-1268-5	10.076	8.525	246.0E6	2111.9E6	512.030	559.930



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 09:07
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 15:40:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:38:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 09:43

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.49	5.49	5.39	5.59	0.00
Aroclor-1016-2	(2)	5.51	5.51	5.41	5.61	0.00
Aroclor-1016-3	(3)	5.57	5.57	5.47	5.67	0.00
Aroclor-1016-4	(4)	5.67	5.67	5.57	5.77	0.00
Aroclor-1016-5	(5)	5.96	5.96	5.86	6.06	0.00
Aroclor-1260-1	(1)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-2	(2)	7.32	7.33	7.23	7.43	0.01
Aroclor-1260-3	(3)	7.68	7.68	7.58	7.78	0.00
Aroclor-1260-4	(4)	7.90	7.90	7.80	8.00	0.00
Aroclor-1260-5	(5)	8.21	8.21	8.11	8.31	0.00
Tetrachloro-m-xylene		4.35	4.35	4.25	4.45	0.00
Decachlorobiphenyl		9.93	9.93	9.83	10.03	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 09:43

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.73	4.63	4.83	0.00
Aroclor-1016-2	(2)	4.75	4.75	4.65	4.85	0.01
Aroclor-1016-3	(3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.17	5.18	5.08	5.28	0.01
Aroclor-1260-1	(1)	6.20	6.21	6.11	6.31	0.01
Aroclor-1260-2	(2)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-3	(3)	6.54	6.55	6.45	6.65	0.01
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.01
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl		8.64	8.64	8.54	8.74	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115067.D **Time Analyzed:** 09:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.486	5.387	5.587	497.130	500.000	-0.6
Aroclor-1016-2	5.508	5.408	5.608	494.520	500.000	-1.1
Aroclor-1016-3	5.570	5.470	5.670	511.420	500.000	2.3
Aroclor-1016-4	5.666	5.566	5.766	498.630	500.000	-0.3
Aroclor-1016-5	5.957	5.858	6.058	488.530	500.000	-2.3
Aroclor-1260-1	7.070	6.971	7.171	466.870	500.000	-6.6
Aroclor-1260-2	7.324	7.225	7.425	505.050	500.000	1.0
Aroclor-1260-3	7.681	7.582	7.782	501.150	500.000	0.2
Aroclor-1260-4	7.903	7.803	8.003	495.530	500.000	-0.9
Aroclor-1260-5	8.210	8.110	8.310	487.850	500.000	-2.4
Decachlorobiphenyl	9.929	9.830	10.030	49.920	50.000	-0.2
Tetrachloro-m-xylene	4.347	4.248	4.448	51.210	50.000	2.4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115067.D **Time Analyzed:** 09:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.628	4.828	486.870	500.000	-2.6
Aroclor-1016-2	4.745	4.647	4.847	490.870	500.000	-1.8
Aroclor-1016-3	4.919	4.821	5.021	496.780	500.000	-0.6
Aroclor-1016-4	4.961	4.864	5.064	496.970	500.000	-0.6
Aroclor-1016-5	5.173	5.076	5.276	493.030	500.000	-1.4
Aroclor-1260-1	6.204	6.106	6.306	467.570	500.000	-6.5
Aroclor-1260-2	6.391	6.294	6.494	462.450	500.000	-7.5
Aroclor-1260-3	6.543	6.446	6.646	467.390	500.000	-6.5
Aroclor-1260-4	7.015	6.917	7.117	475.890	500.000	-4.8
Aroclor-1260-5	7.255	7.158	7.358	467.640	500.000	-6.5
Decachlorobiphenyl	8.638	8.541	8.741	51.050	50.000	2.1
Tetrachloro-m-xylene	3.648	3.548	3.748	49.180	50.000	-1.6

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111325\
 Data File : PO115067.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 09:43
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 01:10:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.648	360.4E6	219.7E6	51.207	49.177
2) SA Decachlor...	9.929	8.638	220.6E6	199.3E6	49.917	51.051
Target Compounds						
3) L1 AR-1016-1	5.486	4.726	122.3E6	92630558	497.126	486.875
4) L1 AR-1016-2	5.508	4.745	196.4E6	133.0E6	494.515	490.870
5) L1 AR-1016-3	5.570	4.919	114.0E6	73510285	511.417	496.781
6) L1 AR-1016-4	5.666	4.961	94179649	58353134	498.635	496.970
7) L1 AR-1016-5	5.957	5.173	87055511	75292074	488.535	493.028
31) L7 AR-1260-1	7.070	6.204	142.7E6	138.5E6	466.868	467.567
32) L7 AR-1260-2	7.324	6.391	199.9E6	198.9E6	505.050	462.448
33) L7 AR-1260-3	7.681	6.543	166.3E6	164.1E6	501.154	467.391
34) L7 AR-1260-4	7.903	7.015	151.1E6	144.2E6	495.530	475.889
35) L7 AR-1260-5	8.210	7.255	347.9E6	384.2E6	487.847	467.640



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO111325\
Data File : PO115067.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 09:43
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 01:10:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 16:09

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.49	5.49	5.39	5.59	0.00
Aroclor-1016-2	(2)	5.51	5.51	5.41	5.61	0.00
Aroclor-1016-3	(3)	5.57	5.57	5.47	5.67	0.00
Aroclor-1016-4	(4)	5.67	5.67	5.57	5.77	0.01
Aroclor-1016-5	(5)	5.96	5.96	5.86	6.06	0.00
Aroclor-1260-1	(1)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-2	(2)	7.32	7.33	7.23	7.43	0.01
Aroclor-1260-3	(3)	7.68	7.68	7.58	7.78	0.00
Aroclor-1260-4	(4)	7.90	7.90	7.80	8.00	0.00
Aroclor-1260-5	(5)	8.21	8.21	8.11	8.31	0.00
Tetrachloro-m-xylene		4.35	4.35	4.25	4.45	0.00
Decachlorobiphenyl		9.93	9.93	9.83	10.03	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 11/04/2025 11/04/2025

Continuing Calib Time: 16:09

Initial Calibration Time(s): 09:39 18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.73	4.63	4.83	0.00
Aroclor-1016-2	(2)	4.75	4.75	4.65	4.85	0.01
Aroclor-1016-3	(3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.17	5.18	5.08	5.28	0.01
Aroclor-1260-1	(1)	6.20	6.21	6.11	6.31	0.01
Aroclor-1260-2	(2)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-3	(3)	6.54	6.55	6.45	6.65	0.01
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.01
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl		8.64	8.64	8.54	8.74	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115082.D **Time Analyzed:** 16:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.485	5.387	5.587	497.520	500.000	-0.5
Aroclor-1016-2	5.507	5.408	5.608	485.700	500.000	-2.9
Aroclor-1016-3	5.569	5.470	5.670	506.150	500.000	1.2
Aroclor-1016-4	5.665	5.566	5.766	492.600	500.000	-1.5
Aroclor-1016-5	5.956	5.858	6.058	478.720	500.000	-4.3
Aroclor-1260-1	7.070	6.971	7.171	469.500	500.000	-6.1
Aroclor-1260-2	7.324	7.225	7.425	504.990	500.000	1.0
Aroclor-1260-3	7.681	7.582	7.782	499.300	500.000	-0.1
Aroclor-1260-4	7.903	7.803	8.003	486.490	500.000	-2.7
Aroclor-1260-5	8.210	8.110	8.310	482.670	500.000	-3.5
Decachlorobiphenyl	9.930	9.830	10.030	48.900	50.000	-2.2
Tetrachloro-m-xylene	4.346	4.248	4.448	50.980	50.000	2.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115082.D **Time Analyzed:** 16:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.628	4.828	485.160	500.000	-3.0
Aroclor-1016-2	4.745	4.647	4.847	486.630	500.000	-2.7
Aroclor-1016-3	4.920	4.821	5.021	496.470	500.000	-0.7
Aroclor-1016-4	4.961	4.864	5.064	493.430	500.000	-1.3
Aroclor-1016-5	5.173	5.076	5.276	495.990	500.000	-0.8
Aroclor-1260-1	6.203	6.106	6.306	460.310	500.000	-7.9
Aroclor-1260-2	6.391	6.294	6.494	452.250	500.000	-9.6
Aroclor-1260-3	6.544	6.446	6.646	457.870	500.000	-8.4
Aroclor-1260-4	7.015	6.917	7.117	466.890	500.000	-6.6
Aroclor-1260-5	7.256	7.158	7.358	466.250	500.000	-6.8
Decachlorobiphenyl	8.639	8.541	8.741	50.550	50.000	1.1
Tetrachloro-m-xylene	3.648	3.548	3.748	49.510	50.000	-1.0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111325\
 Data File : PO115082.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 16:09
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 01:13:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.346	3.648	358.8E6	221.2E6	50.983	49.512
2) SA Decachlor...	9.930	8.639	216.2E6	197.3E6	48.901	50.546
Target Compounds						
3) L1 AR-1016-1	5.485	4.726	122.4E6	92304406	497.517	485.160
4) L1 AR-1016-2	5.507	4.745	192.9E6	131.8E6	485.702	486.633
5) L1 AR-1016-3	5.569	4.920	112.9E6	73464540	506.153	496.472
6) L1 AR-1016-4	5.665	4.961	93040238	57937746	492.602	493.432
7) L1 AR-1016-5	5.956	5.173	85306942	75743900	478.722	495.987
31) L7 AR-1260-1	7.070	6.203	143.5E6	136.3E6	469.496	460.307
32) L7 AR-1260-2	7.324	6.391	199.9E6	194.5E6	504.989	452.252
33) L7 AR-1260-3	7.681	6.544	165.7E6	160.8E6	499.301	457.870
34) L7 AR-1260-4	7.903	7.015	148.4E6	141.5E6	486.490	466.885
35) L7 AR-1260-5	8.210	7.256	344.2E6	383.1E6	482.667	466.249



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111325\
Data File : PO115082.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 16:09
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 01:13:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 10:22

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.81	5.82	5.72	5.92	0.01
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.01
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	7.99	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 10:22

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.30	5.31	5.21	5.41	0.01
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.67	6.68	6.58	6.78	0.01
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.77	3.78	3.68	3.88	0.01
Decachlorobiphenyl		8.77	8.78	8.68	8.88	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076483.D **Time Analyzed:** 10:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.791	5.695	5.895	555.270	500.000	11.1
Aroclor-1016-2	5.813	5.717	5.917	558.660	500.000	11.7
Aroclor-1016-3	5.876	5.780	5.980	566.670	500.000	13.3
Aroclor-1016-4	5.973	5.877	6.077	589.070	500.000	17.8
Aroclor-1016-5	6.265	6.169	6.369	581.940	500.000	16.4
Aroclor-1260-1	7.384	7.287	7.487	597.430	500.000	19.5
Aroclor-1260-2	7.637	7.540	7.740	588.990	500.000	17.8
Aroclor-1260-3	7.993	7.899	8.099	565.600	500.000	13.1
Aroclor-1260-4	8.222	8.127	8.327	587.500	500.000	17.5
Aroclor-1260-5	8.548	8.452	8.652	557.700	500.000	11.5
Decachlorobiphenyl	10.414	10.319	10.519	51.470	50.000	2.9
Tetrachloro-m-xylene	4.638	4.542	4.742	55.020	50.000	10.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076483.D **Time Analyzed:** 10:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.869	4.778	4.978	534.270	500.000	6.9
Aroclor-1016-2	4.927	4.836	5.036	547.220	500.000	9.4
Aroclor-1016-3	5.047	4.954	5.154	471.120	500.000	-5.8
Aroclor-1016-4	5.087	4.996	5.196	553.910	500.000	10.8
Aroclor-1016-5	5.301	5.210	5.410	504.950	500.000	1.0
Aroclor-1260-1	6.329	6.238	6.438	547.410	500.000	9.5
Aroclor-1260-2	6.516	6.425	6.625	570.640	500.000	14.1
Aroclor-1260-3	6.670	6.578	6.778	541.850	500.000	8.4
Aroclor-1260-4	7.138	7.047	7.247	579.430	500.000	15.9
Aroclor-1260-5	7.378	7.287	7.487	549.210	500.000	9.8
Decachlorobiphenyl	8.770	8.682	8.882	46.560	50.000	-6.9
Tetrachloro-m-xylene	3.774	3.681	3.881	54.780	50.000	9.6

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111325\
 Data File : PP076483.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 10:22
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/14/2025
 Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 00:08:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.638	3.774	78624478	362.5E6	55.017	54.777
2)	SA Decachlor...	10.414	8.770	55407228	382.8E6	51.469	46.557
Target Compounds							
3)	L1 AR-1016-1	5.791	4.869	32831234	395.4E6	555.275	534.268
4)	L1 AR-1016-2	5.813	4.927	49410315	187.1E6	558.661	547.220
5)	L1 AR-1016-3	5.876	5.047	31365591	96740435	566.667	471.117
6)	L1 AR-1016-4	5.973	5.087	26767525	103.5E6	589.071	553.908
7)	L1 AR-1016-5	6.265	5.301	25259464	115.1E6	581.941	504.947
31)	L7 AR-1260-1	7.384	6.329	41077227	299.8E6	597.429	547.410
32)	L7 AR-1260-2	7.637	6.516	52760148	464.3E6	588.990	570.636
33)	L7 AR-1260-3	7.993	6.670	39188227	304.9E6	565.596m	541.846
34)	L7 AR-1260-4	8.222	7.138	40285754	300.9E6	587.501m	579.431
35)	L7 AR-1260-5	8.548	7.378	82513162	744.5E6	557.704	549.215

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111325\
Data File : PP076483.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 10:22
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 00:08:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 16:01

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.81	5.82	5.72	5.92	0.01
Aroclor-1016-3	(3)	5.87	5.88	5.78	5.98	0.01
Aroclor-1016-4	(4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.26	6.27	6.17	6.37	0.01
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.63	7.64	7.54	7.74	0.01
Aroclor-1260-3	(3)	7.99	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 16:01

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.30	5.31	5.21	5.41	0.01
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.67	6.68	6.58	6.78	0.01
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.77	8.78	8.68	8.88	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076498.D **Time Analyzed:** 16:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.789	5.695	5.895	530.640	500.000	6.1
Aroclor-1016-2	5.811	5.717	5.917	538.090	500.000	7.6
Aroclor-1016-3	5.873	5.780	5.980	554.420	500.000	10.9
Aroclor-1016-4	5.970	5.877	6.077	563.300	500.000	12.7
Aroclor-1016-5	6.263	6.169	6.369	554.730	500.000	10.9
Aroclor-1260-1	7.381	7.287	7.487	570.430	500.000	14.1
Aroclor-1260-2	7.634	7.540	7.740	567.830	500.000	13.6
Aroclor-1260-3	7.993	7.899	8.099	588.380	500.000	17.7
Aroclor-1260-4	8.220	8.127	8.327	579.720	500.000	15.9
Aroclor-1260-5	8.545	8.452	8.652	545.360	500.000	9.1
Decachlorobiphenyl	10.408	10.319	10.519	49.930	50.000	-0.1
Tetrachloro-m-xylene	4.636	4.542	4.742	52.430	50.000	4.9



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076498.D **Time Analyzed:** 16:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.870	4.778	4.978	505.390	500.000	1.1
Aroclor-1016-2	4.927	4.836	5.036	500.680	500.000	0.1
Aroclor-1016-3	5.047	4.954	5.154	500.710	500.000	0.1
Aroclor-1016-4	5.088	4.996	5.196	509.680	500.000	1.9
Aroclor-1016-5	5.301	5.210	5.410	507.050	500.000	1.4
Aroclor-1260-1	6.329	6.238	6.438	584.480	500.000	16.9
Aroclor-1260-2	6.517	6.425	6.625	566.140	500.000	13.2
Aroclor-1260-3	6.670	6.578	6.778	548.930	500.000	9.8
Aroclor-1260-4	7.138	7.047	7.247	578.240	500.000	15.6
Aroclor-1260-5	7.379	7.287	7.487	535.110	500.000	7.0
Decachlorobiphenyl	8.770	8.682	8.882	45.660	50.000	-8.7
Tetrachloro-m-xylene	3.776	3.681	3.881	51.520	50.000	3.0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111325\
 Data File : PP076498.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 16:01
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 00:11:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.636	3.776	74929074	341.0E6	52.431	51.524
2) SA Decachlor...	10.408	8.770	53756160	375.5E6	49.935	45.661
Target Compounds						
3) L1 AR-1016-1	5.789	4.870	31374582	374.1E6	530.638	505.390
4) L1 AR-1016-2	5.811	4.927	47591222	171.2E6	538.094	500.683
5) L1 AR-1016-3	5.873	5.047	30687684	102.8E6	554.419	500.708
6) L1 AR-1016-4	5.970	5.088	25596558	95252657	563.302	509.684
7) L1 AR-1016-5	6.263	5.301	24078176	115.6E6	554.726	507.048
31) L7 AR-1260-1	7.381	6.329	39220624	320.1E6	570.427	584.482
32) L7 AR-1260-2	7.634	6.517	50864403	460.7E6	567.827	566.135
33) L7 AR-1260-3	7.993	6.670	40766755	308.9E6	588.379	548.927
34) L7 AR-1260-4	8.220	7.138	39752518	300.3E6	579.725	578.235
35) L7 AR-1260-5	8.545	7.379	80687369	725.4E6	545.363	535.113



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111325\
Data File : PP076498.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 16:01
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 00:11:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: ENTACT

SDG No.: Q3618

Project: Whittaker Coatings Site – E9125B

Instrument ID: ECD_O

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 11/04/2025

11/04/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	11/04/2025	09:21	PO114870.D	9.93	4.35
AR1660ICC1000	AR1660ICC1000	11/04/2025	09:39	PO114871.D	9.93	4.35
AR1660ICC750	AR1660ICC750	11/04/2025	09:58	PO114872.D	9.93	4.35
AR1660ICC500	AR1660ICC500	11/04/2025	10:16	PO114873.D	9.93	4.35
AR1660ICC250	AR1660ICC250	11/04/2025	10:34	PO114874.D	9.93	4.35
AR1660ICC050	AR1660ICC050	11/04/2025	10:53	PO114875.D	9.93	4.35
AR1221ICC500	AR1221ICC500	11/04/2025	11:11	PO114876.D	9.93	4.35
AR1232ICC500	AR1232ICC500	11/04/2025	11:30	PO114877.D	9.93	4.35
AR1242ICC1000	AR1242ICC1000	11/04/2025	11:48	PO114878.D	9.93	4.35
AR1242ICC750	AR1242ICC750	11/04/2025	12:06	PO114879.D	9.93	4.35
AR1242ICC500	AR1242ICC500	11/04/2025	12:25	PO114880.D	9.93	4.35
AR1242ICC250	AR1242ICC250	11/04/2025	12:42	PO114881.D	9.93	4.35
AR1242ICC050	AR1242ICC050	11/04/2025	13:00	PO114882.D	9.93	4.35
AR1248ICC1000	AR1248ICC1000	11/04/2025	13:19	PO114883.D	9.93	4.35
AR1248ICC750	AR1248ICC750	11/04/2025	13:37	PO114884.D	9.93	4.35
AR1248ICC500	AR1248ICC500	11/04/2025	13:56	PO114885.D	9.93	4.35
AR1248ICC250	AR1248ICC250	11/04/2025	14:14	PO114886.D	9.93	4.35
AR1248ICC050	AR1248ICC050	11/04/2025	14:33	PO114887.D	9.93	4.35
AR1254ICC1000	AR1254ICC1000	11/04/2025	14:51	PO114888.D	9.93	4.35
AR1254ICC750	AR1254ICC750	11/04/2025	15:09	PO114889.D	9.93	4.35
AR1254ICC500	AR1254ICC500	11/04/2025	15:28	PO114890.D	9.93	4.35
AR1254ICC250	AR1254ICC250	11/04/2025	16:04	PO114891.D	9.93	4.35
AR1254ICC050	AR1254ICC050	11/04/2025	16:23	PO114892.D	9.93	4.35
AR1262ICC500	AR1262ICC500	11/04/2025	17:00	PO114893.D	9.95	4.37
AR1268ICC1000	AR1268ICC1000	11/04/2025	17:18	PO114894.D	9.93	4.35
AR1268ICC750	AR1268ICC750	11/04/2025	17:36	PO114895.D	9.93	4.35
AR1268ICC500	AR1268ICC500	11/04/2025	17:55	PO114896.D	9.93	4.35
AR1268ICC250	AR1268ICC250	11/04/2025	18:13	PO114897.D	9.93	4.35
AR1268ICC050	AR1268ICC050	11/04/2025	18:31	PO114898.D	9.93	4.35
AR1660CCC500	AR1660CCC500	11/13/2025	09:43	PO115067.D	9.93	4.35
IBLK	IBLK	11/13/2025	11:05	PO115071.D	9.93	4.35
PB170526BL	PB170526BL	11/13/2025	12:11	PO115072.D	9.95	4.35
PB170526BS	PB170526BS	11/13/2025	12:29	PO115073.D	9.93	4.35
AU-713-COMP-03MS	Q3609-07MS	11/13/2025	13:42	PO115077.D	9.93	4.35
AU-713-COMP-03MSD	Q3609-07MSD	11/13/2025	14:00	PO115078.D	9.93	4.35
AR1660CCC500	AR1660CCC500	11/13/2025	16:09	PO115082.D	9.93	4.35
IBLK	IBLK	11/13/2025	17:59	PO115086.D	9.93	4.35
IBLK	IBLK	10/21/2025	08:49	PP075887.D	10.42	4.64
AR1660ICC1000	AR1660ICC1000	10/21/2025	09:06	PP075888.D	10.42	4.65
AR1660ICC750	AR1660ICC750	10/21/2025	09:22	PP075889.D	10.42	4.64
AR1660ICC500	AR1660ICC500	10/21/2025	09:38	PP075890.D	10.42	4.64
AR1660ICC250	AR1660ICC250	10/21/2025	09:54	PP075891.D	10.42	4.65

Analytical Sequence

AR1660ICC050	AR1660ICC050	10/21/2025	10:11	PP075892.D	10.42	4.65
AR1221ICC500	AR1221ICC500	10/21/2025	10:27	PP075893.D	10.42	4.64
AR1232ICC500	AR1232ICC500	10/21/2025	10:43	PP075894.D	10.42	4.64
AR1242ICC1000	AR1242ICC1000	10/21/2025	10:59	PP075895.D	10.42	4.65
AR1242ICC750	AR1242ICC750	10/21/2025	11:16	PP075896.D	10.42	4.64
AR1242ICC500	AR1242ICC500	10/21/2025	11:48	PP075897.D	10.42	4.65
AR1242ICC250	AR1242ICC250	10/21/2025	12:04	PP075898.D	10.42	4.65
AR1242ICC050	AR1242ICC050	10/21/2025	12:21	PP075899.D	10.42	4.65
AR1248ICC1000	AR1248ICC1000	10/21/2025	12:41	PP075900.D	10.43	4.65
AR1248ICC750	AR1248ICC750	10/21/2025	12:57	PP075901.D	10.42	4.64
AR1248ICC500	AR1248ICC500	10/21/2025	13:13	PP075902.D	10.42	4.65
AR1248ICC250	AR1248ICC250	10/21/2025	13:30	PP075903.D	10.42	4.65
AR1248ICC050	AR1248ICC050	10/21/2025	13:46	PP075904.D	10.42	4.65
AR1254ICC1000	AR1254ICC1000	10/21/2025	14:02	PP075905.D	10.42	4.65
AR1254ICC750	AR1254ICC750	10/21/2025	14:18	PP075906.D	10.42	4.64
AR1254ICC500	AR1254ICC500	10/21/2025	14:35	PP075907.D	10.42	4.65
AR1254ICC250	AR1254ICC250	10/21/2025	14:51	PP075908.D	10.42	4.65
AR1254ICC050	AR1254ICC050	10/21/2025	15:07	PP075909.D	10.42	4.64
AR1262ICC500	AR1262ICC500	10/21/2025	15:40	PP075910.D	10.43	4.65
AR1268ICC1000	AR1268ICC1000	10/21/2025	15:56	PP075911.D	10.42	4.65
AR1268ICC750	AR1268ICC750	10/21/2025	16:12	PP075912.D	10.42	4.65
AR1268ICC500	AR1268ICC500	10/21/2025	16:29	PP075913.D	10.42	4.65
AR1268ICC250	AR1268ICC250	10/21/2025	16:45	PP075914.D	10.43	4.65
AR1268ICC050	AR1268ICC050	10/21/2025	17:01	PP075915.D	10.42	4.65
AR1660CCC500	AR1660CCC500	11/13/2025	10:22	PP076483.D	10.41	4.64
IBLK	IBLK	11/13/2025	11:32	PP076487.D	10.41	4.64
EME-TOP-SOIL	Q3618-01	11/13/2025	13:09	PP076493.D	10.41	4.64
EME-GENERAL-FILL	Q3618-02	11/13/2025	13:26	PP076494.D	10.41	4.64
AR1660CCC500	AR1660CCC500	11/13/2025	16:01	PP076498.D	10.41	4.64
IBLK	IBLK	11/13/2025	17:39	PP076502.D	10.41	4.64

Analytical Sequence

Client: ENTACT

SDG No.: Q3618

Project: Whittaker Coatings Site – E9125B

Instrument ID: ECD_O

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 11/04/2025

11/04/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	11/04/2025	09:21	PO114870.D	8.64	3.65
AR1660ICC1000	AR1660ICC1000	11/04/2025	09:39	PO114871.D	8.64	3.65
AR1660ICC750	AR1660ICC750	11/04/2025	09:58	PO114872.D	8.64	3.65
AR1660ICC500	AR1660ICC500	11/04/2025	10:16	PO114873.D	8.64	3.65
AR1660ICC250	AR1660ICC250	11/04/2025	10:34	PO114874.D	8.64	3.65
AR1660ICC050	AR1660ICC050	11/04/2025	10:53	PO114875.D	8.64	3.65
AR1221ICC500	AR1221ICC500	11/04/2025	11:11	PO114876.D	8.64	3.65
AR1232ICC500	AR1232ICC500	11/04/2025	11:30	PO114877.D	8.64	3.65
AR1242ICC1000	AR1242ICC1000	11/04/2025	11:48	PO114878.D	8.64	3.65
AR1242ICC750	AR1242ICC750	11/04/2025	12:06	PO114879.D	8.64	3.65
AR1242ICC500	AR1242ICC500	11/04/2025	12:25	PO114880.D	8.64	3.65
AR1242ICC250	AR1242ICC250	11/04/2025	12:42	PO114881.D	8.64	3.65
AR1242ICC050	AR1242ICC050	11/04/2025	13:00	PO114882.D	8.64	3.65
AR1248ICC1000	AR1248ICC1000	11/04/2025	13:19	PO114883.D	8.64	3.65
AR1248ICC750	AR1248ICC750	11/04/2025	13:37	PO114884.D	8.64	3.65
AR1248ICC500	AR1248ICC500	11/04/2025	13:56	PO114885.D	8.64	3.65
AR1248ICC250	AR1248ICC250	11/04/2025	14:14	PO114886.D	8.64	3.65
AR1248ICC050	AR1248ICC050	11/04/2025	14:33	PO114887.D	8.64	3.65
AR1254ICC1000	AR1254ICC1000	11/04/2025	14:51	PO114888.D	8.64	3.65
AR1254ICC750	AR1254ICC750	11/04/2025	15:09	PO114889.D	8.64	3.65
AR1254ICC500	AR1254ICC500	11/04/2025	15:28	PO114890.D	8.64	3.65
AR1254ICC250	AR1254ICC250	11/04/2025	16:04	PO114891.D	8.64	3.65
AR1254ICC050	AR1254ICC050	11/04/2025	16:23	PO114892.D	8.64	3.65
AR1262ICC500	AR1262ICC500	11/04/2025	17:00	PO114893.D	8.64	3.65
AR1268ICC1000	AR1268ICC1000	11/04/2025	17:18	PO114894.D	8.64	3.65
AR1268ICC750	AR1268ICC750	11/04/2025	17:36	PO114895.D	8.64	3.65
AR1268ICC500	AR1268ICC500	11/04/2025	17:55	PO114896.D	8.64	3.65
AR1268ICC250	AR1268ICC250	11/04/2025	18:13	PO114897.D	8.64	3.65
AR1268ICC050	AR1268ICC050	11/04/2025	18:31	PO114898.D	8.64	3.65
AR1660CCC500	AR1660CCC500	11/13/2025	09:43	PO115067.D	8.64	3.65
IBLK	IBLK	11/13/2025	11:05	PO115071.D	8.64	3.65
PB170526BL	PB170526BL	11/13/2025	12:11	PO115072.D	8.64	3.65
PB170526BS	PB170526BS	11/13/2025	12:29	PO115073.D	8.64	3.65
AU-713-COMP-03MS	Q3609-07MS	11/13/2025	13:42	PO115077.D	8.64	3.65
AU-713-COMP-03MSD	Q3609-07MSD	11/13/2025	14:00	PO115078.D	8.64	3.65
AR1660CCC500	AR1660CCC500	11/13/2025	16:09	PO115082.D	8.64	3.65
IBLK	IBLK	11/13/2025	17:59	PO115086.D	8.64	3.65
IBLK	IBLK	10/21/2025	08:49	PP075887.D	8.78	3.78
AR1660ICC1000	AR1660ICC1000	10/21/2025	09:06	PP075888.D	8.78	3.78
AR1660ICC750	AR1660ICC750	10/21/2025	09:22	PP075889.D	8.78	3.78
AR1660ICC500	AR1660ICC500	10/21/2025	09:38	PP075890.D	8.78	3.78
AR1660ICC250	AR1660ICC250	10/21/2025	09:54	PP075891.D	8.78	3.78

Analytical Sequence

AR1660ICC050	AR1660ICC050	10/21/2025	10:11	PP075892.D	8.78	3.78
AR1221ICC500	AR1221ICC500	10/21/2025	10:27	PP075893.D	8.78	3.78
AR1232ICC500	AR1232ICC500	10/21/2025	10:43	PP075894.D	8.78	3.78
AR1242ICC1000	AR1242ICC1000	10/21/2025	10:59	PP075895.D	8.78	3.78
AR1242ICC750	AR1242ICC750	10/21/2025	11:16	PP075896.D	8.78	3.78
AR1242ICC500	AR1242ICC500	10/21/2025	11:48	PP075897.D	8.79	3.78
AR1242ICC250	AR1242ICC250	10/21/2025	12:04	PP075898.D	8.78	3.78
AR1242ICC050	AR1242ICC050	10/21/2025	12:21	PP075899.D	8.78	3.78
AR1248ICC1000	AR1248ICC1000	10/21/2025	12:41	PP075900.D	8.78	3.78
AR1248ICC750	AR1248ICC750	10/21/2025	12:57	PP075901.D	8.78	3.78
AR1248ICC500	AR1248ICC500	10/21/2025	13:13	PP075902.D	8.78	3.78
AR1248ICC250	AR1248ICC250	10/21/2025	13:30	PP075903.D	8.78	3.78
AR1248ICC050	AR1248ICC050	10/21/2025	13:46	PP075904.D	8.78	3.78
AR1254ICC1000	AR1254ICC1000	10/21/2025	14:02	PP075905.D	8.78	3.78
AR1254ICC750	AR1254ICC750	10/21/2025	14:18	PP075906.D	8.78	3.78
AR1254ICC500	AR1254ICC500	10/21/2025	14:35	PP075907.D	8.79	3.79
AR1254ICC250	AR1254ICC250	10/21/2025	14:51	PP075908.D	8.78	3.78
AR1254ICC050	AR1254ICC050	10/21/2025	15:07	PP075909.D	8.78	3.78
AR1262ICC500	AR1262ICC500	10/21/2025	15:40	PP075910.D	8.79	3.79
AR1268ICC1000	AR1268ICC1000	10/21/2025	15:56	PP075911.D	8.79	3.78
AR1268ICC750	AR1268ICC750	10/21/2025	16:12	PP075912.D	8.78	3.78
AR1268ICC500	AR1268ICC500	10/21/2025	16:29	PP075913.D	8.79	3.79
AR1268ICC250	AR1268ICC250	10/21/2025	16:45	PP075914.D	8.79	3.79
AR1268ICC050	AR1268ICC050	10/21/2025	17:01	PP075915.D	8.79	3.79
AR1660CCC500	AR1660CCC500	11/13/2025	10:22	PP076483.D	8.77	3.77
IBLK	IBLK	11/13/2025	11:32	PP076487.D	8.77	3.78
EME-TOP-SOIL	Q3618-01	11/13/2025	13:09	PP076493.D	8.77	3.78
EME-GENERAL-FILL	Q3618-02	11/13/2025	13:26	PP076494.D	8.77	3.78
AR1660CCC500	AR1660CCC500	11/13/2025	16:01	PP076498.D	8.77	3.78
IBLK	IBLK	11/13/2025	17:39	PP076502.D	8.77	3.78



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170526BS

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Lab Sample ID: PB170526BS

Date(s) Analyzed: 11/13/2025 11/13/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115073.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.486	5.436	5.536	153	150	4.78
COLUMN 1	2	5.507	5.457	5.557	150		
	3	5.569	5.519	5.619	155		
	4	5.665	5.615	5.715	150		
	5	5.957	5.907	6.007	144		
	1	4.725	4.675	4.775	143		
COLUMN 2	2	4.745	4.695	4.795	144	143	
	3	4.918	4.868	4.968	146		
	4	4.961	4.911	5.011	143		
	5	5.173	5.123	5.223	141		
Aroclor-1260	1	7.071	7.021	7.121	148	141	3.61
COLUMN 1	2	7.325	7.275	7.375	158		
	3	7.683	7.633	7.733	131		
	4	7.904	7.854	7.954	136		
	5	8.211	8.161	8.261	132		
	1	6.203	6.153	6.253	143		
COLUMN 2	2	6.391	6.341	6.441	141	136	
	3	6.544	6.494	6.594	142		
	4	7.015	6.965	7.065	127		
	5	7.255	7.205	7.305	127		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

AU-713-COMP-03MS

Lab Name: Alliance
Lab Code: ACE
Lab Sample ID: Q3609-07MS
Instrument ID (1): ECD_O

Contract: ENTA05
SDG NO.: Q3618
Date(s) Analyzed: 11/13/2025 11/13/2025
Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115077.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.484	5.434	5.534	164	858	67.5
COLUMN 1	2	5.509	5.459	5.559	279		
	3	5.569	5.519	5.619	201		
	4	5.667	5.617	5.717	411		
	5	5.939	5.889	5.989	3230		
COLUMN 2	1	4.726	4.676	4.776	160	425	
	2	4.745	4.695	4.795	180		
	3	4.919	4.869	4.969	162		
	4	4.962	4.912	5.012	205		
	5	5.186	5.136	5.236	1420		
Aroclor-1260	1	7.068	7.018	7.118	156	244	26.95
COLUMN 1	2	7.319	7.269	7.369	691		
	3	7.68	7.63	7.73	125		
	4	7.902	7.852	7.952	118		
	5	8.209	8.159	8.259	131		
COLUMN 2	1	6.225	6.175	6.275	860	320	
	2	6.389	6.339	6.439	211		
	3	6.544	6.494	6.594	229		
	4	7.015	6.965	7.065	179		
	5	7.255	7.205	7.305	122		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

AU-713-COMP-03MSD

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Lab Sample ID: Q3609-07MSD

Date(s) Analyzed: 11/13/2025 11/13/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115078.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.486	5.436	5.536	166	889	72.7
COLUMN 1	2	5.511	5.461	5.561	281		
	3	5.571	5.521	5.621	204		
	4	5.67	5.62	5.72	473		
	5	5.942	5.892	5.992	3320		
	1	4.726	4.676	4.776	161		
COLUMN 2	2	4.746	4.696	4.796	181	415	
	3	4.919	4.869	4.969	163		
	4	4.963	4.913	5.013	205		
	5	5.187	5.137	5.237	1370		
Aroclor-1260	1	7.07	7.02	7.12	158	247	29.36
COLUMN 1	2	7.322	7.272	7.372	698		
	3	7.682	7.632	7.732	129		
	4	7.904	7.854	7.954	119		
	5	8.211	8.161	8.261	132		
	1	6.226	6.176	6.276	903	332	
COLUMN 2	2	6.39	6.34	6.44	215		
	3	6.544	6.494	6.594	233		
	4	7.016	6.966	7.066	181		
	5	7.256	7.206	7.306	126		



QC SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: ENTACT
Project: Whittaker Coatings Site – E9125B
Client Sample ID: PB170526BL
Lab Sample ID: PB170526BL
Analytical Method: 8082A
Sample Wt/Vol: 30.02 g Final Vol: 10000 uL
Prep Method: SW3541B Prep Date: 11/13/25

Date Collected:
Date Received:
SDG No.: Q3618
Matrix: SOIL
% Solid: 100
Test: PCB

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	3.90	U	1	3.90	17.0	ug/kg	11/13/25 12:11	PB170526
11104-28-2	Aroclor-1221	4.00	U	1	4.00	17.0	ug/kg	11/13/25 12:11	PB170526
11141-16-5	Aroclor-1232	3.70	U	1	3.70	17.0	ug/kg	11/13/25 12:11	PB170526
53469-21-9	Aroclor-1242	4.00	U	1	4.00	17.0	ug/kg	11/13/25 12:11	PB170526
12672-29-6	Aroclor-1248	5.90	U	1	5.90	17.0	ug/kg	11/13/25 12:11	PB170526
11097-69-1	Aroclor-1254	3.20	U	1	3.20	17.0	ug/kg	11/13/25 12:11	PB170526
37324-23-5	Aroclor-1262	5.00	U	1	5.00	17.0	ug/kg	11/13/25 12:11	PB170526
11100-14-4	Aroclor-1268	3.60	U	1	3.60	17.0	ug/kg	11/13/25 12:11	PB170526
11096-82-5	Aroclor-1260	3.20	U	1	3.20	17.0	ug/kg	11/13/25 12:11	PB170526
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.9			21 - 165	109%	SPK: 20		
2051-24-3	Decachlorobiphenyl	23.6			10 - 170	118%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111325\
Data File : PO115072.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 12:11
Operator : YP/AJ
Sample : PB170526BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB170526BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 01:11:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.353	3.647	154.0E6	93448046	21.876	20.921
2) SA Decachlor...	9.948	8.642	99924613	92242776	22.606	23.629

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111325\
Data File : PO115072.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 12:11
Operator : YP/AJ
Sample : PB170526BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB170526BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 01:11:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.353 min
Delta R.T.: 0.005 min
Response: 153969340
Conc: 21.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB170526BL




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: 0.000 min
Response: 93448046
Conc: 20.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.948 min
Delta R.T.: 0.018 min
Response: 99924613
Conc: 22.61 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.642 min
Delta R.T.: 0.002 min
Response: 92242776
Conc: 23.63 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT		Date Collected:	11/04/25
Project:	Whittaker Coatings Site – E9125B		Date Received:	11/04/25
Client Sample ID:	PIBLK-PO114870.D		SDG No.:	Q3618
Lab Sample ID:	I.BLK-PO114870.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/04/25	po110425
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/04/25	po110425
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/04/25	po110425
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/04/25	po110425
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/04/25	po110425
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/04/25	po110425
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/04/25	po110425
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/04/25	po110425
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/04/25	po110425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	15.5			60 - 140	77%	SPK: 20		
2051-24-3	Decachlorobiphenyl	16.8			60 - 140	84%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114870.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:21
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 02:16:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 02:14:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	108.8E6	72481407	15.462	16.227
2) SA Decachlor...	9.930	8.641	74180497	67480996	16.782	17.286

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114870.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:21
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 02:16:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 02:14:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.348 min
Delta R.T.: 0.000 min
Response: 108823881
Conc: 15.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.649 min
Delta R.T.: 0.000 min
Response: 72481407
Conc: 16.23 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.930 min
Delta R.T.: 0.000 min
Response: 74180497
Conc: 16.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.641 min
Delta R.T.: 0.000 min
Response: 67480996
Conc: 17.29 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT		Date Collected:	11/13/25
Project:	Whittaker Coatings Site – E9125B		Date Received:	11/13/25
Client Sample ID:	PIBLK-PO115071.D		SDG No.:	Q3618
Lab Sample ID:	I.BLK-PO115071.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/13/25	po111325
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/13/25	po111325
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/13/25	po111325
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/13/25	po111325
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/13/25	po111325
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/13/25	po111325
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/13/25	po111325
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/13/25	po111325
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/13/25	po111325
SURROGATES									
877-09-8	Tetrachloro-m-xylene	17.3			60 - 140	86%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.4			60 - 140	87%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111325\
Data File : PO115071.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 11:05
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 01:11:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.648	122.0E6	77275300	17.329	17.300
2) SA Decachlor...	9.931	8.639	77092943	72300925	17.441	18.521

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111325\
Data File : PO115071.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 11:05
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 01:11:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.347 min
Delta R.T.: -0.001 min
Response: 121967058
Conc: 17.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: 0.000 min
Response: 77275300
Conc: 17.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.931 min
Delta R.T.: 0.000 min
Response: 77092943
Conc: 17.44 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.639 min
Delta R.T.: -0.002 min
Response: 72300925
Conc: 18.52 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT		Date Collected:	11/13/25
Project:	Whittaker Coatings Site – E9125B		Date Received:	11/13/25
Client Sample ID:	PIBLK-PO115086.D		SDG No.:	Q3618
Lab Sample ID:	I.BLK-PO115086.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/13/25	po111325
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/13/25	po111325
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/13/25	po111325
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/13/25	po111325
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/13/25	po111325
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/13/25	po111325
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/13/25	po111325
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/13/25	po111325
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/13/25	po111325
SURROGATES									
877-09-8	Tetrachloro-m-xylene	14.6			60 - 140	73%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.0			60 - 140	85%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111325\
 Data File : PO115086.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 17:59
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 01:14:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.346	3.647	118.1E6	65403500	16.786	14.642
2)	SA Decachlor...	9.928	8.638	75329240	69451884	17.042	17.791

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO111325\
Data File : PO115086.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 17:59
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 01:14:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.346 min
Delta R.T.: -0.002 min
Response: 118140157
Conc: 16.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: 0.000 min
Response: 65403500
Conc: 14.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.928 min
Delta R.T.: -0.002 min
Response: 75329240
Conc: 17.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.638 min
Delta R.T.: -0.002 min
Response: 69451884
Conc: 17.79 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT		Date Collected:	10/21/25
Project:	Whittaker Coatings Site – E9125B		Date Received:	10/21/25
Client Sample ID:	PIBLK-PP075887.D		SDG No.:	Q3618
Lab Sample ID:	I.BLK-PP075887.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/21/25	PP102125
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/21/25	PP102125
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/21/25	PP102125
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/21/25	PP102125
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/21/25	PP102125
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/21/25	PP102125
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/21/25	PP102125
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/21/25	PP102125
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/21/25	PP102125
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.8			60 - 140	94%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.3			60 - 140	96%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075887.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 08:49
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:26:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.644	3.783	28503905	124.7E6	19.945	18.843
2) SA Decachlor...	10.419	8.784	21974278	158.6E6	20.412m	19.281

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075887.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 08:49
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
I.BLK

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:26:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP075887.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 4.644 min
Delta R.T.: 0.002 min
Response: 28503905
Conc: 19.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025


Time Response_ Signal: PP075887.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.783 min
Delta R.T.: 0.002 min
Response: 124709992
Conc: 18.84 ng/ml


Time Response_ Signal: PP075887.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.419 min
Delta R.T.: 0.000 min
Response: 21974278
Conc: 20.41 ng/ml m

Time Response_ Signal: PP075887.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.784 min
Delta R.T.: 0.002 min
Response: 158550080
Conc: 19.28 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT		Date Collected:	11/13/25
Project:	Whittaker Coatings Site – E9125B		Date Received:	11/13/25
Client Sample ID:	PIBLK-PP076487.D		SDG No.:	Q3618
Lab Sample ID:	I.BLK-PP076487.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/13/25	pp111325
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/13/25	pp111325
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/13/25	pp111325
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/13/25	pp111325
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/13/25	pp111325
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/13/25	pp111325
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/13/25	pp111325
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/13/25	pp111325
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/13/25	pp111325
SURROGATES									
877-09-8	Tetrachloro-m-xylene	20.7			60 - 140	103%	SPK: 20		
2051-24-3	Decachlorobiphenyl	18.7			60 - 140	94%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111325\
Data File : PP076487.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 11:32
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 00:09:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.639	3.776	32351316	136.9E6	22.638	20.679
2)	SA Decachlor...	10.412	8.772	24504701	154.1E6	22.763	18.734

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111325\
Data File : PP076487.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 11:32
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 00:09:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.639 min
Delta R.T.: -0.003 min
Response: 32351316
Conc: 22.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.776 min
Delta R.T.: -0.006 min
Response: 136861084
Conc: 20.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.412 min
Delta R.T.: -0.007 min
Response: 24504701
Conc: 22.76 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.772 min
Delta R.T.: -0.010 min
Response: 154050323
Conc: 18.73 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT		Date Collected:	11/13/25
Project:	Whittaker Coatings Site – E9125B		Date Received:	11/13/25
Client Sample ID:	PIBLK-PP076502.D		SDG No.:	Q3618
Lab Sample ID:	I.BLK-PP076502.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/13/25	pp111325
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/13/25	pp111325
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/13/25	pp111325
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/13/25	pp111325
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/13/25	pp111325
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/13/25	pp111325
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/13/25	pp111325
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/13/25	pp111325
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/13/25	pp111325
SURROGATES									
877-09-8	Tetrachloro-m-xylene	19.1			60 - 140	96%	SPK: 20		
2051-24-3	Decachlorobiphenyl	18.2			60 - 140	91%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111325\
 Data File : PP076502.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 17:39
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 00:12:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.638	3.776	31632188	126.8E6	22.134	19.154
2) SA Decachlor...	10.408	8.771	23285266	149.5E6	21.630	18.177

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111325\
Data File : PP076502.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 17:39
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 00:12:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.638 min
Delta R.T.: -0.004 min
Response: 31632188
Conc: 22.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.776 min
Delta R.T.: -0.005 min
Response: 126769942
Conc: 19.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.408 min
Delta R.T.: -0.011 min
Response: 23285266
Conc: 21.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.771 min
Delta R.T.: -0.011 min
Response: 149467477
Conc: 18.18 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: ENTACT
Project: Whittaker Coatings Site – E9125B
Client Sample ID: PB170526BS
Lab Sample ID: PB170526BS
Analytical Method: 8082A
Sample Wt/Vol: 30.01 g Final Vol: 10000 uL
Prep Method: SW3541B Prep Date: 11/13/25

Date Collected:
Date Received:
SDG No.: Q3618
Matrix: SOIL
% Solid: 100
Test: PCB

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	150		1	3.90	17.0	ug/kg	11/13/25 12:29	PB170526
11104-28-2	Aroclor-1221	4.00	U	1	4.00	17.0	ug/kg	11/13/25 12:29	PB170526
11141-16-5	Aroclor-1232	3.70	U	1	3.70	17.0	ug/kg	11/13/25 12:29	PB170526
53469-21-9	Aroclor-1242	4.00	U	1	4.00	17.0	ug/kg	11/13/25 12:29	PB170526
12672-29-6	Aroclor-1248	5.90	U	1	5.90	17.0	ug/kg	11/13/25 12:29	PB170526
11097-69-1	Aroclor-1254	3.20	U	1	3.20	17.0	ug/kg	11/13/25 12:29	PB170526
37324-23-5	Aroclor-1262	5.00	U	1	5.00	17.0	ug/kg	11/13/25 12:29	PB170526
11100-14-4	Aroclor-1268	3.60	U	1	3.60	17.0	ug/kg	11/13/25 12:29	PB170526
11096-82-5	Aroclor-1260	141		1	3.20	17.0	ug/kg	11/13/25 12:29	PB170526
SURROGATES									
877-09-8	Tetrachloro-m-xylene	23.6			21 - 165	118%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.0			10 - 170	125%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111325\
 Data File : PO115073.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 12:29
 Operator : YP/AJ
 Sample : PB170526BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB170526BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 01:11:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.346	3.647	166.2E6	97482905	23.617	21.824
2) SA Decachlor...	9.934	8.639	104.7E6	97616114	23.685	25.005
Target Compounds						
3) L1 AR-1016-1	5.486	4.725	113.0E6	81463219	459.120	428.178
4) L1 AR-1016-2	5.507	4.745	178.2E6	116.7E6	448.685	430.776
5) L1 AR-1016-3	5.569	4.918	103.9E6	64790427	465.849	437.853
6) L1 AR-1016-4	5.665	4.961	84978812	50295122	449.921	428.343
7) L1 AR-1016-5	5.957	5.173	76874762	64694758	431.403	423.635
31) L7 AR-1260-1	7.071	6.203	135.6E6	127.0E6	443.738	428.851
32) L7 AR-1260-2	7.325	6.391	188.2E6	181.9E6	475.429	423.003
33) L7 AR-1260-3	7.683	6.544	130.5E6	149.6E6	393.180	425.964
34) L7 AR-1260-4	7.904	7.015	124.5E6	115.9E6	408.249	382.338
35) L7 AR-1260-5	8.211	7.255	283.0E6	313.2E6	396.730	381.199



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111325\
Data File : PO115073.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 12:29
Operator : YP/AJ
Sample : PB170526BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB170526BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 01:11:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT		Date Collected:	11/11/25
Project:	Whittaker Coatings Site – E9125B		Date Received:	11/12/25
Client Sample ID:	AU-713-COMP-03MS		SDG No.:	Q3618
Lab Sample ID:	Q3609-07MS		Matrix:	SOIL
Analytical Method:	8082A		% Solid:	85.3
Sample Wt/Vol:	30.05 g	Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B	Prep Date: 11/13/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	858	EP	1	4.60	19.9	ug/kg	11/13/25 13:42	PB170526
11104-28-2	Aroclor-1221	4.70	U	1	4.70	19.9	ug/kg	11/13/25 13:42	PB170526
11141-16-5	Aroclor-1232	4.40	U	1	4.40	19.9	ug/kg	11/13/25 13:42	PB170526
53469-21-9	Aroclor-1242	4.70	U	1	4.70	19.9	ug/kg	11/13/25 13:42	PB170526
12672-29-6	Aroclor-1248	6.90	U	1	6.90	19.9	ug/kg	11/13/25 13:42	PB170526
11097-69-1	Aroclor-1254	3.80	U	1	3.80	19.9	ug/kg	11/13/25 13:42	PB170526
37324-23-5	Aroclor-1262	5.90	U	1	5.90	19.9	ug/kg	11/13/25 13:42	PB170526
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.9	ug/kg	11/13/25 13:42	PB170526
11096-82-5	Aroclor-1260	320		1	3.80	19.9	ug/kg	11/13/25 13:42	PB170526
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.8			21 - 165	109%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.5			10 - 170	97%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO111325\
 Data File : PO115077.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 13:42
 Operator : YP/AJ
 Sample : Q3609-07MS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AU-713-COMP-03MS

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/14/2025
 Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 01:12:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.345	3.648	153.2E6	94400690	21.771	21.134
2) SA Decachlor...	9.929	8.639	81653033	75968008	18.472	19.460
Target Compounds						
3) L1 AR-1016-1	5.484	4.726	103.6E6	78137953	421.167	410.700
4) L1 AR-1016-2	5.509	4.745	283.5E6	125.0E6	714.001	461.435 #
5) L1 AR-1016-3	5.569	4.919	114.9E6	61597810	515.132	416.277
6) L1 AR-1016-4	5.667	4.962	198.9E6	61811880	1053.088	526.427 #
7) L1 AR-1016-5	5.939	5.186	1477.0E6	555.0E6	8288.799	3633.975m#
31) L7 AR-1260-1	7.068	6.225	122.1E6	652.7E6	399.516	2204.234 #
32) L7 AR-1260-2	7.319	6.389	700.7E6	233.1E6	1770.027	541.985 #
33) L7 AR-1260-3	7.680	6.544	106.0E6	206.3E6	319.321	587.380 #
34) L7 AR-1260-4	7.902	7.015	92585424	139.2E6	303.539	459.306 #
35) L7 AR-1260-5	8.209	7.255	240.3E6	257.3E6	336.954	313.129

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111325\
Data File : PO115077.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 13:42
Operator : YP/AJ
Sample : Q3609-07MS
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AU-713-COMP-03MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 01:12:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT		Date Collected:	11/11/25
Project:	Whittaker Coatings Site – E9125B		Date Received:	11/12/25
Client Sample ID:	AU-713-COMP-03MSD		SDG No.:	Q3618
Lab Sample ID:	Q3609-07MSD		Matrix:	SOIL
Analytical Method:	8082A		% Solid:	85.3
Sample Wt/Vol:	30.07 g	Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B	Prep Date: 11/13/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	889	EP	1	4.60	19.9	ug/kg	11/13/25 14:00	PB170526
11104-28-2	Aroclor-1221	4.70	U	1	4.70	19.9	ug/kg	11/13/25 14:00	PB170526
11141-16-5	Aroclor-1232	4.40	U	1	4.40	19.9	ug/kg	11/13/25 14:00	PB170526
53469-21-9	Aroclor-1242	4.70	U	1	4.70	19.9	ug/kg	11/13/25 14:00	PB170526
12672-29-6	Aroclor-1248	6.90	U	1	6.90	19.9	ug/kg	11/13/25 14:00	PB170526
11097-69-1	Aroclor-1254	3.80	U	1	3.80	19.9	ug/kg	11/13/25 14:00	PB170526
37324-23-5	Aroclor-1262	5.90	U	1	5.90	19.9	ug/kg	11/13/25 14:00	PB170526
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.9	ug/kg	11/13/25 14:00	PB170526
11096-82-5	Aroclor-1260	332		1	3.80	19.9	ug/kg	11/13/25 14:00	PB170526
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.9			21 - 165	109%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.4			10 - 170	97%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO111325\
 Data File : PO115078.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 14:00
 Operator : YP/AJ
 Sample : Q3609-07MSD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AU-713-COMP-03MSD

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/14/2025
 Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 01:12:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.648	153.9E6	94138674	21.865	21.075
2) SA Decachlor...	9.931	8.640	81468610	75837531	18.431	19.426
Target Compounds						
3) L1 AR-1016-1	5.486	4.726	105.0E6	78470722	426.732	412.449
4) L1 AR-1016-2	5.511	4.746	285.9E6	125.9E6	720.039	464.824 #
5) L1 AR-1016-3	5.571	4.919	116.7E6	61809837	523.344	417.710
6) L1 AR-1016-4	5.670	4.963	229.1E6	61877398	1213.050	526.985 #
7) L1 AR-1016-5	5.942	5.187	1517.9E6	535.0E6	8518.090	3503.067m#
31) L7 AR-1260-1	7.070	6.226	124.0E6	685.8E6	405.608	2316.210 #
32) L7 AR-1260-2	7.322	6.390	708.9E6	236.7E6	1790.757	550.536 #
33) L7 AR-1260-3	7.682	6.544	109.7E6	210.1E6	330.592	598.189 #
34) L7 AR-1260-4	7.904	7.016	93094660	140.9E6	305.208	464.922 #
35) L7 AR-1260-5	8.211	7.256	241.7E6	264.6E6	338.911	322.004

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111325\
Data File : PO115078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 14:00
Operator : YP/AJ
Sample : Q3609-07MSD
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AU-713-COMP-03MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 01:12:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	po110425	Instrument	ECD_o
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC250	PO114874.D	AR-1260-1 #2	rahul	11/5/2025 8:39:44 AM	yogesh	11/5/2025 10:22:06	Peak Integrated by Software
AR1660ICC250	PO114874.D	AR-1260-2 #2	rahul	11/5/2025 8:39:44 AM	yogesh	11/5/2025 10:22:06	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1016-3 #2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1016-5 #2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-1	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-1 #2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-2 #2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-3	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-4	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1242ICC050	PO114882.D	AR-1242-5 #2	rahul	11/5/2025 8:39:51 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software
AR1254ICC050	PO114892.D	AR-1254-1	rahul	11/5/2025 8:40:06 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software
AR1254ICC050	PO114892.D	AR-1254-2	rahul	11/5/2025 8:40:06 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software

Manual Integration Report

Sequence:

po110425

Instrument

ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254ICC050	PO114892.D	AR-1254-4	rahul	11/5/2025 8:40:06 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software
AR1254ICC050	PO114892.D	Decachlorobiphenyl	rahul	11/5/2025 8:40:06 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software
AR1268ICC250	PO114897.D	AR-1268-1 #2	rahul	11/5/2025 8:40:09 AM	yogesh	11/5/2025 10:22:13	Peak Integrated by Software
AR1242ICV500	PO114900.D	AR-1242-1	rahul	11/5/2025 8:40:12 AM	yogesh	11/5/2025 10:22:14	Peak Integrated by Software
AR1242ICV500	PO114900.D	AR-1242-2	rahul	11/5/2025 8:40:12 AM	yogesh	11/5/2025 10:22:14	Peak Integrated by Software
AR1242ICV500	PO114900.D	AR-1242-3	rahul	11/5/2025 8:40:12 AM	yogesh	11/5/2025 10:22:14	Peak Integrated by Software
AR1242ICV500	PO114900.D	AR-1242-4	rahul	11/5/2025 8:40:12 AM	yogesh	11/5/2025 10:22:14	Peak Integrated by Software

Manual Integration Report

Sequence:	po111325	Instrument	ECD_o
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254CCC500	PO115070.D	AR-1254-2	rahul	11/14/2025 9:11:42 AM	Sohil	11/14/2025 4:22:16	Peak Integrated by Software
Q3609-07MS	PO115077.D	AR-1016-5 #2	rahul	11/14/2025 9:11:51 AM	Sohil	11/14/2025 4:22:26	Peak Integrated by Software
Q3609-07MSD	PO115078.D	AR-1016-5 #2	rahul	11/14/2025 9:11:56 AM	Sohil	11/14/2025 4:22:31	Peak Integrated by Software
AR1660CCC500	PO115089.D	AR-1016-1 #2	rahul	11/14/2025 9:12:06 AM	Sohil	11/14/2025 4:22:41	Peak Integrated by Software

Manual Integration Report

Sequence:	PP102125	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
I.BLK	PP075887.D	Decachlorobiphenyl	yogesh	10/22/2025 12:24:41 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-3	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-3 #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-4	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-4 #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-5 #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1232ICC500	PP075894.D	AR-1232-4 #2	yogesh	10/22/2025 12:24:44 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC100 0	PP075895.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:45 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC750	PP075896.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:46 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC500	PP075897.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:48 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC250	PP075898.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:50 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC250	PP075898.D	AR-1242-5	yogesh	10/22/2025 12:24:50 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software

Manual Integration Report

Sequence:	PP102125	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242ICC050	PP075899.D	AR-1242-5	yogesh	10/22/2025 12:24:52 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC050	PP075899.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:24:52 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC100 0	PP075900.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:53 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC750	PP075901.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:55 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC500	PP075902.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:57 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC250	PP075903.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:58 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-4	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-4 #2	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-5	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-5 #2	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC100 0	PP075905.D	AR-1254-1	yogesh	10/22/2025 12:25:02 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC750	PP075906.D	AR-1254-1	yogesh	10/22/2025 12:25:03 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software

Manual Integration Report

Sequence:

PP102125

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254ICC750	PP075906.D	AR-1254-5 #2	yogesh	10/22/2025 12:25:03 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC500	PP075907.D	AR-1254-1	yogesh	10/22/2025 12:25:05 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC250	PP075908.D	AR-1254-1	yogesh	10/22/2025 12:25:07 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC050	PP075909.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:25:09 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICV500	PP075917.D	AR-1242-4 #2	yogesh	10/24/2025 4:12:30 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICV500	PP075917.D	Decachlorobiphenyl	yogesh	10/24/2025 4:12:30 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICV500	PP075918.D	AR-1248-3 #2	yogesh	10/24/2025 4:12:35 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICV500	PP075919.D	AR-1254-1 #2	yogesh	10/24/2025 4:12:33 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software

Manual Integration Report

Sequence:	PP111325	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP076483.D	AR-1260-3	rahul	11/14/2025 9:07:44 AM	Sohil	11/14/2025 4:20:50	Peak Integrated by Software
AR1660CCC500	PP076483.D	AR-1260-4	rahul	11/14/2025 9:07:44 AM	Sohil	11/14/2025 4:20:50	Peak Integrated by Software
AR1242CCC500	PP076484.D	AR-1242-4 #2	rahul	11/14/2025 9:10:30 AM	Sohil	11/14/2025 4:21:02	Peak Integrated by Software
AR1242CCC500	PP076484.D	AR-1242-5 #2	rahul	11/14/2025 9:10:30 AM	Sohil	11/14/2025 4:21:02	Peak Integrated by Software
AR1242CCC500	PP076484.D	Tetrachloro-m-xylene #2	rahul	11/14/2025 9:10:30 AM	Sohil	11/14/2025 4:21:02	Peak Integrated by Software
AR1248CCC500	PP076485.D	AR-1248-3 #2	rahul	11/14/2025 9:10:34 AM	Sohil	11/14/2025 4:21:07	Peak Integrated by Software
AR1254CCC500	PP076486.D	AR-1254-1	rahul	11/14/2025 9:10:37 AM	Sohil	11/14/2025 4:21:11	Peak Integrated by Software
AR1242CCC500	PP076499.D	AR-1242-4 #2	rahul	11/14/2025 9:11:19 AM	Sohil	11/14/2025 4:21:40	Peak Integrated by Software
AR1248CCC500	PP076500.D	AR-1248-3 #2	rahul	11/14/2025 9:11:24 AM	Sohil	11/14/2025 4:21:44	Peak Integrated by Software
AR1254CCC500	PP076501.D	AR-1254-1	rahul	11/14/2025 9:11:28 AM	Sohil	11/14/2025 4:21:50	Peak Integrated by Software
AR1254CCC500	PP076501.D	AR-1254-5	rahul	11/14/2025 9:11:28 AM	Sohil	11/14/2025 4:21:50	Peak Integrated by Software

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO114869.D	04 Nov 2025 09:03	YP/AJ	Ok
2	I.BLK	PO114870.D	04 Nov 2025 09:21	YP/AJ	Ok
3	AR1660ICC1000	PO114871.D	04 Nov 2025 09:39	YP/AJ	Ok
4	AR1660ICC750	PO114872.D	04 Nov 2025 09:58	YP/AJ	Ok
5	AR1660ICC500	PO114873.D	04 Nov 2025 10:16	YP/AJ	Ok
6	AR1660ICC250	PO114874.D	04 Nov 2025 10:34	YP/AJ	Ok,M
7	AR1660ICC050	PO114875.D	04 Nov 2025 10:53	YP/AJ	Ok,M
8	AR1221ICC500	PO114876.D	04 Nov 2025 11:11	YP/AJ	Ok
9	AR1232ICC500	PO114877.D	04 Nov 2025 11:30	YP/AJ	Ok
10	AR1242ICC1000	PO114878.D	04 Nov 2025 11:48	YP/AJ	Ok
11	AR1242ICC750	PO114879.D	04 Nov 2025 12:06	YP/AJ	Ok
12	AR1242ICC500	PO114880.D	04 Nov 2025 12:25	YP/AJ	Ok
13	AR1242ICC250	PO114881.D	04 Nov 2025 12:42	YP/AJ	Ok
14	AR1242ICC050	PO114882.D	04 Nov 2025 13:00	YP/AJ	Ok,M
15	AR1248ICC1000	PO114883.D	04 Nov 2025 13:19	YP/AJ	Ok
16	AR1248ICC750	PO114884.D	04 Nov 2025 13:37	YP/AJ	Ok
17	AR1248ICC500	PO114885.D	04 Nov 2025 13:56	YP/AJ	Ok
18	AR1248ICC250	PO114886.D	04 Nov 2025 14:14	YP/AJ	Ok
19	AR1248ICC050	PO114887.D	04 Nov 2025 14:33	YP/AJ	Ok
20	AR1254ICC1000	PO114888.D	04 Nov 2025 14:51	YP/AJ	Ok
21	AR1254ICC750	PO114889.D	04 Nov 2025 15:09	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PO114890.D	04 Nov 2025 15:28	YP/AJ	Ok
23	AR1254ICC250	PO114891.D	04 Nov 2025 16:04	YP/AJ	Ok
24	AR1254ICC050	PO114892.D	04 Nov 2025 16:23	YP/AJ	Ok,M
25	AR1262ICC500	PO114893.D	04 Nov 2025 17:00	YP/AJ	Ok
26	AR1268ICC1000	PO114894.D	04 Nov 2025 17:18	YP/AJ	Ok
27	AR1268ICC750	PO114895.D	04 Nov 2025 17:36	YP/AJ	Ok
28	AR1268ICC500	PO114896.D	04 Nov 2025 17:55	YP/AJ	Ok
29	AR1268ICC250	PO114897.D	04 Nov 2025 18:13	YP/AJ	Ok,M
30	AR1268ICC050	PO114898.D	04 Nov 2025 18:31	YP/AJ	Ok
31	PO110425ICV500	PO114899.D	04 Nov 2025 19:08	YP/AJ	Ok
32	AR1242ICV500	PO114900.D	04 Nov 2025 20:21	YP/AJ	Ok,M
33	AR1248ICV500	PO114901.D	04 Nov 2025 20:58	YP/AJ	Ok
34	AR1254ICV500	PO114902.D	04 Nov 2025 21:35	YP/AJ	Ok
35	AR1268ICV500	PO114903.D	04 Nov 2025 22:12	YP/AJ	Ok
36	DDT ANALOG	PO114904.D	04 Nov 2025 22:48	YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO111325

Review By	rahul	Review On	11/14/2025 1:01:17 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:23:00 PM
SubDirectory	PO111325	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO115066.D	13 Nov 2025 09:24	YP/AJ	Ok
2	AR1660CCC500	PO115067.D	13 Nov 2025 09:43	YP/AJ	Ok
3	AR1242CCC500	PO115068.D	13 Nov 2025 10:11	YP/AJ	Ok
4	AR1248CCC500	PO115069.D	13 Nov 2025 10:29	YP/AJ	Ok
5	AR1254CCC500	PO115070.D	13 Nov 2025 10:47	YP/AJ	Ok,M
6	I.BLK	PO115071.D	13 Nov 2025 11:05	YP/AJ	Ok
7	PB170526BL	PO115072.D	13 Nov 2025 12:11	YP/AJ	Ok
8	PB170526BS	PO115073.D	13 Nov 2025 12:29	YP/AJ	Ok
9	Q3609-01	PO115074.D	13 Nov 2025 12:47	YP/AJ	Ok
10	Q3609-04	PO115075.D	13 Nov 2025 13:06	YP/AJ	Ok
11	Q3609-07	PO115076.D	13 Nov 2025 13:24	YP/AJ	Ok,M
12	Q3609-07MS	PO115077.D	13 Nov 2025 13:42	YP/AJ	Ok,M
13	Q3609-07MSD	PO115078.D	13 Nov 2025 14:00	YP/AJ	Ok,M
14	Q3609-10	PO115079.D	13 Nov 2025 14:18	YP/AJ	Ok
15	Q3609-13	PO115080.D	13 Nov 2025 14:37	YP/AJ	Ok
16	Q3614-01	PO115081.D	13 Nov 2025 14:55	YP/AJ	Ok
17	AR1660CCC500	PO115082.D	13 Nov 2025 16:09	YP/AJ	Ok
18	AR1242CCC500	PO115083.D	13 Nov 2025 17:04	YP/AJ	Ok
19	AR1248CCC500	PO115084.D	13 Nov 2025 17:22	YP/AJ	Ok
20	AR1254CCC500	PO115085.D	13 Nov 2025 17:41	YP/AJ	Ok
21	I.BLK	PO115086.D	13 Nov 2025 17:59	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO111325

Review By	rahul	Review On	11/14/2025 1:01:17 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:23:00 PM
SubDirectory	PO111325	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3623-02	PO115087.D	13 Nov 2025 18:18	YP/AJ	Ok
23	Q3624-01	PO115088.D	13 Nov 2025 18:36	YP/AJ	ReRun
24	AR1660CCC500	PO115089.D	13 Nov 2025 19:48	YP/AJ	Ok,M
25	AR1242CCC500	PO115090.D	13 Nov 2025 20:44	YP/AJ	Ok
26	AR1248CCC500	PO115091.D	13 Nov 2025 21:01	YP/AJ	Ok
27	AR1254CCC500	PO115092.D	13 Nov 2025 21:19	YP/AJ	Ok
28	I.BLK	PO115093.D	13 Nov 2025 21:38	YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP075886.D	21 Oct 2025 08:33	YP\AJ	Ok
2	I.BLK	PP075887.D	21 Oct 2025 08:49	YP\AJ	Ok,M
3	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06	YP\AJ	Ok
4	AR1660ICC750	PP075889.D	21 Oct 2025 09:22	YP\AJ	Ok
5	AR1660ICC500	PP075890.D	21 Oct 2025 09:38	YP\AJ	Ok
6	AR1660ICC250	PP075891.D	21 Oct 2025 09:54	YP\AJ	Ok
7	AR1660ICC050	PP075892.D	21 Oct 2025 10:11	YP\AJ	Ok,M
8	AR1221ICC500	PP075893.D	21 Oct 2025 10:27	YP\AJ	Ok
9	AR1232ICC500	PP075894.D	21 Oct 2025 10:43	YP\AJ	Ok,M
10	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59	YP\AJ	Ok,M
11	AR1242ICC750	PP075896.D	21 Oct 2025 11:16	YP\AJ	Ok,M
12	AR1242ICC500	PP075897.D	21 Oct 2025 11:48	YP\AJ	Ok,M
13	AR1242ICC250	PP075898.D	21 Oct 2025 12:04	YP\AJ	Ok,M
14	AR1242ICC050	PP075899.D	21 Oct 2025 12:21	YP\AJ	Ok,M
15	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41	YP\AJ	Ok,M
16	AR1248ICC750	PP075901.D	21 Oct 2025 12:57	YP\AJ	Ok,M
17	AR1248ICC500	PP075902.D	21 Oct 2025 13:13	YP\AJ	Ok,M
18	AR1248ICC250	PP075903.D	21 Oct 2025 13:30	YP\AJ	Ok,M
19	AR1248ICC050	PP075904.D	21 Oct 2025 13:46	YP\AJ	Ok,M
20	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02	YP\AJ	Ok,M
21	AR1254ICC750	PP075906.D	21 Oct 2025 14:18	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP075907.D	21 Oct 2025 14:35	YP\AJ	Ok,M
23	AR1254ICC250	PP075908.D	21 Oct 2025 14:51	YP\AJ	Ok,M
24	AR1254ICC050	PP075909.D	21 Oct 2025 15:07	YP\AJ	Ok,M
25	AR1262ICC500	PP075910.D	21 Oct 2025 15:40	YP\AJ	Ok
26	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56	YP\AJ	Ok
27	AR1268ICC750	PP075912.D	21 Oct 2025 16:12	YP\AJ	Ok
28	AR1268ICC500	PP075913.D	21 Oct 2025 16:29	YP\AJ	Ok
29	AR1268ICC250	PP075914.D	21 Oct 2025 16:45	YP\AJ	Ok
30	AR1268ICC050	PP075915.D	21 Oct 2025 17:01	YP\AJ	Ok
31	PP102125ICV500	PP075916.D	21 Oct 2025 17:34	YP\AJ	Ok
32	AR1242ICV500	PP075917.D	21 Oct 2025 18:06	YP\AJ	Ok,M
33	AR1248ICV500	PP075918.D	21 Oct 2025 18:39	YP\AJ	Ok,M
34	AR1254ICV500	PP075919.D	22 Oct 2025 08:49	YP\AJ	Ok,M
35	AR1268ICV500	PP075920.D	22 Oct 2025 09:07	YP\AJ	Ok
36	DDT ANALOG	PP075921.D	22 Oct 2025 09:25	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111325

Review By	rahul	Review On	11/14/2025 1:00:32 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:22:02 PM
SubDirectory	PP111325	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP076482.D	13 Nov 2025 09:20	YP\AJ	Ok
2	AR1660CCC500	PP076483.D	13 Nov 2025 10:22	YP\AJ	Ok,M
3	AR1242CCC500	PP076484.D	13 Nov 2025 10:43	YP\AJ	Ok,M
4	AR1248CCC500	PP076485.D	13 Nov 2025 10:59	YP\AJ	Ok,M
5	AR1254CCC500	PP076486.D	13 Nov 2025 11:16	YP\AJ	Ok,M
6	I.BLK	PP076487.D	13 Nov 2025 11:32	YP\AJ	Ok
7	PB170501BS	PP076488.D	13 Nov 2025 11:48	YP\AJ	Ok,M
8	Q3601-13	PP076489.D	13 Nov 2025 12:04	YP\AJ	Ok
9	Q3615-01	PP076490.D	13 Nov 2025 12:21	YP\AJ	Ok
10	Q3614-02	PP076491.D	13 Nov 2025 12:37	YP\AJ	Ok,M
11	Q3614-03	PP076492.D	13 Nov 2025 12:53	YP\AJ	Ok
12	Q3618-01	PP076493.D	13 Nov 2025 13:09	YP\AJ	Ok
13	Q3618-02	PP076494.D	13 Nov 2025 13:26	YP\AJ	Ok
14	Q3621-02	PP076495.D	13 Nov 2025 13:42	YP\AJ	ReRun
15	Q3622-01	PP076496.D	13 Nov 2025 13:58	YP\AJ	Ok,M
16	Q3623-01	PP076497.D	13 Nov 2025 14:14	YP\AJ	Ok,M
17	AR1660CCC500	PP076498.D	13 Nov 2025 16:01	YP\AJ	Ok
18	AR1242CCC500	PP076499.D	13 Nov 2025 16:50	YP\AJ	Ok,M
19	AR1248CCC500	PP076500.D	13 Nov 2025 17:06	YP\AJ	Ok,M
20	AR1254CCC500	PP076501.D	13 Nov 2025 17:23	YP\AJ	Ok,M
21	I.BLK	PP076502.D	13 Nov 2025 17:39	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114869.D	04 Nov 2025 09:03		YP/AJ	Ok
2	I.BLK	I.BLK	PO114870.D	04 Nov 2025 09:21		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO114871.D	04 Nov 2025 09:39		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO114872.D	04 Nov 2025 09:58		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO114873.D	04 Nov 2025 10:16		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO114874.D	04 Nov 2025 10:34		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO114875.D	04 Nov 2025 10:53		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO114876.D	04 Nov 2025 11:11		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO114877.D	04 Nov 2025 11:30		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO114878.D	04 Nov 2025 11:48		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO114879.D	04 Nov 2025 12:06		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO114880.D	04 Nov 2025 12:25		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO114881.D	04 Nov 2025 12:42		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO114882.D	04 Nov 2025 13:00		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO114883.D	04 Nov 2025 13:19		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO114884.D	04 Nov 2025 13:37		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO114885.D	04 Nov 2025 13:56		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO114886.D	04 Nov 2025 14:14		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PO114887.D	04 Nov 2025 14:33		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO114888.D	04 Nov 2025 14:51		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO114889.D	04 Nov 2025 15:09		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO114890.D	04 Nov 2025 15:28		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO114891.D	04 Nov 2025 16:04		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO114892.D	04 Nov 2025 16:23		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO114893.D	04 Nov 2025 17:00		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO114894.D	04 Nov 2025 17:18		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO114895.D	04 Nov 2025 17:36		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO114896.D	04 Nov 2025 17:55		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO114897.D	04 Nov 2025 18:13		YP/AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PO114898.D	04 Nov 2025 18:31		YP/AJ	Ok
31	PO110425ICV500	ICVPO110425	PO114899.D	04 Nov 2025 19:08		YP/AJ	Ok
32	AR1242ICV500	ICVPO110425AR1242	PO114900.D	04 Nov 2025 20:21		YP/AJ	Ok,M
33	AR1248ICV500	ICVPO110425AR1248	PO114901.D	04 Nov 2025 20:58		YP/AJ	Ok
34	AR1254ICV500	ICVPO110425AR1254	PO114902.D	04 Nov 2025 21:35		YP/AJ	Ok
35	AR1268ICV500	ICVPO110425AR1268	PO114903.D	04 Nov 2025 22:12		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO114904.D	04 Nov 2025 22:48		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO111325

Review By	rahul	Review On	11/14/2025 1:01:17 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:23:00 PM
SubDirectory	PO111325	HP Acquire Method	HP Processing Method PO110425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO115066.D	13 Nov 2025 09:24		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO115067.D	13 Nov 2025 09:43		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO115068.D	13 Nov 2025 10:11		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO115069.D	13 Nov 2025 10:29		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO115070.D	13 Nov 2025 10:47		YP/AJ	Ok,M
6	I.BLK	I.BLK	PO115071.D	13 Nov 2025 11:05		YP/AJ	Ok
7	PB170526BL	PB170526BL	PO115072.D	13 Nov 2025 12:11		YP/AJ	Ok
8	PB170526BS	PB170526BS	PO115073.D	13 Nov 2025 12:29		YP/AJ	Ok
9	Q3609-01	AU-713-COMP-01	PO115074.D	13 Nov 2025 12:47		YP/AJ	Ok
10	Q3609-04	AU-713-COMP-02	PO115075.D	13 Nov 2025 13:06		YP/AJ	Ok
11	Q3609-07	AU-713-COMP-03	PO115076.D	13 Nov 2025 13:24		YP/AJ	Ok,M
12	Q3609-07MS	AU-713-COMP-03MS	PO115077.D	13 Nov 2025 13:42	AR1016 recovery fail	YP/AJ	Ok,M
13	Q3609-07MSD	AU-713-COMP-03MSD	PO115078.D	13 Nov 2025 14:00	AR1016 recovery fail	YP/AJ	Ok,M
14	Q3609-10	AU-713-COMP-04	PO115079.D	13 Nov 2025 14:18		YP/AJ	Ok
15	Q3609-13	AU-713-COMP-05	PO115080.D	13 Nov 2025 14:37		YP/AJ	Ok
16	Q3614-01	COMP-1	PO115081.D	13 Nov 2025 14:55		YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO115082.D	13 Nov 2025 16:09		YP/AJ	Ok
18	AR1242CCC500	AR1242CCC500	PO115083.D	13 Nov 2025 17:04		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO111325

Review By	rahul	Review On	11/14/2025 1:01:17 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:23:00 PM
SubDirectory	PO111325	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO115084.D	13 Nov 2025 17:22		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO115085.D	13 Nov 2025 17:41		YP/AJ	Ok
21	I.BLK	I.BLK	PO115086.D	13 Nov 2025 17:59		YP/AJ	Ok
22	Q3623-02	RBR200007-SILICA	PO115087.D	13 Nov 2025 18:18		YP/AJ	Ok
23	Q3624-01	OK-02-11122025	PO115088.D	13 Nov 2025 18:36	DCB high in both column	YP/AJ	ReRun
24	AR1660CCC500	AR1660CCC500	PO115089.D	13 Nov 2025 19:48		YP/AJ	Ok,M
25	AR1242CCC500	AR1242CCC500	PO115090.D	13 Nov 2025 20:44		YP/AJ	Ok
26	AR1248CCC500	AR1248CCC500	PO115091.D	13 Nov 2025 21:01		YP/AJ	Ok
27	AR1254CCC500	AR1254CCC500	PO115092.D	13 Nov 2025 21:19		YP/AJ	Ok
28	I.BLK	I.BLK	PO115093.D	13 Nov 2025 21:38		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075886.D	21 Oct 2025 08:33		YP\AJ	Ok
2	I.BLK	I.BLK	PP075887.D	21 Oct 2025 08:49		YP\AJ	Ok,M
3	AR1660ICC1000	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP075889.D	21 Oct 2025 09:22		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP075890.D	21 Oct 2025 09:38		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP075891.D	21 Oct 2025 09:54		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP075892.D	21 Oct 2025 10:11		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP075893.D	21 Oct 2025 10:27		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP075894.D	21 Oct 2025 10:43		YP\AJ	Ok,M
10	AR1242ICC1000	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59		YP\AJ	Ok,M
11	AR1242ICC750	AR1242ICC750	PP075896.D	21 Oct 2025 11:16		YP\AJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP075897.D	21 Oct 2025 11:48		YP\AJ	Ok,M
13	AR1242ICC250	AR1242ICC250	PP075898.D	21 Oct 2025 12:04		YP\AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP075899.D	21 Oct 2025 12:21		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41		YP\AJ	Ok,M
16	AR1248ICC750	AR1248ICC750	PP075901.D	21 Oct 2025 12:57		YP\AJ	Ok,M
17	AR1248ICC500	AR1248ICC500	PP075902.D	21 Oct 2025 13:13		YP\AJ	Ok,M
18	AR1248ICC250	AR1248ICC250	PP075903.D	21 Oct 2025 13:30		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PP075904.D	21 Oct 2025 13:46		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP075906.D	21 Oct 2025 14:18		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP075907.D	21 Oct 2025 14:35		YPIAJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP075908.D	21 Oct 2025 14:51		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP075909.D	21 Oct 2025 15:07		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP075910.D	21 Oct 2025 15:40		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP075912.D	21 Oct 2025 16:12		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP075913.D	21 Oct 2025 16:29		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP075914.D	21 Oct 2025 16:45		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PP075915.D	21 Oct 2025 17:01		YPIAJ	Ok
31	PP102125ICV500	ICVPP102125	PP075916.D	21 Oct 2025 17:34		YPIAJ	Ok
32	AR1242ICV500	ICVPP102125AR1242	PP075917.D	21 Oct 2025 18:06		YPIAJ	Ok,M
33	AR1248ICV500	ICVPP102125AR1248	PP075918.D	21 Oct 2025 18:39		YPIAJ	Ok,M
34	AR1254ICV500	ICVPP102125AR1254	PP075919.D	22 Oct 2025 08:49		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP102125AR1268	PP075920.D	22 Oct 2025 09:07		YPIAJ	Ok
36	DDT ANALOG	DDT ANALOG	PP075921.D	22 Oct 2025 09:25		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111325

Review By	rahul	Review On	11/14/2025 1:00:32 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:22:02 PM
SubDirectory	PP111325	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076482.D	13 Nov 2025 09:20		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076483.D	13 Nov 2025 10:22		YPIAJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP076484.D	13 Nov 2025 10:43		YPIAJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076485.D	13 Nov 2025 10:59		YPIAJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076486.D	13 Nov 2025 11:16		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP076487.D	13 Nov 2025 11:32		YPIAJ	Ok
7	PB170501BS	PB170501BS	PP076488.D	13 Nov 2025 11:48		YPIAJ	Ok,M
8	Q3601-13	TP-4	PP076489.D	13 Nov 2025 12:04	TCMX high in 2nd column	YPIAJ	Ok
9	Q3615-01	East Bathhouse- Concr	PP076490.D	13 Nov 2025 12:21		YPIAJ	Ok
10	Q3614-02	COMP-2	PP076491.D	13 Nov 2025 12:37		YPIAJ	Ok,M
11	Q3614-03	COMP-3	PP076492.D	13 Nov 2025 12:53		YPIAJ	Ok
12	Q3618-01	EME-TOP-SOIL	PP076493.D	13 Nov 2025 13:09		YPIAJ	Ok
13	Q3618-02	EME-GENERAL-FILL	PP076494.D	13 Nov 2025 13:26		YPIAJ	Ok
14	Q3621-02	VNJ-231	PP076495.D	13 Nov 2025 13:42	TCMX high in both column	YPIAJ	ReRun
15	Q3622-01	RT3071-SOLID	PP076496.D	13 Nov 2025 13:58	AR1260 Hit	YPIAJ	Ok,M
16	Q3623-01	RBR200007-SOIL	PP076497.D	13 Nov 2025 14:14	DCB high in 1st column , AR1260 Hit	YPIAJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PP076498.D	13 Nov 2025 16:01		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111325

Review By	rahul	Review On	11/14/2025 1:00:32 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:22:02 PM
SubDirectory	PP111325	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP076499.D	13 Nov 2025 16:50		YPIAJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PP076500.D	13 Nov 2025 17:06		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP076501.D	13 Nov 2025 17:23		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP076502.D	13 Nov 2025 17:39		YPIAJ	Ok

M : Manual Integration



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/13/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 11/12/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 11/13/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137869

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3609-01	AU-713-COMP-01	1	1.15	10.53	11.68	9.74	81.6	
Q3609-02	AU-713-VOC-01	2	1.18	10.57	11.75	10.37	86.9	
Q3609-04	AU-713-COMP-02	3	1.19	10.53	11.72	10.1	84.6	
Q3609-05	AU-713-VOC-02	4	1.15	10.68	11.83	9.23	75.7	
Q3609-07	AU-713-COMP-03	5	1.19	10.33	11.52	10.00	85.3	
Q3609-08	AU-713-VOC-03	6	1.16	10.90	12.06	10.44	85.1	
Q3609-10	AU-713-COMP-04	7	1.18	10.34	11.52	9.51	80.6	
Q3609-11	AU-713-VOC-04	8	1.13	10.46	11.59	10.16	86.3	
Q3609-13	AU-713-COMP-05	9	1.14	10.49	11.63	10.05	84.9	
Q3609-14	AU-713-VOC-05	10	1.13	10.40	11.53	10.00	85.3	
Q3609-15	AU-713-COMP-05	11	1.14	10.49	11.63	10.05	84.9	
Q3610-01	AU-713-TPH-01	12	1.18	10.35	11.53	10.37	88.8	
Q3610-02	AU-713-TPH-02	13	1.19	10.50	11.69	9.48	79.0	
Q3610-03	AU-713-TPH-03	14	1.11	10.33	11.44	9.75	83.6	
Q3610-04	AU-713-TPH-04	15	1.18	10.42	11.6	9.1	76.0	
Q3610-05	AU-713-TPH-05	16	1.18	10.56	11.74	10.33	86.6	
Q3610-06	AU-713-TPH-06	17	1.19	10.37	11.56	10.03	85.2	
Q3610-07	AU-713-TPH-07	18	1.15	10.40	11.55	9.08	76.3	
Q3610-08	AU-713-TPH-08	19	1.19	10.43	11.62	10.04	84.9	
Q3610-09	AU-713-TPH-09	20	1.18	10.71	11.89	9.9	81.4	
Q3610-10	AU-713-TPH-10	21	1.16	10.24	11.4	9.96	85.9	
Q3610-11	AU-713-TPH-11	22	1.14	9.99	11.13	8.96	78.3	
Q3610-12	AU-713-TPH-12	23	1.19	10.49	11.68	9.99	83.9	
Q3610-13	AU-713-TPH-13	24	1.15	10.84	11.99	10.4	85.3	
Q3610-14	AU-713-TPH-14	25	1.18	10.50	11.68	10.44	88.2	
Q3610-15	AU-713-TPH-15	26	1.14	10.48	11.62	10.34	87.8	
Q3611-01	AU-713-01	27	1.13	10.54	11.67	9.83	82.5	
Q3611-02	AU-713-02	28	1.12	11.27	12.39	9.7	76.1	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/13/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 11/12/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 11/13/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137869

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3611-03	AU-713-03	29	1.12	10.07	11.19	9.69	85.1	
Q3611-04	AU-713-04	30	1.16	10.64	11.8	9.27	76.2	
Q3611-05	AU-713-05	31	1.16	10.31	11.47	10.55	91.1	
Q3611-06	AU-713-06	32	1.18	10.70	11.88	10.14	83.7	
Q3611-07	AU-713-07	33	1.15	10.28	11.43	9.16	77.9	
Q3611-08	AU-713-08	34	1.19	10.71	11.9	10.05	82.7	
Q3611-09	AU-713-09	35	1.19	10.66	11.85	9.64	79.3	
Q3611-10	AU-713-10	36	1.17	10.13	11.3	9.79	85.1	
Q3611-11	AU-713-11	37	1.13	11.03	12.16	9.74	78.1	
Q3611-12	AU-713-12	38	1.11	10.13	11.24	9.17	79.6	
Q3611-13	AU-713-13	39	1.13	10.59	11.72	9.54	79.4	
Q3611-14	AU-713-14	40	1.16	10.23	11.39	10.02	86.6	
Q3611-15	AU-713-15	41	1.16	10.44	11.6	10.08	85.4	
Q3614-01	COMP-1	42	1.12	10.94	12.06	10.12	82.3	
Q3614-02	COMP-2	43	1.18	10.44	11.62	9.71	81.7	
Q3614-03	COMP-3	44	1.18	10.30	11.48	9.78	83.5	
Q3615-01	East Bathhouse-Concrete	45	1.00	1.00	2.00	2.00	100.0	Concreate sample
Q3618-01	EME-TOP-SOIL	46	1.18	10.33	11.51	8.44	70.3	
Q3618-02	EME-GENERAL-FILL	47	1.19	10.38	11.57	10.34	88.2	
Q3621-01	VNJ-231	48	1.16	10.38	11.54	11.06	95.4	
Q3621-02	VNJ-231	49	1.19	10.21	11.4	10.7	93.1	
Q3622-01	RT3071-SOLID	50	1.15	10.44	11.59	8.75	72.8	
Q3623-01	RBR200007-SOIL	51	1.13	10.84	11.97	10.5	86.4	
Q3623-02	RBR200007-SILICA	52	1.13	10.86	11.99	3.71	23.8	sludge sample
Q3624-01	OK-02-11122025	53	1.13	10.91	12.04	10.54	86.3	
Q3624-02	OK-02-11122025-E2	54	1.16	10.34	11.5	10.12	86.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

UP 137869

WorkList Name : %1-111225

WorkList ID : 193056

Department : Wet-Chemistry

Date : 11-12-2025 07:31:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3609-01	AU-713-COMP-01	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-02	AU-713-VOC-01	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-04	AU-713-COMP-02	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-05	AU-713-VOC-02	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-07	AU-713-COMP-03	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-08	AU-713-VOC-03	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-10	AU-713-COMP-04	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-11	AU-713-VOC-04	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-13	AU-713-COMP-05	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-14	AU-713-VOC-05	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-15	AU-713-COMP-05	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-01	AU-713-TPH-01	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-02	AU-713-TPH-02	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-03	AU-713-TPH-03	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-04	AU-713-TPH-04	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-05	AU-713-TPH-05	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-06	AU-713-TPH-06	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-07	AU-713-TPH-07	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-08	AU-713-TPH-08	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-09	AU-713-TPH-09	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-10	AU-713-TPH-10	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO

Date/Time 11-12-25 15:30

Raw Sample Received by: JP WCI

Raw Sample Relinquished by: CP

Date/Time 11-12-25

Raw Sample Received by: JP WCI

Raw Sample Relinquished by: JP WCI

WORKLIST(Hardcopy Internal Chain)

137869

WorkList Name : %1-111225

WorkList ID : 193056

Department : Wet-Chemistry

Date : 11-12-2025 07:31:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3610-11	AU-713-TPH-11	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-12	AU-713-TPH-12	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-13	AU-713-TPH-13	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-14	AU-713-TPH-14	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-15	AU-713-TPH-15	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3611-01	AU-713-01	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-02	AU-713-02	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-03	AU-713-03	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-04	AU-713-04	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-05	AU-713-05	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-06	AU-713-06	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-07	AU-713-07	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-08	AU-713-08	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-09	AU-713-09	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-10	AU-713-10	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-11	AU-713-11	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-12	AU-713-12	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-13	AU-713-13	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-14	AU-713-14	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-15	AU-713-15	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3614-01	COMP-1	Solid	Percent Solids	Cool 4 deg C	POWE02	J21	11/11/2025	Chemtech -SO

Date/Time 11-12-25 13:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 11-12-25 17:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

137869

WorkList Name : %1-111225

WorkList ID : 193056

Department : Wet-Chemistry

Date : 11-12-2025 07:31:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3614-02	COMP-2	Solid	Percent Solids	Cool 4 deg C	POWE02	J21	11/11/2025	Chemtech -SO
Q3614-03	COMP-3	Solid	Percent Solids	Cool 4 deg C	POWE02	J21	11/11/2025	Chemtech -SO
Q3615-01	East Bathroom- Concrete	Solid	Percent Solids	Cool 4 deg C	CORE02	J32	11/12/2025	Chemtech -SO
Q3618-01	EME-TOP-SOIL	Solid	Percent Solids	Cool 4 deg C	ENTA05	D41	11/12/2025	Chemtech -SO
Q3618-02	EME-GENERAL-FILL	Solid	Percent Solids	Cool 4 deg C	ENTA05	D41	11/12/2025	Chemtech -SO
Q3621-01	VNJ-231	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/12/2025	Chemtech -SO
Q3621-02	VNJ-231	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/12/2025	Chemtech -SO
Q3622-01	RT3071-SOLID	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/12/2025	Chemtech -SO
Q3623-01	RBR200007-SOIL	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/12/2025	Chemtech -SO
Q3623-02	RBR200007-SILICA	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/12/2025	Chemtech -SO
Q3624-01	OK-02-11122025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	11/12/2025	Chemtech -SO
Q3624-02	OK-02-11122025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	11/12/2025	Chemtech -SO

Date/Time 11-12-25 15:30

Raw Sample Received by: SO (WC)

Raw Sample Relinquished by: [Signature]

Date/Time 11-12-25

Raw Sample Received by: CP [Signature]

Raw Sample Relinquished by: [Signature]

1740

SO (WC)

SOP ID: M3541-ASE Extraction-15

Clean Up SOP #: Acid Cleanup

Extraction Start Date : 11/13/2025

Matrix : Solid

Extraction Start Time : 08:50

Weigh By: EH

Extraction By: RJ

Extraction End Date : 11/13/2025

Balance check: EH

Filter By: RJ

Extraction End Time : 12:00

Balance ID: EX-SC-2

pH Meter ID: N/A

Concentration By: EH

pH Strip Lot#: N/A

Hood ID: 3,7

Supervisor By : RUPESH

Extraction Method: ☐ Separatory Funnel ☐ Continuous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24930
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2660
Baked Na2SO4	N/A	EP2655
Hexane	N/A	E3981
H2SO4 1:1	N/A	EP2657
Sand	N/A	E3951
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721.

KD Bath ID: N/A

Envap ID: N-EVAP-02

KD Bath Temperature: N/A

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/13/25	RS (Ext Lab)	Y-P-PEH1P9B
12:05	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-15

Concentration Date: 11/13/2025

Sample ID	Client Sample ID	Test	g/ mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170526BL	ABLK526	PCB	30.02	N/A	ritesh	Evelyn	10			U3-1
PB170526BS	ALCS526	PCB	30.01	N/A	ritesh	Evelyn	10			2
Q3577-07	WC-1A	PCB	30.06	N/A	ritesh	Evelyn	10			3
Q3609-01	AU-713-COMP-01	PCB	30.04	N/A	ritesh	Evelyn	10	E		4
Q3609-04	AU-713-COMP-02	PCB	30.08	N/A	ritesh	Evelyn	10	E		5
Q3609-07	AU-713-COMP-03	PCB	30.03	N/A	ritesh	Evelyn	10	E		6
Q3609-07MS	AU-713-COMP-03MS	PCB	30.05	N/A	ritesh	Evelyn	10	E		U4-1
Q3609-07MS D	AU-713-COMP-03MSD	PCB	30.07	N/A	ritesh	Evelyn	10	E		2
Q3609-10	AU-713-COMP-04	PCB	30.01	N/A	ritesh	Evelyn	10	E		3
Q3609-13	AU-713-COMP-05	PCB	30.06	N/A	ritesh	Evelyn	10	E		4
Q3614-01	COMP-1	PCB Group1	30.09	N/A	ritesh	Evelyn	10	E		5
Q3614-02	COMP-2	PCB Group1	30.05	N/A	ritesh	Evelyn	10	E		6
Q3614-03	COMP-3	PCB Group1	30.03	N/A	ritesh	Evelyn	10	E		U6-1
Q3615-01	EAST BATHHOUSE- CONCRETE	PCB	30.04	N/A	ritesh	Evelyn	10	E	Concrete	2
Q3618-01	EME-TOP-SOIL	PCB	30.08	N/A	ritesh	Evelyn	10	E		3
Q3618-02	EME-GENERAL-FILL	PCB	30.02	N/A	ritesh	Evelyn	10	F		4
Q3621-02	VNJ-231	PCB	30.06	N/A	ritesh	Evelyn	10	B		5
Q3622-01	RT3071-SOLID	PCB	30.02	N/A	ritesh	Evelyn	10	A	Gel Mat.	6
Q3623-01	RBR200007-SOIL	PCB	30.04	N/A	ritesh	Evelyn	10	B		U7-1
Q3623-02	RBR200007-SILICA	PCB	30.07	N/A	ritesh	Evelyn	10		Gel Mat.	2
Q3624-01	OK-02-11122025	PCB	30.05	N/A	ritesh	Evelyn	10	E		3

* Extracts relinquished on the same date as received.

RS
11/13

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3609 WorkList ID : 193096 Department : Extraction Date : 11-13-2025 08:39:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3577-07	WC-1A	Solid	PCB	Cool 4 deg C	PSEG03		11/07/2025	8082A
Q3609-01	AU-713-COMP-01	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/11/2025	8082A
Q3609-04	AU-713-COMP-02	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/11/2025	8082A
Q3609-07	AU-713-COMP-03	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/11/2025	8082A
Q3609-10	AU-713-COMP-04	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/11/2025	8082A
Q3609-13	AU-713-COMP-05	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/11/2025	8082A
Q3614-01	COMP-1	Solid	PCB Group1	Cool 4 deg C	POWE02	J21	11/11/2025	8082A
Q3614-02	COMP-2	Solid	PCB Group1	Cool 4 deg C	POWE02	J21	11/11/2025	8082A
Q3614-03	COMP-3	Solid	PCB Group1	Cool 4 deg C	POWE02	J21	11/11/2025	8082A
Q3615-01	East Bathroom- Concrete	Solid	PCB	Cool 4 deg C	CORE02	J32	11/12/2025	8082A
Q3618-01	EME-TOP-SOIL	Solid	PCB	Cool 4 deg C	ENTA05	D41	11/12/2025	8082A
Q3618-02	EME-GENERAL-FILL	Solid	PCB	Cool 4 deg C	ENTA05	D41	11/12/2025	8082A
Q3621-02	VNJ-231	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/12/2025	8082A
Q3622-01	RT3071-SOLID	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/12/2025	8082A
Q3623-01	RBR200007-SOIL	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/12/2025	8082A
Q3623-02	RBR200007-SILICA	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/12/2025	8082A
Q3624-01	OK-02-11122025	Solid	PCB	Cool 4 deg C	PSEG05	D41	11/12/2025	8082A

Date/Time 11/13/25 8:15
Raw Sample Received by: RJ (GTHub)
Raw Sample Relinquished by: CJ

Date/Time 11/13/25 9:15
Raw Sample Received by: RJ (GTHub)
Raw Sample Relinquished by: RJ (GTHub)



SHIPPING DOCUMENTS

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:
COMPANY: ENTACT LLC
ADDRESS: 150 Bay St.
CITY Jersey City STATE: NJ ZIP: 07032
ATTENTION: Austin Farmerie
PHONE: FAX:

PROJECT NAME: Whitaker Coatings
PROJECT NO.: E9125 LOCATION: North Brunswick
PROJECT MANAGER:
e-mail:
PHONE: FAX:

BILL TO: ENTACT LLC PO#: E9125B
ADDRESS: 999 Oakmont Plaza drive 300
CITY Westmont STATE: FL ZIP: 60559
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

DATA DELIVERABLE INFORMATION

FAX (RUSH) DAYS*
HARDCOPY (DATA PACKAGE): 5 TAT DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

1. TCL VOCs 2. SVOCs, PCBs, Pest. 3. WJDER EPH 4. Herb. 5. TH. Met. Heavy 6. Chemo. - T24000 7. Hex. Chemo. 8. SEE Attached 9.

ALLIANCE
SAMPLE
ID

PROJECT
SAMPLE IDENTIFICATION

SAMPLE
MATRIX

SAMPLE
TYPE
COMP GRAB

SAMPLE
COLLECTION
DATE TIME

OF BOTTLES

PRESERVATIVES

COMMENTS

Specify Preservatives
A-HCl D-NaOH
B-HNO3 E-ICE
C-H2SO4 F-OTHER

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP GRAB	SAMPLE COLLECTION DATE TIME	# OF BOTTLES	E/F	E	E	E	E	E	E	E	E	COMMENTS
1.	EME - Top Soil	S.	X	11-12	1000	9	X	X	X	X	X	X	X	X	
2.	EME - General Fill	S.	X	11-12	1015	9	X	X	X	X	X	X	X	X	
3.															
4.															
5.															
6.															
7.															
8.															
9.															
10.															

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. [Signature]	DATE/TIME: 11-12-25	RECEIVED BY: 1. [Signature]	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP _____ °C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2.	Comments: No cooler - ROC in folder
RELINQUISHED BY SAMPLER: 3.	DATE/TIME:	RECEIVED BY: 3.	Page 1 of 1

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO

No problem, please confirm this is a bottle order request?

Can you please provide project name so we can enter.

Regards,

Jordan



Jordan Hedvat
Account Executive, Environmental Laboratories
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3144
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

From: Nathan Bennett <nbennett@entact.com>
Sent: Wednesday, November 12, 2025 8:55 AM
To: Jordan Hedvat <jordan.hedvat@alliancetg.com>
Cc: Austin Farmerie <AFarmerie@entact.com>
Subject: Fill Material Sampling - ENTACT

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Jordan-

Austin, CC'd, is collecting 2 different fill sources that we will need to perform the following tests:

- a. Target Compound List (TCL) VOCs (USEPA SW 846 Method 8260)
- b. TCL SVOCs (USEPA SW-846 Method 8270)
- c. PCBs (USEPA SW-846 Method 8082)
- d. Extractable Petroleum Hydrocarbons (EPH) (NJDEP EPH Method Version 3.0)
- e. Pesticides (USEPA SW-846 Method 8081)
- f. Herbicides (USEPA SW-846 Method 8151)
- g. Target analyte list (TAL) metals (USEPA SW-846 Method 6010)
- h. Mercury (USEPA SW846 Method 7471)
- i. Chromium - trivalent and hexavalent (USEPA SW846 Method 7196)
- j. Cyanide (USEPA SW-846 Method 9014)
- k. Per- and Polyfluoroalkyl substances (USEPA Method 537.1), and 1,4-dioxane (USEPA Method SW-846 8270d SIM)
- l. Total organic carbon (USEPA SW 846 9060 Modified)

He does not have any bottles or a COC. Last second here. Would you be able to help us out on that? He should be there sometime around 12-1.

Thanks,

From: Jordan Hedvat <Jordan.Hedvat@alliancetg.com>
Sent: Wednesday, November 12, 2025 2:02 PM
Subject: Fw: Fill Material Samling - ENTACT

Incoming today ENTA05 - Q2511072

Regards,

Jordan



Jordan Hedvat
Account Executive, Environmental Laboratories
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3144
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

From: Nathan Bennett <nbennett@entact.com>
Sent: Wednesday, November 12, 2025 12:53 PM
To: Jordan Hedvat <Jordan.Hedvat@alliancetg.com>
Cc: Austin Farmerie <AFarmerie@entact.com>
Subject: RE: Fill Material Samling - ENTACT

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Jordan-

As discussed, Austin, will be there within the hour with the material.

Project name will be Whittaker Coatings Site – E9125B.

Thanks,
Nate

From: Jordan Hedvat <Jordan.Hedvat@alliancetg.com>
Sent: Wednesday, November 12, 2025 1:08 PM
To: Nathan Bennett <nbennett@entact.com>
Cc: Austin Farmerie <AFarmerie@entact.com>
Subject: Re: Fill Material Samling - ENTACT

Hi Nathan,

No problem, please confirm this is a bottle order request?

Can you please provide project name so we can enter.

Regards,

Jordan



Jordan Hedvat
Account Executive, Environmental Laboratories
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3144
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

From: Nathan Bennett <nbennett@entact.com>
Sent: Wednesday, November 12, 2025 8:55 AM
To: Jordan Hedvat <jordan.hedvat@alliancetg.com>
Cc: Austin Farmerie <AFarmerie@entact.com>
Subject: Fill Material Samling - ENTACT

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Jordan-

Austin, CC'd, is collecting 2 different fill sources that we will need to perform the following tests:

- a. Target Compound List (TCL) VOCs (USEPA SW 846 Method 8260)
- b. TCL SVOCs (USEPA SW-846 Method 8270)
- c. PCBs (USEPA SW-846 Method 8082)
- d. Extractable Petroleum Hydrocarbons (EPH) (NJDEP EPH Method Version 3.0)
- e. Pesticides (USEPA SW-846 Method 8081)
- f. Herbicides (USEPA SW-846 Method 8151)
- g. Target analyte list (TAL) metals (USEPA SW-846 Method 6010)
- h. Mercury (USEPA SW846 Method 7471)
- i. Chromium - trivalent and hexavalent (USEPA SW846 Method 7196)
- j. Cyanide (USEPA SW-846 Method 9014)
- k. Per- and Polyfluoroalkyl substances (USEPA Method 537.1), and 1,4-dioxane (USEPA Method SW-846 8270d SIM)
- l. Total organic carbon (USEPA SW 846 9060 Modified)

He does not have any bottles or a COC. Last second here. Would you be able to help us out on that? He should be there sometime around 12-1.

Thanks,

Nate Bennett
ENTACT, LLC
703 S. Elmer St. | Suite 117 | Sayre, PA 18840

Cell: 607.743.2666

nbennett@entact.com



Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3618	ENTA05	Order Date : 11/12/2025 2:35:43 PM	Project Mgr :
Client Name : ENTACT		Project Name : Whittaker Coatings Site – E	Report Type : Level 1
Client Contact : Accounts Payable		Receive DateTime : 11/12/2025 2:25:00 PM	EDD Type : EXCEL NOCLEANUP
Invoice Name : ENTACT		Purchase Order :	Hard Copy Date :
Invoice Contact : Accounts Payable			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3618-01	EME-TOP-SOIL	Solid	11/12/2025	11:00 10:00	VOC-TCLVOA-10		8260D	5 Bus. Days	
Q3618-02	EME-GENERAL-FILL	Solid	11/12/2025	11:00 10:15	VOC-TCLVOA-10		8260D	5 Bus. Days	

Relinquished By :

Date / Time :



11-12-25

15:00

Received By :

Date / Time :



11/12/25

Storage Area : VOA Refridgerator Room

15:00