

Surrogate Summary

SDG No.: Q3618

Client: ENTACT

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		31	149
		Tetrachloro-m-xyl	1	20	19.4	97		41	134
		Decachlorobiphen	2	20	20.3	102		31	149
		Tetrachloro-m-xyl	2	20	20.0	100		41	134
I.BLK-PD091141.D	PIBLK-PD091141.D	Decachlorobiphen	1	20	21.3	107		31	149
		Tetrachloro-m-xyl	1	20	21.9	109		41	134
		Decachlorobiphen	2	20	24.7	123		31	149
		Tetrachloro-m-xyl	2	20	24.7	123		41	134
PB170527BL	PB170527BL	Decachlorobiphen	1	20	18.8	94		10	143
		Tetrachloro-m-xyl	1	20	17.8	89		10	131
		Decachlorobiphen	2	20	21.6	108		10	143
		Tetrachloro-m-xyl	2	20	20.9	104		10	131
Q3609-13MS	AU-713-COMP-05MS	Decachlorobiphen	1	20	16.6	83		10	143
		Tetrachloro-m-xyl	1	20	17.1	86		10	131
		Decachlorobiphen	2	20	19.3	96		10	143
		Tetrachloro-m-xyl	2	20	20.7	104		10	131
Q3609-13MSD	AU-713-COMP-05MSD	Decachlorobiphen	1	20	15.3	76		10	143
		Tetrachloro-m-xyl	1	20	16.5	83		10	131
		Decachlorobiphen	2	20	18.6	93		10	143
		Tetrachloro-m-xyl	2	20	19.6	98		10	131
Q3618-01	EME-TOP-SOIL	Decachlorobiphen	1	20	5.44	27		10	143
		Tetrachloro-m-xyl	1	20	9.50	48		10	131
		Decachlorobiphen	2	20	6.85	34		10	143
		Tetrachloro-m-xyl	2	20	10.7	54		10	131
Q3618-02	EME-GENERAL-FILL	Decachlorobiphen	1	20	15.2	76		10	143
		Tetrachloro-m-xyl	1	20	17.8	89		10	131
		Decachlorobiphen	2	20	18.3	92		10	143
		Tetrachloro-m-xyl	2	20	21.0	105		10	131
I.BLK-PD091160.D	PIBLK-PD091160.D	Decachlorobiphen	1	20	20.4	102		31	149
		Tetrachloro-m-xyl	1	20	20.6	103		41	134
		Decachlorobiphen	2	20	21.1	105		31	149
		Tetrachloro-m-xyl	2	20	23.4	117		41	134
I.BLK-PD091193.D	PIBLK-PD091193.D	Decachlorobiphen	1	20	20.8	104		31	149
		Tetrachloro-m-xyl	1	20	19.5	98		41	134
		Decachlorobiphen	2	20	20.1	100		31	149
		Tetrachloro-m-xyl	2	20	20.0	100		41	134
I.BLK-PD091218.D	PIBLK-PD091218.D	Decachlorobiphen	1	20	20.0	100		31	149
		Tetrachloro-m-xyl	1	20	19.6	98		41	134
		Decachlorobiphen	2	20	19.3	96		31	149
		Tetrachloro-m-xyl	2	20	22.6	113		41	134
PB170527BS	PB170527BS	Decachlorobiphen	1	20	23.5	117		10	143

Surrogate Summary

SDG No.: Q3618

Client: ENTACT

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
PB170527BS	PB170527BS	Tetrachloro-m-xyl	1	20	21.5	108		10	131
		Decachlorobiphen	2	20	24.9	124		10	143
		Tetrachloro-m-xyl	2	20	24.1	120		10	131
I.BLK-PD091234.D	PIBLK-PD091234.D	Decachlorobiphen	1	20	20.8	104		31	149
		Tetrachloro-m-xyl	1	20	20.0	100		41	134
		Decachlorobiphen	2	20	18.9	94		31	149
		Tetrachloro-m-xyl	2	20	21.4	107		41	134