

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ENTA05</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3618</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>12/01/2025 10:23</u>
Lab File ID:	<u>BN038222.D</u>	Init. Calib. Date(s):	<u>11/26/2025 11/27/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4</u>	Init. Calib. Time(s):	<u>20:35 00:13</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.568	0.586		3.2	20.0
Fluoranthene-d10	0.868	0.913		5.2	20.0
2-Fluorophenol	0.935	0.918		-1.8	20.0
Phenol-d6	1.168	1.147		-1.8	20.0
Nitrobenzene-d5	0.320	0.313		-2.2	20.0
2-Fluorobiphenyl	1.545	1.642		6.3	20.0
2,4,6-Tribromophenol	0.126	0.135		7.1	20.0
Terphenyl-d14	0.924	0.969		4.9	20.0
1,4-Dioxane	0.421	0.434		3.1	20.0

All other compounds must meet a minimum RRF of 0.010.