

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN112725

Review By	Jagrut	Review On	12/2/2025 3:36:38 PM
Supervise By		Supervise On	
SubDirectory	BN112725	HP Acquire Method	BNA_N, 8270_HP Processing Method BN112725

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6907,10ul/1000ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038211.D	26 Nov 2025 19:55		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN038212.D	26 Nov 2025 20:35		RC/JU	Ok,NS
3	SSTDICC0.2	SSTDICC0.2	BN038213.D	26 Nov 2025 21:12		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN038214.D	26 Nov 2025 21:48	Compound #20 kept on QR.	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN038215.D	26 Nov 2025 22:25		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN038216.D	26 Nov 2025 23:01		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN038217.D	26 Nov 2025 23:37		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN038218.D	27 Nov 2025 00:13	Compound #14,24 removed from 5.0 level	RC/JU	Ok
9	SSTDICV0.4	ICVBN112725	BN038219.D	27 Nov 2025 00:49	ICV failed for compound #05,20,24	RC/JU	Ok
10	PB170483BL	PB170483BL	BN038220.D	27 Nov 2025 02:01		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN120125

Review By	Jagrut	Review On	12/2/2025 3:40:28 PM
Supervise By		Supervise On	
SubDirectory	BN120125	HP Acquire Method	BNA_N, 8270_HP Processing Method BN112725

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6907,10ul/1000ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038221.D	01 Dec 2025 09:44		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038222.D	01 Dec 2025 10:23		RC/JU	Ok
3	PB170590BL	PB170590BL	BN038223.D	01 Dec 2025 10:59		RC/JU	Ok
4	Q3483-12	HW1025-PT-TRIAZINE	BN038224.D	01 Dec 2025 11:35	HW1025-PT-TRIAZINE-SOIL Need 100X Dilution	RC/JU	Dilution
5	Q3483-12DL	HW1025-PT-TRIAZINE	BN038225.D	01 Dec 2025 14:20		RC/JU	Ok
6	Q3483-13	HW1025-PT-PAH-SOIL	BN038226.D	01 Dec 2025 14:56	Need 10X Dilution	RC/JU	Dilution
7	Q3483-13DL	HW1025-PT-PAH-SOIL	BN038227.D	01 Dec 2025 15:49		RC/JU	Ok
8	SSTDCCC0.4	SSTDCCC0.4EC	BN038228.D	01 Dec 2025 16:35		RC/JU	Ok
9	DFTPP	DFTPP	BN038229.D	01 Dec 2025 17:11		RC/JU	Ok
10	SSTDCCC0.4	SSTDCCC0.4	BN038230.D	01 Dec 2025 17:51		RC/JU	Ok
11	PB170712BL	PB170712BL	BN038231.D	01 Dec 2025 18:28		RC/JU	Ok
12	PB170712BS	PB170712BS	BN038232.D	01 Dec 2025 19:04		RC/JU	Ok,NS
13	Q3618-02	EME-GENERAL-FILL	BN038233.D	01 Dec 2025 19:40	Regular Internal standard added	RC/JU	Ok
14	Q3618-02MS	EME-GENERAL-FILLM	BN038234.D	01 Dec 2025 20:17	Regular Internal standard added.	RC/JU	Ok,NS
15	Q3618-02MSD	EME-GENERAL-FILLM	BN038235.D	01 Dec 2025 20:53	Regular Internal standard added	RC/JU	Ok,NS
16	Q3618-01	EME-TOP-SOIL	BN038236.D	01 Dec 2025 21:29	Regular Internal standard added	RC/JU	Ok
17	PB170590BS	PB170590BS	BN038237.D	01 Dec 2025 22:05		RC/JU	Ok,NS

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN120125

Review By	Jagrut	Review On	12/2/2025 3:40:28 PM
Supervise By		Supervise On	
SubDirectory	BN120125	HP Acquire Method	BNA_N, 8270_HP Processing Method BN112725
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904		
CCC	SP6902		
Internal Standard/PEM	SP6907,10ul/1000ul sample		
ICV/I.BLK	SP6906		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	PB170590BSD	PB170590BSD	BN038238.D	01 Dec 2025 22:42	PB170590BS LOADED AGIN	RC/JU	Not Ok
19	Q3690-01	RW8-SP100-20251118	BN038239.D	01 Dec 2025 23:17		RC/JU	Ok
20	Q3690-02	RW8-SP303-20251118	BN038240.D	01 Dec 2025 23:54		RC/JU	Ok
21	Q3691-01	RW7-SP100-20251118	BN038241.D	02 Dec 2025 00:30		RC/JU	Ok
22	Q3691-02	RW7-SP201-20251118	BN038242.D	02 Dec 2025 01:07		RC/JU	Ok
23	Q3691-03	RW7-SP302-20251118	BN038243.D	02 Dec 2025 01:43		RC/JU	Ok
24	Q3691-04	RW7-SP303-20251118	BN038244.D	02 Dec 2025 02:19		RC/JU	Ok
25	Q3692-01	RW5-SP100-20251118	BN038245.D	02 Dec 2025 02:55		RC/JU	Ok
26	Q3692-02	RW5-SP201-20251118	BN038246.D	02 Dec 2025 03:32		RC/JU	Ok
27	Q3692-03	RW5-SP303-20251118	BN038247.D	02 Dec 2025 04:08		RC/JU	Ok
28	SSTDCCC0.4	SSTDCCC0.4EC	BN038248.D	02 Dec 2025 04:44		RC/JU	Ok

M : Manual Integration