

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3618

Analytical Method: 8270-Modified

Client: ENTACT

DataFile: BN038237.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits	
								Qual	Low	High
PB170590BS	1,4-Dioxane	13.3	10.9	ug/Kg	82				70	130