

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3618

Client ID : SSTDCCC0.4

Date Analyzed: 12/01/2025

Lab File ID: BN038222.D

Time Analyzed: 10:23

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	7356	7.768	22525	10.56	12376	14.43
UPPER LIMIT	14712	8.268	45050	11.062	24752	14.93
LOWER LIMIT	3678	7.268	11262.5	10.062	6188	13.93
EPA SAMPLE NO.						
01 PB170590BL	8151	7.77	22998	10.57	11468	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.