

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111325\
 Data File : BP026126.D
 Acq On : 13 Nov 2025 13:16
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 13 13:45:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 30 11:51:34 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	112	-0.02
2	1,4-Dioxane	40.000	39.365	1.6	109	-0.01
3	Pyridine	40.000	39.830	0.4	109	0.00
4	n-Nitrosodimethylamine	40.000	36.436	8.9	101	-0.01
5 S	2-Fluorophenol	80.000	83.320	-4.1	113	0.00
6	Aniline	40.000	39.569	1.1	108	-0.02
7 S	Phenol-d6	80.000	83.352	-4.2	112	0.00
8	2-Chlorophenol	40.000	43.328	-8.3	116	-0.01
9	Benzaldehyde	40.000	36.790	8.0	102	-0.01
10 C	Phenol	40.000	40.899	-2.2	110	-0.01
11	bis(2-Chloroethyl)ether	40.000	39.933	0.2	110	-0.02
12	1,3-Dichlorobenzene	40.000	40.647	-1.6	111	-0.02
13 C	1,4-Dichlorobenzene	40.000	39.982	0.0	110	-0.02
14	1,2-Dichlorobenzene	40.000	40.454	-1.1	111	-0.01
15	Benzyl Alcohol	40.000	40.745	-1.9	111	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.805	10.5	97	-0.01
17	2-Methylphenol	40.000	41.277	-3.2	110	0.00
18	Hexachloroethane	40.000	42.580	-6.4	115	-0.02
19 P	n-Nitroso-di-n-propylamine	40.000	40.378	-0.9	105	-0.02
20	3+4-Methylphenols	40.000	40.797	-2.0	110	-0.02
21 I	Naphthalene-d8	20.000	20.000	0.0	111	-0.02
22	Acetophenone	40.000	39.730	0.7	107	-0.02
23 S	Nitrobenzene-d5	80.000	87.521	-9.4	115	-0.02
24	Nitrobenzene	40.000	42.427	-6.1	111	-0.02
25	Isophorone	40.000	39.822	0.4	105	-0.01
26 C	2-Nitrophenol	40.000	58.213	-45.5#	159	0.00
27	2,4-Dimethylphenol	40.000	40.177	-0.4	106	-0.01
28	bis(2-Chloroethoxy)methane	40.000	39.964	0.1	107	0.00
29 C	2,4-Dichlorophenol	40.000	44.086	-10.2	114	0.00
30	1,2,4-Trichlorobenzene	40.000	40.811	-2.0	110	-0.01
31	Naphthalene	40.000	40.034	-0.1	108	-0.01
32	Benzoic acid	40.000	52.145	-30.4#	155	-0.03
33	4-Chloroaniline	40.000	40.182	-0.5	107	0.00
34 C	Hexachlorobutadiene	40.000	41.799	-4.5	113	-0.01
35	Caprolactam	40.000	42.293	-5.7	111	-0.01
36 C	4-Chloro-3-methylphenol	40.000	42.917	-7.3	110	0.00
37	2-Methylnaphthalene	40.000	39.803	0.5	106	-0.01
38	1-Methylnaphthalene	40.000	39.711	0.7	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	109	-0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	41.391	-3.5	111	-0.01
41 P	Hexachlorocyclopentadiene	40.000	44.066	-10.2	116	-0.01
42 S	2,4,6-Tribromophenol	80.000	97.739	-22.2	122	0.00
43 C	2,4,6-Trichlorophenol	40.000	46.769	-16.9	117	0.00
44	2,4,5-Trichlorophenol	40.000	45.420	-13.6	114	0.00
45 S	2-Fluorobiphenyl	80.000	79.396	0.8	106	0.00
46	1,1'-Biphenyl	40.000	39.736	0.7	106	0.00
47	2-Chloronaphthalene	40.000	40.062	-0.2	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111325\
 Data File : BP026126.D
 Acq On : 13 Nov 2025 13:16
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 13 13:45:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 30 11:51:34 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.322	-8.3	120	0.00
49	Acenaphthylene	40.000	40.472	-1.2	106	0.00
50	Dimethylphthalate	40.000	41.189	-3.0	107	0.00
51	2,6-Dinitrotoluene	40.000	45.430	-13.6	127	0.00
52 C	Acenaphthene	40.000	39.191	2.0	103	0.00
53	3-Nitroaniline	40.000	42.978	-7.4	118	0.00
54 P	2,4-Dinitrophenol	40.000	56.374	-40.9#	170	0.00
55	Dibenzofuran	40.000	39.980	0.1	106	0.00
56 P	4-Nitrophenol	40.000	43.408	-8.5	120	0.01
57	2,4-Dinitrotoluene	40.000	45.808	-14.5	127	0.00
58	Fluorene	40.000	39.955	0.1	107	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	48.499	-21.2	120	-0.01
60	Diethylphthalate	40.000	41.293	-3.2	106	0.00
61	4-Chlorophenyl-phenylether	40.000	40.200	-0.5	107	0.00
62	4-Nitroaniline	40.000	46.736	-16.8	116	0.00
63	Azobenzene	40.000	38.934	2.7	102	-0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	40.000	55.655	-39.1#	167	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.423	1.4	103	0.00
67	4-Bromophenyl-phenylether	40.000	41.140	-2.9	108	0.00
68	Hexachlorobenzene	40.000	41.793	-4.5	111	0.00
69	Atrazine	40.000	41.376	-3.4	108	0.00
70 C	Pentachlorophenol	40.000	46.341	-15.9	120	0.00
71	Phenanthrene	40.000	39.394	1.5	104	0.00
72	Anthracene	40.000	40.308	-0.8	107	0.00
73	Carbazole	40.000	40.114	-0.3	105	0.00
74	Di-n-butylphthalate	40.000	43.249	-8.1	107	0.00
75 C	Fluoranthene	40.000	40.450	-1.1	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	108	0.02
77	Benzidine	40.000	32.557	18.6	86	0.00
78	Pyrene	40.000	40.117	-0.3	102	0.01
79 S	Terphenyl-d14	80.000	81.385	-1.7	106	0.00
80	Butylbenzylphthalate	40.000	46.020	-15.1	126	0.00
81	Benzo(a)anthracene	40.000	40.235	-0.6	105	0.02
82	3,3'-Dichlorobenzidine	40.000	42.685	-6.7	110	0.00
83	Chrysene	40.000	39.577	1.1	103	0.02
84	Bis(2-ethylhexyl)phthalate	40.000	43.123	-7.8	112	0.01
85 c	Di-n-octyl phthalate	40.000	50.778	-26.9#	127	0.01
86 I	Perylene-d12	20.000	20.000	0.0	112	0.04
87	Indeno(1,2,3-cd)pyrene	40.000	41.681	-4.2	110	0.05
88	Benzo(b)fluoranthene	40.000	40.842	-2.1	109	0.00
89	Benzo(k)fluoranthene	40.000	39.681	0.8	106	0.01
90 C	Benzo(a)pyrene	40.000	41.445	-3.6	109	0.01
91	Dibenzo(a,h)anthracene	40.000	41.717	-4.3	110	0.04
92	Benzo(g,h,i)perylene	40.000	41.601	-4.0	111	0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111325\
Data File : BP026126.D
Acq On : 13 Nov 2025 13:16
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 13 13:45:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 30 11:51:34 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 2