

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3618

Client ID : SSTDCCC040

Date Analyzed: 11/13/2025

Lab File ID: BP026126.D

Time Analyzed: 13:16

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	326955	7.672	1332770	10.44	834342	14.31
UPPER LIMIT	653910	8.172	2665540	10.937	1668680	14.807
LOWER LIMIT	163478	7.172	666385	9.937	417171	13.807
EPA SAMPLE NO.						
01 EME-TOP-SOIL	330501	7.67	1294500	10.44	805747	14.31

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.