



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

## **CASE NARRATIVE**

### **Ambric Technology Corporation**

**Project Name: 6506 Haverford Road**

**Project # N/A**

**Order ID # Q3619**

**Test Name: VOCMS Group10, Metals Group10**

### **A. Number of Samples and Date of Receipt:**

6 Solid samples were received on 11/12/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: VOCMS Group10, Metals Group10. This data package contains results for VOCMS Group10(8260D), Metals Group10(6010D).

### **C. Analytical Techniques:**

VOCMS Group10 : The analysis performed on instrument MSVOA\_W were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOCMS Group10 was based on method 8260D.

Metals Group10 : The analysis of Metals Group10 was based on method 6010D and digestion based on method 3050 (soils).

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the acceptable requirements.

The RPD recoveries met criteria.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

The Duplicate analysis met criteria for all samples.

The Serial Dilution met the acceptable requirements.

### **E. Additional Comments:**

This data package has been revised to report VOC data with mg\Kg.

The soil samples results are based on a dry weight basis.



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

VOCMS Group10 : Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_