

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Ambric Technology Corporation

SDG No.: Q3619

Contract: AMBR01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3920	4000	98	90 - 110	P	11/13/2025	13:12	LB137889

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Ambric Technology Corporation

SDG No.: Q3619

Contract: AMBR01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	12.0	12.0	100	80 - 120	P	11/13/2025	13:16	LB137889

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Ambric Technology Corporation

SDG No.: Q3619

Contract: AMBR01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4930	5000	99	90 - 110	P	11/13/2025	13:46	LB137889
CCV02	Lead	4830	5000	96	90 - 110	P	11/13/2025	14:35	LB137889
CCV03	Lead	4860	5000	97	90 - 110	P	11/13/2025	15:48	LB137889
CCV04	Lead	5010	5000	100	90 - 110	P	11/13/2025	16:36	LB137889
CCV05	Lead	4860	5000	97	90 - 110	P	11/13/2025	17:27	LB137889
CCV06	Lead	4900	5000	98	90 - 110	P	11/13/2025	18:17	LB137889
CCV07	Lead	4960	5000	99	90 - 110	P	11/13/2025	18:38	LB137889