

Process Report

Order ID : Q3626

Date	Bottle ID	Lab Sample Number	InComing Vacuum	Canister ID	Can Size	Reg ID	Reg Size	Batch Code	Certification ID
11/14/2025	B2510052	Q3626-05	-10.0	10464	1.4 L	10698		C2509004	VL042989.D Seq :VL092325 TUNE : VL042982.D ICAL : VL042955.D CCAL : VL042983.D MB : VL042984.D
11/14/2025	B2510052	Q3626-06	-8.0	10435	1.4 L	10557		C2509004	VL042989.D Seq :VL092325 TUNE : VL042982.D ICAL : VL042955.D CCAL : VL042983.D MB : VL042984.D
11/14/2025	B2510051	Q3626-08	-10.8	10046	1.4 L	10165		C2504011	VL042492.D Seq :VL050625 TUNE : VL042486.D ICAL : VL042466.D CCAL : VL042487.D MB : VL042488.D
11/14/2025	B2510051	Q3626-12	-9.2	10794	1.4 L	10475		C2504011	VL042492.D Seq :VL050625 TUNE : VL042486.D ICAL : VL042466.D CCAL : VL042487.D MB : VL042488.D
11/14/2025	B2510051	Q3626-01	-8.0	10682	1.4 L	10205		C2504011	VL042492.D Seq :VL050625 TUNE : VL042486.D ICAL : VL042466.D CCAL : VL042487.D MB : VL042488.D
11/14/2025	B2510051	Q3626-04	-9.6	10677	1.4 L	10689		C2504011	VL042492.D Seq :VL050625 TUNE : VL042486.D ICAL : VL042466.D CCAL : VL042487.D MB : VL042488.D

Process Report

Order ID : Q3626

Date	Bottle ID	Lab Sample Number	InComing Vacuum	Canister ID	Can Size	Reg ID	Reg Size	Batch Code	Certification ID
11/14/2025	B2510051	Q3626-09	-10.8	10673	1.4 L	10623		C2504011	VL042492.D Seq :vl050625 TUNE : VL042486.D ICAL : VL042466.D CCAL : VL042487.D MB : VL042488.D
11/14/2025	B2510051	Q3626-07	-10.8	10471	1.4 L	10569		C2504011	VL042492.D Seq :vl050625 TUNE : VL042486.D ICAL : VL042466.D CCAL : VL042487.D MB : VL042488.D
11/14/2025	B2510051	Q3626-03	-9.6	10450	1.4 L	10694		C2504011	VL042492.D Seq :vl050625 TUNE : VL042486.D ICAL : VL042466.D CCAL : VL042487.D MB : VL042488.D
11/14/2025	B2510051	Q3626-10	-7.5	10416	1.4 L	10518		C2504011	VL042492.D Seq :vl050625 TUNE : VL042486.D ICAL : VL042466.D CCAL : VL042487.D MB : VL042488.D
11/14/2025	B2510051	Q3626-02	-10.4	10413	1.4 L	10512		C2504011	VL042492.D Seq :vl050625 TUNE : VL042486.D ICAL : VL042466.D CCAL : VL042487.D MB : VL042488.D
11/14/2025	B2510051	Q3626-11	-10.4	10123	1.4 L	10242		C2509004	VL042989.D Seq :VL092325 TUNE : VL042982.D ICAL : VL042955.D CCAL : VL042983.D MB : VL042984.D



Canister Cleaning

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Canister Batch : C2509004

Type of Canister : 1.4 L

Cleaning Date : 02/26/2025

Analyst : sam

Oven Temperature : 125.0 c

Canister	Vaccum After Cleaning ("Hg)	Vaccum After Cleaning Date	Analysis Date	VAC 24Hr ("Hg)	Vaccum After 24hr Date	New Sample Number	QC Canister	File
10747	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3285-07	10747	VL042989.D
10752	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3556-04	10747	VL042989.D
10470	-30	02/26/2025	09/23/2025	-30	02/27/2025		10747	VL042989.D
10464	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3626-05	10747	VL042989.D
10435	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3626-06	10747	VL042989.D
10123	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3626-11	10747	VL042989.D
10737	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3556-03	10747	VL042989.D
10420	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3285-03	10420	VL042990.D
10743	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3285-06	10743	VL042991.D
10451	-30	02/26/2025	09/23/2025	-30	02/27/2025		10747	VL042989.D
10428	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3557-05	10747	VL042989.D
10423	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3555-04	10747	VL042989.D
10668	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3285-04	10668	VL042992.D
10122	-30	02/26/2025	09/23/2025	-30	02/27/2025		10747	VL042989.D
10136	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3285-01	10136	VL042987.D
10147	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3285-02	10147	VL042986.D
10674	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3557-04	10747	VL042989.D
10684	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3285-05	10684	VL042988.D
10653	-30	02/26/2025	09/23/2025	-30	02/27/2025	Q3555-03	10747	VL042989.D
10746	-30	02/26/2025	09/23/2025	-30	02/27/2025		10747	VL042989.D



Canister Cleaning

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Canister Batch : C2504011

Type of Canister : 1.4 L

Cleaning Date : 04/22/2025

Analyst : SAM

Oven Temperature : 125.0 c



Canister	Vaccum After Cleaning ("Hg)	Vaccum After Cleaning Date	Analysis Date	VAC 24Hr ("Hg)	Vaccum After 24hr Date	New Sample Number	QC Canister	File
10682	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-01	10669	VL042492.D
10793	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3627-05	10669	VL042492.D
10795	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3655-02	10669	VL042492.D
10794	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-12	10669	VL042492.D
10763	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10677	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-04	10669	VL042492.D
10669	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3655-01	10669	VL042492.D
10673	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-09	10669	VL042492.D
10046	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-08	10669	VL042492.D
10450	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-03	10669	VL042492.D
10416	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-10	10669	VL042492.D
10118	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10679	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3655-06	10669	VL042492.D
10471	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-07	10669	VL042492.D
10123	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10454	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10455	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10413	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-02	10669	VL042492.D
10652	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10452	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10650	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3627-04	10669	VL042492.D

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
Data File : VL042457.D
Acq On : 01 May 2025 08:19
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri May 02 03:16:31 2025



AutoFind: Scans 3178, 3179, 3180; Background Corrected with Scan 3165

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	25.5	33424	PASS
75	95	30	66	62.8	82325	PASS
95	95	100	100	100.0	131013	PASS
96	95	5	9	6.8	8845	PASS
173	174	0.00	2	0.6	483	PASS
174	95	50	120	57.4	75195	PASS
175	174	4	9	7.5	5628	PASS
176	174	93	101	97.9	73643	PASS
177	176	5	9	6.3	4652	PASS

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL050125AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Fri May 02 03:16:31 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL042466.D 0.1 =VL042462.D 0.5 =VL042461.D 1 =VL042460.D 2 =VL042459.D 10 =VL042458.D 15 =VL042465.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.887	1.650	1.629	1.470	1.639	1.655	9.01
3) Chlorodifluoro...			1.886	1.871	1.820	1.569	1.602	1.750	8.71
4) Chloromethane			0.738	0.696	0.734	0.624	0.666	0.692	6.90
5) T Vinyl Chloride	0.756	0.964	0.708	0.625	0.623	0.567	0.631	0.696	19.18
6) T Bromomethane			0.349	0.359	0.327	0.286	0.324	0.329	8.64
7) Chloroethane			0.250	0.272	0.271	0.231	0.261	0.257	6.55
8) T Dichlorotetra...			1.445	1.371	1.402	1.235	1.368	1.364	5.77
9) T Propene			0.585	0.606	0.631	0.525	0.515	0.572	8.84
10) T Heptane			1.276	1.634	1.593	1.445	1.474	1.484	9.48
11) T Trichlorofluor...			1.660	1.489	1.528	1.376	1.582	1.527	6.96
12) T 1,1,2-Trichlor...			1.318	1.234	1.245	1.099	1.201	1.219	6.52
13) Ethanol			0.138	0.136	0.143	0.080	0.084	0.116	27.19
14) T Bromoethene			0.488	0.443	0.457	0.398	0.466	0.450	7.39
15) T Acetone			2.229	2.094	2.022	1.318	1.465	1.826	22.28
16) T 1,3-Butadiene			0.883	0.874	0.836	0.681	0.751	0.805	10.77
17) tert-Butyl alc...			1.816	1.615	1.620	1.489	1.502	1.608	8.15
18) T 1,1-Dichloroet...			0.650	0.587	0.567	0.495	0.547	0.569	9.98
19) T Isopropyl Alcohol			1.221	1.087	0.976	0.876	0.891	1.010	14.33
20) T Methylene Chlo...			0.753	0.547	0.518	0.449	0.477	0.549	21.88
21) T Allyl Chloride			1.283	1.181	1.164	1.076	1.166	1.174	6.25
22) T trans-1,2-Dich...			0.741	0.742	0.645	0.552	0.644	0.665	11.96
23) T Vinyl Acetate			1.898	1.521	1.810	1.726	1.756	1.742	8.03
24) T 1,1-Dichloroet...			1.432	1.465	1.368	1.196	1.322	1.357	7.79
25) T Ethyl Acetate			3.002	3.044	2.945	2.832	2.792	2.923	3.69
26) T Hexane			1.278	1.290	1.345	1.215	1.239	1.273	3.91
27) T Carbon Disulfide			1.509	1.446	1.499	1.418	1.558	1.486	3.69
28) T Methyl tert-Bu...			0.802	0.853	0.759	0.671	0.740	0.765	8.91
29) T Chloroform			2.551	2.482	2.399	2.198	2.301	2.386	5.89
30) T Cyclohexane			1.135	0.998	1.096	1.045	1.048	1.065	4.92
31) T cis-1,2-Dichlo...			1.551	1.604	1.545	1.342	1.386	1.486	7.71
32) T 1,1,1-Trichlor...	2.980	2.842	2.567	2.501	2.511	2.288	2.433	2.589	9.28

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.802	0.732	0.774	0.659	0.650	0.723	9.32
35) T Carbon Tetrach...	1.166	0.799	0.856	0.935	0.875	0.827	0.902	0.908	13.44
36) T Benzene			0.924	0.974	1.036	0.892	0.892	0.944	6.53
37) T 1,2-Dichloroet...			0.738	0.755	0.769	0.666	0.709	0.727	5.66
38) T Trichloroethene	0.241	0.419	0.389	0.364	0.404	0.347	0.355	0.360	16.31
39) T 1,2-Dichloropr...			0.368	0.405	0.362	0.317	0.315	0.353	10.71

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\ Method File : VL050125AIR.M											
40)	T	1,4-Dioxane			0.176	0.185	0.172	0.151	0.144	0.165	10.60
41)	T	Tetrahydrofuran			0.329	0.354	0.387	0.326	0.322	0.344	7.98
42)	T	Bromodichlorom...			0.848	0.921	0.931	0.839	0.904	0.889	4.77
43)		Methyl Methacr...			0.328	0.341	0.362	0.363	0.366	0.352	4.72
44)	T	2,2,4-Trimethy...			1.490	1.590	1.633	1.437	1.498	1.530	5.23
45)	T	t-1,3-Dichloro...			0.438	0.423	0.460	0.480	0.510	0.462	7.46
46)	T	cis-1,3-Dichlo...			0.526	0.539	0.549	0.555	0.577	0.549	3.47
47)	T	1,1,2-Trichlor...			0.372	0.409	0.428	0.357	0.371	0.387	7.74
48)	T	Dibromochlorom...			0.651	0.643	0.673	0.630	0.674	0.654	2.96
49)	T	Bromoform			0.540	0.535	0.573	0.527	0.568	0.549	3.73
50)	T	4-Methyl-2-Pen...			0.852	0.980	0.985	0.870	0.887	0.915	6.92
51)	T	2-Hexanone			0.649	0.754	0.797	0.722	0.736	0.731	7.42
52)	T	Tetrachloroethene	0.509	0.380	0.325	0.339	0.338	0.305	0.318	0.359	19.49
53)	T	Toluene			0.935	0.969	1.025	1.016	1.050	0.999	4.64
54)	T	1,2-Dibromoethane		0.625	0.589	0.599	0.610	0.554	0.577	0.592	4.26
-----ISTD-----											
55)	I	Chlorobenzene-d5									
56)		1,1,1,2-Tetrac...			0.618	0.633	0.596	0.542	0.564	0.591	6.38
57)	T	Chlorobenzene			1.101	1.060	1.036	0.904	0.923	1.005	8.64
58)	T	Ethyl Benzene			1.492	1.634	1.752	1.725	1.754	1.671	6.68
59)	T	m/p-Xylene			1.212	1.378	1.506	1.429	1.479	1.401	8.31
60)	T	o-Xylene			1.197	1.393	1.519	1.423	1.481	1.403	8.90
61)	T	Styrene			0.415	0.500	0.538	0.583	0.582	0.523	13.35
62)		Isopropylbenzene			1.781	2.030	2.165	2.033	2.108	2.023	7.26
63)	T	1,1,2,2-Tetrac...	1.140	0.945	0.976	0.985	0.990	0.846	0.887	0.967	9.69
64)		n-propylbenzene			0.400	0.524	0.549	0.482	0.518	0.495	11.75
65)		tert-Butylbenzene			1.519	1.791	2.028	1.817	1.873	1.806	10.23
66)	T	Benzyl Chloride			0.158	0.204	0.195	0.232	0.255	0.209	17.73
67)		sec-Butylbenzene			2.149	2.560	2.905	2.541	2.675	2.566	10.69
68)	S	1-Bromo-4-Fluo...	0.920	0.911	0.902	0.928	0.915	0.951	0.963	0.927	2.39
69)		p-Isopropyltol...			1.805	2.166	2.406	2.150	2.283	2.162	10.39
70)		n-Butylbenzene			1.802	2.218	2.560	2.198	2.396	2.235	12.66
71)		2-Chlorotoluene			1.405	1.569	1.697	1.597	1.667	1.587	7.20
72)	T	4-Ethyltoluene			1.281	1.647	1.828	1.722	1.792	1.654	13.28
73)	T	1,3,5-Trimethy...			1.177	1.467	1.618	1.475	1.535	1.455	11.44
74)	T	1,2,4-Trimethy...			1.568	1.788	2.075	1.646	1.723	1.760	11.07
75)	T	1,3-Dichlorobe...			0.975	1.022	1.056	0.893	0.952	0.980	6.42
76)	T	1,4-Dichlorobe...			0.869	0.945	0.966	0.907	0.949	0.927	4.22
77)	T	1,2-Dichlorobe...			0.977	1.007	1.033	0.864	0.914	0.959	7.23
78)	T	Hexachloro-1,3...			1.028	1.081	1.070	0.772	0.838	0.958	14.93
79)	T	Naphthalene		0.720	1.168	1.657	2.155	1.527	1.661	1.482	33.01
80)	T	Naphthalene,2-...			0.498	0.925	1.573	1.172	1.338	1.101	37.38
81)	T	1,2,4-Trichlor...			0.599	0.798	0.947	0.778	0.828	0.790	15.84

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042458.D
 Acq On : 01 May 2025 09:03
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:39:48 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	71402	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	201290	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	174168	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 165650 10.258 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.600%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	104948	8.882	ppbv	100
3) Chlorodifluoromethane	1.476	51	112008	8.966	ppbv	100
4) Chloromethane	1.531	50	44585	9.028	ppbv	100
5) Vinyl Chloride	1.580	62	40466	8.138	ppbv	100
6) Bromomethane	1.667	94	20392	8.680	ppbv	100
7) Chloroethane	1.703	64	16528	9.002	ppbv	100
8) Dichlorotetrafluoroethane	1.554	85	88162	9.051	ppbv	100
9) Propene	1.483	41	37491	9.175	ppbv	100
10) Heptane	4.978	43	103167	9.733	ppbv	100
11) Trichlorofluoromethane	1.874	101	98215	9.320	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.136	101	78482	9.015	ppbv	100
13) Ethanol	1.719	45	5679	6.849	ppbv	100
14) Bromoethene	1.780	108	28440	8.849	ppbv	100
15) Acetone	1.835	43	94092	7.219	ppbv	100
16) 1,3-Butadiene	1.609	39	48641	8.461	ppbv	100
17) tert-Butyl alcohol	2.039	59	106304	9.257	ppbv	100
18) 1,1-Dichloroethene	2.033	96	35320	8.689	ppbv	100
19) Isopropyl Alcohol	1.884	45	62533	8.668	ppbv	100
20) Methylene Chloride	2.059	84	32039	8.176	ppbv	100
21) Allyl Chloride	2.094	41	76859	9.169	ppbv	100
22) trans-1,2-Dichloroethene	2.340	96	39427	8.305	ppbv	100
23) Vinyl Acetate	2.460	43	123215	10.146	ppbv	100
24) 1,1-Dichloroethane	2.408	63	85367	8.814	ppbv	100
25) Ethyl Acetate	2.835	43	202215	9.689	ppbv	100
26) Hexane	2.826	57	86764	9.543	ppbv	100
27) Carbon Disulfide	2.146	76	101248	9.542	ppbv #	92
28) Methyl tert-Butyl Ether	2.437	73	47941	8.777	ppbv	100
29) Chloroform	2.848	83	156976	9.213	ppbv	100
30) Cyclohexane	3.855	84	74614	9.815	ppbv	100
31) cis-1,2-Dichloroethene	2.719	61	95797	9.031	ppbv	100
32) 1,1,1-Trichloroethane	3.366	97	163334	8.836	ppbv	100
34) 2-Butanone	2.557	43	132685	9.113	ppbv	100
35) Carbon Tetrachloride	3.758	117	166490	9.511	ppbv	100
36) Benzene	3.654	78	179532	9.451	ppbv	100
37) 1,2-Dichloroethane	3.217	62	133963	9.150	ppbv	100
38) Trichloroethene	4.544	130	69846	9.642	ppbv	100
39) 1,2-Dichloropropane	4.311	63	63894	8.979	ppbv	100
40) 1,4-Dioxane	4.583	88	30303m	9.161	ppbv	
41) Tetrahydrofuran	3.052	42	65713	9.516	ppbv	100
42) Bromodichloromethane	4.486	83	168957	9.445	ppbv	100
43) Methyl Methacrylate	4.881	69	73117	10.339	ppbv	100
44) 2,2,4-Trimethylpentane	4.635	57	289161	9.420	ppbv	100
45) t-1,3-Dichloropropene	6.412	75	96692	10.385	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042458.D
 Acq On : 01 May 2025 09:03
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:39:48 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.600	75	111768	10.111	ppbv	100
47) 1,1,2-Trichloroethane	6.580	97	71805	9.207	ppbv	100
48) Dibromochloromethane	7.386	129	126754	9.628	ppbv	100
49) Bromoform	9.480	173	106108	9.610	ppbv	100
50) 4-Methyl-2-Pentanone	5.748	43	175193	9.518	ppbv	100
51) 2-Hexanone	7.435	43	145272	9.855	ppbv	100
52) Tetrachloroethene	8.224	164	61456	8.500	ppbv	100
53) Toluene	6.930	91	204494	10.171	ppbv	100
54) 1,2-Dibromoethane	7.648	107	111506	9.300	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.924	131	94346	9.172	ppbv	99
57) Chlorobenzene	8.920	112	157401	8.994	ppbv	100
58) Ethyl Benzene	9.354	91	300412	10.321	ppbv	100
59) m/p-Xylene	9.529	91	497922m	20.404	ppbv	
60) o-Xylene	9.963	91	247784	10.157	ppbv	100
61) Styrene	9.872	104	101509	11.134	ppbv	100
62) Isopropylbenzene	10.539	105	354055	10.047	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.963	83	147295	9.305	ppbv	100
64) n-propylbenzene	10.989	120	84006	9.787	ppbv	100
65) tert-Butylbenzene	11.529	119	316532	10.065	ppbv	100
66) Benzyl Chloride	11.613	91	40331	11.096	ppbv	100
67) sec-Butylbenzene	11.769	105	442600	9.903	ppbv	100
69) p-Isopropyltoluene	11.924	119	374540	9.947	ppbv	100
70) n-Butylbenzene	12.280	91	382883	9.836	ppbv	100
71) 2-Chlorotoluene	10.898	91	278215	10.065	ppbv	100
72) 4-Ethyltoluene	11.124	105	299831	10.409	ppbv	100
73) 1,3,5-Trimethylbenzene	11.202	105	256952	10.143	ppbv	100
74) 1,2,4-Trimethylbenzene	11.536	105	286615	9.351	ppbv	100
75) 1,3-Dichlorobenzene	11.607	146	155598	9.118	ppbv	100
76) 1,4-Dichlorobenzene	11.671	146	157966	9.782	ppbv	100
77) 1,2-Dichlorobenzene	11.947	146	150432	9.007	ppbv	100
78) Hexachloro-1,3-Butadiene	13.879	225	134402	8.058	ppbv	100
79) Naphthalene	13.513	128	266036	10.352	ppbv	100
80) Naphthalene,2-methyl-	14.458	142	204042	10.641	ppbv	100
81) 1,2,4-Trichlorobenzene	13.439	180	135473	9.844	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations APPROVED

Quant Time: May 02 01:39:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri May 02 01:37:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042459.D
 Acq On : 01 May 2025 09:37
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/02/2025
 Supervised By : Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:40:35 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	73220	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	201951	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	174624	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 159777 9.869 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 98.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	23859	1.969	ppbv	100
3) Chlorodifluoromethane	1.476	51	26655	2.081	ppbv	95
4) Chloromethane	1.531	50	10751	2.123	ppbv	95
5) Vinyl Chloride	1.580	62	9122	1.789	ppbv	92
6) Bromomethane	1.667	94	4793	1.990	ppbv #	62
7) Chloroethane	1.703	64	3972	2.110	ppbv	94
8) Dichlorotetrafluoroethane	1.554	85	20529	2.055	ppbv	100
9) Propene	1.482	41	9234	2.204	ppbv	95
10) Heptane	4.978	43	23329m	2.146	ppbv	
11) Trichlorofluoromethane	1.877	101	22375	2.070	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.140	101	18230	2.042	ppbv	97
13) Ethanol	1.722	45	2099	2.468	ppbv	94
14) Bromoethene	1.780	108	6686	2.029	ppbv #	96
15) Acetone	1.835	43	29607	2.215	ppbv	99
16) 1,3-Butadiene	1.609	39	12245	2.077	ppbv	92
17) tert-Butyl alcohol	2.042	59	23722	2.014	ppbv #	93
18) 1,1-Dichloroethene	2.033	96	8303	1.992	ppbv	92
19) Isopropyl Alcohol	1.884	45	14288	1.931	ppbv	99
20) Methylene Chloride	2.062	84	7590	1.889	ppbv	96
21) Allyl Chloride	2.094	41	17050	1.983	ppbv	96
22) trans-1,2-Dichloroethene	2.340	96	9444	1.940	ppbv	97
23) Vinyl Acetate	2.463	43	26502	2.128	ppbv	98
24) 1,1-Dichloroethane	2.408	63	20027	2.016	ppbv	98
25) Ethyl Acetate	2.839	43	43122	2.015	ppbv	97
26) Hexane	2.822	57	19691	2.112	ppbv	95
27) Carbon Disulfide	2.149	76	21945	2.017	ppbv #	43
28) Methyl tert-Butyl Ether	2.441	73	11119	1.985	ppbv	94
29) Chloroform	2.848	83	35127	2.010	ppbv	86
30) Cyclohexane	3.858	84	16056	2.060	ppbv	95
31) cis-1,2-Dichloroethene	2.719	61	22619	2.079	ppbv	95
32) 1,1,1-Trichloroethane	3.369	97	36768	1.940	ppbv	99
34) 2-Butanone	2.557	43	31242	2.139	ppbv	95
35) Carbon Tetrachloride	3.764	117	35326	2.011	ppbv	97
36) Benzene	3.661	78	41848	2.196	ppbv #	88
37) 1,2-Dichloroethane	3.217	62	31049	2.114	ppbv	99
38) Trichloroethene	4.544	130	16311m	2.244	ppbv	
39) 1,2-Dichloropropane	4.311	63	14611	2.047	ppbv	91
40) 1,4-Dioxane	4.593	88	6957m	2.096	ppbv	
41) Tetrahydrofuran	3.059	42	15641m	2.258	ppbv	
42) Bromodichloromethane	4.489	83	37615	2.096	ppbv #	99
43) Methyl Methacrylate	4.884	69	14610m	2.059	ppbv	
44) 2,2,4-Trimethylpentane	4.638	57	65967	2.142	ppbv	99
45) t-1,3-Dichloropropene	6.409	75	18560m	1.987	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042459.D
 Acq On : 01 May 2025 09:37
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:40:35 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	22184	2.000	ppbv	93
47) 1,1,2-Trichloroethane	6.587	97	17307m	2.212	ppbv	
48) Dibromochloromethane	7.386	129	27174	2.057	ppbv	99
49) Bromoform	9.483	173	23126	2.088	ppbv	98
50) 4-Methyl-2-Pentanone	5.758	43	39798	2.155	ppbv	99
51) 2-Hexanone	7.441	43	32196	2.177	ppbv #	97
52) Tetrachloroethene	8.224	164	13670	1.885	ppbv	93
53) Toluene	6.930	91	41392	2.052	ppbv	97
54) 1,2-Dibromoethane	7.652	107	24650	2.049	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.927	131	20822	2.019	ppbv	95
57) Chlorobenzene	8.920	112	36187	2.062	ppbv #	88
58) Ethyl Benzene	9.354	91	61177	2.096	ppbv	91
59) m/p-Xylene	9.551	91	105185m	4.299	ppbv	
60) o-Xylene	9.963	91	53051	2.169	ppbv	99
61) Styrene	9.872	104	18776	2.054	ppbv	97
62) Isopropylbenzene	10.539	105	75626	2.140	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.963	83	34562	2.178	ppbv	98
64) n-propylbenzene	10.989	120	19160	2.226	ppbv	94
65) tert-Butylbenzene	11.532	119	70813	2.246	ppbv	97
66) Benzyl Chloride	11.613	91	6794	1.864	ppbv	97
67) sec-Butylbenzene	11.769	105	101458	2.264	ppbv	99
69) p-Isopropyltoluene	11.924	119	84026	2.226	ppbv	98
70) n-Butylbenzene	12.280	91	89400	2.291	ppbv	97
71) 2-Chlorotoluene	10.901	91	59271	2.139	ppbv	97
72) 4-Ethyltoluene	11.124	105	63833	2.210	ppbv	98
73) 1,3,5-Trimethylbenzene	11.205	105	56516	2.225	ppbv	94
74) 1,2,4-Trimethylbenzene	11.536	105	72485	2.359	ppbv	94
75) 1,3-Dichlorobenzene	11.607	146	36885	2.156	ppbv	97
76) 1,4-Dichlorobenzene	11.671	146	33728	2.083	ppbv	94
77) 1,2-Dichlorobenzene	11.947	146	36078	2.154	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	37360	2.234	ppbv	99
79) Naphthalene	13.513	128	75248	2.920	ppbv	96
80) Naphthalene,2-methyl-	14.458	142	54922	2.857	ppbv	96
81) 1,2,4-Trichlorobenzene	13.442	180	33063	2.396	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042460.D
 Acq On : 01 May 2025 10:08
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:41:24 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	71499	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	200111	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	170872	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.374 95 158600 10.011 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.100%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	11796	0.997	ppbv	96
3) Chlorodifluoromethane	1.476	51	13375	1.069	ppbv	96
4) Chloromethane	1.531	50	4975	1.006	ppbv	95
5) Vinyl Chloride	1.576	62	4469	0.898	ppbv	100
6) Bromomethane	1.667	94	2570	1.092	ppbv	86
7) Chloroethane	1.703	64	1943	1.057	ppbv	94
8) Dichlorotetrafluoroethane	1.554	85	9804	1.005	ppbv	91
9) Propene	1.482	41	4331	1.058	ppbv	88
10) Heptane	4.975	43	11681m	1.101	ppbv	
11) Trichlorofluoromethane	1.877	101	10649m	1.009	ppbv	
12) 1,1,2-Trichlorotrifluo...	2.140	101	8821	1.012	ppbv	97
13) Ethanol	1.719	45	975	1.174	ppbv #	52
14) Bromoethene	1.780	108	3164	0.983	ppbv #	91
15) Acetone	1.835	43	14974	1.147	ppbv	98
16) 1,3-Butadiene	1.609	39	6248	1.085	ppbv	96
17) tert-Butyl alcohol	2.039	59	11546	1.004	ppbv #	91
18) 1,1-Dichloroethene	2.033	96	4200	1.032	ppbv	86
19) Isopropyl Alcohol	1.884	45	7775	1.076	ppbv	99
20) Methylene Chloride	2.059	84	3909	0.996	ppbv	86
21) Allyl Chloride	2.094	41	8443	1.006	ppbv	94
22) trans-1,2-Dichloroethene	2.337	96	5305	1.116	ppbv	88
23) Vinyl Acetate	2.460	43	10876	0.894	ppbv #	94
24) 1,1-Dichloroethane	2.405	63	10477	1.080	ppbv	97
25) Ethyl Acetate	2.835	43	21761	1.041	ppbv	99
26) Hexane	2.822	57	9223	1.013	ppbv	97
27) Carbon Disulfide	2.149	76	10342	0.973	ppbv #	21
28) Methyl tert-Butyl Ether	2.441	73	6101	1.115	ppbv	93
29) Chloroform	2.848	83	17744	1.040	ppbv	97
30) Cyclohexane	3.852	84	7139	0.938	ppbv	88
31) cis-1,2-Dichloroethene	2.719	61	11472	1.080	ppbv	87
32) 1,1,1-Trichloroethane	3.366	97	17881	0.966	ppbv	98
34) 2-Butanone	2.560	43	14641	1.012	ppbv	94
35) Carbon Tetrachloride	3.755	117	18704	1.075	ppbv #	76
36) Benzene	3.654	78	19493	1.032	ppbv #	94
37) 1,2-Dichloroethane	3.214	62	15116	1.039	ppbv	98
38) Trichloroethene	4.541	130	7278m	1.011	ppbv	
39) 1,2-Dichloropropane	4.302	63	8103m	1.145	ppbv	
40) 1,4-Dioxane	4.593	88	3703m	1.126	ppbv	
41) Tetrahydrofuran	3.062	42	7091m	1.033	ppbv	
42) Bromodichloromethane	4.486	83	18427	1.036	ppbv #	81
43) Methyl Methacrylate	4.874	69	6829m	0.971	ppbv	
44) 2,2,4-Trimethylpentane	4.635	57	31820m	1.043	ppbv	
45) t-1,3-Dichloropropene	6.409	75	8461m	0.914	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042460.D
 Acq On : 01 May 2025 10:08
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:41:24 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.596	75	10781m	0.981	ppbv	
47) 1,1,2-Trichloroethane	6.583	97	8183m	1.055	ppbv	
48) Dibromochloromethane	7.386	129	12859m	0.982	ppbv	
49) Bromoform	9.483	173	10709	0.976	ppbv	97
50) 4-Methyl-2-Pentanone	5.768	43	19618m	1.072	ppbv	
51) 2-Hexanone	7.448	43	15091m	1.030	ppbv	
52) Tetrachloroethene	8.221	164	6780	0.943	ppbv	86
53) Toluene	6.930	91	19390	0.970	ppbv	97
54) 1,2-Dibromoethane	7.655	107	11980	1.005	ppbv	91
56) 1,1,1,2-Tetrachloroethane	8.920	131	10811	1.071	ppbv	94
57) Chlorobenzene	8.917	112	18111	1.055	ppbv	90
58) Ethyl Benzene	9.354	91	27914	0.978	ppbv	97
59) m/p-Xylene	9.548	91	47105m	1.968	ppbv	
60) o-Xylene	9.962	91	23806	0.995	ppbv	98
61) Styrene	9.872	104	8537	0.954	ppbv	98
62) Isopropylbenzene	10.539	105	34679	1.003	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.962	83	16837	1.084	ppbv	98
64) n-propylbenzene	10.985	120	8958	1.064	ppbv	86
65) tert-Butylbenzene	11.529	119	30604	0.992	ppbv	95
66) Benzyl Chloride	11.613	91	3486	0.978	ppbv #	96
67) sec-Butylbenzene	11.765	105	43740	0.998	ppbv	98
69) p-Isopropyltoluene	11.927	119	37011	1.002	ppbv	99
70) n-Butylbenzene	12.280	91	37892	0.992	ppbv	97
71) 2-Chlorotoluene	10.898	91	26813	0.989	ppbv	99
72) 4-Ethyltoluene	11.124	105	28149	0.996	ppbv	96
73) 1,3,5-Trimethylbenzene	11.202	105	25073	1.009	ppbv	97
74) 1,2,4-Trimethylbenzene	11.535	105	30550	1.016	ppbv	92
75) 1,3-Dichlorobenzene	11.607	146	17456	1.043	ppbv	98
76) 1,4-Dichlorobenzene	11.671	146	16151	1.019	ppbv	99
77) 1,2-Dichlorobenzene	11.947	146	17207	1.050	ppbv	94
78) Hexachloro-1,3-Butadiene	13.879	225	18472	1.129	ppbv	98
79) Naphthalene	13.513	128	28316	1.123	ppbv #	94
80) Naphthalene,2-methyl-	14.458	142	15804	0.840	ppbv	93
81) 1,2,4-Trichlorobenzene	13.439	180	13644	1.011	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042460.D
 Acq On : 01 May 2025 10:08
 Operator : SY/MD
 Sample : VSTDIC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC001

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025


Quant Time: May 02 01:41:24 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042461.D
 Acq On : 01 May 2025 10:40
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:42:12 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	68913	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	193701	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	166625	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 150277 9.728 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 97.300%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	6501	0.570	ppbv	94
3) Chlorodifluoromethane	1.476	51	6500	0.539	ppbv	99
4) Chloromethane	1.535	50	2542	0.533	ppbv	95
5) Vinyl Chloride	1.580	62	2441	0.509	ppbv	92
6) Bromomethane	1.667	94	1202	0.530	ppbv	97
7) Chloroethane	1.703	64	862	0.486	ppbv #	68
8) Dichlorotetrafluoroethane	1.554	85	4979	0.530	ppbv	93
9) Propene	1.486	41	2017	0.511	ppbv	96
10) Heptane	4.982	43	4397m	0.430	ppbv	
11) Trichlorofluoromethane	1.878	101	5720	0.562	ppbv	95
12) 1,1,2-Trichlorotrifluo...	2.137	101	4540	0.540	ppbv	94
13) Ethanol	1.722	45	474m	0.592	ppbv	
14) Bromoethene	1.784	108	1680	0.542	ppbv #	87
15) Acetone	1.839	43	7680	0.610	ppbv #	82
16) 1,3-Butadiene	1.609	39	3044	0.549	ppbv	86
17) tert-Butyl alcohol	2.046	59	6256	0.564	ppbv #	91
18) 1,1-Dichloroethene	2.036	96	2240	0.571	ppbv	94
19) Isopropyl Alcohol	1.887	45	4208	0.604	ppbv #	96
20) Methylene Chloride	2.062	84	2594	0.686	ppbv #	80
21) Allyl Chloride	2.098	41	4420	0.546	ppbv #	93
22) trans-1,2-Dichloroethene	2.340	96	2554	0.557	ppbv	95
23) Vinyl Acetate	2.460	43	6541	0.558	ppbv #	95
24) 1,1-Dichloroethane	2.412	63	4933	0.528	ppbv	98
25) Ethyl Acetate	2.842	43	10344	0.514	ppbv #	96
26) Hexane	2.829	57	4403	0.502	ppbv	95
27) Carbon Disulfide	2.153	76	5201	0.508	ppbv #	1
28) Methyl tert-Butyl Ether	2.441	73	2762	0.524	ppbv #	82
29) Chloroform	2.852	83	8791	0.535	ppbv	92
30) Cyclohexane	3.855	84	3910m	0.533	ppbv	
31) cis-1,2-Dichloroethene	2.726	61	5345m	0.522	ppbv	
32) 1,1,1-Trichloroethane	3.366	97	8846m	0.496	ppbv	
34) 2-Butanone	2.564	43	7766	0.554	ppbv	93
35) Carbon Tetrachloride	3.768	117	8286	0.492	ppbv	86
36) Benzene	3.655	78	8951	0.490	ppbv #	92
37) 1,2-Dichloroethane	3.224	62	7149	0.507	ppbv #	95
38) Trichloroethene	4.548	130	3769m	0.541	ppbv	
39) 1,2-Dichloropropane	4.321	63	3561m	0.520	ppbv	
40) 1,4-Dioxane	4.619	88	1701m	0.534	ppbv	
41) Tetrahydrofuran	3.069	42	3187m	0.480	ppbv	
42) Bromodichloromethane	4.496	83	8212	0.477	ppbv	93
43) Methyl Methacrylate	4.891	69	3179m	0.467	ppbv	
44) 2,2,4-Trimethylpentane	4.648	57	14435m	0.489	ppbv	
45) t-1,3-Dichloropropene	6.415	75	4238m	0.473	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042461.D
 Acq On : 01 May 2025 10:40
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:42:12 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.600	75	5091m	0.479	ppbv	
47) 1,1,2-Trichloroethane	6.584	97	3606m	0.480	ppbv	
48) Dibromochloromethane	7.390	129	6304m	0.498	ppbv	
49) Bromoform	9.484	173	5228	0.492	ppbv	97
50) 4-Methyl-2-Pentanone	5.771	43	8248m	0.466	ppbv	
51) 2-Hexanone	7.454	43	6283m	0.443	ppbv	
52) Tetrachloroethene	8.221	164	3151m	0.453	ppbv	
53) Toluene	6.933	91	9052m	0.468	ppbv	
54) 1,2-Dibromoethane	7.652	107	5705	0.494	ppbv	83
56) 1,1,1,2-Tetrachloroethane	8.924	131	5152	0.524	ppbv #	92
57) Chlorobenzene	8.924	112	9173	0.548	ppbv #	86
58) Ethyl Benzene	9.357	91	12427	0.446	ppbv	90
59) m/p-Xylene	9.552	91	20195m	0.865	ppbv	
60) o-Xylene	9.966	91	9976m	0.427	ppbv	
61) Styrene	9.872	104	3457m	0.396	ppbv	
62) Isopropylbenzene	10.545	105	14837	0.440	ppbv	95
63) 1,1,2,2-Tetrachloroethane	9.963	83	8135	0.537	ppbv	96
64) n-propylbenzene	10.992	120	3329m	0.405	ppbv	
65) tert-Butylbenzene	11.536	119	12655	0.421	ppbv	99
66) Benzyl Chloride	11.613	91	1317m	0.379	ppbv	
67) sec-Butylbenzene	11.769	105	17905	0.419	ppbv	97
69) p-Isopropyltoluene	11.927	119	15036	0.417	ppbv	94
70) n-Butylbenzene	12.283	91	15016	0.403	ppbv	93
71) 2-Chlorotoluene	10.905	91	11702	0.443	ppbv	99
72) 4-Ethyltoluene	11.128	105	10671	0.387	ppbv #	91
73) 1,3,5-Trimethylbenzene	11.206	105	9808	0.405	ppbv	89
74) 1,2,4-Trimethylbenzene	11.539	105	13060	0.445	ppbv #	83
75) 1,3-Dichlorobenzene	11.607	146	8127	0.498	ppbv	96
76) 1,4-Dichlorobenzene	11.675	146	7238	0.469	ppbv	96
77) 1,2-Dichlorobenzene	11.950	146	8141	0.509	ppbv	94
78) Hexachloro-1,3-Butadiene	13.886	225	8565	0.537	ppbv	99
79) Naphthalene	13.517	128	9735m	0.396	ppbv	
80) Naphthalene,2-methyl-	14.458	142	4153	0.226	ppbv	98
81) 1,2,4-Trichlorobenzene	13.445	180	4992	0.379	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042462.D
 Acq On : 01 May 2025 11:12
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:43:00 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	65449	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	186907	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	157618	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	143644	9.830	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.300%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	631	0.138	ppbv #	53
32) 1,1,1-Trichloroethane	3.363	97	1860m	0.110	ppbv	
35) Carbon Tetrachloride	3.758	117	1493	0.092	ppbv #	55
38) Trichloroethene	4.551	130	784m	0.117	ppbv	
52) Tetrachloroethene	8.221	164	710m	0.106	ppbv	
54) 1,2-Dibromoethane	7.665	107	1169m	0.105	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.966	83	1490	0.104	ppbv #	75
79) Naphthalene	13.517	128	1135m	0.049	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
Data File : VL042462.D
Acq On : 01 May 2025 11:12
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:43:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 02 01:37:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042465.D
 Acq On : 01 May 2025 12:49
 Operator : SY/MD
 Sample : VSTDICC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:45:25 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	64314	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	178877	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	157673	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 151837 10.387 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 103.900%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	158069	14.852	ppbv	99
3) Chlorodifluoromethane	1.476	51	154556	13.735	ppbv	100
4) Chloromethane	1.535	50	64267	14.447	ppbv	95
5) Vinyl Chloride	1.580	62	60873	13.592	ppbv	98
6) Bromomethane	1.667	94	31246	14.766	ppbv	87
7) Chloroethane	1.703	64	25181	15.227	ppbv	92
8) Dichlorotetrafluoroethane	1.554	85	131995	15.044	ppbv	99
9) Propene	1.486	41	49647	13.489	ppbv	99
10) Heptane	4.982	43	142235	14.898	ppbv	100
11) Trichlorofluoromethane	1.878	101	152655	16.082	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.140	101	115855	14.775	ppbv	96
13) Ethanol	1.722	45	8090	10.832	ppbv	98
14) Bromoethene	1.781	108	44928	15.519	ppbv	96
15) Acetone	1.836	43	141306	12.036	ppbv	99
16) 1,3-Butadiene	1.609	39	72455	13.992	ppbv	98
17) tert-Butyl alcohol	2.043	59	144935	14.012	ppbv	99
18) 1,1-Dichloroethene	2.033	96	52813	14.424	ppbv	98
19) Isopropyl Alcohol	1.887	45	86000	13.235	ppbv #	1
20) Methylene Chloride	2.062	84	46063	13.050	ppbv	94
21) Allyl Chloride	2.098	41	112451	14.893	ppbv	99
22) trans-1,2-Dichloroethene	2.341	96	62153	14.534	ppbv	100
23) Vinyl Acetate	2.464	43	169418	15.489	ppbv	99
24) 1,1-Dichloroethane	2.409	63	127567	14.622	ppbv	97
25) Ethyl Acetate	2.836	43	269372	14.329	ppbv	100
26) Hexane	2.826	57	119552	14.598	ppbv	97
27) Carbon Disulfide	2.150	76	150258	15.722	ppbv #	92
28) Methyl tert-Butyl Ether	2.441	73	71341	14.500	ppbv	99
29) Chloroform	2.849	83	221976	14.464	ppbv	99
30) Cyclohexane	3.859	84	101146	14.771	ppbv	99
31) cis-1,2-Dichloroethene	2.723	61	133735	13.997	ppbv	93
32) 1,1,1-Trichloroethane	3.370	97	234752	14.099	ppbv	99
34) 2-Butanone	2.557	43	174521	13.488	ppbv	97
35) Carbon Tetrachloride	3.765	117	242144	15.566	ppbv	99
36) Benzene	3.658	78	239352	14.179	ppbv	98
37) 1,2-Dichloroethane	3.221	62	190179	14.618	ppbv	98
38) Trichloroethene	4.551	130	95257	14.798	ppbv	94
39) 1,2-Dichloropropane	4.312	63	84414	13.349	ppbv	93
40) 1,4-Dioxane	4.584	88	38551	13.115	ppbv #	100
41) Tetrahydrofuran	3.056	42	86296	14.063	ppbv	98
42) Bromodichloromethane	4.493	83	242556	15.259	ppbv	99
43) Methyl Methacrylate	4.885	69	98279	15.639	ppbv	100
44) 2,2,4-Trimethylpentane	4.642	57	401920	14.734	ppbv	99
45) t-1,3-Dichloropropene	6.419	75	136762	16.529	ppbv	92

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042465.D
 Acq On : 01 May 2025 12:49
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:45:25 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025


Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.600	75	154705m	15.749	ppbv	
47) 1,1,2-Trichloroethane	6.587	97	99496	14.356	ppbv	93
48) Dibromochloromethane	7.393	129	180970	15.468	ppbv	98
49) Bromoform	9.487	173	152440	15.535	ppbv	99
50) 4-Methyl-2-Pentanone	5.755	43	237985	14.550	ppbv	97
51) 2-Hexanone	7.438	43	197423	15.071	ppbv	99
52) Tetrachloroethene	8.228	164	85271	13.272	ppbv	90
53) Toluene	6.937	91	281731	15.768	ppbv	100
54) 1,2-Dibromoethane	7.655	107	154739	14.523	ppbv	94
56) 1,1,1,2-Tetrachloroethane	8.927	131	133410	14.326	ppbv	98
57) Chlorobenzene	8.927	112	218370	13.783	ppbv	99
58) Ethyl Benzene	9.358	91	414805	15.743	ppbv	98
59) m/p-Xylene	9.555	91	699743m	31.674	ppbv	
60) o-Xylene	9.966	91	350328	15.863	ppbv	100
61) Styrene	9.876	104	137720	16.686	ppbv	96
62) Isopropylbenzene	10.542	105	498619	15.629	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.966	83	209726	14.635	ppbv	99
64) n-propylbenzene	10.992	120	122452	15.759	ppbv	97
65) tert-Butylbenzene	11.536	119	443042	15.562	ppbv	97
66) Benzyl Chloride	11.617	91	60376	18.348	ppbv	97
67) sec-Butylbenzene	11.769	105	632600	15.636	ppbv	100
69) p-Isopropyltoluene	11.928	119	539860	15.837	ppbv	99
70) n-Butylbenzene	12.284	91	566780	16.084	ppbv	99
71) 2-Chlorotoluene	10.905	91	394227	15.755	ppbv	100
72) 4-Ethyltoluene	11.128	105	423746	16.250	ppbv	99
73) 1,3,5-Trimethylbenzene	11.206	105	362951	15.826	ppbv	96
74) 1,2,4-Trimethylbenzene	11.539	105	407416	14.683	ppbv	96
75) 1,3-Dichlorobenzene	11.610	146	225215	14.579	ppbv	98
76) 1,4-Dichlorobenzene	11.675	146	224505	15.357	ppbv	99
77) 1,2-Dichlorobenzene	11.950	146	216147	14.295	ppbv	99
78) Hexachloro-1,3-Butadiene	13.886	225	198112	13.121	ppbv	99
79) Naphthalene	13.517	128	392900	16.888	ppbv	98
80) Naphthalene,2-methyl-	14.459	142	316493	18.232	ppbv	99
81) 1,2,4-Trichlorobenzene	13.442	180	195942	15.728	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations APPROVED

Quant Time: May 02 01:45:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri May 02 01:37:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042466.D
 Acq On : 01 May 2025 14:10
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:46:15 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.791	49	66546	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	186737	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	161050	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	148088	9.918	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.200%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	151	0.033	ppbv #	87
32) 1,1,1-Trichloroethane	3.370	97	595m	0.035	ppbv	
35) Carbon Tetrachloride	3.771	117	653m	0.040	ppbv	
38) Trichloroethene	4.551	130	135	0.020	ppbv #	40
52) Tetrachloroethene	8.231	164	285m	0.042	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.973	83	551m	0.038	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
Data File : VL042466.D
Acq On : 01 May 2025 14:10
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:46:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 02 01:37:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICSVVL050125

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 03:48:56 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	63539	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	183679	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	156982	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.380 95 150522 10.342 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 103.400%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	108782	10.346	ppbv	97
3) Chlorodifluoromethane	1.476	51	110290	9.921	ppbv	98
4) Chloromethane	1.534	50	43950	10.001	ppbv	96
5) Vinyl Chloride	1.580	62	41181	9.307	ppbv	99
6) Bromomethane	1.667	94	21021	10.055	ppbv	91
7) Chloroethane	1.706	64	16630	10.179	ppbv	90
8) Dichlorotetrafluoroethane	1.554	85	89345	10.307	ppbv	99
9) Propene	1.486	41	36920	10.153	ppbv	97
10) Heptane	4.985	43	94631	10.033	ppbv	95
11) Trichlorofluoromethane	1.878	101	100713	10.380	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.140	101	81143	10.474	ppbv	99
13) Ethanol	1.722	45	5736	7.773	ppbv	99
14) Bromoethene	1.780	108	30094	10.522	ppbv	97
15) Acetone	1.835	43	95460	8.230	ppbv	98
16) 1,3-Butadiene	1.609	39	49566	9.689	ppbv	100
17) tert-Butyl alcohol	2.043	59	102047	9.986	ppbv	100
18) 1,1-Dichloroethene	2.033	96	35784	9.892	ppbv	98
19) Isopropyl Alcohol	1.887	45	60977	9.499	ppbv	99
20) Methylene Chloride	2.062	84	31212	8.951	ppbv	94
21) Allyl Chloride	2.098	41	73365	9.835	ppbv	94
22) trans-1,2-Dichloroethene	2.340	96	42922	10.160	ppbv	96
23) Vinyl Acetate	2.463	43	113131	10.220	ppbv #	98
24) 1,1-Dichloroethane	2.408	63	89117	10.340	ppbv	97
25) Ethyl Acetate	2.839	43	186185	10.025	ppbv	99
26) Hexane	2.826	57	83132	10.275	ppbv	96
27) Carbon Disulfide	2.149	76	100196	10.612	ppbv	96
28) Methyl tert-Butyl Ether	2.441	73	48326	9.942	ppbv	100
29) Chloroform	2.852	83	152381	10.050	ppbv	96
30) Cyclohexane	3.858	84	70581	10.434	ppbv	98
31) cis-1,2-Dichloroethene	2.722	61	95793	10.148	ppbv	96
32) 1,1,1-Trichloroethane	3.370	97	161465	9.816	ppbv	99
34) 2-Butanone	2.560	43	125883	9.475	ppbv	97
35) Carbon Tetrachloride	3.764	117	166814	9.998	ppbv	97
36) Benzene	3.658	78	171384	9.888	ppbv	98
37) 1,2-Dichloroethane	3.221	62	130063	9.736	ppbv	100
38) Trichloroethene	4.551	130	65940	9.976	ppbv	96
39) 1,2-Dichloropropane	4.315	63	60010	9.248	ppbv	95
40) 1,4-Dioxane	4.583	88	28306m	9.316	ppbv	
41) Tetrahydrofuran	3.059	42	59604	9.440	ppbv	98
42) Bromodichloromethane	4.489	83	163210	9.999	ppbv	95
43) Methyl Methacrylate	4.881	69	67516	10.438	ppbv	99
44) 2,2,4-Trimethylpentane	4.642	57	277217	9.867	ppbv	100
45) t-1,3-Dichloropropene	6.415	75	89638m	10.563	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL050125

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 03:48:56 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.609	75	104665	10.378	ppbv	97
47) 1,1,2-Trichloroethane	6.584	97	67821	9.530	ppbv	95
48) Dibromochloromethane	7.393	129	121381	10.103	ppbv	97
49) Bromoform	9.487	173	102150	10.138	ppbv	100
50) 4-Methyl-2-Pentanone	5.752	43	159280	9.478	ppbv	99
51) 2-Hexanone	7.438	43	126396	9.407	ppbv #	98
52) Tetrachloroethene	8.228	164	58272	8.832	ppbv	96
53) Toluene	6.940	91	191922	10.461	ppbv	99
54) 1,2-Dibromoethane	7.658	107	105149	9.664	ppbv	95
56) 1,1,1,2-Tetrachloroethane	8.927	131	92337	9.959	ppbv	97
57) Chlorobenzene	8.927	112	148759	9.431	ppbv	92
58) Ethyl Benzene	9.357	91	280660	10.698	ppbv	99
59) m/p-Xylene	9.555	91	468984m	21.324	ppbv	
60) o-Xylene	9.966	91	238347	10.824	ppbv	97
61) Styrene	9.872	104	91128	11.090	ppbv	97
62) Isopropylbenzene	10.542	105	337992	10.641	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.966	83	140306	9.242	ppbv	100
64) n-propylbenzene	10.992	120	79983	10.303	ppbv	99
65) tert-Butylbenzene	11.536	119	302506	10.672	ppbv	99
66) Benzyl Chloride	11.617	91	34933	10.663	ppbv	97
67) sec-Butylbenzene	11.769	105	425444	10.562	ppbv	99
69) p-Isopropyltoluene	11.927	119	363563	10.712	ppbv	98
70) n-Butylbenzene	12.283	91	371369	10.585	ppbv	99
71) 2-Chlorotoluene	10.905	91	264063	10.599	ppbv	99
72) 4-Ethyltoluene	11.128	105	287375	11.069	ppbv	99
73) 1,3,5-Trimethylbenzene	11.206	105	247128	10.823	ppbv	93
74) 1,2,4-Trimethylbenzene	11.539	105	278306	10.074	ppbv	98
75) 1,3-Dichlorobenzene	11.610	146	150026	9.754	ppbv	99
76) 1,4-Dichlorobenzene	11.675	146	151740	10.425	ppbv	97
77) 1,2-Dichlorobenzene	11.950	146	146314	9.719	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	133688	8.893	ppbv	99
79) Naphthalene	13.516	128	251318	10.806	ppbv	98
80) Naphthalene,2-methyl-	14.458	142	198604	11.489	ppbv	99
81) 1,2,4-Trichlorobenzene	13.442	180	129406	10.433	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICSVVL050125

Quant Time: May 02 03:48:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	1.655	1.712	-3.4	104	0.00
3	Chlorodifluoromethane	1.750	1.736	0.8	98	0.00
4	Chloromethane	0.692	0.692	0.0	99	0.00
5 T	Vinyl Chloride	0.696	0.648	6.9	102	0.00
6 T	Bromomethane	0.329	0.331	-0.6	103	0.00
7	Chloroethane	0.257	0.262	-1.9	101	0.00
8 T	Dichlorotetrafluoroethane	1.364	1.406	-3.1	101	0.00
9 T	Propene	0.572	0.581	-1.6	98	0.00
10 T	Heptane	1.484	1.489	-0.3	92	0.00
11 T	Trichlorofluoromethane	1.527	1.585	-3.8	103	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.219	1.277	-4.8	103	0.00
13	Ethanol	0.116	0.090	22.4	101	0.00
14 T	Bromoethene	0.450	0.474	-5.3	106	0.00
15 T	Acetone	1.826	1.502	17.7	101	0.00
16 T	1,3-Butadiene	0.805	0.780	3.1	102	0.00
17	tert-Butyl alcohol	1.608	1.606	0.1	96	0.00
18 T	1,1-Dichloroethene	0.569	0.563	1.1	101	0.00
19 T	Isopropyl Alcohol	1.010	0.960	5.0	98	0.00
20 T	Methylene Chloride	0.549	0.491	10.6	97	0.00
21 T	Allyl Chloride	1.174	1.155	1.6	95	0.00
22 T	trans-1,2-Dichloroethene	0.665	0.676	-1.7	109	0.00
23 T	Vinyl Acetate	1.742	1.780	-2.2	92	0.00
24 T	1,1-Dichloroethane	1.357	1.403	-3.4	104	0.00
25 T	Ethyl Acetate	2.923	2.930	-0.2	92	0.00
26 T	Hexane	1.273	1.308	-2.7	96	0.00
27 T	Carbon Disulfide	1.486	1.577	-6.1	99	0.00
28 T	Methyl tert-Butyl Ether	0.765	0.761	0.5	101	0.00
29 T	Chloroform	2.386	2.398	-0.5	97	0.00
30 T	Cyclohexane	1.065	1.111	-4.3	95	0.00
31 T	cis-1,2-Dichloroethene	1.486	1.508	-1.5	100	0.00
32 T	1,1,1-Trichloroethane	2.589	2.541	1.9	99	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
34 T	2-Butanone	0.723	0.685	5.3	95	0.00
35 T	Carbon Tetrachloride	0.908	0.908	0.0	100	0.00
36 T	Benzene	0.944	0.933	1.2	95	0.00
37 T	1,2-Dichloroethane	0.727	0.708	2.6	97	0.00
38 T	Trichloroethene	0.360	0.359	0.3	94	0.00
39 T	1,2-Dichloropropane	0.353	0.327	7.4	94	0.00
40 T	1,4-Dioxane	0.165	0.154	6.7	93	0.00
41 T	Tetrahydrofuran	0.344	0.325	5.5	91	0.00
42 T	Bromodichloromethane	0.889	0.889	0.0	97	0.00
43	Methyl Methacrylate	0.352	0.368	-4.5	92	0.00
44 T	2,2,4-Trimethylpentane	1.530	1.509	1.4	96	0.00
45 T	t-1,3-Dichloropropene	0.462	0.488	-5.6	93	0.00
46 T	cis-1,3-Dichloropropene	0.549	0.570	-3.8	94	0.00
47 T	1,1,2-Trichloroethane	0.387	0.369	4.7	94	0.00
48 T	Dibromochloromethane	0.654	0.661	-1.1	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL050125

Quant Time: May 02 03:48:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.549	0.556	-1.3	96	0.00
50 T	4-Methyl-2-Pentanone	0.915	0.867	5.2	91	0.00
51 T	2-Hexanone	0.731	0.688	5.9	87	0.00
52 T	Tetrachloroethene	0.359	0.317	11.7	95	0.00
53 T	Toluene	0.999	1.045	-4.6	94	0.00
54 T	1,2-Dibromoethane	0.592	0.572	3.4	94	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
56	1,1,1,2-Tetrachloroethane	0.591	0.588	0.5	98	0.00
57 T	Chlorobenzene	1.005	0.948	5.7	95	0.00
58 T	Ethyl Benzene	1.671	1.788	-7.0	93	0.00
59 T	m/p-Xylene	1.401	1.494	-6.6	94	0.03
60 T	o-Xylene	1.403	1.518	-8.2	96	0.00
61 T	Styrene	0.523	0.580	-10.9	90	0.00
62	Isopropylbenzene	2.023	2.153	-6.4	95	0.00
63 T	1,1,2,2-Tetrachloroethane	0.967	0.894	7.5	95	0.00
64	n-propylbenzene	0.495	0.510	-3.0	95	0.00
65	tert-Butylbenzene	1.806	1.927	-6.7	96	0.00
66 T	Benzyl Chloride	0.209	0.223	-6.7	87	0.00
67	sec-Butylbenzene	2.566	2.710	-5.6	96	0.00
68 S	1-Bromo-4-Fluorobenzene	0.927	0.959	-3.5	91	0.00
69	p-Isopropyltoluene	2.162	2.316	-7.1	97	0.00
70	n-Butylbenzene	2.235	2.366	-5.9	97	0.00
71	2-Chlorotoluene	1.587	1.682	-6.0	95	0.00
72 T	4-Ethyltoluene	1.654	1.831	-10.7	96	0.00
73 T	1,3,5-Trimethylbenzene	1.455	1.574	-8.2	96	0.00
74 T	1,2,4-Trimethylbenzene	1.760	1.773	-0.7	97	0.00
75 T	1,3-Dichlorobenzene	0.980	0.956	2.4	96	0.00
76 T	1,4-Dichlorobenzene	0.927	0.967	-4.3	96	0.00
77 T	1,2-Dichlorobenzene	0.959	0.932	2.8	97	0.00
78 T	Hexachloro-1,3-Butadiene	0.958	0.852	11.1	99	0.00
79 T	Naphthalene	1.482	1.601	-8.0	94	0.00
80 T	Naphthalene,2-methyl-	1.101	1.265	-14.9	97	0.00
81 T	1,2,4-Trichlorobenzene	0.790	0.824	-4.3	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICSVVL050125

Quant Time: May 02 03:48:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	10.000	10.346	-3.5	104	0.00
3	Chlorodifluoromethane	10.000	9.921	0.8	98	0.00
4	Chloromethane	10.000	10.001	-0.0	99	0.00
5 T	Vinyl Chloride	10.000	9.307	6.9	102	0.00
6 T	Bromomethane	10.000	10.055	-0.5	103	0.00
7	Chloroethane	10.000	10.179	-1.8	101	0.00
8 T	Dichlorotetrafluoroethane	10.000	10.307	-3.1	101	0.00
9 T	Propene	10.000	10.153	-1.5	98	0.00
10 T	Heptane	10.000	10.033	-0.3	92	0.00
11 T	Trichlorofluoromethane	10.000	10.380	-3.8	103	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	10.474	-4.7	103	0.00
13	Ethanol	10.000	7.773	22.3	101	0.00
14 T	Bromoethene	10.000	10.522	-5.2	106	0.00
15 T	Acetone	10.000	8.230	17.7	101	0.00
16 T	1,3-Butadiene	10.000	9.689	3.1	102	0.00
17	tert-Butyl alcohol	10.000	9.986	0.1	96	0.00
18 T	1,1-Dichloroethene	10.000	9.892	1.1	101	0.00
19 T	Isopropyl Alcohol	10.000	9.499	5.0	98	0.00
20 T	Methylene Chloride	10.000	8.951	10.5	97	0.00
21 T	Allyl Chloride	10.000	9.835	1.6	95	0.00
22 T	trans-1,2-Dichloroethene	10.000	10.160	-1.6	109	0.00
23 T	Vinyl Acetate	10.000	10.220	-2.2	92	0.00
24 T	1,1-Dichloroethane	10.000	10.340	-3.4	104	0.00
25 T	Ethyl Acetate	10.000	10.025	-0.3	92	0.00
26 T	Hexane	10.000	10.275	-2.8	96	0.00
27 T	Carbon Disulfide	10.000	10.612	-6.1	99	0.00
28 T	Methyl tert-Butyl Ether	10.000	9.942	0.6	101	0.00
29 T	Chloroform	10.000	10.050	-0.5	97	0.00
30 T	Cyclohexane	10.000	10.434	-4.3	95	0.00
31 T	cis-1,2-Dichloroethene	10.000	10.148	-1.5	100	0.00
32 T	1,1,1-Trichloroethane	10.000	9.816	1.8	99	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	91	0.00
34 T	2-Butanone	10.000	9.475	5.3	95	0.00
35 T	Carbon Tetrachloride	10.000	9.998	0.0	100	0.00
36 T	Benzene	10.000	9.888	1.1	95	0.00
37 T	1,2-Dichloroethane	10.000	9.736	2.6	97	0.00
38 T	Trichloroethene	10.000	9.976	0.2	94	0.00
39 T	1,2-Dichloropropane	10.000	9.248	7.5	94	0.00
40 T	1,4-Dioxane	10.000	9.316	6.8	93	0.00
41 T	Tetrahydrofuran	10.000	9.440	5.6	91	0.00
42 T	Bromodichloromethane	10.000	9.999	0.0	97	0.00
43	Methyl Methacrylate	10.000	10.438	-4.4	92	0.00
44 T	2,2,4-Trimethylpentane	10.000	9.867	1.3	96	0.00
45 T	t-1,3-Dichloropropene	10.000	10.563	-5.6	93	0.00
46 T	cis-1,3-Dichloropropene	10.000	10.378	-3.8	94	0.00
47 T	1,1,2-Trichloroethane	10.000	9.530	4.7	94	0.00
48 T	Dibromochloromethane	10.000	10.103	-1.0	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL050125

Quant Time: May 02 03:48:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	Bromoform	10.000	10.138	-1.4	96	0.00
50 T	4-Methyl-2-Pentanone	10.000	9.478	5.2	91	0.00
51 T	2-Hexanone	10.000	9.407	5.9	87	0.00
52 T	Tetrachloroethene	10.000	8.832	11.7	95	0.00
53 T	Toluene	10.000	10.461	-4.6	94	0.00
54 T	1,2-Dibromoethane	10.000	9.664	3.4	94	0.00

55 I	Chlorobenzene-d5	10.000	10.000	0.0	90	0.00
56	1,1,1,2-Tetrachloroethane	10.000	9.959	0.4	98	0.00
57 T	Chlorobenzene	10.000	9.431	5.7	95	0.00
58 T	Ethyl Benzene	10.000	10.698	-7.0	93	0.00
59 T	m/p-Xylene	20.000	21.324	-6.6	94	0.03
60 T	o-Xylene	10.000	10.824	-8.2	96	0.00
61 T	Styrene	10.000	11.090	-10.9	90	0.00
62	Isopropylbenzene	10.000	10.641	-6.4	95	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	9.242	7.6	95	0.00
64	n-propylbenzene	10.000	10.303	-3.0	95	0.00
65	tert-Butylbenzene	10.000	10.672	-6.7	96	0.00
66 T	Benzyl Chloride	10.000	10.663	-6.6	87	0.00
67	sec-Butylbenzene	10.000	10.562	-5.6	96	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	10.342	-3.4	91	0.00
69	p-Isopropyltoluene	10.000	10.712	-7.1	97	0.00
70	n-Butylbenzene	10.000	10.585	-5.9	97	0.00
71	2-Chlorotoluene	10.000	10.599	-6.0	95	0.00
72 T	4-Ethyltoluene	10.000	11.069	-10.7	96	0.00
73 T	1,3,5-Trimethylbenzene	10.000	10.823	-8.2	96	0.00
74 T	1,2,4-Trimethylbenzene	10.000	10.074	-0.7	97	0.00
75 T	1,3-Dichlorobenzene	10.000	9.754	2.5	96	0.00
76 T	1,4-Dichlorobenzene	10.000	10.425	-4.3	96	0.00
77 T	1,2-Dichlorobenzene	10.000	9.719	2.8	97	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	8.893	11.1	99	0.00
79 T	Naphthalene	10.000	10.806	-8.1	94	0.00
80 T	Naphthalene,2-methyl-	10.000	11.489	-14.9	97	0.00
81 T	1,2,4-Trichlorobenzene	10.000	10.433	-4.3	96	0.00

(#) = Out of Range



SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
Data File : VL042486.D
Acq On : 06 May 2025 08:16
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri May 02 03:16:31 2025



AutoFind: Scans 3180, 3181, 3182; Background Corrected with Scan 3165

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	24.9	28768	PASS
75	95	30	66	63.6	73411	PASS
95	95	100	100	100.0	115408	PASS
96	95	5	9	6.2	7151	PASS
173	174	0.00	2	0.6	392	PASS
174	95	50	120	56.1	64757	PASS
175	174	4	9	7.6	4952	PASS
176	174	93	101	94.3	61040	PASS
177	176	5	9	6.1	3717	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042487.D
 Acq On : 06 May 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/07/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 07 01:35:37 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.791	49	69320	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	188157	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	161366	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.381 95 165339 11.051 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 110.500%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	131525	11.466	ppbv	96
3) Chlorodifluoromethane	1.477	51	119859	9.883	ppbv	98
4) Chloromethane	1.535	50	51848	10.814	ppbv	91
5) Vinyl Chloride	1.580	62	47536	9.847	ppbv	95
6) Bromomethane	1.667	94	25689	11.263	ppbv	99
7) Chloroethane	1.706	64	18985	10.651	ppbv	97
8) Dichlorotetrafluoroethane	1.554	85	105839	11.192	ppbv	97
9) Propene	1.486	41	34471	8.689	ppbv	96
10) Heptane	4.982	43	103210	10.030	ppbv	96
11) Trichlorofluoromethane	1.878	101	123855	11.700	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.140	101	93934	11.114	ppbv	98
13) Ethanol	1.722	45	6351	7.889	ppbv	98
14) Bromoethene	1.784	108	36335	11.645	ppbv	92
15) Acetone	1.836	43	111680	8.825	ppbv	99
16) 1,3-Butadiene	1.609	39	55608	9.963	ppbv	95
17) tert-Butyl alcohol	2.043	59	118491	10.628	ppbv	97
18) 1,1-Dichloroethene	2.036	96	41887	10.614	ppbv	98
19) Isopropyl Alcohol	1.888	45	69030	9.856	ppbv	100
20) Methylene Chloride	2.062	84	37393	9.829	ppbv	97
21) Allyl Chloride	2.098	41	84167	10.342	ppbv	97
22) trans-1,2-Dichloroethene	2.341	96	47256	10.253	ppbv	95
23) Vinyl Acetate	2.464	43	118463	9.809	ppbv	97
24) 1,1-Dichloroethane	2.409	63	100156	10.651	ppbv	99
25) Ethyl Acetate	2.839	43	201213	9.931	ppbv	99
26) Hexane	2.826	57	88211	9.993	ppbv	97
27) Carbon Disulfide	2.150	76	120690	11.716	ppbv	93
28) Methyl tert-Butyl Ether	2.441	73	51653	9.740	ppbv	98
29) Chloroform	2.852	83	172606	10.435	ppbv	93
30) Cyclohexane	3.859	84	74929	10.153	ppbv	98
31) cis-1,2-Dichloroethene	2.723	61	97864	9.503	ppbv	93
32) 1,1,1-Trichloroethane	3.370	97	182668	10.179	ppbv	99
34) 2-Butanone	2.561	43	131262	9.645	ppbv	97
35) Carbon Tetrachloride	3.765	117	190178	11.127	ppbv	100
36) Benzene	3.661	78	181690	10.233	ppbv	99
37) 1,2-Dichloroethane	3.221	62	146010	10.669	ppbv	98
38) Trichloroethene	4.555	130	71976	10.630	ppbv	90
39) 1,2-Dichloropropane	4.322	63	63612	9.570	ppbv #	87
40) 1,4-Dioxane	4.590	88	29247	9.396	ppbv #	100
41) Tetrahydrofuran	3.059	42	64241	9.932	ppbv	98
42) Bromodichloromethane	4.496	83	185825	11.113	ppbv	91
43) Methyl Methacrylate	4.891	69	72618	10.960	ppbv	96
44) 2,2,4-Trimethylpentane	4.645	57	311522	10.824	ppbv	99
45) t-1,3-Dichloropropene	6.422	75	93178	10.719	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042487.D
 Acq On : 06 May 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/07/2025
 Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 07 01:35:37 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.606	75	108203	10.473	ppbv	95
47) 1,1,2-Trichloroethane	6.587	97	75237	10.320	ppbv	93
48) Dibromochloromethane	7.393	129	134135	10.899	ppbv	99
49) Bromoform	9.490	173	119371	11.565	ppbv	98
50) 4-Methyl-2-Pentanone	5.759	43	177017	10.283	ppbv	96
51) 2-Hexanone	7.442	43	141871	10.308	ppbv	99
52) Tetrachloroethene	8.228	164	63220	9.354	ppbv	86
53) Toluene	6.940	91	205094	10.913	ppbv	99
54) 1,2-Dibromoethane	7.659	107	117565	10.548	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.934	131	104985	11.016	ppbv	98
57) Chlorobenzene	8.927	112	169084	10.428	ppbv	92
58) Ethyl Benzene	9.361	91	308243	11.431	ppbv	100
59) m/p-Xylene	9.555	91	540767m	23.920	ppbv	
60) o-Xylene	9.969	91	271925	12.013	ppbv	99
61) Styrene	9.876	104	99907	11.828	ppbv	97
62) Isopropylbenzene	10.546	105	384743	11.784	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.969	83	167724	10.748	ppbv	99
64) n-propylbenzene	10.992	120	91898	11.517	ppbv	95
65) tert-Butylbenzene	11.536	119	349935	12.010	ppbv	98
66) Benzyl Chloride	11.620	91	43100	12.798	ppbv	99
67) sec-Butylbenzene	11.772	105	496533	11.992	ppbv	99
69) p-Isopropyltoluene	11.931	119	416243	11.931	ppbv	99
70) n-Butylbenzene	12.287	91	437958	12.144	ppbv	100
71) 2-Chlorotoluene	10.905	91	303775	11.862	ppbv	100
72) 4-Ethyltoluene	11.131	105	331569	12.424	ppbv	99
73) 1,3,5-Trimethylbenzene	11.209	105	287473	12.248	ppbv	98
74) 1,2,4-Trimethylbenzene	11.542	105	324889	11.441	ppbv	94
75) 1,3-Dichlorobenzene	11.610	146	175086	11.074	ppbv	99
76) 1,4-Dichlorobenzene	11.675	146	178127	11.906	ppbv	98
77) 1,2-Dichlorobenzene	11.950	146	170747	11.034	ppbv	100
78) Hexachloro-1,3-Butadiene	13.886	225	152349	9.859	ppbv	99
79) Naphthalene	13.517	128	292585	12.239	ppbv	97
80) Naphthalene,2-methyl-	14.462	142	225180	12.673	ppbv	99
81) 1,2,4-Trichlorobenzene	13.446	180	142121	11.147	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042487.D
 Acq On : 06 May 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 07 01:35:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	1.655	1.897	-14.6	125	0.00
3	Chlorodifluoromethane	1.750	1.729	1.2	107	0.00
4	Chloromethane	0.692	0.748	-8.1	116	0.00
5 T	Vinyl Chloride	0.696	0.686	1.4	117	0.00
6 T	Bromomethane	0.329	0.371	-12.8	126	0.00
7	Chloroethane	0.257	0.274	-6.6	115	0.00
8 T	Dichlorotetrafluoroethane	1.364	1.527	-12.0	120	0.00
9 T	Propene	0.572	0.497	13.1	92	0.00
10 T	Heptane	1.484	1.489	-0.3	100	0.00
11 T	Trichlorofluoromethane	1.527	1.787	-17.0	126	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.219	1.355	-11.2	120	0.00
13	Ethanol	0.116	0.092	20.7	112	0.00
14 T	Bromoethene	0.450	0.524	-16.4	128	0.00
15 T	Acetone	1.826	1.611	11.8	119	0.00
16 T	1,3-Butadiene	0.805	0.802	0.4	114	0.00
17	tert-Butyl alcohol	1.608	1.709	-6.3	111	0.00
18 T	1,1-Dichloroethene	0.569	0.604	-6.2	119	0.00
19 T	Isopropyl Alcohol	1.010	0.996	1.4	110	0.00
20 T	Methylene Chloride	0.549	0.539	1.8	117	0.00
21 T	Allyl Chloride	1.174	1.214	-3.4	110	0.00
22 T	trans-1,2-Dichloroethene	0.665	0.682	-2.6	120	0.00
23 T	Vinyl Acetate	1.742	1.709	1.9	96	0.00
24 T	1,1-Dichloroethane	1.357	1.445	-6.5	117	0.00
25 T	Ethyl Acetate	2.923	2.903	0.7	100	0.00
26 T	Hexane	1.273	1.273	0.0	102	0.00
27 T	Carbon Disulfide	1.486	1.741	-17.2	119	0.00
28 T	Methyl tert-Butyl Ether	0.765	0.745	2.6	108	0.00
29 T	Chloroform	2.386	2.490	-4.4	110	0.00
30 T	Cyclohexane	1.065	1.081	-1.5	100	0.00
31 T	cis-1,2-Dichloroethene	1.486	1.412	5.0	102	0.00
32 T	1,1,1-Trichloroethane	2.589	2.635	-1.8	112	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
34 T	2-Butanone	0.723	0.698	3.5	99	0.00
35 T	Carbon Tetrachloride	0.908	1.011	-11.3	114	0.00
36 T	Benzene	0.944	0.966	-2.3	101	0.00
37 T	1,2-Dichloroethane	0.727	0.776	-6.7	109	0.00
38 T	Trichloroethene	0.360	0.383	-6.4	103	0.01
39 T	1,2-Dichloropropane	0.353	0.338	4.2	100	0.01
40 T	1,4-Dioxane	0.165	0.155	6.1	97	0.00
41 T	Tetrahydrofuran	0.344	0.341	0.9	98	0.00
42 T	Bromodichloromethane	0.889	0.988	-11.1	110	0.01
43	Methyl Methacrylate	0.352	0.386	-9.7	99	0.01
44 T	2,2,4-Trimethylpentane	1.530	1.656	-8.2	108	0.01
45 T	t-1,3-Dichloropropene	0.462	0.495	-7.1	96	0.01
46 T	cis-1,3-Dichloropropene	0.549	0.575	-4.7	97	0.00
47 T	1,1,2-Trichloroethane	0.387	0.400	-3.4	105	0.00
48 T	Dibromochloromethane	0.654	0.713	-9.0	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042487.D
 Acq On : 06 May 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 07 01:35:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.549	0.634	-15.5	112	0.01
50 T	4-Methyl-2-Pentanone	0.915	0.941	-2.8	101	0.01
51 T	2-Hexanone	0.731	0.754	-3.1	98	0.00
52 T	Tetrachloroethene	0.359	0.336	6.4	103	0.00
53 T	Toluene	0.999	1.090	-9.1	100	0.01
54 T	1,2-Dibromoethane	0.592	0.625	-5.6	105	0.01
55 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
56	1,1,1,2-Tetrachloroethane	0.591	0.651	-10.2	111	0.01
57 T	Chlorobenzene	1.005	1.048	-4.3	107	0.00
58 T	Ethyl Benzene	1.671	1.910	-14.3	103	0.00
59 T	m/p-Xylene	1.401	1.676	-19.6	109	0.03
60 T	o-Xylene	1.403	1.685	-20.1	110	0.00
61 T	Styrene	0.523	0.619	-18.4	98	0.00
62	Isopropylbenzene	2.023	2.384	-17.8	109	0.00
63 T	1,1,2,2-Tetrachloroethane	0.967	1.039	-7.4	114	0.00
64	n-propylbenzene	0.495	0.570	-15.2	109	0.00
65	tert-Butylbenzene	1.806	2.169	-20.1	111	0.00
66 T	Benzyl Chloride	0.209	0.267	-27.8	107	0.00
67	sec-Butylbenzene	2.566	3.077	-19.9	112	0.00
68 S	1-Bromo-4-Fluorobenzene	0.927	1.025	-10.6	100	0.00
69	p-Isopropyltoluene	2.162	2.579	-19.3	111	0.00
70	n-Butylbenzene	2.235	2.714	-21.4	114	0.00
71	2-Chlorotoluene	1.587	1.883	-18.7	109	0.00
72 T	4-Ethyltoluene	1.654	2.055	-24.2	111	0.00
73 T	1,3,5-Trimethylbenzene	1.455	1.781	-22.4	112	0.00
74 T	1,2,4-Trimethylbenzene	1.760	2.013	-14.4	113	0.00
75 T	1,3-Dichlorobenzene	0.980	1.085	-10.7	113	0.00
76 T	1,4-Dichlorobenzene	0.927	1.104	-19.1	113	0.00
77 T	1,2-Dichlorobenzene	0.959	1.058	-10.3	114	0.00
78 T	Hexachloro-1,3-Butadiene	0.958	0.944	1.5	113	0.00
79 T	Naphthalene	1.482	1.813	-22.3	110	0.00
80 T	Naphthalene,2-methyl-	1.101	1.395	-26.7	110	0.00
81 T	1,2,4-Trichlorobenzene	0.790	0.881	-11.5	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042487.D
 Acq On : 06 May 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 07 01:35:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	10.000	11.466	-14.7	125	0.00
3	Chlorodifluoromethane	10.000	9.883	1.2	107	0.00
4	Chloromethane	10.000	10.814	-8.1	116	0.00
5 T	Vinyl Chloride	10.000	9.847	1.5	117	0.00
6 T	Bromomethane	10.000	11.263	-12.6	126	0.00
7	Chloroethane	10.000	10.651	-6.5	115	0.00
8 T	Dichlorotetrafluoroethane	10.000	11.192	-11.9	120	0.00
9 T	Propene	10.000	8.689	13.1	92	0.00
10 T	Heptane	10.000	10.030	-0.3	100	0.00
11 T	Trichlorofluoromethane	10.000	11.700	-17.0	126	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	11.114	-11.1	120	0.00
13	Ethanol	10.000	7.889	21.1	112	0.00
14 T	Bromoethene	10.000	11.645	-16.4	128	0.00
15 T	Acetone	10.000	8.825	11.8	119	0.00
16 T	1,3-Butadiene	10.000	9.963	0.4	114	0.00
17	tert-Butyl alcohol	10.000	10.628	-6.3	111	0.00
18 T	1,1-Dichloroethene	10.000	10.614	-6.1	119	0.00
19 T	Isopropyl Alcohol	10.000	9.856	1.4	110	0.00
20 T	Methylene Chloride	10.000	9.829	1.7	117	0.00
21 T	Allyl Chloride	10.000	10.342	-3.4	110	0.00
22 T	trans-1,2-Dichloroethene	10.000	10.253	-2.5	120	0.00
23 T	Vinyl Acetate	10.000	9.809	1.9	96	0.00
24 T	1,1-Dichloroethane	10.000	10.651	-6.5	117	0.00
25 T	Ethyl Acetate	10.000	9.931	0.7	100	0.00
26 T	Hexane	10.000	9.993	0.1	102	0.00
27 T	Carbon Disulfide	10.000	11.716	-17.2	119	0.00
28 T	Methyl tert-Butyl Ether	10.000	9.740	2.6	108	0.00
29 T	Chloroform	10.000	10.435	-4.4	110	0.00
30 T	Cyclohexane	10.000	10.153	-1.5	100	0.00
31 T	cis-1,2-Dichloroethene	10.000	9.503	5.0	102	0.00
32 T	1,1,1-Trichloroethane	10.000	10.179	-1.8	112	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	93	0.00
34 T	2-Butanone	10.000	9.645	3.6	99	0.00
35 T	Carbon Tetrachloride	10.000	11.127	-11.3	114	0.00
36 T	Benzene	10.000	10.233	-2.3	101	0.00
37 T	1,2-Dichloroethane	10.000	10.669	-6.7	109	0.00
38 T	Trichloroethene	10.000	10.630	-6.3	103	0.01
39 T	1,2-Dichloropropane	10.000	9.570	4.3	100	0.01
40 T	1,4-Dioxane	10.000	9.396	6.0	97	0.00
41 T	Tetrahydrofuran	10.000	9.932	0.7	98	0.00
42 T	Bromodichloromethane	10.000	11.113	-11.1	110	0.01
43	Methyl Methacrylate	10.000	10.960	-9.6	99	0.01
44 T	2,2,4-Trimethylpentane	10.000	10.824	-8.2	108	0.01
45 T	t-1,3-Dichloropropene	10.000	10.719	-7.2	96	0.01
46 T	cis-1,3-Dichloropropene	10.000	10.473	-4.7	97	0.00
47 T	1,1,2-Trichloroethane	10.000	10.320	-3.2	105	0.00
48 T	Dibromochloromethane	10.000	10.899	-9.0	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042487.D
 Acq On : 06 May 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 07 01:35:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	Bromoform	10.000	11.565	-15.6	112	0.01
50 T	4-Methyl-2-Pentanone	10.000	10.283	-2.8	101	0.01
51 T	2-Hexanone	10.000	10.308	-3.1	98	0.00
52 T	Tetrachloroethene	10.000	9.354	6.5	103	0.00
53 T	Toluene	10.000	10.913	-9.1	100	0.01
54 T	1,2-Dibromoethane	10.000	10.548	-5.5	105	0.01
55 I	Chlorobenzene-d5	10.000	10.000	0.0	93	0.00
56	1,1,1,2-Tetrachloroethane	10.000	11.016	-10.2	111	0.01
57 T	Chlorobenzene	10.000	10.428	-4.3	107	0.00
58 T	Ethyl Benzene	10.000	11.431	-14.3	103	0.00
59 T	m/p-Xylene	20.000	23.920	-19.6	109	0.03
60 T	o-Xylene	10.000	12.013	-20.1	110	0.00
61 T	Styrene	10.000	11.828	-18.3	98	0.00
62	Isopropylbenzene	10.000	11.784	-17.8	109	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	10.748	-7.5	114	0.00
64	n-propylbenzene	10.000	11.517	-15.2	109	0.00
65	tert-Butylbenzene	10.000	12.010	-20.1	111	0.00
66 T	Benzyl Chloride	10.000	12.798	-28.0	107	0.00
67	sec-Butylbenzene	10.000	11.992	-19.9	112	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	11.051	-10.5	100	0.00
69	p-Isopropyltoluene	10.000	11.931	-19.3	111	0.00
70	n-Butylbenzene	10.000	12.144	-21.4	114	0.00
71	2-Chlorotoluene	10.000	11.862	-18.6	109	0.00
72 T	4-Ethyltoluene	10.000	12.424	-24.2	111	0.00
73 T	1,3,5-Trimethylbenzene	10.000	12.248	-22.5	112	0.00
74 T	1,2,4-Trimethylbenzene	10.000	11.441	-14.4	113	0.00
75 T	1,3-Dichlorobenzene	10.000	11.074	-10.7	113	0.00
76 T	1,4-Dichlorobenzene	10.000	11.906	-19.1	113	0.00
77 T	1,2-Dichlorobenzene	10.000	11.034	-10.3	114	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	9.859	1.4	113	0.00
79 T	Naphthalene	10.000	12.239	-22.4	110	0.00
80 T	Naphthalene,2-methyl-	10.000	12.673	-26.7	110	0.00
81 T	1,2,4-Trichlorobenzene	10.000	11.147	-11.5	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042488.D
 Acq On : 06 May 2025 09:49
 Operator : SY/MD
 Sample : VL0506ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0506ABL01

Quant Time: May 07 01:36:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	71766	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	194806	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	158449	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	146316	9.960	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.600%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
Data File : VL042488.D
Acq On : 06 May 2025 09:49
Operator : SY/MD
Sample : VL0506ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0506ABL01

Quant Time: May 07 01:36:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 02 03:16:31 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042489.D
 Acq On : 06 May 2025 10:44
 Operator : SY/MD
 Sample : VL0506ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0506ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/07/2025
 Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 07 01:36:51 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	68512	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	189766	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	162176	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.380 95 162347 10.797 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 108.000%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	128425	11.327	ppbv	99
3) Chlorodifluoromethane	1.476	51	120834	10.080	ppbv	98
4) Chloromethane	1.531	50	51363	10.839	ppbv	94
5) Vinyl Chloride	1.580	62	50552	10.596	ppbv	95
6) Bromomethane	1.667	94	24844	11.021	ppbv #	81
7) Chloroethane	1.706	64	19253	10.929	ppbv	98
8) Dichlorotetrafluoroethane	1.554	85	108256	11.583	ppbv	99
9) Propene	1.483	41	36687	9.357	ppbv	100
10) Heptane	4.985	43	108748	10.693	ppbv	99
11) Trichlorofluoromethane	1.878	101	118964	11.371	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.140	101	92491	11.072	ppbv	97
13) Ethanol	1.722	45	5717	7.185	ppbv	96
14) Bromoethene	1.781	108	35036	11.361	ppbv	99
15) Acetone	1.836	43	110911	8.868	ppbv	98
16) 1,3-Butadiene	1.609	39	57769	10.473	ppbv	100
17) tert-Butyl alcohol	2.043	59	121629	11.038	ppbv	98
18) 1,1-Dichloroethene	2.033	96	41573	10.658	ppbv	91
19) Isopropyl Alcohol	1.887	45	70143	10.133	ppbv	100
20) Methylene Chloride	2.062	84	38025	10.113	ppbv	97
21) Allyl Chloride	2.095	41	87030	10.820	ppbv	98
22) trans-1,2-Dichloroethene	2.341	96	46194	10.140	ppbv	92
23) Vinyl Acetate	2.463	43	122507	10.263	ppbv	99
24) 1,1-Dichloroethane	2.408	63	99878	10.747	ppbv	98
25) Ethyl Acetate	2.839	43	207809	10.377	ppbv	99
26) Hexane	2.826	57	94005	10.775	ppbv	95
27) Carbon Disulfide	2.150	76	118732	11.662	ppbv #	89
28) Methyl tert-Butyl Ether	2.441	73	52295	9.977	ppbv	97
29) Chloroform	2.849	83	178140	10.896	ppbv	98
30) Cyclohexane	3.858	84	78419	10.751	ppbv	98
31) cis-1,2-Dichloroethene	2.722	61	102999	10.119	ppbv	97
32) 1,1,1-Trichloroethane	3.370	97	185042	10.433	ppbv	99
34) 2-Butanone	2.561	43	135534	9.874	ppbv	95
35) Carbon Tetrachloride	3.765	117	192452	11.164	ppbv	96
36) Benzene	3.658	78	188082	10.503	ppbv	96
37) 1,2-Dichloroethane	3.221	62	150741	10.922	ppbv	99
38) Trichloroethene	4.554	130	75153	11.005	ppbv	97
39) 1,2-Dichloropropane	4.315	63	67616	10.086	ppbv	92
40) 1,4-Dioxane	4.583	88	30720	9.786	ppbv #	100
41) Tetrahydrofuran	3.056	42	66397	10.179	ppbv	98
42) Bromodichloromethane	4.496	83	188247	11.163	ppbv	96
43) Methyl Methacrylate	4.888	69	73710	11.030	ppbv	95
44) 2,2,4-Trimethylpentane	4.642	57	317017	10.921	ppbv	99
45) t-1,3-Dichloropropene	6.419	75	95245	10.864	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042489.D
 Acq On : 06 May 2025 10:44
 Operator : SY/MD
 Sample : VL0506ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0506ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/07/2025
 Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 07 01:36:51 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.603	75	112024	10.751	ppbv	96
47) 1,1,2-Trichloroethane	6.590	97	80000	10.880	ppbv	99
48) Dibromochloromethane	7.390	129	137019	11.039	ppbv	99
49) Bromoform	9.487	173	115649	11.110	ppbv	99
50) 4-Methyl-2-Pentanone	5.755	43	179961	10.365	ppbv	97
51) 2-Hexanone	7.441	43	143801	10.359	ppbv #	99
52) Tetrachloroethene	8.228	164	61852	9.074	ppbv	94
53) Toluene	6.940	91	211514	11.159	ppbv	99
54) 1,2-Dibromoethane	7.655	107	119785	10.656	ppbv	90
56) 1,1,1,2-Tetrachloroethane	8.927	131	107810	11.256	ppbv	99
57) Chlorobenzene	8.927	112	170186	10.444	ppbv	98
58) Ethyl Benzene	9.361	91	316187	11.667	ppbv	100
59) m/p-Xylene	9.555	91	546871m	24.069	ppbv	
60) o-Xylene	9.966	91	275574	12.114	ppbv	98
61) Styrene	9.875	104	100944	11.891	ppbv	95
62) Isopropylbenzene	10.542	105	386743	11.786	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.966	83	168155	10.722	ppbv	99
64) n-propylbenzene	10.992	120	95485	11.906	ppbv	98
65) tert-Butylbenzene	11.536	119	347665	11.872	ppbv	97
66) Benzyl Chloride	11.617	91	44409	13.121	ppbv	96
67) sec-Butylbenzene	11.772	105	500648	12.031	ppbv	99
69) p-Isopropyltoluene	11.931	119	418184	11.927	ppbv	99
70) n-Butylbenzene	12.283	91	439757	12.133	ppbv	99
71) 2-Chlorotoluene	10.905	91	304635	11.836	ppbv	99
72) 4-Ethyltoluene	11.128	105	332254	12.388	ppbv	99
73) 1,3,5-Trimethylbenzene	11.209	105	285001	12.082	ppbv	94
74) 1,2,4-Trimethylbenzene	11.539	105	326996	11.457	ppbv	95
75) 1,3-Dichlorobenzene	11.610	146	176191	11.089	ppbv	99
76) 1,4-Dichlorobenzene	11.675	146	173576	11.543	ppbv	98
77) 1,2-Dichlorobenzene	11.950	146	171125	11.003	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	153031	9.854	ppbv	99
79) Naphthalene	13.517	128	289175	12.036	ppbv	98
80) Naphthalene,2-methyl-	14.458	142	236741	13.257	ppbv	99
81) 1,2,4-Trichlorobenzene	13.442	180	142190	11.096	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042489.D
 Acq On : 06 May 2025 10:44
 Operator : SY/MD
 Sample : VL0506ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0506ABS01

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 05/07/2025
 Supervised By : Mahesh Dadoda 05/09/2025


Quant Time: May 07 01:36:51 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042492.D
 Acq On : 06 May 2025 12:19
 Operator : SY/MD
 Sample : QC042225 CAN 10669
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC042225 CAN 10669

Quant Time: May 07 01:38:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	63187	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	172595	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	143665	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.384	95	131884	9.901	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.000%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
Data File : VL042492.D
Acq On : 06 May 2025 12:19
Operator : SY/MD
Sample : QC042225 CAN 10669
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC042225 CAN 10669

Quant Time: May 07 01:38:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 02 03:16:31 2025
Response via : Initial Calibration





Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
Data File : VL042949.D
Acq On : 22 Sep 2025 08:55
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Tue Sep 23 02:39:16 2025



AutoFind: Scans 3176, 3177, 3178; Background Corrected with Scan 3156

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	22.5	81605	PASS
75	95	30	66	54.7	198315	PASS
95	95	100	100	100.0	362228	PASS
96	95	5	9	6.6	23817	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	66.2	239765	PASS
175	174	4	9	7.4	17693	PASS
176	174	93	101	96.6	231701	PASS
177	176	5	9	6.6	15367	PASS

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL092225AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Tue Sep 23 02:39:16 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL042955.D 0.1 =VL042954.D 0.5 =VL042953.D 1 =VL042952.D 2 =VL042951.D 10 =VL042950.D 15 =VL042956.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.391	1.452	1.401	1.248	1.232	1.345	7.33
3) Chlorodifluoro...			1.782	1.821	1.710	1.574	1.494	1.676	8.27
4) Chloromethane			0.693	0.622	0.624	0.560	0.545	0.609	9.66
5) T Vinyl Chloride	0.729	0.562	0.529	0.567	0.524	0.470	0.472	0.550	15.95
6) T Bromomethane			0.224	0.215	0.200	0.166	0.167	0.194	13.76
7) Chloroethane			0.234	0.238	0.212	0.196	0.189	0.214	10.33
8) T Dichlorotetra...			1.227	1.239	1.180	1.047	1.047	1.148	8.24
9) T Propene			0.784	0.787	0.762	0.684	0.631	0.729	9.47
10) T Heptane			2.102	2.060	1.961	1.870	1.738	1.946	7.57
11) T Trichlorofluor...			1.354	1.376	1.346	1.199	1.210	1.297	6.58
12) T 1,1,2-Trichlor...			1.130	1.117	1.068	0.963	0.973	1.050	7.48
13) Ethanol			0.076	0.090	0.088	0.053	0.047	0.071	27.75
14) T Bromoethene			0.430	0.417	0.415	0.362	0.362	0.397	8.23
15) T Acetone			1.490	1.552	1.475	0.993	1.002	1.302	21.48
16) T 1,3-Butadiene			0.751	0.675	0.607	0.520	0.519	0.614	16.41
17) tert-Butyl alc...			1.587	1.544	1.398	1.351	1.262	1.429	9.47
18) T 1,1-Dichloroet...			0.549	0.503	0.485	0.446	0.447	0.486	8.86
19) T Isopropyl Alcohol			0.898	0.813	0.818	0.718	0.667	0.783	11.61
20) T Methylene Chlo...			0.832	0.617	0.510	0.388	0.380	0.545	34.40
21) T Allyl Chloride			0.904	0.971	0.966	0.857	0.837	0.907	6.72
22) T trans-1,2-Dich...			0.589	0.612	0.556	0.507	0.499	0.552	8.98
23) T Vinyl Acetate			2.208	1.897	1.888	1.594	1.716	1.861	12.45
24) T 1,1-Dichloroet...			1.147	1.172	1.109	0.993	0.996	1.083	7.76
25) T Ethyl Acetate			3.405	3.494	3.443	3.153	2.966	3.292	6.83
26) T Hexane			1.621	1.614	1.553	1.398	1.317	1.501	9.08
27) T Carbon Disulfide			1.395	1.436	1.396	1.271	1.232	1.346	6.59
28) T Methyl tert-Bu...			0.953	0.773	0.742	0.638	0.794	0.780	14.61
29) T Chloroform			2.078	2.042	2.038	1.848	1.804	1.962	6.42
30) T Cyclohexane			1.312	1.333	1.280	1.193	1.154	1.255	6.19
31) T cis-1,2-Dichlo...			1.443	1.464	1.465	1.357	1.300	1.406	5.26
32) T 1,1,1-Trichlor...	2.724	2.087	2.017	2.164	2.071	1.902	1.868	2.119	13.51

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.744	0.731	0.741	0.655	0.633	0.701	7.52
35) T Carbon Tetrach...	0.734	0.616	0.660	0.665	0.657	0.610	0.617	0.651	6.63
36) T Benzene			0.974	0.970	0.972	0.898	0.881	0.939	4.84
37) T 1,2-Dichloroet...			0.502	0.517	0.507	0.470	0.473	0.494	4.28
38) T Trichloroethene	0.527	0.495	0.481	0.467	0.452	0.426	0.419	0.467	8.20
39) T 1,2-Dichloropr...			0.344	0.361	0.356	0.327	0.325	0.343	4.85

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\										
Method File : VL092225AIR.M										
40)	T	1,4-Dioxane		0.173	0.172	0.163	0.148	0.141	0.159	8.95
41)	T	Tetrahydrofuran		0.427	0.407	0.419	0.389	0.372	0.403	5.57
42)	T	Bromodichlorom...		0.689	0.699	0.702	0.681	0.667	0.688	2.07
43)		Methyl Methacr...		0.406	0.396	0.396	0.368	0.355	0.384	5.61
44)	T	2,2,4-Trimethy...		1.705	1.660	1.690	1.563	1.492	1.622	5.64
45)	T	t-1,3-Dichloro...		0.403	0.414	0.421	0.409	0.402	0.410	1.99
46)	T	cis-1,3-Dichlo...		0.522	0.528	0.537	0.523	0.501	0.522	2.55
47)	T	1,1,2-Trichlor...		0.382	0.377	0.371	0.344	0.336	0.362	5.70
48)	T	Dibromochlorom...		0.572	0.586	0.600	0.571	0.568	0.580	2.36
49)	T	Bromoform		0.477	0.507	0.534	0.508	0.495	0.504	4.18
50)	T	4-Methyl-2-Pen...		0.928	0.949	0.988	0.934	0.897	0.939	3.51
51)	T	2-Hexanone		0.719	0.774	0.769	0.751	0.730	0.749	3.19
52)	T	Tetrachloroethene	0.436 0.404	0.356	0.355	0.351	0.325	0.315	0.363	11.76
53)	T	Toluene		1.155	1.137	1.156	1.078	1.031	1.111	4.95
54)	T	1,2-Dibromoethane	0.520	0.520	0.520	0.525	0.490	0.485	0.510	3.47
-----ISTD-----										
55)	I	Chlorobenzene-d5								
56)		1,1,1,2-Tetrac...		0.497	0.486	0.505	0.458	0.447	0.479	5.28
57)	T	Chlorobenzene		0.968	0.978	0.944	0.853	0.822	0.913	7.75
58)	T	Ethyl Benzene		1.657	1.715	1.737	1.570	1.531	1.642	5.45
59)	T	m/p-Xylene		1.332	1.365	1.361	1.225	1.190	1.295	6.29
60)	T	o-Xylene		1.335	1.379	1.359	1.221	1.176	1.294	6.96
61)	T	Styrene		0.605	0.619	0.648	0.602	0.585	0.612	3.89
62)		Isopropylbenzene		2.010	2.059	2.028	1.803	1.744	1.929	7.49
63)	T	1,1,2,2-Tetrac...	0.806 0.722	0.688	0.675	0.673	0.610	0.586	0.680	10.66
64)		n-propylbenzene		0.531	0.529	0.523	0.473	0.463	0.504	6.51
65)		tert-Butylbenzene		1.836	1.839	1.842	1.557	1.503	1.715	9.91
66)	T	Benzyl Chloride		0.170	0.173	0.196	0.195	0.165	0.180	8.14
67)		sec-Butylbenzene		2.508	2.559	2.579	2.191	2.115	2.390	9.19
68)	S	1-Bromo-4-Fluo...	0.736 0.738	0.749	0.743	0.747	0.763	0.750	0.746	1.20
69)		p-Isopropyltol...		2.105	2.168	2.179	1.856	1.788	2.019	9.11
70)		n-Butylbenzene		1.929	2.108	2.136	1.881	1.803	1.971	7.36
71)		2-Chlorotoluene		1.543	1.504	1.528	1.373	1.333	1.456	6.61
72)	T	4-Ethyltoluene		1.715	1.686	1.703	1.523	1.476	1.621	6.92
73)	T	1,3,5-Trimethy...		1.333	1.346	1.399	1.263	1.231	1.314	5.12
74)	T	1,2,4-Trimethy...		1.534	1.522	1.558	1.349	1.294	1.451	8.34
75)	T	1,3-Dichlorobe...		0.918	0.954	0.956	0.845	0.827	0.900	6.72
76)	T	1,4-Dichlorobe...		0.922	0.914	0.959	0.845	0.829	0.894	6.12
77)	T	1,2-Dichlorobe...		0.874	0.914	0.918	0.796	0.780	0.856	7.61
78)	T	Hexachloro-1,3...		0.755	0.742	0.728	0.575	0.549	0.670	14.79
79)	T	Naphthalene	1.093	1.102	1.320	1.474	1.373	1.339	1.283	11.96
80)	T	Naphthalene,2- ...		0.384	0.365	0.428	0.402	0.404	0.396	5.94
81)	T	1,2,4-Trichlor...		0.557	0.691	0.777	0.718	0.689	0.686	11.73

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042950.D
 Acq On : 22 Sep 2025 09:39
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/23/2025
 Supervised By : Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:08:12 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 23 02:07:33 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	170899	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	526679	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	486724	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.371 95 371195 10.217 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.200%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	213351	9.282	ppbv	100
3) Chlorodifluoromethane	1.473	51	269073	9.392	ppbv	100
4) Chloromethane	1.531	50	95679	9.196	ppbv	100
5) Vinyl Chloride	1.577	62	80350	8.543	ppbv	100
6) Bromomethane	1.667	94	28301	8.530	ppbv	100
7) Chloroethane	1.703	64	33446	9.158	ppbv	100
8) Dichlorotetrafluoroethane	1.554	85	178992	9.122	ppbv	100
9) Propene	1.483	41	116849	9.376	ppbv	100
10) Heptane	4.969	43	319577	9.560	ppbv	100
11) Trichlorofluoromethane	1.874	101	204850	9.243	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.137	101	164598	9.171	ppbv	100
13) Ethanol	1.719	45	9094	7.613	ppbv	100
14) Bromoethene	1.780	108	61824	9.111	ppbv	100
15) Acetone	1.832	43	169676	7.623	ppbv	100
16) 1,3-Butadiene	1.606	39	88812	8.459	ppbv	100
17) tert-Butyl alcohol	2.036	59	230853	9.456	ppbv	100
18) 1,1-Dichloroethene	2.030	96	76184	9.170	ppbv	100
19) Isopropyl Alcohol	1.884	45	122753	9.174	ppbv	100
20) Methylene Chloride	2.059	84	66314	9.992	ppbv	100
21) Allyl Chloride	2.094	41	146473	9.449	ppbv	100
22) trans-1,2-Dichloroethene	2.337	96	86580	9.172	ppbv	100
23) Vinyl Acetate	2.460	43	272378	8.566	ppbv	100
24) 1,1-Dichloroethane	2.405	63	169721	9.168	ppbv	100
25) Ethyl Acetate	2.829	43	538761	9.577	ppbv	100
26) Hexane	2.823	57	238896	9.316	ppbv	100
27) Carbon Disulfide	2.146	76	217259	9.446	ppbv	100
28) Methyl tert-Butyl Ether	2.434	73	108987	8.176	ppbv	100
29) Chloroform	2.845	83	315779	9.418	ppbv	100
30) Cyclohexane	3.852	84	203803	9.506	ppbv	100
31) cis-1,2-Dichloroethene	2.716	61	231907	9.654	ppbv	100
32) 1,1,1-Trichloroethane	3.363	97	325118	8.947	ppbv	100
34) 2-Butanone	2.551	43	344874	9.343	ppbv	100
35) Carbon Tetrachloride	3.755	117	321267	9.480	ppbv	100
36) Benzene	3.651	78	472855	9.561	ppbv	100
37) 1,2-Dichloroethane	3.214	62	247754	9.525	ppbv	100
38) Trichloroethene	4.541	130	224516	9.134	ppbv	100
39) 1,2-Dichloropropane	4.305	63	172059	9.523	ppbv	100
40) 1,4-Dioxane	4.570	88	77889	9.283	ppbv #	100
41) Tetrahydrofuran	3.046	42	204724	9.649	ppbv	100
42) Bromodichloromethane	4.480	83	358858	9.903	ppbv	100
43) Methyl Methacrylate	4.871	69	193676	9.646	ppbv	100
44) 2,2,4-Trimethylpentane	4.632	57	823031	9.631	ppbv	100
45) t-1,3-Dichloropropene	6.406	75	215233	9.969	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042950.D
 Acq On : 22 Sep 2025 09:39
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/23/2025
 Supervised By : Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:08:12 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Tue Sep 23 02:07:33 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.590	75	275210	10.006	ppbv	100
47) 1,1,2-Trichloroethane	6.571	97	181329	9.498	ppbv	100
48) Dibromochloromethane	7.383	129	300497	9.845	ppbv	100
49) Bromoform	9.480	173	267814	10.087	ppbv	100
50) 4-Methyl-2-Pentanone	5.736	43	492068	9.920	ppbv	100
51) 2-Hexanone	7.428	43	395744	10.036	ppbv #	100
52) Tetrachloroethene	8.218	164	171294	8.951	ppbv	100
53) Toluene	6.923	91	567779	9.699	ppbv	100
54) 1,2-Dibromoethane	7.645	107	257904	9.605	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.921	131	223066	9.575	ppbv	100
57) Chlorobenzene	8.917	112	415238	9.345	ppbv	100
58) Ethyl Benzene	9.348	91	764249	9.562	ppbv	100
59) m/p-Xylene	9.526	91	1192566m	18.913	ppbv	
60) o-Xylene	9.959	91	594329	9.437	ppbv	100
61) Styrene	9.866	104	293111	9.840	ppbv	100
62) Isopropylbenzene	10.536	105	877401	9.347	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.956	83	296678	8.963	ppbv	100
64) n-propylbenzene	10.985	120	230432	9.394	ppbv	100
65) tert-Butylbenzene	11.526	119	757958	9.078	ppbv	100
66) Benzyl Chloride	11.610	91	94982	10.842	ppbv	100
67) sec-Butylbenzene	11.762	105	1066409	9.166	ppbv	100
69) p-Isopropyltoluene	11.921	119	903376	9.192	ppbv	100
70) n-Butylbenzene	12.277	91	915624	9.543	ppbv	100
71) 2-Chlorotoluene	10.895	91	668239	9.428	ppbv	100
72) 4-Ethyltoluene	11.121	105	741298	9.398	ppbv	100
73) 1,3,5-Trimethylbenzene	11.199	105	614632	9.608	ppbv	100
74) 1,2,4-Trimethylbenzene	11.532	105	656535	9.294	ppbv	100
75) 1,3-Dichlorobenzene	11.604	146	411473	9.392	ppbv	100
76) 1,4-Dichlorobenzene	11.668	146	411464	9.456	ppbv	100
77) 1,2-Dichlorobenzene	11.944	146	387219	9.289	ppbv	100
78) Hexachloro-1,3-Butadiene	13.876	225	280080	8.591	ppbv	100
79) Naphthalene	13.510	128	668171	10.696	ppbv	100
80) Naphthalene,2-methyl-	14.452	142	195695	10.142	ppbv	100
81) 1,2,4-Trichlorobenzene	13.436	180	349393	10.461	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042951.D
 Acq On : 22 Sep 2025 10:14
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Quant Time: Sep 23 02:09:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 23 02:07:33 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 09/23/2025
 Supervised By : Mahesh Dadoda 09/24/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	170758	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	524389	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	483821	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	361586	10.012	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	= 100.100%	

Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.499	85	47851	2.083	ppbv	97
3) Chlorodifluoromethane	1.476	51	58395	2.040	ppbv	95
4) Chloromethane	1.534	50	21326	2.051	ppbv	99
5) Vinyl Chloride	1.580	62	17892	1.904	ppbv	98
6) Bromomethane	1.667	94	6816	2.056	ppbv	96
7) Chloroethane	1.706	64	7250	1.987	ppbv #	85
8) Dichlorotetrafluoroethane	1.554	85	40306	2.056	ppbv	98
9) Propene	1.483	41	26007	2.088	ppbv	99
10) Heptane	4.978	43	66978	2.005	ppbv	99
11) Trichlorofluoromethane	1.877	101	45973	2.076	ppbv	94
12) 1,1,2-Trichlorotrifluo...	2.140	101	36490	2.035	ppbv	98
13) Ethanol	1.722	45	2989	2.504	ppbv	93
14) Bromoethene	1.784	108	14189	2.093	ppbv #	96
15) Acetone	1.835	43	50357	2.264	ppbv	99
16) 1,3-Butadiene	1.609	39	20729	1.976	ppbv	94
17) tert-Butyl alcohol	2.042	59	47749	1.957	ppbv #	93
18) 1,1-Dichloroethene	2.036	96	16572	1.996	ppbv	95
19) Isopropyl Alcohol	1.887	45	27930	2.089	ppbv #	79
20) Methylene Chloride	2.062	84	17413	2.097	ppbv #	84
21) Allyl Chloride	2.098	41	32985	2.130	ppbv	99
22) trans-1,2-Dichloroethene	2.340	96	18991	2.013	ppbv	98
23) Vinyl Acetate	2.463	43	64479	2.029	ppbv #	98
24) 1,1-Dichloroethane	2.408	63	37865	2.047	ppbv	99
25) Ethyl Acetate	2.835	43	117567	2.092	ppbv	99
26) Hexane	2.826	57	53034	2.070	ppbv	98
27) Carbon Disulfide	2.153	76	47660	2.074	ppbv #	60
28) Methyl tert-Butyl Ether	2.441	73	25348	1.903	ppbv	92
29) Chloroform	2.848	83	69599	2.077	ppbv	96
30) Cyclohexane	3.855	84	43725	2.041	ppbv	99
31) cis-1,2-Dichloroethene	2.719	61	50019	2.084	ppbv	95
32) 1,1,1-Trichloroethane	3.366	97	70729	1.948	ppbv	97
34) 2-Butanone	2.557	43	77689	2.114	ppbv	98
35) Carbon Tetrachloride	3.761	117	68858	2.041	ppbv	95
36) Benzene	3.658	78	101901	2.069	ppbv	97
37) 1,2-Dichloroethane	3.217	62	53130	2.052	ppbv	98
38) Trichloroethene	4.548	130	47424	1.938	ppbv	94
39) 1,2-Dichloropropane	4.308	63	37381m	2.078	ppbv	
40) 1,4-Dioxane	4.586	88	17074m	2.044	ppbv	
41) Tetrahydrofuran	3.056	42	43962	2.081	ppbv	98
42) Bromodichloromethane	4.486	83	73675	2.042	ppbv	98
43) Methyl Methacrylate	4.881	69	41527	2.077	ppbv	99
44) 2,2,4-Trimethylpentane	4.638	57	177242	2.083	ppbv	99
45) t-1,3-Dichloropropene	6.415	75	44171	2.055	ppbv	91

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042951.D
 Acq On : 22 Sep 2025 10:14
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/23/2025
 Supervised By :Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:09:14 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 23 02:07:33 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.596	75	56335	2.057	ppbv	97
47) 1,1,2-Trichloroethane	6.577	97	38961m	2.050	ppbv	
48) Dibromochloromethane	7.383	129	62968	2.072	ppbv	98
49) Bromoform	9.480	173	56021	2.119	ppbv	99
50) 4-Methyl-2-Pentanone	5.755	43	103605m	2.098	ppbv	
51) 2-Hexanone	7.438	43	80620	2.053	ppbv	99
52) Tetrachloroethene	8.224	164	36853	1.934	ppbv	94
53) Toluene	6.933	91	121268	2.081	ppbv	100
54) 1,2-Dibromoethane	7.652	107	55022	2.058	ppbv	94
56) 1,1,1,2-Tetrachloroethane	8.920	131	48901	2.112	ppbv	98
57) Chlorobenzene	8.924	112	91325	2.068	ppbv	94
58) Ethyl Benzene	9.354	91	168063	2.115	ppbv	99
59) m/p-Xylene	9.532	91	263362m	4.202	ppbv	
60) o-Xylene	9.963	91	131507	2.101	ppbv	98
61) Styrene	9.869	104	62748	2.119	ppbv	98
62) Isopropylbenzene	10.539	105	196198	2.103	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.963	83	65089	1.978	ppbv	98
64) n-propylbenzene	10.985	120	50569	2.074	ppbv	93
65) tert-Butylbenzene	11.529	119	178218	2.147	ppbv	98
66) Benzyl Chloride	11.613	91	18992	2.181	ppbv	98
67) sec-Butylbenzene	11.765	105	249510	2.158	ppbv	100
69) p-Isopropyltoluene	11.924	119	210808	2.158	ppbv	97
70) n-Butylbenzene	12.280	91	206691	2.167	ppbv	100
71) 2-Chlorotoluene	10.898	91	147901	2.099	ppbv	100
72) 4-Ethyltoluene	11.124	105	164770	2.102	ppbv	99
73) 1,3,5-Trimethylbenzene	11.202	105	135344	2.128	ppbv	100
74) 1,2,4-Trimethylbenzene	11.536	105	150772	2.147	ppbv	98
75) 1,3-Dichlorobenzene	11.607	146	92487	2.124	ppbv	98
76) 1,4-Dichlorobenzene	11.671	146	92794	2.145	ppbv	99
77) 1,2-Dichlorobenzene	11.947	146	88825	2.144	ppbv	100
78) Hexachloro-1,3-Butadiene	13.879	225	70481	2.175	ppbv	99
79) Naphthalene	13.510	128	142653	2.297	ppbv	98
80) Naphthalene,2-methyl-	14.458	142	41369	2.157	ppbv	98
81) 1,2,4-Trichlorobenzene	13.439	180	75139	2.263	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042952.D
 Acq On : 22 Sep 2025 10:48
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/23/2025
 Supervised By : Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:10:27 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 23 02:07:33 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	162797	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	511767	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	469265	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.370 95 348878 9.960 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 99.600%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	23644	1.080	ppbv	97
3) Chlorodifluoromethane	1.476	51	29646	1.086	ppbv	99
4) Chloromethane	1.531	50	10120	1.021	ppbv	97
5) Vinyl Chloride	1.580	62	9228	1.030	ppbv	99
6) Bromomethane	1.667	94	3493	1.105	ppbv	96
7) Chloroethane	1.703	64	3872	1.113	ppbv #	81
8) Dichlorotetrafluoroethane	1.554	85	20172	1.079	ppbv	97
9) Propene	1.482	41	12808	1.079	ppbv	92
10) Heptane	4.972	43	33540m	1.053	ppbv	
11) Trichlorofluoromethane	1.874	101	22397	1.061	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.136	101	18180	1.063	ppbv	99
13) Ethanol	1.719	45	1463	1.286	ppbv #	33
14) Bromoethene	1.780	108	6783	1.049	ppbv #	87
15) Acetone	1.832	43	25273	1.192	ppbv	99
16) 1,3-Butadiene	1.609	39	10993	1.099	ppbv	88
17) tert-Butyl alcohol	2.039	59	25141	1.081	ppbv #	88
18) 1,1-Dichloroethene	2.033	96	8192	1.035	ppbv	96
19) Isopropyl Alcohol	1.884	45	13241	1.039	ppbv #	41
20) Methylene Chloride	2.059	84	10045	0.985	ppbv	97
21) Allyl Chloride	2.094	41	15801	1.070	ppbv	95
22) trans-1,2-Dichloroethene	2.337	96	9960	1.108	ppbv	96
23) Vinyl Acetate	2.457	43	30884	1.020	ppbv #	96
24) 1,1-Dichloroethane	2.405	63	19077	1.082	ppbv	99
25) Ethyl Acetate	2.832	43	56877	1.061	ppbv	97
26) Hexane	2.822	57	26276	1.076	ppbv	96
27) Carbon Disulfide	2.149	76	23373	1.067	ppbv #	32
28) Methyl tert-Butyl Ether	2.441	73	12580	0.991	ppbv #	86
29) Chloroform	2.842	83	33240	1.041	ppbv	99
30) Cyclohexane	3.855	84	21707	1.063	ppbv	98
31) cis-1,2-Dichloroethene	2.716	61	23828	1.041	ppbv	96
32) 1,1,1-Trichloroethane	3.360	97	35233	1.018	ppbv	97
34) 2-Butanone	2.554	43	37431	1.044	ppbv	97
35) Carbon Tetrachloride	3.758	117	34016	1.033	ppbv	96
36) Benzene	3.651	78	49652	1.033	ppbv	100
37) 1,2-Dichloroethane	3.214	62	26482	1.048	ppbv	98
38) Trichloroethene	4.538	130	23906	1.001	ppbv	98
39) 1,2-Dichloropropane	4.311	63	18493m	1.053	ppbv	
40) 1,4-Dioxane	4.583	88	8804m	1.080	ppbv	
41) Tetrahydrofuran	3.055	42	20845	1.011	ppbv	98
42) Bromodichloromethane	4.480	83	35768m	1.016	ppbv	
43) Methyl Methacrylate	4.871	69	20289	1.040	ppbv	98
44) 2,2,4-Trimethylpentane	4.635	57	84970	1.023	ppbv	96
45) t-1,3-Dichloropropene	6.402	75	21210m	1.011	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042952.D
 Acq On : 22 Sep 2025 10:48
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/23/2025
 Supervised By :Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:10:27 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Sep 23 02:07:33 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.587	75	27025m	1.011	ppbv	
47) 1,1,2-Trichloroethane	6.574	97	19311	1.041	ppbv	96
48) Dibromochloromethane	7.383	129	30015	1.012	ppbv	99
49) Bromoform	9.477	173	25929	1.005	ppbv	98
50) 4-Methyl-2-Pentanone	5.748	43	48546	1.007	ppbv	98
51) 2-Hexanone	7.431	43	39627	1.034	ppbv #	99
52) Tetrachloroethene	8.218	164	18188	0.978	ppbv	97
53) Toluene	6.926	91	58200	1.023	ppbv	100
54) 1,2-Dibromoethane	7.645	107	26608	1.020	ppbv	94
56) 1,1,1,2-Tetrachloroethane	8.920	131	22802	1.015	ppbv	94
57) Chlorobenzene	8.914	112	45873	1.071	ppbv	98
58) Ethyl Benzene	9.351	91	80473	1.044	ppbv	97
59) m/p-Xylene	9.529	91	128139m	2.108	ppbv	
60) o-Xylene	9.959	91	64703	1.066	ppbv	99
61) Styrene	9.869	104	29061	1.012	ppbv	98
62) Isopropylbenzene	10.535	105	96621	1.068	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.959	83	31695	0.993	ppbv	96
64) n-propylbenzene	10.985	120	24820	1.049	ppbv	98
65) tert-Butylbenzene	11.526	119	86303	1.072	ppbv	99
66) Benzyl Chloride	11.610	91	8140	0.964	ppbv	100
67) sec-Butylbenzene	11.762	105	120070	1.070	ppbv	100
69) p-Isopropyltoluene	11.921	119	101749	1.074	ppbv	99
70) n-Butylbenzene	12.277	91	98908	1.069	ppbv	99
71) 2-Chlorotoluene	10.895	91	70577	1.033	ppbv	98
72) 4-Ethyltoluene	11.121	105	79095	1.040	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	63174	1.024	ppbv	98
74) 1,2,4-Trimethylbenzene	11.532	105	71411	1.049	ppbv	95
75) 1,3-Dichlorobenzene	11.603	146	44769	1.060	ppbv	99
76) 1,4-Dichlorobenzene	11.668	146	42879	1.022	ppbv	97
77) 1,2-Dichlorobenzene	11.940	146	42911	1.068	ppbv	98
78) Hexachloro-1,3-Butadiene	13.879	225	34798	1.107	ppbv	99
79) Naphthalene	13.510	128	61925	1.028	ppbv	98
80) Naphthalene,2-methyl-	14.455	142	17116	0.920	ppbv	97
81) 1,2,4-Trichlorobenzene	13.435	180	32411	1.007	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042953.D
 Acq On : 22 Sep 2025 11:22
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/23/2025
 Supervised By : Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:11:22 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 23 02:07:33 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	165435	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	509306	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	464091	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.374 95 347374 10.027 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.300%

Target Compounds

					Qvalue	
2) Dichlorodifluoromethane	1.499	85	11508	0.517	ppbv	96
3) Chlorodifluoromethane	1.476	51	14744	0.532	ppbv #	90
4) Chloromethane	1.534	50	5729	0.569	ppbv	97
5) Vinyl Chloride	1.580	62	4373	0.480	ppbv	93
6) Bromomethane	1.667	94	1850	0.576	ppbv	100
7) Chloroethane	1.703	64	1936	0.548	ppbv #	87
8) Dichlorotetrafluoroethane	1.554	85	10149	0.534	ppbv	96
9) Propene	1.483	41	6482	0.537	ppbv	94
10) Heptane	4.978	43	17388m	0.537	ppbv	
11) Trichlorofluoromethane	1.877	101	11200	0.522	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.140	101	9346	0.538	ppbv	100
13) Ethanol	1.719	45	629m	0.544	ppbv	
14) Bromoethene	1.784	108	3553	0.541	ppbv #	84
15) Acetone	1.835	43	12326	0.572	ppbv #	86
16) 1,3-Butadiene	1.609	39	6215	0.611	ppbv #	82
17) tert-Butyl alcohol	2.043	59	13128	0.555	ppbv #	77
18) 1,1-Dichloroethene	2.033	96	4543	0.565	ppbv	93
19) Isopropyl Alcohol	1.887	45	7431	0.574	ppbv #	46
20) Methylene Chloride	2.062	84	6881	0.430	ppbv	87
21) Allyl Chloride	2.098	41	7481	0.499	ppbv #	69
22) trans-1,2-Dichloroethene	2.337	96	4869m	0.533	ppbv	
23) Vinyl Acetate	2.460	43	18264	0.593	ppbv #	95
24) 1,1-Dichloroethane	2.408	63	9484	0.529	ppbv #	98
25) Ethyl Acetate	2.836	43	28162	0.517	ppbv #	95
26) Hexane	2.823	57	13406	0.540	ppbv	98
27) Carbon Disulfide	2.149	76	11535	0.518	ppbv #	1
28) Methyl tert-Butyl Ether	2.441	73	7886	0.611	ppbv #	65
29) Chloroform	2.845	83	17188	0.530	ppbv	100
30) Cyclohexane	3.855	84	10856	0.523	ppbv	93
31) cis-1,2-Dichloroethene	2.719	61	11939	0.513	ppbv	99
32) 1,1,1-Trichloroethane	3.366	97	16684	0.474	ppbv	96
34) 2-Butanone	2.557	43	18948	0.531	ppbv	96
35) Carbon Tetrachloride	3.761	117	16799	0.513	ppbv	94
36) Benzene	3.654	78	24807	0.519	ppbv #	95
37) 1,2-Dichloroethane	3.217	62	12788	0.508	ppbv	100
38) Trichloroethene	4.544	130	12243	0.515	ppbv #	88
39) 1,2-Dichloropropane	4.315	63	8772	0.502	ppbv #	86
40) 1,4-Dioxane	4.599	88	4405m	0.543	ppbv	
41) Tetrahydrofuran	3.059	42	10871	0.530	ppbv	95
42) Bromodichloromethane	4.486	83	17552	0.501	ppbv	96
43) Methyl Methacrylate	4.881	69	10330m	0.532	ppbv	
44) 2,2,4-Trimethylpentane	4.642	57	43414m	0.525	ppbv	
45) t-1,3-Dichloropropene	6.409	75	10262m	0.492	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042953.D
 Acq On : 22 Sep 2025 11:22
 Operator : SY/MD
 Sample : VSTDIC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/23/2025
 Supervised By :Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:11:22 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Sep 23 02:07:33 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	13304m	0.500	ppbv	
47) 1,1,2-Trichloroethane	6.580	97	9725m	0.527	ppbv	
48) Dibromochloromethane	7.383	129	14563	0.493	ppbv	99
49) Bromoform	9.480	173	12136	0.473	ppbv	97
50) 4-Methyl-2-Pentanone	5.761	43	23639	0.493	ppbv	95
51) 2-Hexanone	7.441	43	18322	0.480	ppbv #	97
52) Tetrachloroethene	8.218	164	9073m	0.490	ppbv	
53) Toluene	6.927	91	29404	0.519	ppbv	99
54) 1,2-Dibromoethane	7.648	107	13233	0.510	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.924	131	11535	0.519	ppbv #	1
57) Chlorobenzene	8.920	112	22463	0.530	ppbv	91
58) Ethyl Benzene	9.351	91	38459	0.505	ppbv	96
59) m/p-Xylene	9.545	91	61801m	1.028	ppbv	
60) o-Xylene	9.959	91	30976	0.516	ppbv	98
61) Styrene	9.865	104	14048	0.495	ppbv	98
62) Isopropylbenzene	10.535	105	46640	0.521	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.959	83	15974	0.506	ppbv	97
64) n-propylbenzene	10.985	120	12333	0.527	ppbv	100
65) tert-Butylbenzene	11.529	119	42595	0.535	ppbv	99
66) Benzyl Chloride	11.613	91	3942m	0.472	ppbv	
67) sec-Butylbenzene	11.765	105	58190	0.525	ppbv	99
69) p-Isopropyltoluene	11.924	119	48855	0.521	ppbv	100
70) n-Butylbenzene	12.280	91	44764	0.489	ppbv	100
71) 2-Chlorotoluene	10.898	91	35797	0.530	ppbv	96
72) 4-Ethyltoluene	11.125	105	39803	0.529	ppbv	99
73) 1,3,5-Trimethylbenzene	11.202	105	30932	0.507	ppbv	96
74) 1,2,4-Trimethylbenzene	11.532	105	35601	0.529	ppbv	97
75) 1,3-Dichlorobenzene	11.607	146	21307	0.510	ppbv	98
76) 1,4-Dichlorobenzene	11.668	146	21404	0.516	ppbv	98
77) 1,2-Dichlorobenzene	11.943	146	20279	0.510	ppbv	95
78) Hexachloro-1,3-Butadiene	13.879	225	17515	0.563	ppbv	98
79) Naphthalene	13.510	128	25577	0.429	ppbv	97
80) Naphthalene,2-methyl-	14.458	142	8906	0.484	ppbv	92
81) 1,2,4-Trichlorobenzene	13.439	180	12920	0.406	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
Data File : VL042953.D
Acq On : 22 Sep 2025 11:22
Operator : SY/MD
Sample : VSTDIC0.5
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/23/2025
Supervised By :Mahesh Dadoda 09/24/2025


Quant Time: Sep 23 02:11:22 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug

QLast Update : Tue Sep 23 02:07:33 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042954.D
 Acq On : 22 Sep 2025 11:56
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/23/2025
 Supervised By : Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:12:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 22 2025
 QLast Update : Tue Sep 23 02:07:33 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	163771	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	498673	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	457171	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	337196	9.881	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.800%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	920	0.102	ppbv #	75
32) 1,1,1-Trichloroethane	3.370	97	3418	0.098	ppbv	96
35) Carbon Tetrachloride	3.764	117	3071m	0.096	ppbv	
38) Trichloroethene	4.541	130	2470m	0.106	ppbv	
52) Tetrachloroethene	8.228	164	2013m	0.111	ppbv	
54) 1,2-Dibromoethane	7.655	107	2595m	0.102	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.966	83	3301	0.106	ppbv	97
79) Naphthalene	13.516	128	4997m	0.085	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
Data File : VL042954.D
Acq On : 22 Sep 2025 11:56
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/23/2025
Supervised By :Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:12:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Tue Sep 23 02:07:33 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042955.D
 Acq On : 22 Sep 2025 12:30
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/23/2025
 Supervised By :Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:13:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 23 02:07:33 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	163615	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	504423	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	454013	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	334002	9.855	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.600%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	358	0.040	ppbv	# 53
32) 1,1,1-Trichloroethane	3.373	97	1337m	0.038	ppbv	
35) Carbon Tetrachloride	3.761	117	1110m	0.034	ppbv	
38) Trichloroethene	4.551	130	797m	0.034	ppbv	
52) Tetrachloroethene	8.231	164	660m	0.036	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.959	83	1098m	0.036	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
Data File : VL042955.D
Acq On : 22 Sep 2025 12:30
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Sep 23 02:13:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Sep 23 02:07:33 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 09/23/2025
Supervised By : Mahesh Dadoda 09/24/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042956.D
 Acq On : 22 Sep 2025 13:05
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/23/2025
 Supervised By :Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:14:14 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Sep 23 02:07:33 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	162846	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	494779	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	458256	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 343752 10.049 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.500%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.502	85	300946	13.740	ppbv	97
3) Chlorodifluoromethane	1.476	51	364959	13.369	ppbv	100
4) Chloromethane	1.534	50	133240	13.439	ppbv	99
5) Vinyl Chloride	1.583	62	115257	12.860	ppbv	100
6) Bromomethane	1.670	94	40861	12.925	ppbv	94
7) Chloroethane	1.706	64	46081	13.241	ppbv	99
8) Dichlorotetrafluoroethane	1.557	85	255815	13.682	ppbv	98
9) Propene	1.486	41	154069	12.973	ppbv	94
10) Heptane	4.985	43	424500	13.327	ppbv	98
11) Trichlorofluoromethane	1.881	101	295479	13.991	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.143	101	237615	13.894	ppbv	99
13) Ethanol	1.722	45	11502	10.105	ppbv	99
14) Bromoethene	1.784	108	88375	13.668	ppbv	100
15) Acetone	1.835	43	244861	11.545	ppbv	97
16) 1,3-Butadiene	1.612	39	126670	12.661	ppbv	97
17) tert-Butyl alcohol	2.039	59	308347	13.255	ppbv	99
18) 1,1-Dichloroethene	2.036	96	109211	13.796	ppbv	99
19) Isopropyl Alcohol	1.887	45	162952	12.780	ppbv	100
20) Methylene Chloride	2.062	84	92708	14.996	ppbv	95
21) Allyl Chloride	2.098	41	204552	13.848	ppbv	100
22) trans-1,2-Dichloroethene	2.343	96	121815	13.542	ppbv	98
23) Vinyl Acetate	2.463	43	419251	13.837	ppbv	99
24) 1,1-Dichloroethane	2.411	63	243229	13.789	ppbv	99
25) Ethyl Acetate	2.835	43	724418	13.514	ppbv	99
26) Hexane	2.829	57	321710	13.166	ppbv	99
27) Carbon Disulfide	2.153	76	300931	13.731	ppbv	97
28) Methyl tert-Butyl Ether	2.444	73	193915	15.267	ppbv	94
29) Chloroform	2.852	83	440772	13.796	ppbv	98
30) Cyclohexane	3.858	84	281932	13.800	ppbv	98
31) cis-1,2-Dichloroethene	2.722	61	317456	13.869	ppbv	98
32) 1,1,1-Trichloroethane	3.369	97	456335	13.178	ppbv	99
34) 2-Butanone	2.557	43	469872	13.551	ppbv	99
35) Carbon Tetrachloride	3.764	117	457985	14.385	ppbv	97
36) Benzene	3.658	78	654205	14.080	ppbv	99
37) 1,2-Dichloroethane	3.221	62	350841	14.358	ppbv	100
38) Trichloroethene	4.548	130	310602	13.450	ppbv	97
39) 1,2-Dichloropropane	4.315	63	241323	14.218	ppbv	95
40) 1,4-Dioxane	4.580	88	104884	13.306	ppbv #	98
41) Tetrahydrofuran	3.052	42	276157	13.855	ppbv	99
42) Bromodichloromethane	4.489	83	495091	14.544	ppbv	98
43) Methyl Methacrylate	4.881	69	263659	13.979	ppbv	99
44) 2,2,4-Trimethylpentane	4.641	57	1107180	13.791	ppbv	99
45) t-1,3-Dichloropropene	6.415	75	298019	14.694	ppbv	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042956.D
 Acq On : 22 Sep 2025 13:05
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/23/2025
 Supervised By :Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:14:14 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 23 02:07:33 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.603	75	371770	14.389	ppbv	99
47) 1,1,2-Trichloroethane	6.583	97	249453	13.909	ppbv	96
48) Dibromochloromethane	7.389	129	421755	14.709	ppbv	99
49) Bromoform	9.487	173	367120	14.719	ppbv	99
50) 4-Methyl-2-Pentanone	5.748	43	666049	14.293	ppbv	100
51) 2-Hexanone	7.435	43	541611	14.620	ppbv	100
52) Tetrachloroethene	8.228	164	233980	13.015	ppbv	99
53) Toluene	6.936	91	765216	13.915	ppbv	98
54) 1,2-Dibromoethane	7.652	107	359750	14.261	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.927	131	306952	13.994	ppbv	99
57) Chlorobenzene	8.924	112	565065	13.507	ppbv	98
58) Ethyl Benzene	9.357	91	1052303	13.985	ppbv	99
59) m/p-Xylene	9.555	91	1636030m	27.557	ppbv	
60) o-Xylene	9.966	91	808098	13.629	ppbv	100
61) Styrene	9.872	104	401899	14.330	ppbv	99
62) Isopropylbenzene	10.542	105	1198767	13.564	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.966	83	403038	12.932	ppbv	100
64) n-propylbenzene	10.992	120	318541	13.793	ppbv	97
65) tert-Butylbenzene	11.532	119	1033382	13.146	ppbv	100
66) Benzyl Chloride	11.616	91	113549	13.767	ppbv	99
67) sec-Butylbenzene	11.769	105	1454158	13.276	ppbv	99
69) p-Isopropyltoluene	11.927	119	1228773	13.280	ppbv	100
70) n-Butylbenzene	12.283	91	1239042	13.716	ppbv	99
71) 2-Chlorotoluene	10.904	91	916456	13.733	ppbv	100
72) 4-Ethyltoluene	11.128	105	1014567	13.662	ppbv	99
73) 1,3,5-Trimethylbenzene	11.205	105	845958	14.046	ppbv	98
74) 1,2,4-Trimethylbenzene	11.539	105	889339	13.372	ppbv	97
75) 1,3-Dichlorobenzene	11.610	146	568597	13.785	ppbv	100
76) 1,4-Dichlorobenzene	11.675	146	570104	13.916	ppbv	99
77) 1,2-Dichlorobenzene	11.950	146	536286	13.665	ppbv	100
78) Hexachloro-1,3-Butadiene	13.882	225	377435	12.296	ppbv	99
79) Naphthalene	13.513	128	920123	15.645	ppbv	100
80) Naphthalene,2-methyl-	14.458	142	277792	15.290	ppbv	99
81) 1,2,4-Trichlorobenzene	13.442	180	473767	15.066	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042957.D
 Acq On : 22 Sep 2025 14:15
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL092225

Quant Time: Sep 23 02:49:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 09/23/2025
 Supervised By : Mahesh Dadoda 09/24/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	159826	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.956	114	493072	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	450930	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.377	95	338071	10.043	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.400%

Target Compounds	Qvalue					
2) Dichlorodifluoromethane	1.499	85	197861	9.204	ppbv	98
3) Chlorodifluoromethane	1.476	51	246838	9.213	ppbv	99
4) Chloromethane	1.535	50	85932	8.831	ppbv	100
5) Vinyl Chloride	1.580	62	74819	8.506	ppbv	98
6) Bromomethane	1.667	94	26463	8.529	ppbv	91
7) Chloroethane	1.706	64	30377	8.894	ppbv	100
8) Dichlorotetrafluoroethane	1.554	85	164255	8.951	ppbv	98
9) Propene	1.483	41	105822	9.079	ppbv	95
10) Heptane	4.982	43	284333	9.141	ppbv	99
11) Trichlorofluoromethane	1.878	101	189800	9.157	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.140	101	151398	9.020	ppbv	99
13) Ethanol	1.722	45	6606	5.842	ppbv	90
14) Bromoethene	1.781	108	55748	8.785	ppbv	98
15) Acetone	1.836	43	157306	7.557	ppbv	98
16) 1,3-Butadiene	1.609	39	80854	8.234	ppbv	98
17) tert-Butyl alcohol	2.040	59	206466	9.043	ppbv	100
18) 1,1-Dichloroethene	2.036	96	69181	8.904	ppbv	94
19) Isopropyl Alcohol	1.884	45	108171	8.644	ppbv	100
20) Methylene Chloride	2.062	84	61743	9.945	ppbv	98
21) Allyl Chloride	2.098	41	132878	9.166	ppbv	100
22) trans-1,2-Dichloroethene	2.341	96	78085	8.845	ppbv	98
23) Vinyl Acetate	2.464	43	280339	9.427	ppbv	99
24) 1,1-Dichloroethane	2.409	63	155861	9.003	ppbv	100
25) Ethyl Acetate	2.833	43	486993	9.256	ppbv	99
26) Hexane	2.826	57	216458	9.026	ppbv	99
27) Carbon Disulfide	2.150	76	196191	9.121	ppbv	98
28) Methyl tert-Butyl Ether	2.441	73	121564	9.752	ppbv	95
29) Chloroform	2.849	83	300507	9.583	ppbv	100
30) Cyclohexane	3.855	84	186186	9.286	ppbv	98
31) cis-1,2-Dichloroethene	2.719	61	213362	9.497	ppbv	99
32) 1,1,1-Trichloroethane	3.370	97	307237	9.071	ppbv	100
34) 2-Butanone	2.558	43	313801	9.081	ppbv	100
35) Carbon Tetrachloride	3.762	117	304293	9.479	ppbv	97
36) Benzene	3.655	78	435981	9.416	ppbv	98
37) 1,2-Dichloroethane	3.218	62	233305	9.581	ppbv	99
38) Trichloroethene	4.548	130	205431	8.927	ppbv	95
39) 1,2-Dichloropropane	4.312	63	160275	9.482	ppbv	95
40) 1,4-Dioxane	4.577	88	69659	8.863	ppbv #	95
41) Tetrahydrofuran	3.053	42	184258	9.277	ppbv	100
42) Bromodichloromethane	4.490	83	329680	9.721	ppbv	99
43) Methyl Methacrylate	4.878	69	172867	9.125	ppbv	99
44) 2,2,4-Trimethylpentane	4.642	57	738876	9.239	ppbv	99
45) t-1,3-Dichloropropene	6.416	75	188332	9.321	ppbv	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042957.D
 Acq On : 22 Sep 2025 14:15
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL092225

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/23/2025
 Supervised By : Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:49:01 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 23 02:39:16 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.603	75	244367	9.490	ppbv	99
47) 1,1,2-Trichloroethane	6.577	97	165721	9.279	ppbv	97
48) Dibromochloromethane	7.387	129	278090	9.732	ppbv	99
49) Bromoform	9.484	173	238337	9.589	ppbv	99
50) 4-Methyl-2-Pentanone	5.746	43	435845	9.411	ppbv	99
51) 2-Hexanone	7.435	43	352962	9.561	ppbv	100
52) Tetrachloroethene	8.225	164	155469	8.678	ppbv	98
53) Toluene	6.933	91	509968	9.306	ppbv	99
54) 1,2-Dibromoethane	7.652	107	237604	9.452	ppbv	96
56) 1,1,1,2-Tetrachloroethane	8.924	131	205811	9.536	ppbv	99
57) Chlorobenzene	8.924	112	379110	9.209	ppbv	99
58) Ethyl Benzene	9.354	91	698086	9.428	ppbv	100
59) m/p-Xylene	9.552	91	1092496m	18.715	ppbv	
60) o-Xylene	9.966	91	538861	9.236	ppbv	99
61) Styrene	9.872	104	264257	9.575	ppbv	99
62) Isopropylbenzene	10.539	105	802723	9.230	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.963	83	272416	8.883	ppbv	100
64) n-propylbenzene	10.989	120	214374	9.433	ppbv	99
65) tert-Butylbenzene	11.533	119	711319	9.196	ppbv	99
66) Benzyl Chloride	11.617	91	71896	8.858	ppbv	99
67) sec-Butylbenzene	11.769	105	995407	9.235	ppbv	99
69) p-Isopropyltoluene	11.928	119	848196	9.316	ppbv	99
70) n-Butylbenzene	12.280	91	834882	9.392	ppbv	100
71) 2-Chlorotoluene	10.902	91	611697	9.315	ppbv	100
72) 4-Ethyltoluene	11.125	105	686229	9.391	ppbv	100
73) 1,3,5-Trimethylbenzene	11.203	105	567874	9.582	ppbv	100
74) 1,2,4-Trimethylbenzene	11.536	105	607159	9.277	ppbv	92
75) 1,3-Dichlorobenzene	11.607	146	380822	9.382	ppbv	100
76) 1,4-Dichlorobenzene	11.672	146	380109	9.429	ppbv	99
77) 1,2-Dichlorobenzene	11.947	146	359995	9.322	ppbv	100
78) Hexachloro-1,3-Butadiene	13.883	225	249142	8.248	ppbv	100
79) Naphthalene	13.514	128	590759	10.208	ppbv	99
80) Naphthalene,2-methyl-	14.459	142	173023	9.678	ppbv	98
81) 1,2,4-Trichlorobenzene	13.439	180	304647	9.845	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042957.D
 Acq On : 22 Sep 2025 14:15
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL092225

Quant Time: Sep 23 02:49:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	1.345	1.238	8.0	93	0.00
3	Chlorodifluoromethane	1.676	1.544	7.9	92	0.00
4	Chloromethane	0.609	0.538	11.7	90	0.00
5 T	Vinyl Chloride	0.550	0.468	14.9	93	0.00
6 T	Bromomethane	0.194	0.166	14.4	94	0.00
7	Chloroethane	0.214	0.190	11.2	91	0.00
8 T	Dichlorotetrafluoroethane	1.148	1.028	10.5	92	0.00
9 T	Propene	0.729	0.662	9.2	91	0.00
10 T	Heptane	1.946	1.779	8.6	89	0.01
11 T	Trichlorofluoromethane	1.297	1.188	8.4	93	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.050	0.947	9.8	92	0.00
13	Ethanol	0.071	0.041#	42.3#	73	0.00
14 T	Bromoethene	0.397	0.349	12.1	90	0.00
15 T	Acetone	1.302	0.984	24.4	93	0.00
16 T	1,3-Butadiene	0.614	0.506	17.6	91	0.00
17	tert-Butyl alcohol	1.429	1.292	9.6	89	0.00
18 T	1,1-Dichloroethene	0.486	0.433	10.9	91	0.00
19 T	Isopropyl Alcohol	0.783	0.677	13.5	88	0.00
20 T	Methylene Chloride	0.545	0.386	29.2	93	0.00
21 T	Allyl Chloride	0.907	0.831	8.4	91	0.00
22 T	trans-1,2-Dichloroethene	0.552	0.489	11.4	90	0.00
23 T	Vinyl Acetate	1.861	1.754	5.7	103	0.00
24 T	1,1-Dichloroethane	1.083	0.975	10.0	92	0.00
25 T	Ethyl Acetate	3.292	3.047	7.4	90	0.00
26 T	Hexane	1.501	1.354	9.8	91	0.00
27 T	Carbon Disulfide	1.346	1.228	8.8	90	0.00
28 T	Methyl tert-Butyl Ether	0.780	0.761	2.4	112	0.00
29 T	Chloroform	1.962	1.880	4.2	95	0.00
30 T	Cyclohexane	1.255	1.165	7.2	91	0.00
31 T	cis-1,2-Dichloroethene	1.406	1.335	5.0	92	0.00
32 T	1,1,1-Trichloroethane	2.119	1.922	9.3	95	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
34 T	2-Butanone	0.701	0.636	9.3	91	0.00
35 T	Carbon Tetrachloride	0.651	0.617	5.2	95	0.00
36 T	Benzene	0.939	0.884	5.9	92	0.00
37 T	1,2-Dichloroethane	0.494	0.473	4.3	94	0.00
38 T	Trichloroethene	0.467	0.417	10.7	91	0.00
39 T	1,2-Dichloropropane	0.343	0.325	5.2	93	0.00
40 T	1,4-Dioxane	0.159	0.141	11.3	89	0.00
41 T	Tetrahydrofuran	0.403	0.374	7.2	90	0.00
42 T	Bromodichloromethane	0.688	0.669	2.8	92	0.00
43	Methyl Methacrylate	0.384	0.351	8.6	89	0.00
44 T	2,2,4-Trimethylpentane	1.622	1.499	7.6	90	0.00
45 T	t-1,3-Dichloropropene	0.410	0.382	6.8	88	0.00
46 T	cis-1,3-Dichloropropene	0.522	0.496	5.0	89	0.01
47 T	1,1,2-Trichloroethane	0.362	0.336	7.2	91	0.00
48 T	Dibromochloromethane	0.580	0.564	2.8	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042957.D
 Acq On : 22 Sep 2025 14:15
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL092225

Quant Time: Sep 23 02:49:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.504	0.483	4.2	89	0.00
50 T	4-Methyl-2-Pentanone	0.939	0.884	5.9	89	0.00
51 T	2-Hexanone	0.749	0.716	4.4	89	0.00
52 T	Tetrachloroethene	0.363	0.315	13.2	91	0.00
53 T	Toluene	1.111	1.034	6.9	90	0.00
54 T	1,2-Dibromoethane	0.510	0.482	5.5	92	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
56	1,1,1,2-Tetrachloroethane	0.479	0.456	4.8	92	0.00
57 T	Chlorobenzene	0.913	0.841	7.9	91	0.00
58 T	Ethyl Benzene	1.642	1.548	5.7	91	0.00
59 T	m/p-Xylene	1.295	1.211	6.5	92	0.03
60 T	o-Xylene	1.294	1.195	7.7	91	0.00
61 T	Styrene	0.612	0.586	4.2	90	0.00
62	Isopropylbenzene	1.929	1.780	7.7	91	0.00
63 T	1,1,2,2-Tetrachloroethane	0.680	0.604	11.2	92	0.00
64	n-propylbenzene	0.504	0.475	5.8	93	0.00
65	tert-Butylbenzene	1.715	1.577	8.0	94	0.00
66 T	Benzyl Chloride	0.180	0.159	11.7	76	0.00
67	sec-Butylbenzene	2.390	2.207	7.7	93	0.00
68 S	1-Bromo-4-Fluorobenzene	0.746	0.750	-0.5	91	0.00
69	p-Isopropyltoluene	2.019	1.881	6.8	94	0.00
70	n-Butylbenzene	1.971	1.851	6.1	91	0.00
71	2-Chlorotoluene	1.456	1.357	6.8	92	0.00
72 T	4-Ethyltoluene	1.621	1.522	6.1	93	0.00
73 T	1,3,5-Trimethylbenzene	1.314	1.259	4.2	92	0.00
74 T	1,2,4-Trimethylbenzene	1.451	1.346	7.2	92	0.00
75 T	1,3-Dichlorobenzene	0.900	0.845	6.1	93	0.00
76 T	1,4-Dichlorobenzene	0.894	0.843	5.7	92	0.00
77 T	1,2-Dichlorobenzene	0.856	0.798	6.8	93	0.00
78 T	Hexachloro-1,3-Butadiene	0.670	0.553	17.5	89	0.00
79 T	Naphthalene	1.283	1.310	-2.1	88	0.00
80 T	Naphthalene,2-methyl-	0.396	0.384	3.0	88	0.00
81 T	1,2,4-Trichlorobenzene	0.686	0.676	1.5	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042957.D
 Acq On : 22 Sep 2025 14:15
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL092225

Quant Time: Sep 23 02:49:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	10.000	9.204	8.0	93	0.00
3	Chlorodifluoromethane	10.000	9.213	7.9	92	0.00
4	Chloromethane	10.000	8.831	11.7	90	0.00
5 T	Vinyl Chloride	10.000	8.506	14.9	93	0.00
6 T	Bromomethane	10.000	8.529	14.7	94	0.00
7	Chloroethane	10.000	8.894	11.1	91	0.00
8 T	Dichlorotetrafluoroethane	10.000	8.951	10.5	92	0.00
9 T	Propene	10.000	9.079	9.2	91	0.00
10 T	Heptane	10.000	9.141	8.6	89	0.01
11 T	Trichlorofluoromethane	10.000	9.157	8.4	93	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	9.020	9.8	92	0.00
13	Ethanol	10.000	5.842	41.6#	73	0.00
14 T	Bromoethene	10.000	8.785	12.1	90	0.00
15 T	Acetone	10.000	7.557	24.4	93	0.00
16 T	1,3-Butadiene	10.000	8.234	17.7	91	0.00
17	tert-Butyl alcohol	10.000	9.043	9.6	89	0.00
18 T	1,1-Dichloroethene	10.000	8.904	11.0	91	0.00
19 T	Isopropyl Alcohol	10.000	8.644	13.6	88	0.00
20 T	Methylene Chloride	10.000	9.945	0.5	93	0.00
21 T	Allyl Chloride	10.000	9.166	8.3	91	0.00
22 T	trans-1,2-Dichloroethene	10.000	8.845	11.5	90	0.00
23 T	Vinyl Acetate	10.000	9.427	5.7	103	0.00
24 T	1,1-Dichloroethane	10.000	9.003	10.0	92	0.00
25 T	Ethyl Acetate	10.000	9.256	7.4	90	0.00
26 T	Hexane	10.000	9.026	9.7	91	0.00
27 T	Carbon Disulfide	10.000	9.121	8.8	90	0.00
28 T	Methyl tert-Butyl Ether	10.000	9.752	2.5	112	0.00
29 T	Chloroform	10.000	9.583	4.2	95	0.00
30 T	Cyclohexane	10.000	9.286	7.1	91	0.00
31 T	cis-1,2-Dichloroethene	10.000	9.497	5.0	92	0.00
32 T	1,1,1-Trichloroethane	10.000	9.071	9.3	95	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	94	0.00
34 T	2-Butanone	10.000	9.081	9.2	91	0.00
35 T	Carbon Tetrachloride	10.000	9.479	5.2	95	0.00
36 T	Benzene	10.000	9.416	5.8	92	0.00
37 T	1,2-Dichloroethane	10.000	9.581	4.2	94	0.00
38 T	Trichloroethene	10.000	8.927	10.7	91	0.00
39 T	1,2-Dichloropropane	10.000	9.482	5.2	93	0.00
40 T	1,4-Dioxane	10.000	8.863	11.4	89	0.00
41 T	Tetrahydrofuran	10.000	9.277	7.2	90	0.00
42 T	Bromodichloromethane	10.000	9.721	2.8	92	0.00
43	Methyl Methacrylate	10.000	9.125	8.8	89	0.00
44 T	2,2,4-Trimethylpentane	10.000	9.239	7.6	90	0.00
45 T	t-1,3-Dichloropropene	10.000	9.321	6.8	88	0.00
46 T	cis-1,3-Dichloropropene	10.000	9.490	5.1	89	0.01
47 T	1,1,2-Trichloroethane	10.000	9.279	7.2	91	0.00
48 T	Dibromochloromethane	10.000	9.732	2.7	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042957.D
 Acq On : 22 Sep 2025 14:15
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL092225

Quant Time: Sep 23 02:49:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	Bromoform	10.000	9.589	4.1	89	0.00
50 T	4-Methyl-2-Pentanone	10.000	9.411	5.9	89	0.00
51 T	2-Hexanone	10.000	9.561	4.4	89	0.00
52 T	Tetrachloroethene	10.000	8.678	13.2	91	0.00
53 T	Toluene	10.000	9.306	6.9	90	0.00
54 T	1,2-Dibromoethane	10.000	9.452	5.5	92	0.00

55 I	Chlorobenzene-d5	10.000	10.000	0.0	93	0.00
56	1,1,1,2-Tetrachloroethane	10.000	9.536	4.6	92	0.00
57 T	Chlorobenzene	10.000	9.209	7.9	91	0.00
58 T	Ethyl Benzene	10.000	9.428	5.7	91	0.00
59 T	m/p-Xylene	20.000	18.715	6.4	92	0.03
60 T	o-Xylene	10.000	9.236	7.6	91	0.00
61 T	Styrene	10.000	9.575	4.3	90	0.00
62	Isopropylbenzene	10.000	9.230	7.7	91	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	8.883	11.2	92	0.00
64	n-propylbenzene	10.000	9.433	5.7	93	0.00
65	tert-Butylbenzene	10.000	9.196	8.0	94	0.00
66 T	Benzyl Chloride	10.000	8.858	11.4	76	0.00
67	sec-Butylbenzene	10.000	9.235	7.7	93	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	10.043	-0.4	91	0.00
69	p-Isopropyltoluene	10.000	9.316	6.8	94	0.00
70	n-Butylbenzene	10.000	9.392	6.1	91	0.00
71	2-Chlorotoluene	10.000	9.315	6.9	92	0.00
72 T	4-Ethyltoluene	10.000	9.391	6.1	93	0.00
73 T	1,3,5-Trimethylbenzene	10.000	9.582	4.2	92	0.00
74 T	1,2,4-Trimethylbenzene	10.000	9.277	7.2	92	0.00
75 T	1,3-Dichlorobenzene	10.000	9.382	6.2	93	0.00
76 T	1,4-Dichlorobenzene	10.000	9.429	5.7	92	0.00
77 T	1,2-Dichlorobenzene	10.000	9.322	6.8	93	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	8.248	17.5	89	0.00
79 T	Naphthalene	10.000	10.208	-2.1	88	0.00
80 T	Naphthalene,2-methyl-	10.000	9.678	3.2	88	0.00
81 T	1,2,4-Trichlorobenzene	10.000	9.845	1.5	87	0.00

(#) = Out of Range



SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
Data File : VL042982.D
Acq On : 23 Sep 2025 08:30
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Tue Sep 23 02:39:16 2025



AutoFind: Scans 3178, 3179, 3180; Background Corrected with Scan 3165

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	24.4	52027	PASS
75	95	30	66	57.3	122083	PASS
95	95	100	100	100.0	213141	PASS
96	95	5	9	6.9	14633	PASS
173	174	0.00	2	1.1	1487	PASS
174	95	50	120	62.2	132544	PASS
175	174	4	9	7.1	9369	PASS
176	174	93	101	94.3	124939	PASS
177	176	5	9	6.3	7833	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042983.D
 Acq On : 23 Sep 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/24/2025
 Supervised By : Mahesh Dadoda 09/26/2025

Quant Time: Sep 24 02:04:07 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 23 02:39:16 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.793	49	140159	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	397606	10.000	ppbv	0.01
55) Chlorobenzene-d5	8.888	117	356928	10.000	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.383 95 285330 10.709 ppbv 0.01
 Spiked Amount 10.000 Range 65 - 135 Recovery = 107.100%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.502	85	187032	9.921	ppbv	95
3) Chlorodifluoromethane	1.479	51	226561	9.643	ppbv	97
4) Chloromethane	1.534	50	78049	9.147	ppbv	99
5) Vinyl Chloride	1.583	62	67278	8.722	ppbv	100
6) Bromomethane	1.670	94	24422	8.975	ppbv	96
7) Chloroethane	1.709	64	26802	8.948	ppbv	98
8) Dichlorotetrafluoroethane	1.557	85	151817	9.434	ppbv	100
9) Propene	1.486	41	83019	8.122	ppbv	96
10) Heptane	4.991	43	242117	8.876	ppbv	99
11) Trichlorofluoromethane	1.881	101	185063	10.181	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.143	101	144385	9.809	ppbv	99
13) Ethanol	1.725	45	6504	6.559	ppbv	96
14) Bromoethene	1.787	108	51728	9.295	ppbv	98
15) Acetone	1.839	43	146605	8.031	ppbv	98
16) 1,3-Butadiene	1.612	39	74090	8.604	ppbv	97
17) tert-Butyl alcohol	2.043	59	167840	8.383	ppbv	100
18) 1,1-Dichloroethene	2.039	96	63829	9.368	ppbv	99
19) Isopropyl Alcohol	1.890	45	94735	8.633	ppbv	95
20) Methylene Chloride	2.065	84	55766	10.264	ppbv	96
21) Allyl Chloride	2.101	41	117413	9.235	ppbv	97
22) trans-1,2-Dichloroethene	2.344	96	71690	9.260	ppbv	97
23) Vinyl Acetate	2.467	43	230988	8.857	ppbv #	98
24) 1,1-Dichloroethane	2.415	63	145685	9.596	ppbv	99
25) Ethyl Acetate	2.839	43	423103	9.170	ppbv	99
26) Hexane	2.832	57	188410	8.959	ppbv	98
27) Carbon Disulfide	2.156	76	170121	9.019	ppbv	96
28) Methyl tert-Butyl Ether	2.444	73	91757	8.393	ppbv	95
29) Chloroform	2.855	83	281560	10.239	ppbv	99
30) Cyclohexane	3.862	84	158650	9.022	ppbv	97
31) cis-1,2-Dichloroethene	2.725	61	188239	9.555	ppbv	100
32) 1,1,1-Trichloroethane	3.373	97	274087	9.228	ppbv	99
34) 2-Butanone	2.560	43	279284	10.023	ppbv	99
35) Carbon Tetrachloride	3.771	117	279346	10.791	ppbv	98
36) Benzene	3.664	78	373905	10.014	ppbv	99
37) 1,2-Dichloroethane	3.224	62	224338	11.424	ppbv	97
38) Trichloroethene	4.557	130	174937	9.427	ppbv	92
39) 1,2-Dichloropropane	4.324	63	138847	10.186	ppbv	87
40) 1,4-Dioxane	4.587	88	58847	9.285	ppbv #	97
41) Tetrahydrofuran	3.059	42	152129	9.498	ppbv	100
42) Bromodichloromethane	4.499	83	297856	10.891	ppbv	97
43) Methyl Methacrylate	4.891	69	143508	9.394	ppbv	99
44) 2,2,4-Trimethylpentane	4.648	57	637574	9.887	ppbv	97
45) t-1,3-Dichloropropene	6.428	75	149384	9.169	ppbv	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042983.D
 Acq On : 23 Sep 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/24/2025
 Supervised By :Mahesh Dadoda 09/26/2025

Quant Time: Sep 24 02:04:07 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 23 02:39:16 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.613	75	197209	9.498	ppbv	96
47) 1,1,2-Trichloroethane	6.590	97	146175	10.149	ppbv	97
48) Dibromochloromethane	7.396	129	241948	10.500	ppbv	99
49) Bromoform	9.490	173	201373	10.047	ppbv	100
50) 4-Methyl-2-Pentanone	5.758	43	371641	9.951	ppbv	99
51) 2-Hexanone	7.444	43	299396	10.057	ppbv #	97
52) Tetrachloroethene	8.231	164	130498	9.033	ppbv	97
53) Toluene	6.943	91	424755	9.612	ppbv	100
54) 1,2-Dibromoethane	7.661	107	210744	10.396	ppbv	96
56) 1,1,1,2-Tetrachloroethane	8.933	131	181159	10.604	ppbv	98
57) Chlorobenzene	8.930	112	332694	10.210	ppbv	99
58) Ethyl Benzene	9.364	91	602544	10.281	ppbv	98
59) m/p-Xylene	9.558	91	964478m	20.873	ppbv	
60) o-Xylene	9.972	91	484224	10.485	ppbv	99
61) Styrene	9.878	104	212504	9.728	ppbv	97
62) Isopropylbenzene	10.545	105	703444	10.219	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.969	83	246998	10.175	ppbv	100
64) n-propylbenzene	10.995	120	181920	10.113	ppbv	94
65) tert-Butylbenzene	11.536	119	612510	10.004	ppbv	97
66) Benzyl Chloride	11.623	91	38429	5.982	ppbv	98
67) sec-Butylbenzene	11.772	105	877873	10.290	ppbv	98
69) p-Isopropyltoluene	11.930	119	733573	10.179	ppbv	98
70) n-Butylbenzene	12.286	91	744721	10.584	ppbv	99
71) 2-Chlorotoluene	10.908	91	531933	10.234	ppbv	98
72) 4-Ethyltoluene	11.131	105	592733	10.248	ppbv	98
73) 1,3,5-Trimethylbenzene	11.209	105	491234	10.472	ppbv	97
74) 1,2,4-Trimethylbenzene	11.542	105	545008	10.521	ppbv	96
75) 1,3-Dichlorobenzene	11.613	146	327787	10.203	ppbv	99
76) 1,4-Dichlorobenzene	11.678	146	328439	10.293	ppbv	98
77) 1,2-Dichlorobenzene	11.953	146	313228	10.247	ppbv	99
78) Hexachloro-1,3-Butadiene	13.885	225	214126	8.956	ppbv	99
79) Naphthalene	13.516	128	503948	11.001	ppbv	99
80) Naphthalene,2-methyl-	14.462	142	125827	8.892	ppbv	95
81) 1,2,4-Trichlorobenzene	13.445	180	261685	10.684	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042983.D
 Acq On : 23 Sep 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 24 02:04:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	1.345	1.334	0.8	88	0.00
3	Chlorodifluoromethane	1.676	1.616	3.6	84	0.00
4	Chloromethane	0.609	0.557	8.5	82	0.00
5 T	Vinyl Chloride	0.550	0.480	12.7	84	0.00
6 T	Bromomethane	0.194	0.174	10.3	86	0.00
7	Chloroethane	0.214	0.191	10.7	80	0.00
8 T	Dichlorotetrafluoroethane	1.148	1.083	5.7	85	0.00
9 T	Propene	0.729	0.592	18.8	71	0.00
10 T	Heptane	1.946	1.727	11.3	76	0.02
11 T	Trichlorofluoromethane	1.297	1.320	-1.8	90	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.050	1.030	1.9	88	0.00
13	Ethanol	0.071	0.046#	35.2#	72	0.00
14 T	Bromoethene	0.397	0.369	7.1	84	0.00
15 T	Acetone	1.302	1.046	19.7	86	0.00
16 T	1,3-Butadiene	0.614	0.529	13.8	83	0.00
17	tert-Butyl alcohol	1.429	1.197	16.2	73	0.00
18 T	1,1-Dichloroethene	0.486	0.455	6.4	84	0.00
19 T	Isopropyl Alcohol	0.783	0.676	13.7	77	0.00
20 T	Methylene Chloride	0.545	0.398	27.0	84	0.00
21 T	Allyl Chloride	0.907	0.838	7.6	80	0.00
22 T	trans-1,2-Dichloroethene	0.552	0.511	7.4	83	0.00
23 T	Vinyl Acetate	1.861	1.648	11.4	85	0.00
24 T	1,1-Dichloroethane	1.083	1.039	4.1	86	0.00
25 T	Ethyl Acetate	3.292	3.019	8.3	79	0.00
26 T	Hexane	1.501	1.344	10.5	79	0.00
27 T	Carbon Disulfide	1.346	1.214	9.8	78	0.00
28 T	Methyl tert-Butyl Ether	0.780	0.655	16.0	84	0.00
29 T	Chloroform	1.962	2.009	-2.4	89	0.00
30 T	Cyclohexane	1.255	1.132	9.8	78	0.00
31 T	cis-1,2-Dichloroethene	1.406	1.343	4.5	81	0.00
32 T	1,1,1-Trichloroethane	2.119	1.956	7.7	84	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	75	0.01
34 T	2-Butanone	0.701	0.702	-0.1	81	0.00
35 T	Carbon Tetrachloride	0.651	0.703	-8.0	87	0.02
36 T	Benzene	0.939	0.940	-0.1	79	0.01
37 T	1,2-Dichloroethane	0.494	0.564	-14.2	91	0.00
38 T	Trichloroethene	0.467	0.440	5.8	78	0.02
39 T	1,2-Dichloropropane	0.343	0.349	-1.7	81	0.02
40 T	1,4-Dioxane	0.159	0.148	6.9	76	0.02
41 T	Tetrahydrofuran	0.403	0.383	5.0	74	0.01
42 T	Bromodichloromethane	0.688	0.749	-8.9	83	0.02
43	Methyl Methacrylate	0.384	0.361	6.0	74	0.02
44 T	2,2,4-Trimethylpentane	1.622	1.604	1.1	77	0.02
45 T	t-1,3-Dichloropropene	0.410	0.376	8.3	69	0.02
46 T	cis-1,3-Dichloropropene	0.522	0.496	5.0	72	0.02
47 T	1,1,2-Trichloroethane	0.362	0.368	-1.7	81	0.02
48 T	Dibromochloromethane	0.580	0.609	-5.0	81	0.01

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042983.D
 Acq On : 23 Sep 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 24 02:04:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.504	0.506	-0.4	75	0.00
50 T	4-Methyl-2-Pentanone	0.939	0.935	0.4	76	0.02
51 T	2-Hexanone	0.749	0.753	-0.5	76	0.02
52 T	Tetrachloroethene	0.363	0.328	9.6	76	0.01
53 T	Toluene	1.111	1.068	3.9	75	0.02
54 T	1,2-Dibromoethane	0.510	0.530	-3.9	82	0.02
55 I	Chlorobenzene-d5	1.000	1.000	0.0	73	0.01
56	1,1,1,2-Tetrachloroethane	0.479	0.508	-6.1	81	0.01
57 T	Chlorobenzene	0.913	0.932	-2.1	80	0.01
58 T	Ethyl Benzene	1.642	1.688	-2.8	79	0.02
59 T	m/p-Xylene	1.295	1.351	-4.3	81	0.03
60 T	o-Xylene	1.294	1.357	-4.9	81	0.01
61 T	Styrene	0.612	0.595	2.8	72	0.01
62	Isopropylbenzene	1.929	1.971	-2.2	80	0.00
63 T	1,1,2,2-Tetrachloroethane	0.680	0.692	-1.8	83	0.01
64	n-propylbenzene	0.504	0.510	-1.2	79	0.00
65	tert-Butylbenzene	1.715	1.716	-0.1	81	0.00
66 T	Benzyl Chloride	0.180	0.108	40.0#	40	0.01
67	sec-Butylbenzene	2.390	2.460	-2.9	82	0.00
68 S	1-Bromo-4-Fluorobenzene	0.746	0.799	-7.1	77	0.01
69	p-Isopropyltoluene	2.019	2.055	-1.8	81	0.00
70	n-Butylbenzene	1.971	2.086	-5.8	81	0.00
71	2-Chlorotoluene	1.456	1.490	-2.3	80	0.01
72 T	4-Ethyltoluene	1.621	1.661	-2.5	80	0.00
73 T	1,3,5-Trimethylbenzene	1.314	1.376	-4.7	80	0.00
74 T	1,2,4-Trimethylbenzene	1.451	1.527	-5.2	83	0.00
75 T	1,3-Dichlorobenzene	0.900	0.918	-2.0	80	0.00
76 T	1,4-Dichlorobenzene	0.894	0.920	-2.9	80	0.00
77 T	1,2-Dichlorobenzene	0.856	0.878	-2.6	81	0.00
78 T	Hexachloro-1,3-Butadiene	0.670	0.600	10.4	76	0.00
79 T	Naphthalene	1.283	1.412	-10.1	75	0.00
80 T	Naphthalene,2-methyl-	0.396	0.353	10.9	64	0.00
81 T	1,2,4-Trichlorobenzene	0.686	0.733	-6.9	75	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042983.D
 Acq On : 23 Sep 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 24 02:04:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	10.000	9.921	0.8	88	0.00
3	Chlorodifluoromethane	10.000	9.643	3.6	84	0.00
4	Chloromethane	10.000	9.147	8.5	82	0.00
5 T	Vinyl Chloride	10.000	8.722	12.8	84	0.00
6 T	Bromomethane	10.000	8.975	10.3	86	0.00
7	Chloroethane	10.000	8.948	10.5	80	0.00
8 T	Dichlorotetrafluoroethane	10.000	9.434	5.7	85	0.00
9 T	Propene	10.000	8.122	18.8	71	0.00
10 T	Heptane	10.000	8.876	11.2	76	0.02
11 T	Trichlorofluoromethane	10.000	10.181	-1.8	90	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	9.809	1.9	88	0.00
13	Ethanol	10.000	6.559	34.4#	72	0.00
14 T	Bromoethene	10.000	9.295	7.1	84	0.00
15 T	Acetone	10.000	8.031	19.7	86	0.00
16 T	1,3-Butadiene	10.000	8.604	14.0	83	0.00
17	tert-Butyl alcohol	10.000	8.383	16.2	73	0.00
18 T	1,1-Dichloroethene	10.000	9.368	6.3	84	0.00
19 T	Isopropyl Alcohol	10.000	8.633	13.7	77	0.00
20 T	Methylene Chloride	10.000	10.264	-2.6	84	0.00
21 T	Allyl Chloride	10.000	9.235	7.7	80	0.00
22 T	trans-1,2-Dichloroethene	10.000	9.260	7.4	83	0.00
23 T	Vinyl Acetate	10.000	8.857	11.4	85	0.00
24 T	1,1-Dichloroethane	10.000	9.596	4.0	86	0.00
25 T	Ethyl Acetate	10.000	9.170	8.3	79	0.00
26 T	Hexane	10.000	8.959	10.4	79	0.00
27 T	Carbon Disulfide	10.000	9.019	9.8	78	0.00
28 T	Methyl tert-Butyl Ether	10.000	8.393	16.1	84	0.00
29 T	Chloroform	10.000	10.239	-2.4	89	0.00
30 T	Cyclohexane	10.000	9.022	9.8	78	0.00
31 T	cis-1,2-Dichloroethene	10.000	9.555	4.5	81	0.00
32 T	1,1,1-Trichloroethane	10.000	9.228	7.7	84	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	75	0.01
34 T	2-Butanone	10.000	10.023	-0.2	81	0.00
35 T	Carbon Tetrachloride	10.000	10.791	-7.9	87	0.02
36 T	Benzene	10.000	10.014	-0.1	79	0.01
37 T	1,2-Dichloroethane	10.000	11.424	-14.2	91	0.00
38 T	Trichloroethene	10.000	9.427	5.7	78	0.02
39 T	1,2-Dichloropropane	10.000	10.186	-1.9	81	0.02
40 T	1,4-Dioxane	10.000	9.285	7.1	76	0.02
41 T	Tetrahydrofuran	10.000	9.498	5.0	74	0.01
42 T	Bromodichloromethane	10.000	10.891	-8.9	83	0.02
43	Methyl Methacrylate	10.000	9.394	6.1	74	0.02
44 T	2,2,4-Trimethylpentane	10.000	9.887	1.1	77	0.02
45 T	t-1,3-Dichloropropene	10.000	9.169	8.3	69	0.02
46 T	cis-1,3-Dichloropropene	10.000	9.498	5.0	72	0.02
47 T	1,1,2-Trichloroethane	10.000	10.149	-1.5	81	0.02
48 T	Dibromochloromethane	10.000	10.500	-5.0	81	0.01

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042983.D
 Acq On : 23 Sep 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Sep 24 02:04:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	Bromoform	10.000	10.047	-0.5	75	0.00
50 T	4-Methyl-2-Pentanone	10.000	9.951	0.5	76	0.02
51 T	2-Hexanone	10.000	10.057	-0.6	76	0.02
52 T	Tetrachloroethene	10.000	9.033	9.7	76	0.01
53 T	Toluene	10.000	9.612	3.9	75	0.02
54 T	1,2-Dibromoethane	10.000	10.396	-4.0	82	0.02
55 I	Chlorobenzene-d5	10.000	10.000	0.0	73	0.01
56	1,1,1,2-Tetrachloroethane	10.000	10.604	-6.0	81	0.01
57 T	Chlorobenzene	10.000	10.210	-2.1	80	0.01
58 T	Ethyl Benzene	10.000	10.281	-2.8	79	0.02
59 T	m/p-Xylene	20.000	20.873	-4.4	81	0.03
60 T	o-Xylene	10.000	10.485	-4.8	81	0.01
61 T	Styrene	10.000	9.728	2.7	72	0.01
62	Isopropylbenzene	10.000	10.219	-2.2	80	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	10.175	-1.8	83	0.01
64	n-propylbenzene	10.000	10.113	-1.1	79	0.00
65	tert-Butylbenzene	10.000	10.004	-0.0	81	0.00
66 T	Benzyl Chloride	10.000	5.982	40.2#	40	0.01
67	sec-Butylbenzene	10.000	10.290	-2.9	82	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	10.709	-7.1	77	0.01
69	p-Isopropyltoluene	10.000	10.179	-1.8	81	0.00
70	n-Butylbenzene	10.000	10.584	-5.8	81	0.00
71	2-Chlorotoluene	10.000	10.234	-2.3	80	0.01
72 T	4-Ethyltoluene	10.000	10.248	-2.5	80	0.00
73 T	1,3,5-Trimethylbenzene	10.000	10.472	-4.7	80	0.00
74 T	1,2,4-Trimethylbenzene	10.000	10.521	-5.2	83	0.00
75 T	1,3-Dichlorobenzene	10.000	10.203	-2.0	80	0.00
76 T	1,4-Dichlorobenzene	10.000	10.293	-2.9	80	0.00
77 T	1,2-Dichlorobenzene	10.000	10.247	-2.5	81	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	8.956	10.4	76	0.00
79 T	Naphthalene	10.000	11.001	-10.0	75	0.00
80 T	Naphthalene,2-methyl-	10.000	8.892	11.1	64	0.00
81 T	1,2,4-Trichlorobenzene	10.000	10.684	-6.8	75	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042984.D
 Acq On : 23 Sep 2025 10:01
 Operator : SY/MD
 Sample : VL0923ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0923ABL01

Quant Time: Sep 24 02:05:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.794	49	140749	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	395903	10.000	ppbv	0.01
55) Chlorobenzene-d5	8.888	117	343517	10.000	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.383	95	271935	10.605	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.000%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
Data File : VL042984.D
Acq On : 23 Sep 2025 10:01
Operator : SY/MD
Sample : VL0923ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0923ABL01

Quant Time: Sep 24 02:05:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Sep 23 02:39:16 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042985.D
 Acq On : 23 Sep 2025 10:46
 Operator : SY/MD
 Sample : VL0923ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0923ABS01

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/24/2025
 Supervised By : Mahesh Dadoda 09/26/2025

Quant Time: Sep 24 02:05:30 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 23 02:39:16 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	141345	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.958	114	400881	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	355060	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 285118 10.757 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 107.600%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	189707	9.979	ppbv	98
3) Chlorodifluoromethane	1.476	51	229022	9.665	ppbv	96
4) Chloromethane	1.534	50	80187	9.319	ppbv	99
5) Vinyl Chloride	1.580	62	68503	8.806	ppbv	98
6) Bromomethane	1.667	94	24823	9.046	ppbv	97
7) Chloroethane	1.706	64	27737	9.182	ppbv	99
8) Dichlorotetrafluoroethane	1.554	85	155561	9.585	ppbv	99
9) Propene	1.482	41	83859	8.135	ppbv	96
10) Heptane	4.981	43	244830	8.900	ppbv	99
11) Trichlorofluoromethane	1.877	101	187304	10.218	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.140	101	145246	9.785	ppbv	100
13) Ethanol	1.722	45	6053	6.053	ppbv	93
14) Bromoethene	1.783	108	52420	9.341	ppbv	99
15) Acetone	1.835	43	151197	8.213	ppbv	95
16) 1,3-Butadiene	1.609	39	75511	8.696	ppbv	95
17) tert-Butyl alcohol	2.039	59	167211	8.281	ppbv	99
18) 1,1-Dichloroethene	2.036	96	65099	9.475	ppbv	98
19) Isopropyl Alcohol	1.887	45	91978	8.311	ppbv	95
20) Methylene Chloride	2.062	84	56301	10.276	ppbv	99
21) Allyl Chloride	2.097	41	119358	9.310	ppbv	95
22) trans-1,2-Dichloroethene	2.340	96	74438	9.534	ppbv	97
23) Vinyl Acetate	2.463	43	213995	8.137	ppbv #	98
24) 1,1-Dichloroethane	2.408	63	146400	9.562	ppbv	98
25) Ethyl Acetate	2.835	43	424873	9.132	ppbv	99
26) Hexane	2.826	57	189786	8.948	ppbv	98
27) Carbon Disulfide	2.149	76	171644	9.023	ppbv	98
28) Methyl tert-Butyl Ether	2.441	73	84694	7.682	ppbv	95
29) Chloroform	2.848	83	283787	10.233	ppbv	99
30) Cyclohexane	3.855	84	158294	8.927	ppbv	98
31) cis-1,2-Dichloroethene	2.722	61	189377	9.532	ppbv	98
32) 1,1,1-Trichloroethane	3.369	97	273236	9.122	ppbv	98
34) 2-Butanone	2.557	43	277525	9.878	ppbv	100
35) Carbon Tetrachloride	3.764	117	284218	10.890	ppbv	99
36) Benzene	3.657	78	374729	9.954	ppbv	97
37) 1,2-Dichloroethane	3.217	62	225333	11.381	ppbv	97
38) Trichloroethene	4.548	130	175469	9.379	ppbv	95
39) 1,2-Dichloropropane	4.315	63	137775	10.025	ppbv	94
40) 1,4-Dioxane	4.577	88	56926	8.908	ppbv #	99
41) Tetrahydrofuran	3.052	42	153071	9.479	ppbv	99
42) Bromodichloromethane	4.489	83	301564	10.937	ppbv	98
43) Methyl Methacrylate	4.878	69	143822	9.338	ppbv	99
44) 2,2,4-Trimethylpentane	4.641	57	644168	9.907	ppbv	97
45) t-1,3-Dichloropropene	6.415	75	146403	8.913	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042985.D
 Acq On : 23 Sep 2025 10:46
 Operator : SY/MD
 Sample : VL0923ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0923ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/24/2025
 Supervised By :Mahesh Dadoda 09/26/2025

Quant Time: Sep 24 02:05:30 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Sep 23 02:39:16 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.603	75	199882	9.548	ppbv	98
47) 1,1,2-Trichloroethane	6.580	97	146862	10.114	ppbv	98
48) Dibromochloromethane	7.389	129	239646	10.315	ppbv	100
49) Bromoform	9.487	173	200469	9.920	ppbv	99
50) 4-Methyl-2-Pentanone	5.748	43	376317	9.994	ppbv	98
51) 2-Hexanone	7.438	43	296793	9.888	ppbv #	96
52) Tetrachloroethene	8.228	164	129889	8.917	ppbv	97
53) Toluene	6.933	91	429817	9.647	ppbv	100
54) 1,2-Dibromoethane	7.655	107	207545	10.155	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.927	131	178090	10.479	ppbv	99
57) Chlorobenzene	8.924	112	334709	10.326	ppbv	97
58) Ethyl Benzene	9.357	91	594235	10.192	ppbv	99
59) m/p-Xylene	9.551	91	960667m	20.900	ppbv	
60) o-Xylene	9.966	91	478503	10.416	ppbv	99
61) Styrene	9.872	104	213951	9.846	ppbv	99
62) Isopropylbenzene	10.542	105	701023	10.237	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.962	83	246097	10.192	ppbv	99
64) n-propylbenzene	10.992	120	182650	10.207	ppbv	96
65) tert-Butylbenzene	11.532	119	617982	10.146	ppbv	99
66) Benzyl Chloride	11.616	91	37909	5.932	ppbv	96
67) sec-Butylbenzene	11.769	105	870866	10.261	ppbv	98
69) p-Isopropyltoluene	11.927	119	731903	10.209	ppbv	98
70) n-Butylbenzene	12.283	91	755608	10.795	ppbv	98
71) 2-Chlorotoluene	10.901	91	536204	10.370	ppbv	97
72) 4-Ethyltoluene	11.124	105	591673	10.283	ppbv	99
73) 1,3,5-Trimethylbenzene	11.205	105	488186	10.461	ppbv	99
74) 1,2,4-Trimethylbenzene	11.539	105	537697	10.434	ppbv	94
75) 1,3-Dichlorobenzene	11.607	146	320019	10.013	ppbv	98
76) 1,4-Dichlorobenzene	11.671	146	325338	10.250	ppbv	99
77) 1,2-Dichlorobenzene	11.947	146	313262	10.302	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	213384	8.972	ppbv	99
79) Naphthalene	13.513	128	492107	10.799	ppbv	99
80) Naphthalene,2-methyl-	14.458	142	129885	9.227	ppbv	98
81) 1,2,4-Trichlorobenzene	13.439	180	258061	10.592	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042985.D
 Acq On : 23 Sep 2025 10:46
 Operator : SY/MD
 Sample : VL0923ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0923ABS01

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 09/24/2025
 Supervised By :Mahesh Dadoda 09/26/2025


Quant Time: Sep 24 02:05:30 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug

QLast Update : Tue Sep 23 02:39:16 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
 Data File : VL042989.D
 Acq On : 23 Sep 2025 13:26
 Operator : SY/MD
 Sample : CAN 10747
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10747

Quant Time: Sep 24 02:07:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Sep 23 02:39:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.793	49	149203	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.968	114	404080	10.000	ppbv	0.02
55) Chlorobenzene-d5	8.888	117	359428	10.000	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.383	95	280412	10.451	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.500%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092325\
Data File : VL042989.D
Acq On : 23 Sep 2025 13:26
Operator : SY/MD
Sample : CAN 10747
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10747

Quant Time: Sep 24 02:07:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Sep 23 02:39:16 2025
Response via : Initial Calibration



Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL050125

Review By	Semsettin Yesilyurt	Review On	5/2/2025 7:59:23 AM		
Supervise By	Maresh Dadoda	Supervise On	5/5/2025 2:55:39 PM		
SubDirectory	VL050125	HP Acquire Method	TO15AIR.M	HP Processing Method	VL050125AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2603			
Initial Calibration Stds		AP2606,AP2607,AP2608			
CCC		AP2606			
Internal Standard/PEM		AP2603			
ICV/I.BLK		AP2609			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042457.D	01 May 2025 08:19	SY/MD	Ok
2	VSTDICCC010	VL042458.D	01 May 2025 09:03	SY/MD	Ok,M
3	VSTDICCC002	VL042459.D	01 May 2025 09:37	SY/MD	Ok,M
4	VSTDICCC001	VL042460.D	01 May 2025 10:08	SY/MD	Ok,M
5	VSTDICCC0.5	VL042461.D	01 May 2025 10:40	SY/MD	Ok,M
6	VSTDICCC0.1	VL042462.D	01 May 2025 11:12	SY/MD	Ok,M
7	VSTDICCC0.03	VL042463.D	01 May 2025 11:44	SY/MD	Not Ok
8	VSTDICCC0.03	VL042464.D	01 May 2025 12:15	SY/MD	Not Ok
9	VSTDICCC015	VL042465.D	01 May 2025 12:49	SY/MD	Ok,M
10	VSTDICCC0.03	VL042466.D	01 May 2025 14:10	SY/MD	Ok,M
11	VSTDICV010	VL042467.D	01 May 2025 15:46	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL050625

Review By	Semsettin Yesilyurt	Review On	5/9/2025 2:37:46 PM
Supervise By	Maresh Dadoda	Supervise On	5/9/2025 2:51:31 PM
SubDirectory	VL050625	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL050125AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2603		
Initial Calibration Stds	AP2606,AP2607,AP2608		
CCC	AP2606		
Internal Standard/PEM	AP2603		
ICV/I.BLK	AP2609		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042486.D	06 May 2025 08:16	SY/MD	Ok
2	VSTDCCC010	VL042487.D	06 May 2025 09:00	SY/MD	Ok,M
3	VL0506ABL01	VL042488.D	06 May 2025 09:49	SY/MD	Ok
4	VL0506ABS01	VL042489.D	06 May 2025 10:44	SY/MD	Ok,M
5	QC043025 CAN 10605	VL042490.D	06 May 2025 11:15	SY/MD	Ok
6	QC042925 CAN 10590	VL042491.D	06 May 2025 11:47	SY/MD	Ok
7	QC042225 CAN 10669	VL042492.D	06 May 2025 12:19	SY/MD	Ok
8	QC042325 CAN 10683	VL042493.D	06 May 2025 12:51	SY/MD	Ok
9	QC0429325 CAN 10458	VL042494.D	06 May 2025 13:23	SY/MD	Ok
10	MDL1 0.45	VL042495.D	06 May 2025 14:26	SY/MD	Ok,M
11	MDL2 0.45	VL042496.D	06 May 2025 14:57	SY/MD	Ok,M
12	MDL3 0.45	VL042497.D	06 May 2025 15:29	SY/MD	Not Ok
13	MDL1 0.03	VL042498.D	06 May 2025 16:01	SY/MD	Ok,M
14	MDL2 0.03	VL042499.D	06 May 2025 16:32	SY/MD	Not Ok
15	MDL3 0.03	VL042500.D	06 May 2025 17:04	SY/MD	Ok,M
16	MDL1 0.40	VL042501.D	06 May 2025 17:36	SY/MD	Ok,M
17	MDL2 0.40	VL042502.D	06 May 2025 18:07	SY/MD	Ok,M
18	MDL3 0.40	VL042503.D	06 May 2025 18:39	SY/MD	Ok,M
19	MDL1 0.1	VL042504.D	06 May 2025 19:10	SY/MD	Ok,M
20	MDL2 0.1	VL042505.D	06 May 2025 19:42	SY/MD	Ok,M
21	MDL3 0.1	VL042506.D	06 May 2025 20:13	SY/MD	Ok,M

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL050625

Review By	Semsettin Yesilyurt	Review On	5/9/2025 2:37:46 PM		
Supervise By	Mahesh Dadoda	Supervise On	5/9/2025 2:51:31 PM		
SubDirectory	VL050625	HP Acquire Method	TO15AIR.M	HP Processing Method	VL050125AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2603			
Initial Calibration Stds		AP2606,AP2607,AP2608			
CCC		AP2606			
Internal Standard/PEM		AP2603			
ICV/I.BLK		AP2609			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	VSTDCCC010	VL042507.D	06 May 2025 20:45	SY/MD	Ok,M
----	------------	------------	-------------------	-------	------

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL092225

Review By	Mahesh Dadoda	Review On	9/24/2025 2:18:14 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/24/2025 2:26:00 AM		
SubDirectory	VL092225	HP Acquire Method	TO15AIR	HP Processing Method	VL092225AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2676			
Initial Calibration Stds		AP2673,AP2674,AP2675			
CCC		AP2673			
Internal Standard/PEM		AP2676			
ICV/I.BLK		AP2671			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042949.D	22 Sep 2025 08:55	SY/MD	Ok
2	VSTDICCC010	VL042950.D	22 Sep 2025 09:39	SY/MD	Ok,M
3	VSTDICCC002	VL042951.D	22 Sep 2025 10:14	SY/MD	Ok,M
4	VSTDICCC001	VL042952.D	22 Sep 2025 10:48	SY/MD	Ok,M
5	VSTDICCC0.5	VL042953.D	22 Sep 2025 11:22	SY/MD	Ok,M
6	VSTDICCC0.1	VL042954.D	22 Sep 2025 11:56	SY/MD	Ok,M
7	VSTDICCC0.03	VL042955.D	22 Sep 2025 12:30	SY/MD	Ok,M
8	VSTDICCC015	VL042956.D	22 Sep 2025 13:05	SY/MD	Ok,M
9	VSTDICV010	VL042957.D	22 Sep 2025 14:15	SY/MD	Ok,M
10	VL0922ABL01	VL042958.D	22 Sep 2025 15:06	SY/MD	Ok
11	VL0922ABS01	VL042959.D	22 Sep 2025 15:52	SY/MD	Ok,M
12	Q3111-01	VL042960.D	22 Sep 2025 16:53	SY/MD	Dilution
13	Q3111-01DL	VL042961.D	22 Sep 2025 17:28	SY/MD	Ok,M
14	Q3112-01	VL042962.D	22 Sep 2025 18:04	SY/MD	Ok,M
15	Q3112-01DL	VL042963.D	22 Sep 2025 18:39	SY/MD	Not Ok
16	Q3112-02	VL042964.D	22 Sep 2025 19:14	SY/MD	Dilution
17	Q3112-02DL	VL042965.D	22 Sep 2025 19:49	SY/MD	Ok
18	Q3112-03	VL042966.D	22 Sep 2025 20:24	SY/MD	Ok,M
19	Q3112-03DL	VL042967.D	22 Sep 2025 20:59	SY/MD	Not Ok
20	VIBLK	VL042968.D	22 Sep 2025 21:35	SY/MD	Ok
21	Q3130-02	VL042969.D	22 Sep 2025 22:11	SY/MD	Ok,M

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL092225

Review By	Maresh Dadoda	Review On	9/24/2025 2:18:14 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/24/2025 2:26:00 AM		
SubDirectory	VL092225	HP Acquire Method	TO15AIR	HP Processing Method	VL092225AIR.M
STD. NAME	STD REF.#				
Tune/Reschk	AP2676				
Initial Calibration Stds	AP2673,AP2674,AP2675				
CCC	AP2673				
Internal Standard/PEM	AP2676				
ICV/I.BLK	AP2671				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q3130-02DUP	VL042970.D	22 Sep 2025 22:48	SY/MD	Ok,M
23	Q3130-02DL	VL042971.D	22 Sep 2025 23:23	SY/MD	Not Ok
24	Q3130-01	VL042972.D	23 Sep 2025 00:00	SY/MD	Ok,M
25	Q3130-01DL	VL042973.D	23 Sep 2025 00:35	SY/MD	Not Ok
26	Q3131-01	VL042974.D	23 Sep 2025 01:11	SY/MD	Dilution
27	Q3131-01DL	VL042975.D	23 Sep 2025 01:46	SY/MD	Ok,M
28	Q3131-02	VL042976.D	23 Sep 2025 02:22	SY/MD	Dilution
29	Q3131-02DL	VL042977.D	23 Sep 2025 02:57	SY/MD	Ok,M
30	Q3132-01	VL042978.D	23 Sep 2025 03:33	SY/MD	Dilution
31	Q3132-01DL	VL042979.D	23 Sep 2025 04:08	SY/MD	Ok,M
32	Q3132-02	VL042980.D	23 Sep 2025 04:44	SY/MD	Ok,M
33	Q3132-02DL	VL042981.D	23 Sep 2025 05:19	SY/MD	Not Ok

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL092325

Review By	Semsettin Yesilyurt	Review On	9/24/2025 9:13:51 AM
Supervise By	Maresh Dadoda	Supervise On	9/26/2025 1:40:16 AM
SubDirectory	VL092325	HP Acquire Method	TO15AIR
		HP Processing Method	VL092225AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2676		
Initial Calibration Stds	AP2673,AP2674,AP2675		
CCC	AP2673		
Internal Standard/PEM	AP2676		
ICV/I.BLK	AP2671		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042982.D	23 Sep 2025 08:30	SY/MD	Ok
2	VSTDCCC010	VL042983.D	23 Sep 2025 09:14	SY/MD	Ok,M
3	VL0923ABL01	VL042984.D	23 Sep 2025 10:01	SY/MD	Ok
4	VL0923ABS01	VL042985.D	23 Sep 2025 10:46	SY/MD	Ok,M
5	CAN 10147	VL042986.D	23 Sep 2025 11:40	SY/MD	Ok
6	CAN 10136	VL042987.D	23 Sep 2025 12:15	SY/MD	Ok
7	CAN 10684	VL042988.D	23 Sep 2025 12:50	SY/MD	Ok
8	CAN 10747	VL042989.D	23 Sep 2025 13:26	SY/MD	Ok
9	CAN 10420	VL042990.D	23 Sep 2025 14:01	SY/MD	Ok
10	CAN 10743	VL042991.D	23 Sep 2025 14:37	SY/MD	Ok
11	CAN 10668	VL042992.D	23 Sep 2025 15:13	SY/MD	Ok
12	Q2893-14 0.5 PPB LOQ	VL042993.D	23 Sep 2025 15:49	SY/MD	ReRun
13	Q2893-14 0.1 PPB LOQ	VL042994.D	23 Sep 2025 16:25	SY/MD	Ok,M
14	Q2893-13 0.45	VL042995.D	23 Sep 2025 17:01	SY/MD	Ok,M
15	Q2893-13 0.40	VL042996.D	23 Sep 2025 17:36	SY/MD	Ok,M
16	Q2893-13 0.1	VL042997.D	23 Sep 2025 18:12	SY/MD	Ok,M
17	Q2893-13 0.03	VL042998.D	23 Sep 2025 18:48	SY/MD	Ok,M
18	Q2893-14 0.03	VL042999.D	23 Sep 2025 19:24	SY/MD	ReRun

M : Manual Integration