

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: RTPE01

Lab Code: ACE

SDG No.: Q3626

Instrument ID: MSVOA_L

Calibration Date(s): 11/18/2025 11/18/2025

Heated Purge: (Y/N) N

Calibration Time(s): 09:29 13:24

GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF0.5 = VL043203.D		RRF010 = VL043204.D		RRF002 = VL043205.D		
		RRF001 = VL043206.D		RRF0.1 = VL043208.D		RRF.03 = VL043209.D		
COMPOUND	RRF0.5	RRF010	RRF002	RRF001	RRF0.1	RRF.03	RRF	% RSD
Dichlorodifluoromethane	1.398	1.340	1.476	1.505			1.410	5.5
Chloromethane	0.521	0.498	0.568	0.559			0.527	6.7
Vinyl Chloride	0.533	0.486	0.575	0.580	0.655	0.724	0.577	15.3
Bromomethane	0.300	0.253	0.286	0.302			0.280	8.3
Chloroethane	0.246	0.194	0.236	0.248			0.225	11.6
Tetrahydrofuran	0.418	0.367	0.402	0.401			0.390	6.3
Trichlorofluoromethane	1.238	1.240	1.487	1.515			1.367	9.6
1,1,2-Trichlorotrifluoroethane	1.061	0.987	1.115	1.142			1.067	5.9
Dichlorotetrafluoroethane	1.156	1.091	1.262	1.241			1.173	6.4
tert-Butyl alcohol	1.784	1.705	1.880	2.014			1.824	6.9
Heptane	2.106	1.604	1.728	1.804			1.753	12.9
1,1-Dichloroethene	0.555	0.441	0.515	0.530			0.497	10.5
Acetone	1.721	1.032	1.519	1.591			1.389	22.4
Carbon Disulfide	1.303	1.257	1.432	1.424			1.341	6
Methyl tert-Butyl Ether	0.642	0.591	0.674	0.704			0.645	7
Methylene Chloride	0.906	0.398	0.585	0.710			0.600	35.9
trans-1,2-Dichloroethene	0.605	0.499	0.599	0.614			0.568	9.4
1,1-Dichloroethane	1.050	1.018	1.194	1.173			1.096	7.4
Cyclohexane	1.135	1.024	1.130	1.194			1.101	6.8
2-Butanone	0.744	0.644	0.736	0.751			0.706	7.4
Carbon Tetrachloride	0.779	0.708	0.737	0.771	0.765	0.832	0.759	5.5
cis-1,2-Dichloroethene	1.399	1.240	1.322	1.351			1.306	5.8
Chloroform	1.955	1.785	1.928	1.898			1.867	4.5
1,1,1-Trichloroethane	1.971	1.884	2.033	2.018	2.046	2.116	1.990	4.5
2,2,4-Trimethylpentane	1.746	1.555	1.621	1.765			1.645	6.5
Benzene	0.996	0.905	0.986	1.001			0.958	5.3
1,2-Dichloroethane	0.572	0.531	0.565	0.580			0.560	3.4
Trichloroethene	0.390	0.353	0.371	0.388	0.475	0.293	0.375	14.6
1,2-Dichloropropane	0.372	0.335	0.350	0.364			0.351	5
Bromodichloromethane	0.782	0.745	0.783	0.795			0.773	2.6

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: RTPE01

Lab Code: ACE

SDG No.: Q3626

Instrument ID: MSVOA_L

Calibration Date(s): 11/18/2025 11/18/2025

Heated Purge: (Y/N) N

Calibration Time(s): 09:29 13:24

GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF0.5 = VL043203.D		RRF010 = VL043204.D		RRF002 = VL043205.D		RRF001 = VL043206.D		RRF0.1 = VL043208.D		RRF.03 = VL043209.D	
COMPOUND		RRF0.5	RRF010	RRF002	RRF001	RRF0.1	RRF.03	RRF	% RSD				
4-Methyl-2-Pentanone		1.025	0.946	1.027	1.061			0.994	6.2				
Toluene		1.198	1.075	1.146	1.182			1.134	5.3				
t-1,3-Dichloropropene		0.495	0.455	0.489	0.472			0.474	3.8				
cis-1,3-Dichloropropene		0.599	0.534	0.562	0.557			0.556	5				
1,1,2-Trichloroethane		0.424	0.344	0.378	0.372			0.372	8.9				
Dibromochloromethane		0.629	0.618	0.663	0.666			0.640	3.6				
1,2-Dibromoethane		0.587	0.525	0.549	0.569	0.528		0.547	4.8				
Tetrachloroethene		0.344	0.329	0.349	0.362	0.424	0.363	0.356	9.4				
Chlorobenzene		1.035	0.969	1.089	1.088			1.027	6.2				
Ethyl Benzene		1.927	1.821	2.008	2.038			1.925	5.2				
m/p-Xylene		1.618	1.434	1.607	1.611			1.543	6.1				
o-Xylene		1.584	1.419	1.659	1.625			1.542	7.4				
Styrene		0.669	0.648	0.719	0.690			0.675	4.5				
Bromoform		0.583	0.555	0.581	0.569			0.566	2.9				
1,1,2,2-Tetrachloroethane		1.016	0.891	1.028	1.012	1.034	0.999	0.981	6.6				
2-Chlorotoluene		2.198	1.960	2.210	2.257			2.120	6.6				
1,3,5-Trimethylbenzene		1.623	1.456	1.662	1.682			1.575	7.1				
1,2,4-Trimethylbenzene		1.796	1.548	1.871	1.844			1.719	9.5				
1,3-Dichlorobenzene		1.043	0.937	1.061	1.053			1.003	6.8				
1,4-Dichlorobenzene		1.056	0.940	1.060	1.033			1.003	6.3				
1,2-Dichlorobenzene		0.974	0.872	1.050	1.064			0.965	9.8				
1,2,4-Trichlorobenzene		0.549	0.678	0.813	0.692			0.683	13.7				
Hexachloro-1,3-Butadiene		0.854	0.686	0.920	0.899			0.805	14.9				
1,3-Butadiene		0.745	0.551	0.626	0.629			0.619	13				
Naphthalene		1.483	1.496	1.860	1.599	1.230		1.527	13.4				
4-Ethyltoluene		1.934	1.749	2.009	1.948			1.877	6.5				
1-Bromo-4-Fluorobenzene		0.729	0.735	0.741	0.733	0.725	0.712	0.733	1.9				
Hexane		1.374	1.235	1.375	1.424			1.314	8.4				
Allyl Chloride		1.139	0.849	0.936	0.895			0.935	12.8				
1,4-Dioxane		180.443	142.483	155.097	165.607			157.566	10				

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: RTPE01
 Lab Code: ACE SDG No.: Q3626
 Instrument ID: MSVOA_L Calibration Date(s): 11/18/2025 11/18/2025
 Heated Purge: (Y/N) N Calibration Time(s): 09:29 13:24
 GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:								
RRF0.5 = VL043203.D			RRF010 = VL043204.D			RRF002 = VL043205.D		
RRF001 = VL043206.D			RRF0.1 = VL043208.D			RRF.03 = VL043209.D		
COMPOUND	RRF0.5	RRF010	RRF002	RRF001	RRF0.1	RRF.03	RRF	% RSD
Methyl Methacrylate	0.576	0.462	0.511	0.532			0.508	9.5

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.