



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: RTPE01
Lab Code: ACE SDG No.: Q3626
Instrument ID: MSVOA_L Calibration Date/Time: 11/19/2025 09:31
Lab File ID: VL043232.D Init. Calib. Date(s): 11/18/2025 11/18/2025
Heated Purge: (Y/N) N Init. Calib. Time(s): 09:29 13:24
GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	10.000	11.44	-14.4	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.