

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: RTPE01
 Lab Code: ACE SDG No.: Q3626
 Instrument ID: MSVOA_L Calibration Date/Time: 11/20/2025 10:14
 Lab File ID: VL043257.D Init. Calib. Date(s): 11/18/2025 11/18/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 09:29 13:24
 GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	10.000	11.66	-16.59	20