

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Contract: RTPE01

Lab Code: ACE

SDG NO.: Q3626

Lab File ID: VL043232.D

Date Analyzed: 11/19/2025

Instrument ID: MSVOA L

Time Analyzed: 09:31

GC Column: RTX-1 ID: 0.32 (mm)

Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	92858	2.78	261062	3.95	210840	8.88
UPPER LIMIT	130001	3.11	365487	4.28	295176	9.21
LOWER LIMIT	55714.8	2.45	156637	3.62	126504	8.55
EPA SAMPLE NO.						
9A-SG-1	106377	2.79	292792	3.96	229139	8.88
9A-SG-1DL	101838	2.78	273802	3.95	211877	8.87
9A-SG-2	105026	2.78	292365	3.95	227868	8.88
9A-SG-2DUP	99705	2.78	276826	3.95	209824	8.87
9B-SG-1	102616	2.78	290774	3.95	231142	8.87
9B-SG-2	103574	2.78	291331	3.96	223974	8.88
9B-SG-2DL	98441	2.78	273742	3.95	208879	8.87
10-SG-1	103410	2.79	285776	3.96	221762	8.88
10-SG-1DL	98540	2.78	260350	3.95	200769	8.87
11-SG-1	105671	2.78	288442	3.95	236337	8.87
12-SG-1	103097	2.78	289304	3.96	221774	8.87
12-SG-2	105972	2.78	291718	3.95	225662	8.87
12-SG-2DL	100214	2.78	269502	3.95	209321	8.87
13-SG-1	101617	2.79	283781	3.96	218981	8.88
14-SG-1	106615	2.78	284605	3.95	227663	8.87
15-SG-1	105038	2.78	285629	3.95	229034	8.87
15-SG-1DL	97568	2.78	261435	3.95	198573	8.87
16-SG-1	107506	2.78	305536	3.96	246745	8.87
16-SG-1DL	102231	2.78	281627	3.95	224034	8.87
16-SG-1DL2	101416	2.78	271635	3.95	204048	8.87
VL1119ABL01	100515	2.79	277905	3.96	219972	8.88
VL1119ABS01	103562	2.78	284672	3.95	231008	8.87

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: RTPE01
 Lab Code: ACE SDG NO.: Q3626
 Lab File ID: VL043232.D Date Analyzed: 11/19/2025
 Instrument ID: MSVOA_L Time Analyzed: 09:31
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	92858	2.78	261062	3.95	210840	8.88
UPPER LIMIT	130001	3.11	365487	4.28	295176	9.21
LOWER LIMIT	55714.8	2.45	156637	3.62	126504	8.55
EPA SAMPLE NO.						

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.