

Process Report

Order ID : Q3627

Date	Bottle ID	Lab Sample Number	InComing Vacuum	Canister ID	Can Size	Reg ID	Reg Size	Batch Code	Certification ID
11/14/2025	B2510051	Q3627-05	-10.4	10793	1.4 L	10213		C2504011	VL042492.D Seq :vl050625 TUNE : VL042486.D ICAL : VL042466.D CCAL : VL042487.D MB : VL042488.D
11/14/2025	B2510051	Q3627-04	-11.2	10650	1.4 L	10712		C2504011	VL042492.D Seq :vl050625 TUNE : VL042486.D ICAL : VL042466.D CCAL : VL042487.D MB : VL042488.D
11/14/2025	B2510052	Q3627-01	-8.0	10608	6 L	10249		C2509003	VL043048.D Seq :VL100625 TUNE : VL043043.D ICAL : VL043037.D CCAL : VL043044.D MB : VL043045.D
11/14/2025	B2510052	Q3627-03	-2.7	10590	6 L	10240		C2509003	VL043048.D Seq :VL100625 TUNE : VL043043.D ICAL : VL043037.D CCAL : VL043044.D MB : VL043045.D
11/14/2025	B2510052	Q3627-02	-6.7	10321	6 L	10771		C2509003	VL043048.D Seq :VL100625 TUNE : VL043043.D ICAL : VL043037.D CCAL : VL043044.D MB : VL043045.D



Canister Cleaning

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Canister Batch : C2504011

Type of Canister : 1.4 L

Cleaning Date : 04/22/2025

Analyst : SAM

Oven Temperature : 125.0 c

Canister	Vaccum After Cleaning ("Hg)	Vaccum After Cleaning Date	Analysis Date	VAC 24Hr ("Hg)	Vaccum After 24hr Date	New Sample Number	QC Canister	File
10682	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-01	10669	VL042492.D
10793	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3627-05	10669	VL042492.D
10795	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3655-02	10669	VL042492.D
10794	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-12	10669	VL042492.D
10763	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10677	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-04	10669	VL042492.D
10669	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3655-01	10669	VL042492.D
10673	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-09	10669	VL042492.D
10046	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-08	10669	VL042492.D
10450	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-03	10669	VL042492.D
10416	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-10	10669	VL042492.D
10118	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10679	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3655-06	10669	VL042492.D
10471	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-07	10669	VL042492.D
10123	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10454	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10455	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10413	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3626-02	10669	VL042492.D
10652	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10452	-30	04/23/2025	05/06/2025	-30	04/24/2025		10669	VL042492.D
10650	-30	04/23/2025	05/06/2025	-30	04/24/2025	Q3627-04	10669	VL042492.D



Canister Cleaning

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Canister Batch : C2509003

Type of Canister : 6 L

Cleaning Date : 09/24/2025

Analyst : sam

Oven Temperature : 125.0 c



Canister	Vaccum After Cleaning ("Hg)	Vaccum After Cleaning Date	Analysis Date	VAC 24Hr ("Hg)	Vaccum After 24hr Date	New Sample Number	QC Canister	File
10321	-30	09/25/2025	10/07/2025	-30	09/26/2025	Q3627-02	10590	VL043048.D
10280	-30	09/25/2025	10/07/2025	-29	09/26/2025	Q3651-06	10590	VL043048.D
10315	-30	09/25/2025	10/07/2025	-30	09/26/2025	Q3651-05	10590	VL043048.D
10590	-30	09/25/2025	10/07/2025	-30	09/26/2025	Q3627-03	10590	VL043048.D
10608	-30	09/25/2025	10/07/2025	-30	09/26/2025	Q3627-01	10590	VL043048.D
10320	-30	09/25/2025	10/07/2025	-30	09/26/2025	Q3651-04	10590	VL043048.D

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
Data File : VL042457.D
Acq On : 01 May 2025 08:19
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri May 02 03:16:31 2025



AutoFind: Scans 3178, 3179, 3180; Background Corrected with Scan 3165

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	25.5	33424	PASS
75	95	30	66	62.8	82325	PASS
95	95	100	100	100.0	131013	PASS
96	95	5	9	6.8	8845	PASS
173	174	0.00	2	0.6	483	PASS
174	95	50	120	57.4	75195	PASS
175	174	4	9	7.5	5628	PASS
176	174	93	101	97.9	73643	PASS
177	176	5	9	6.3	4652	PASS

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL050125AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Fri May 02 03:16:31 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL042466.D 0.1 =VL042462.D 0.5 =VL042461.D 1 =VL042460.D 2 =VL042459.D 10 =VL042458.D 15 =VL042465.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.887	1.650	1.629	1.470	1.639	1.655	9.01
3) Chlorodifluoro...			1.886	1.871	1.820	1.569	1.602	1.750	8.71
4) Chloromethane			0.738	0.696	0.734	0.624	0.666	0.692	6.90
5) T Vinyl Chloride	0.756	0.964	0.708	0.625	0.623	0.567	0.631	0.696	19.18
6) T Bromomethane			0.349	0.359	0.327	0.286	0.324	0.329	8.64
7) Chloroethane			0.250	0.272	0.271	0.231	0.261	0.257	6.55
8) T Dichlorotetra...			1.445	1.371	1.402	1.235	1.368	1.364	5.77
9) T Propene			0.585	0.606	0.631	0.525	0.515	0.572	8.84
10) T Heptane			1.276	1.634	1.593	1.445	1.474	1.484	9.48
11) T Trichlorofluor...			1.660	1.489	1.528	1.376	1.582	1.527	6.96
12) T 1,1,2-Trichlor...			1.318	1.234	1.245	1.099	1.201	1.219	6.52
13) Ethanol			0.138	0.136	0.143	0.080	0.084	0.116	27.19
14) T Bromoethene			0.488	0.443	0.457	0.398	0.466	0.450	7.39
15) T Acetone			2.229	2.094	2.022	1.318	1.465	1.826	22.28
16) T 1,3-Butadiene			0.883	0.874	0.836	0.681	0.751	0.805	10.77
17) tert-Butyl alc...			1.816	1.615	1.620	1.489	1.502	1.608	8.15
18) T 1,1-Dichloroet...			0.650	0.587	0.567	0.495	0.547	0.569	9.98
19) T Isopropyl Alcohol			1.221	1.087	0.976	0.876	0.891	1.010	14.33
20) T Methylene Chlo...			0.753	0.547	0.518	0.449	0.477	0.549	21.88
21) T Allyl Chloride			1.283	1.181	1.164	1.076	1.166	1.174	6.25
22) T trans-1,2-Dich...			0.741	0.742	0.645	0.552	0.644	0.665	11.96
23) T Vinyl Acetate			1.898	1.521	1.810	1.726	1.756	1.742	8.03
24) T 1,1-Dichloroet...			1.432	1.465	1.368	1.196	1.322	1.357	7.79
25) T Ethyl Acetate			3.002	3.044	2.945	2.832	2.792	2.923	3.69
26) T Hexane			1.278	1.290	1.345	1.215	1.239	1.273	3.91
27) T Carbon Disulfide			1.509	1.446	1.499	1.418	1.558	1.486	3.69
28) T Methyl tert-Bu...			0.802	0.853	0.759	0.671	0.740	0.765	8.91
29) T Chloroform			2.551	2.482	2.399	2.198	2.301	2.386	5.89
30) T Cyclohexane			1.135	0.998	1.096	1.045	1.048	1.065	4.92
31) T cis-1,2-Dichlo...			1.551	1.604	1.545	1.342	1.386	1.486	7.71
32) T 1,1,1-Trichlor...	2.980	2.842	2.567	2.501	2.511	2.288	2.433	2.589	9.28

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.802	0.732	0.774	0.659	0.650	0.723	9.32
35) T Carbon Tetrach...	1.166	0.799	0.856	0.935	0.875	0.827	0.902	0.908	13.44
36) T Benzene			0.924	0.974	1.036	0.892	0.892	0.944	6.53
37) T 1,2-Dichloroet...			0.738	0.755	0.769	0.666	0.709	0.727	5.66
38) T Trichloroethene	0.241	0.419	0.389	0.364	0.404	0.347	0.355	0.360	16.31
39) T 1,2-Dichloropr...			0.368	0.405	0.362	0.317	0.315	0.353	10.71

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\ Method File : VL050125AIR.M											
40)	T	1,4-Dioxane			0.176	0.185	0.172	0.151	0.144	0.165	10.60
41)	T	Tetrahydrofuran			0.329	0.354	0.387	0.326	0.322	0.344	7.98
42)	T	Bromodichlorom...			0.848	0.921	0.931	0.839	0.904	0.889	4.77
43)		Methyl Methacr...			0.328	0.341	0.362	0.363	0.366	0.352	4.72
44)	T	2,2,4-Trimethy...			1.490	1.590	1.633	1.437	1.498	1.530	5.23
45)	T	t-1,3-Dichloro...			0.438	0.423	0.460	0.480	0.510	0.462	7.46
46)	T	cis-1,3-Dichlo...			0.526	0.539	0.549	0.555	0.577	0.549	3.47
47)	T	1,1,2-Trichlor...			0.372	0.409	0.428	0.357	0.371	0.387	7.74
48)	T	Dibromochlorom...			0.651	0.643	0.673	0.630	0.674	0.654	2.96
49)	T	Bromoform			0.540	0.535	0.573	0.527	0.568	0.549	3.73
50)	T	4-Methyl-2-Pen...			0.852	0.980	0.985	0.870	0.887	0.915	6.92
51)	T	2-Hexanone			0.649	0.754	0.797	0.722	0.736	0.731	7.42
52)	T	Tetrachloroethene	0.509	0.380	0.325	0.339	0.338	0.305	0.318	0.359	19.49
53)	T	Toluene			0.935	0.969	1.025	1.016	1.050	0.999	4.64
54)	T	1,2-Dibromoethane		0.625	0.589	0.599	0.610	0.554	0.577	0.592	4.26
-----ISTD-----											
55)	I	Chlorobenzene-d5									
56)		1,1,1,2-Tetrac...			0.618	0.633	0.596	0.542	0.564	0.591	6.38
57)	T	Chlorobenzene			1.101	1.060	1.036	0.904	0.923	1.005	8.64
58)	T	Ethyl Benzene			1.492	1.634	1.752	1.725	1.754	1.671	6.68
59)	T	m/p-Xylene			1.212	1.378	1.506	1.429	1.479	1.401	8.31
60)	T	o-Xylene			1.197	1.393	1.519	1.423	1.481	1.403	8.90
61)	T	Styrene			0.415	0.500	0.538	0.583	0.582	0.523	13.35
62)		Isopropylbenzene			1.781	2.030	2.165	2.033	2.108	2.023	7.26
63)	T	1,1,2,2-Tetrac...	1.140	0.945	0.976	0.985	0.990	0.846	0.887	0.967	9.69
64)		n-propylbenzene			0.400	0.524	0.549	0.482	0.518	0.495	11.75
65)		tert-Butylbenzene			1.519	1.791	2.028	1.817	1.873	1.806	10.23
66)	T	Benzyl Chloride			0.158	0.204	0.195	0.232	0.255	0.209	17.73
67)		sec-Butylbenzene			2.149	2.560	2.905	2.541	2.675	2.566	10.69
68)	S	1-Bromo-4-Fluo...	0.920	0.911	0.902	0.928	0.915	0.951	0.963	0.927	2.39
69)		p-Isopropyltol...			1.805	2.166	2.406	2.150	2.283	2.162	10.39
70)		n-Butylbenzene			1.802	2.218	2.560	2.198	2.396	2.235	12.66
71)		2-Chlorotoluene			1.405	1.569	1.697	1.597	1.667	1.587	7.20
72)	T	4-Ethyltoluene			1.281	1.647	1.828	1.722	1.792	1.654	13.28
73)	T	1,3,5-Trimethy...			1.177	1.467	1.618	1.475	1.535	1.455	11.44
74)	T	1,2,4-Trimethy...			1.568	1.788	2.075	1.646	1.723	1.760	11.07
75)	T	1,3-Dichlorobe...			0.975	1.022	1.056	0.893	0.952	0.980	6.42
76)	T	1,4-Dichlorobe...			0.869	0.945	0.966	0.907	0.949	0.927	4.22
77)	T	1,2-Dichlorobe...			0.977	1.007	1.033	0.864	0.914	0.959	7.23
78)	T	Hexachloro-1,3...			1.028	1.081	1.070	0.772	0.838	0.958	14.93
79)	T	Naphthalene		0.720	1.168	1.657	2.155	1.527	1.661	1.482	33.01
80)	T	Naphthalene,2-...			0.498	0.925	1.573	1.172	1.338	1.101	37.38
81)	T	1,2,4-Trichlor...			0.599	0.798	0.947	0.778	0.828	0.790	15.84

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042458.D
 Acq On : 01 May 2025 09:03
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/02/2025
 Supervised By : Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:39:48 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	71402	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	201290	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	174168	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 165650 10.258 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.600%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	104948	8.882	ppbv	100
3) Chlorodifluoromethane	1.476	51	112008	8.966	ppbv	100
4) Chloromethane	1.531	50	44585	9.028	ppbv	100
5) Vinyl Chloride	1.580	62	40466	8.138	ppbv	100
6) Bromomethane	1.667	94	20392	8.680	ppbv	100
7) Chloroethane	1.703	64	16528	9.002	ppbv	100
8) Dichlorotetrafluoroethane	1.554	85	88162	9.051	ppbv	100
9) Propene	1.483	41	37491	9.175	ppbv	100
10) Heptane	4.978	43	103167	9.733	ppbv	100
11) Trichlorofluoromethane	1.874	101	98215	9.320	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.136	101	78482	9.015	ppbv	100
13) Ethanol	1.719	45	5679	6.849	ppbv	100
14) Bromoethene	1.780	108	28440	8.849	ppbv	100
15) Acetone	1.835	43	94092	7.219	ppbv	100
16) 1,3-Butadiene	1.609	39	48641	8.461	ppbv	100
17) tert-Butyl alcohol	2.039	59	106304	9.257	ppbv	100
18) 1,1-Dichloroethene	2.033	96	35320	8.689	ppbv	100
19) Isopropyl Alcohol	1.884	45	62533	8.668	ppbv	100
20) Methylene Chloride	2.059	84	32039	8.176	ppbv	100
21) Allyl Chloride	2.094	41	76859	9.169	ppbv	100
22) trans-1,2-Dichloroethene	2.340	96	39427	8.305	ppbv	100
23) Vinyl Acetate	2.460	43	123215	10.146	ppbv	100
24) 1,1-Dichloroethane	2.408	63	85367	8.814	ppbv	100
25) Ethyl Acetate	2.835	43	202215	9.689	ppbv	100
26) Hexane	2.826	57	86764	9.543	ppbv	100
27) Carbon Disulfide	2.146	76	101248	9.542	ppbv #	92
28) Methyl tert-Butyl Ether	2.437	73	47941	8.777	ppbv	100
29) Chloroform	2.848	83	156976	9.213	ppbv	100
30) Cyclohexane	3.855	84	74614	9.815	ppbv	100
31) cis-1,2-Dichloroethene	2.719	61	95797	9.031	ppbv	100
32) 1,1,1-Trichloroethane	3.366	97	163334	8.836	ppbv	100
34) 2-Butanone	2.557	43	132685	9.113	ppbv	100
35) Carbon Tetrachloride	3.758	117	166490	9.511	ppbv	100
36) Benzene	3.654	78	179532	9.451	ppbv	100
37) 1,2-Dichloroethane	3.217	62	133963	9.150	ppbv	100
38) Trichloroethene	4.544	130	69846	9.642	ppbv	100
39) 1,2-Dichloropropane	4.311	63	63894	8.979	ppbv	100
40) 1,4-Dioxane	4.583	88	30303m	9.161	ppbv	
41) Tetrahydrofuran	3.052	42	65713	9.516	ppbv	100
42) Bromodichloromethane	4.486	83	168957	9.445	ppbv	100
43) Methyl Methacrylate	4.881	69	73117	10.339	ppbv	100
44) 2,2,4-Trimethylpentane	4.635	57	289161	9.420	ppbv	100
45) t-1,3-Dichloropropene	6.412	75	96692	10.385	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042458.D
 Acq On : 01 May 2025 09:03
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:39:48 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.600	75	111768	10.111	ppbv	100
47) 1,1,2-Trichloroethane	6.580	97	71805	9.207	ppbv	100
48) Dibromochloromethane	7.386	129	126754	9.628	ppbv	100
49) Bromoform	9.480	173	106108	9.610	ppbv	100
50) 4-Methyl-2-Pentanone	5.748	43	175193	9.518	ppbv	100
51) 2-Hexanone	7.435	43	145272	9.855	ppbv	100
52) Tetrachloroethene	8.224	164	61456	8.500	ppbv	100
53) Toluene	6.930	91	204494	10.171	ppbv	100
54) 1,2-Dibromoethane	7.648	107	111506	9.300	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.924	131	94346	9.172	ppbv	99
57) Chlorobenzene	8.920	112	157401	8.994	ppbv	100
58) Ethyl Benzene	9.354	91	300412	10.321	ppbv	100
59) m/p-Xylene	9.529	91	497922m	20.404	ppbv	
60) o-Xylene	9.963	91	247784	10.157	ppbv	100
61) Styrene	9.872	104	101509	11.134	ppbv	100
62) Isopropylbenzene	10.539	105	354055	10.047	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.963	83	147295	9.305	ppbv	100
64) n-propylbenzene	10.989	120	84006	9.787	ppbv	100
65) tert-Butylbenzene	11.529	119	316532	10.065	ppbv	100
66) Benzyl Chloride	11.613	91	40331	11.096	ppbv	100
67) sec-Butylbenzene	11.769	105	442600	9.903	ppbv	100
69) p-Isopropyltoluene	11.924	119	374540	9.947	ppbv	100
70) n-Butylbenzene	12.280	91	382883	9.836	ppbv	100
71) 2-Chlorotoluene	10.898	91	278215	10.065	ppbv	100
72) 4-Ethyltoluene	11.124	105	299831	10.409	ppbv	100
73) 1,3,5-Trimethylbenzene	11.202	105	256952	10.143	ppbv	100
74) 1,2,4-Trimethylbenzene	11.536	105	286615	9.351	ppbv	100
75) 1,3-Dichlorobenzene	11.607	146	155598	9.118	ppbv	100
76) 1,4-Dichlorobenzene	11.671	146	157966	9.782	ppbv	100
77) 1,2-Dichlorobenzene	11.947	146	150432	9.007	ppbv	100
78) Hexachloro-1,3-Butadiene	13.879	225	134402	8.058	ppbv	100
79) Naphthalene	13.513	128	266036	10.352	ppbv	100
80) Naphthalene,2-methyl-	14.458	142	204042	10.641	ppbv	100
81) 1,2,4-Trichlorobenzene	13.439	180	135473	9.844	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042459.D
 Acq On : 01 May 2025 09:37
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/02/2025
 Supervised By : Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:40:35 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	73220	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	201951	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	174624	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 159777 9.869 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 98.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	23859	1.969	ppbv	100
3) Chlorodifluoromethane	1.476	51	26655	2.081	ppbv	95
4) Chloromethane	1.531	50	10751	2.123	ppbv	95
5) Vinyl Chloride	1.580	62	9122	1.789	ppbv	92
6) Bromomethane	1.667	94	4793	1.990	ppbv #	62
7) Chloroethane	1.703	64	3972	2.110	ppbv	94
8) Dichlorotetrafluoroethane	1.554	85	20529	2.055	ppbv	100
9) Propene	1.482	41	9234	2.204	ppbv	95
10) Heptane	4.978	43	23329m	2.146	ppbv	
11) Trichlorofluoromethane	1.877	101	22375	2.070	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.140	101	18230	2.042	ppbv	97
13) Ethanol	1.722	45	2099	2.468	ppbv	94
14) Bromoethene	1.780	108	6686	2.029	ppbv #	96
15) Acetone	1.835	43	29607	2.215	ppbv	99
16) 1,3-Butadiene	1.609	39	12245	2.077	ppbv	92
17) tert-Butyl alcohol	2.042	59	23722	2.014	ppbv #	93
18) 1,1-Dichloroethene	2.033	96	8303	1.992	ppbv	92
19) Isopropyl Alcohol	1.884	45	14288	1.931	ppbv	99
20) Methylene Chloride	2.062	84	7590	1.889	ppbv	96
21) Allyl Chloride	2.094	41	17050	1.983	ppbv	96
22) trans-1,2-Dichloroethene	2.340	96	9444	1.940	ppbv	97
23) Vinyl Acetate	2.463	43	26502	2.128	ppbv	98
24) 1,1-Dichloroethane	2.408	63	20027	2.016	ppbv	98
25) Ethyl Acetate	2.839	43	43122	2.015	ppbv	97
26) Hexane	2.822	57	19691	2.112	ppbv	95
27) Carbon Disulfide	2.149	76	21945	2.017	ppbv #	43
28) Methyl tert-Butyl Ether	2.441	73	11119	1.985	ppbv	94
29) Chloroform	2.848	83	35127	2.010	ppbv	86
30) Cyclohexane	3.858	84	16056	2.060	ppbv	95
31) cis-1,2-Dichloroethene	2.719	61	22619	2.079	ppbv	95
32) 1,1,1-Trichloroethane	3.369	97	36768	1.940	ppbv	99
34) 2-Butanone	2.557	43	31242	2.139	ppbv	95
35) Carbon Tetrachloride	3.764	117	35326	2.011	ppbv	97
36) Benzene	3.661	78	41848	2.196	ppbv #	88
37) 1,2-Dichloroethane	3.217	62	31049	2.114	ppbv	99
38) Trichloroethene	4.544	130	16311m	2.244	ppbv	
39) 1,2-Dichloropropane	4.311	63	14611	2.047	ppbv	91
40) 1,4-Dioxane	4.593	88	6957m	2.096	ppbv	
41) Tetrahydrofuran	3.059	42	15641m	2.258	ppbv	
42) Bromodichloromethane	4.489	83	37615	2.096	ppbv #	99
43) Methyl Methacrylate	4.884	69	14610m	2.059	ppbv	
44) 2,2,4-Trimethylpentane	4.638	57	65967	2.142	ppbv	99
45) t-1,3-Dichloropropene	6.409	75	18560m	1.987	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042459.D
 Acq On : 01 May 2025 09:37
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:40:35 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	22184	2.000	ppbv	93
47) 1,1,2-Trichloroethane	6.587	97	17307m	2.212	ppbv	
48) Dibromochloromethane	7.386	129	27174	2.057	ppbv	99
49) Bromoform	9.483	173	23126	2.088	ppbv	98
50) 4-Methyl-2-Pentanone	5.758	43	39798	2.155	ppbv	99
51) 2-Hexanone	7.441	43	32196	2.177	ppbv #	97
52) Tetrachloroethene	8.224	164	13670	1.885	ppbv	93
53) Toluene	6.930	91	41392	2.052	ppbv	97
54) 1,2-Dibromoethane	7.652	107	24650	2.049	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.927	131	20822	2.019	ppbv	95
57) Chlorobenzene	8.920	112	36187	2.062	ppbv #	88
58) Ethyl Benzene	9.354	91	61177	2.096	ppbv	91
59) m/p-Xylene	9.551	91	105185m	4.299	ppbv	
60) o-Xylene	9.963	91	53051	2.169	ppbv	99
61) Styrene	9.872	104	18776	2.054	ppbv	97
62) Isopropylbenzene	10.539	105	75626	2.140	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.963	83	34562	2.178	ppbv	98
64) n-propylbenzene	10.989	120	19160	2.226	ppbv	94
65) tert-Butylbenzene	11.532	119	70813	2.246	ppbv	97
66) Benzyl Chloride	11.613	91	6794	1.864	ppbv	97
67) sec-Butylbenzene	11.769	105	101458	2.264	ppbv	99
69) p-Isopropyltoluene	11.924	119	84026	2.226	ppbv	98
70) n-Butylbenzene	12.280	91	89400	2.291	ppbv	97
71) 2-Chlorotoluene	10.901	91	59271	2.139	ppbv	97
72) 4-Ethyltoluene	11.124	105	63833	2.210	ppbv	98
73) 1,3,5-Trimethylbenzene	11.205	105	56516	2.225	ppbv	94
74) 1,2,4-Trimethylbenzene	11.536	105	72485	2.359	ppbv	94
75) 1,3-Dichlorobenzene	11.607	146	36885	2.156	ppbv	97
76) 1,4-Dichlorobenzene	11.671	146	33728	2.083	ppbv	94
77) 1,2-Dichlorobenzene	11.947	146	36078	2.154	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	37360	2.234	ppbv	99
79) Naphthalene	13.513	128	75248	2.920	ppbv	96
80) Naphthalene,2-methyl-	14.458	142	54922	2.857	ppbv	96
81) 1,2,4-Trichlorobenzene	13.442	180	33063	2.396	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042460.D
 Acq On : 01 May 2025 10:08
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:41:24 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	71499	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	200111	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	170872	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.374 95 158600 10.011 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.100%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	11796	0.997	ppbv	96
3) Chlorodifluoromethane	1.476	51	13375	1.069	ppbv	96
4) Chloromethane	1.531	50	4975	1.006	ppbv	95
5) Vinyl Chloride	1.576	62	4469	0.898	ppbv	100
6) Bromomethane	1.667	94	2570	1.092	ppbv	86
7) Chloroethane	1.703	64	1943	1.057	ppbv	94
8) Dichlorotetrafluoroethane	1.554	85	9804	1.005	ppbv	91
9) Propene	1.482	41	4331	1.058	ppbv	88
10) Heptane	4.975	43	11681m	1.101	ppbv	
11) Trichlorofluoromethane	1.877	101	10649m	1.009	ppbv	
12) 1,1,2-Trichlorotrifluo...	2.140	101	8821	1.012	ppbv	97
13) Ethanol	1.719	45	975	1.174	ppbv #	52
14) Bromoethene	1.780	108	3164	0.983	ppbv #	91
15) Acetone	1.835	43	14974	1.147	ppbv	98
16) 1,3-Butadiene	1.609	39	6248	1.085	ppbv	96
17) tert-Butyl alcohol	2.039	59	11546	1.004	ppbv #	91
18) 1,1-Dichloroethene	2.033	96	4200	1.032	ppbv	86
19) Isopropyl Alcohol	1.884	45	7775	1.076	ppbv	99
20) Methylene Chloride	2.059	84	3909	0.996	ppbv	86
21) Allyl Chloride	2.094	41	8443	1.006	ppbv	94
22) trans-1,2-Dichloroethene	2.337	96	5305	1.116	ppbv	88
23) Vinyl Acetate	2.460	43	10876	0.894	ppbv #	94
24) 1,1-Dichloroethane	2.405	63	10477	1.080	ppbv	97
25) Ethyl Acetate	2.835	43	21761	1.041	ppbv	99
26) Hexane	2.822	57	9223	1.013	ppbv	97
27) Carbon Disulfide	2.149	76	10342	0.973	ppbv #	21
28) Methyl tert-Butyl Ether	2.441	73	6101	1.115	ppbv	93
29) Chloroform	2.848	83	17744	1.040	ppbv	97
30) Cyclohexane	3.852	84	7139	0.938	ppbv	88
31) cis-1,2-Dichloroethene	2.719	61	11472	1.080	ppbv	87
32) 1,1,1-Trichloroethane	3.366	97	17881	0.966	ppbv	98
34) 2-Butanone	2.560	43	14641	1.012	ppbv	94
35) Carbon Tetrachloride	3.755	117	18704	1.075	ppbv #	76
36) Benzene	3.654	78	19493	1.032	ppbv #	94
37) 1,2-Dichloroethane	3.214	62	15116	1.039	ppbv	98
38) Trichloroethene	4.541	130	7278m	1.011	ppbv	
39) 1,2-Dichloropropane	4.302	63	8103m	1.145	ppbv	
40) 1,4-Dioxane	4.593	88	3703m	1.126	ppbv	
41) Tetrahydrofuran	3.062	42	7091m	1.033	ppbv	
42) Bromodichloromethane	4.486	83	18427	1.036	ppbv #	81
43) Methyl Methacrylate	4.874	69	6829m	0.971	ppbv	
44) 2,2,4-Trimethylpentane	4.635	57	31820m	1.043	ppbv	
45) t-1,3-Dichloropropene	6.409	75	8461m	0.914	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042460.D
 Acq On : 01 May 2025 10:08
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:41:24 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.596	75	10781m	0.981	ppbv	
47) 1,1,2-Trichloroethane	6.583	97	8183m	1.055	ppbv	
48) Dibromochloromethane	7.386	129	12859m	0.982	ppbv	
49) Bromoform	9.483	173	10709	0.976	ppbv	97
50) 4-Methyl-2-Pentanone	5.768	43	19618m	1.072	ppbv	
51) 2-Hexanone	7.448	43	15091m	1.030	ppbv	
52) Tetrachloroethene	8.221	164	6780	0.943	ppbv	86
53) Toluene	6.930	91	19390	0.970	ppbv	97
54) 1,2-Dibromoethane	7.655	107	11980	1.005	ppbv	91
56) 1,1,1,2-Tetrachloroethane	8.920	131	10811	1.071	ppbv	94
57) Chlorobenzene	8.917	112	18111	1.055	ppbv	90
58) Ethyl Benzene	9.354	91	27914	0.978	ppbv	97
59) m/p-Xylene	9.548	91	47105m	1.968	ppbv	
60) o-Xylene	9.962	91	23806	0.995	ppbv	98
61) Styrene	9.872	104	8537	0.954	ppbv	98
62) Isopropylbenzene	10.539	105	34679	1.003	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.962	83	16837	1.084	ppbv	98
64) n-propylbenzene	10.985	120	8958	1.064	ppbv	86
65) tert-Butylbenzene	11.529	119	30604	0.992	ppbv	95
66) Benzyl Chloride	11.613	91	3486	0.978	ppbv #	96
67) sec-Butylbenzene	11.765	105	43740	0.998	ppbv	98
69) p-Isopropyltoluene	11.927	119	37011	1.002	ppbv	99
70) n-Butylbenzene	12.280	91	37892	0.992	ppbv	97
71) 2-Chlorotoluene	10.898	91	26813	0.989	ppbv	99
72) 4-Ethyltoluene	11.124	105	28149	0.996	ppbv	96
73) 1,3,5-Trimethylbenzene	11.202	105	25073	1.009	ppbv	97
74) 1,2,4-Trimethylbenzene	11.535	105	30550	1.016	ppbv	92
75) 1,3-Dichlorobenzene	11.607	146	17456	1.043	ppbv	98
76) 1,4-Dichlorobenzene	11.671	146	16151	1.019	ppbv	99
77) 1,2-Dichlorobenzene	11.947	146	17207	1.050	ppbv	94
78) Hexachloro-1,3-Butadiene	13.879	225	18472	1.129	ppbv	98
79) Naphthalene	13.513	128	28316	1.123	ppbv #	94
80) Naphthalene,2-methyl-	14.458	142	15804	0.840	ppbv	93
81) 1,2,4-Trichlorobenzene	13.439	180	13644	1.011	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042461.D
 Acq On : 01 May 2025 10:40
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:42:12 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	68913	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	193701	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	166625	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 150277 9.728 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 97.300%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	6501	0.570	ppbv	94
3) Chlorodifluoromethane	1.476	51	6500	0.539	ppbv	99
4) Chloromethane	1.535	50	2542	0.533	ppbv	95
5) Vinyl Chloride	1.580	62	2441	0.509	ppbv	92
6) Bromomethane	1.667	94	1202	0.530	ppbv	97
7) Chloroethane	1.703	64	862	0.486	ppbv #	68
8) Dichlorotetrafluoroethane	1.554	85	4979	0.530	ppbv	93
9) Propene	1.486	41	2017	0.511	ppbv	96
10) Heptane	4.982	43	4397m	0.430	ppbv	
11) Trichlorofluoromethane	1.878	101	5720	0.562	ppbv	95
12) 1,1,2-Trichlorotrifluo...	2.137	101	4540	0.540	ppbv	94
13) Ethanol	1.722	45	474m	0.592	ppbv	
14) Bromoethene	1.784	108	1680	0.542	ppbv #	87
15) Acetone	1.839	43	7680	0.610	ppbv #	82
16) 1,3-Butadiene	1.609	39	3044	0.549	ppbv	86
17) tert-Butyl alcohol	2.046	59	6256	0.564	ppbv #	91
18) 1,1-Dichloroethene	2.036	96	2240	0.571	ppbv	94
19) Isopropyl Alcohol	1.887	45	4208	0.604	ppbv #	96
20) Methylene Chloride	2.062	84	2594	0.686	ppbv #	80
21) Allyl Chloride	2.098	41	4420	0.546	ppbv #	93
22) trans-1,2-Dichloroethene	2.340	96	2554	0.557	ppbv	95
23) Vinyl Acetate	2.460	43	6541	0.558	ppbv #	95
24) 1,1-Dichloroethane	2.412	63	4933	0.528	ppbv	98
25) Ethyl Acetate	2.842	43	10344	0.514	ppbv #	96
26) Hexane	2.829	57	4403	0.502	ppbv	95
27) Carbon Disulfide	2.153	76	5201	0.508	ppbv #	1
28) Methyl tert-Butyl Ether	2.441	73	2762	0.524	ppbv #	82
29) Chloroform	2.852	83	8791	0.535	ppbv	92
30) Cyclohexane	3.855	84	3910m	0.533	ppbv	
31) cis-1,2-Dichloroethene	2.726	61	5345m	0.522	ppbv	
32) 1,1,1-Trichloroethane	3.366	97	8846m	0.496	ppbv	
34) 2-Butanone	2.564	43	7766	0.554	ppbv	93
35) Carbon Tetrachloride	3.768	117	8286	0.492	ppbv	86
36) Benzene	3.655	78	8951	0.490	ppbv #	92
37) 1,2-Dichloroethane	3.224	62	7149	0.507	ppbv #	95
38) Trichloroethene	4.548	130	3769m	0.541	ppbv	
39) 1,2-Dichloropropane	4.321	63	3561m	0.520	ppbv	
40) 1,4-Dioxane	4.619	88	1701m	0.534	ppbv	
41) Tetrahydrofuran	3.069	42	3187m	0.480	ppbv	
42) Bromodichloromethane	4.496	83	8212	0.477	ppbv	93
43) Methyl Methacrylate	4.891	69	3179m	0.467	ppbv	
44) 2,2,4-Trimethylpentane	4.648	57	14435m	0.489	ppbv	
45) t-1,3-Dichloropropene	6.415	75	4238m	0.473	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042461.D
 Acq On : 01 May 2025 10:40
 Operator : SY/MD
 Sample : VSTDIC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:42:12 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.600	75	5091m	0.479	ppbv	
47) 1,1,2-Trichloroethane	6.584	97	3606m	0.480	ppbv	
48) Dibromochloromethane	7.390	129	6304m	0.498	ppbv	
49) Bromoform	9.484	173	5228	0.492	ppbv	97
50) 4-Methyl-2-Pentanone	5.771	43	8248m	0.466	ppbv	
51) 2-Hexanone	7.454	43	6283m	0.443	ppbv	
52) Tetrachloroethene	8.221	164	3151m	0.453	ppbv	
53) Toluene	6.933	91	9052m	0.468	ppbv	
54) 1,2-Dibromoethane	7.652	107	5705	0.494	ppbv	83
56) 1,1,1,2-Tetrachloroethane	8.924	131	5152	0.524	ppbv #	92
57) Chlorobenzene	8.924	112	9173	0.548	ppbv #	86
58) Ethyl Benzene	9.357	91	12427	0.446	ppbv	90
59) m/p-Xylene	9.552	91	20195m	0.865	ppbv	
60) o-Xylene	9.966	91	9976m	0.427	ppbv	
61) Styrene	9.872	104	3457m	0.396	ppbv	
62) Isopropylbenzene	10.545	105	14837	0.440	ppbv	95
63) 1,1,2,2-Tetrachloroethane	9.963	83	8135	0.537	ppbv	96
64) n-propylbenzene	10.992	120	3329m	0.405	ppbv	
65) tert-Butylbenzene	11.536	119	12655	0.421	ppbv	99
66) Benzyl Chloride	11.613	91	1317m	0.379	ppbv	
67) sec-Butylbenzene	11.769	105	17905	0.419	ppbv	97
69) p-Isopropyltoluene	11.927	119	15036	0.417	ppbv	94
70) n-Butylbenzene	12.283	91	15016	0.403	ppbv	93
71) 2-Chlorotoluene	10.905	91	11702	0.443	ppbv	99
72) 4-Ethyltoluene	11.128	105	10671	0.387	ppbv #	91
73) 1,3,5-Trimethylbenzene	11.206	105	9808	0.405	ppbv	89
74) 1,2,4-Trimethylbenzene	11.539	105	13060	0.445	ppbv #	83
75) 1,3-Dichlorobenzene	11.607	146	8127	0.498	ppbv	96
76) 1,4-Dichlorobenzene	11.675	146	7238	0.469	ppbv	96
77) 1,2-Dichlorobenzene	11.950	146	8141	0.509	ppbv	94
78) Hexachloro-1,3-Butadiene	13.886	225	8565	0.537	ppbv	99
79) Naphthalene	13.517	128	9735m	0.396	ppbv	
80) Naphthalene,2-methyl-	14.458	142	4153	0.226	ppbv	98
81) 1,2,4-Trichlorobenzene	13.445	180	4992	0.379	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
Data File : VL042461.D
Acq On : 01 May 2025 10:40
Operator : SY/MD
Sample : VSTDICC0.5
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
Supervised By :Mahesh Dadoda 05/05/2025


Quant Time: May 02 01:42:12 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042462.D
 Acq On : 01 May 2025 11:12
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:43:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri May 02 01:37:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	65449	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	186907	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	157618	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	143644	9.830	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.300%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	631	0.138	ppbv #	53
32) 1,1,1-Trichloroethane	3.363	97	1860m	0.110	ppbv	
35) Carbon Tetrachloride	3.758	117	1493	0.092	ppbv #	55
38) Trichloroethene	4.551	130	784m	0.117	ppbv	
52) Tetrachloroethene	8.221	164	710m	0.106	ppbv	
54) 1,2-Dibromoethane	7.665	107	1169m	0.105	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.966	83	1490	0.104	ppbv #	75
79) Naphthalene	13.517	128	1135m	0.049	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
Data File : VL042462.D
Acq On : 01 May 2025 11:12
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:43:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 02 01:37:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042465.D
 Acq On : 01 May 2025 12:49
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/02/2025
 Supervised By : Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:45:25 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	64314	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	178877	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	157673	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 151837 10.387 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 103.900%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	158069	14.852	ppbv	99
3) Chlorodifluoromethane	1.476	51	154556	13.735	ppbv	100
4) Chloromethane	1.535	50	64267	14.447	ppbv	95
5) Vinyl Chloride	1.580	62	60873	13.592	ppbv	98
6) Bromomethane	1.667	94	31246	14.766	ppbv	87
7) Chloroethane	1.703	64	25181	15.227	ppbv	92
8) Dichlorotetrafluoroethane	1.554	85	131995	15.044	ppbv	99
9) Propene	1.486	41	49647	13.489	ppbv	99
10) Heptane	4.982	43	142235	14.898	ppbv	100
11) Trichlorofluoromethane	1.878	101	152655	16.082	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.140	101	115855	14.775	ppbv	96
13) Ethanol	1.722	45	8090	10.832	ppbv	98
14) Bromoethene	1.781	108	44928	15.519	ppbv	96
15) Acetone	1.836	43	141306	12.036	ppbv	99
16) 1,3-Butadiene	1.609	39	72455	13.992	ppbv	98
17) tert-Butyl alcohol	2.043	59	144935	14.012	ppbv	99
18) 1,1-Dichloroethene	2.033	96	52813	14.424	ppbv	98
19) Isopropyl Alcohol	1.887	45	86000	13.235	ppbv #	1
20) Methylene Chloride	2.062	84	46063	13.050	ppbv	94
21) Allyl Chloride	2.098	41	112451	14.893	ppbv	99
22) trans-1,2-Dichloroethene	2.341	96	62153	14.534	ppbv	100
23) Vinyl Acetate	2.464	43	169418	15.489	ppbv	99
24) 1,1-Dichloroethane	2.409	63	127567	14.622	ppbv	97
25) Ethyl Acetate	2.836	43	269372	14.329	ppbv	100
26) Hexane	2.826	57	119552	14.598	ppbv	97
27) Carbon Disulfide	2.150	76	150258	15.722	ppbv #	92
28) Methyl tert-Butyl Ether	2.441	73	71341	14.500	ppbv	99
29) Chloroform	2.849	83	221976	14.464	ppbv	99
30) Cyclohexane	3.859	84	101146	14.771	ppbv	99
31) cis-1,2-Dichloroethene	2.723	61	133735	13.997	ppbv	93
32) 1,1,1-Trichloroethane	3.370	97	234752	14.099	ppbv	99
34) 2-Butanone	2.557	43	174521	13.488	ppbv	97
35) Carbon Tetrachloride	3.765	117	242144	15.566	ppbv	99
36) Benzene	3.658	78	239352	14.179	ppbv	98
37) 1,2-Dichloroethane	3.221	62	190179	14.618	ppbv	98
38) Trichloroethene	4.551	130	95257	14.798	ppbv	94
39) 1,2-Dichloropropane	4.312	63	84414	13.349	ppbv	93
40) 1,4-Dioxane	4.584	88	38551	13.115	ppbv #	100
41) Tetrahydrofuran	3.056	42	86296	14.063	ppbv	98
42) Bromodichloromethane	4.493	83	242556	15.259	ppbv	99
43) Methyl Methacrylate	4.885	69	98279	15.639	ppbv	100
44) 2,2,4-Trimethylpentane	4.642	57	401920	14.734	ppbv	99
45) t-1,3-Dichloropropene	6.419	75	136762	16.529	ppbv	92

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042465.D
 Acq On : 01 May 2025 12:49
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:45:25 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.600	75	154705m	15.749	ppbv	
47) 1,1,2-Trichloroethane	6.587	97	99496	14.356	ppbv	93
48) Dibromochloromethane	7.393	129	180970	15.468	ppbv	98
49) Bromoform	9.487	173	152440	15.535	ppbv	99
50) 4-Methyl-2-Pentanone	5.755	43	237985	14.550	ppbv	97
51) 2-Hexanone	7.438	43	197423	15.071	ppbv	99
52) Tetrachloroethene	8.228	164	85271	13.272	ppbv	90
53) Toluene	6.937	91	281731	15.768	ppbv	100
54) 1,2-Dibromoethane	7.655	107	154739	14.523	ppbv	94
56) 1,1,1,2-Tetrachloroethane	8.927	131	133410	14.326	ppbv	98
57) Chlorobenzene	8.927	112	218370	13.783	ppbv	99
58) Ethyl Benzene	9.358	91	414805	15.743	ppbv	98
59) m/p-Xylene	9.555	91	699743m	31.674	ppbv	
60) o-Xylene	9.966	91	350328	15.863	ppbv	100
61) Styrene	9.876	104	137720	16.686	ppbv	96
62) Isopropylbenzene	10.542	105	498619	15.629	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.966	83	209726	14.635	ppbv	99
64) n-propylbenzene	10.992	120	122452	15.759	ppbv	97
65) tert-Butylbenzene	11.536	119	443042	15.562	ppbv	97
66) Benzyl Chloride	11.617	91	60376	18.348	ppbv	97
67) sec-Butylbenzene	11.769	105	632600	15.636	ppbv	100
69) p-Isopropyltoluene	11.928	119	539860	15.837	ppbv	99
70) n-Butylbenzene	12.284	91	566780	16.084	ppbv	99
71) 2-Chlorotoluene	10.905	91	394227	15.755	ppbv	100
72) 4-Ethyltoluene	11.128	105	423746	16.250	ppbv	99
73) 1,3,5-Trimethylbenzene	11.206	105	362951	15.826	ppbv	96
74) 1,2,4-Trimethylbenzene	11.539	105	407416	14.683	ppbv	96
75) 1,3-Dichlorobenzene	11.610	146	225215	14.579	ppbv	98
76) 1,4-Dichlorobenzene	11.675	146	224505	15.357	ppbv	99
77) 1,2-Dichlorobenzene	11.950	146	216147	14.295	ppbv	99
78) Hexachloro-1,3-Butadiene	13.886	225	198112	13.121	ppbv	99
79) Naphthalene	13.517	128	392900	16.888	ppbv	98
80) Naphthalene,2-methyl-	14.459	142	316493	18.232	ppbv	99
81) 1,2,4-Trichlorobenzene	13.442	180	195942	15.728	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042465.D
 Acq On : 01 May 2025 12:49
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:45:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
 QLast Update : Fri May 02 01:37:48 2025
 Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042466.D
 Acq On : 01 May 2025 14:10
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:46:15 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.791	49	66546	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	186737	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	161050	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	148088	9.918	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.200%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	151	0.033	ppbv #	87
32) 1,1,1-Trichloroethane	3.370	97	595m	0.035	ppbv	
35) Carbon Tetrachloride	3.771	117	653m	0.040	ppbv	
38) Trichloroethene	4.551	130	135	0.020	ppbv #	40
52) Tetrachloroethene	8.231	164	285m	0.042	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.973	83	551m	0.038	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
Data File : VL042466.D
Acq On : 01 May 2025 14:10
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:46:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 02 01:37:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL050125

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 03:48:56 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	63539	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	183679	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	156982	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.380 95 150522 10.342 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 103.400%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	108782	10.346	ppbv	97
3) Chlorodifluoromethane	1.476	51	110290	9.921	ppbv	98
4) Chloromethane	1.534	50	43950	10.001	ppbv	96
5) Vinyl Chloride	1.580	62	41181	9.307	ppbv	99
6) Bromomethane	1.667	94	21021	10.055	ppbv	91
7) Chloroethane	1.706	64	16630	10.179	ppbv	90
8) Dichlorotetrafluoroethane	1.554	85	89345	10.307	ppbv	99
9) Propene	1.486	41	36920	10.153	ppbv	97
10) Heptane	4.985	43	94631	10.033	ppbv	95
11) Trichlorofluoromethane	1.878	101	100713	10.380	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.140	101	81143	10.474	ppbv	99
13) Ethanol	1.722	45	5736	7.773	ppbv	99
14) Bromoethene	1.780	108	30094	10.522	ppbv	97
15) Acetone	1.835	43	95460	8.230	ppbv	98
16) 1,3-Butadiene	1.609	39	49566	9.689	ppbv	100
17) tert-Butyl alcohol	2.043	59	102047	9.986	ppbv	100
18) 1,1-Dichloroethene	2.033	96	35784	9.892	ppbv	98
19) Isopropyl Alcohol	1.887	45	60977	9.499	ppbv	99
20) Methylene Chloride	2.062	84	31212	8.951	ppbv	94
21) Allyl Chloride	2.098	41	73365	9.835	ppbv	94
22) trans-1,2-Dichloroethene	2.340	96	42922	10.160	ppbv	96
23) Vinyl Acetate	2.463	43	113131	10.220	ppbv #	98
24) 1,1-Dichloroethane	2.408	63	89117	10.340	ppbv	97
25) Ethyl Acetate	2.839	43	186185	10.025	ppbv	99
26) Hexane	2.826	57	83132	10.275	ppbv	96
27) Carbon Disulfide	2.149	76	100196	10.612	ppbv	96
28) Methyl tert-Butyl Ether	2.441	73	48326	9.942	ppbv	100
29) Chloroform	2.852	83	152381	10.050	ppbv	96
30) Cyclohexane	3.858	84	70581	10.434	ppbv	98
31) cis-1,2-Dichloroethene	2.722	61	95793	10.148	ppbv	96
32) 1,1,1-Trichloroethane	3.370	97	161465	9.816	ppbv	99
34) 2-Butanone	2.560	43	125883	9.475	ppbv	97
35) Carbon Tetrachloride	3.764	117	166814	9.998	ppbv	97
36) Benzene	3.658	78	171384	9.888	ppbv	98
37) 1,2-Dichloroethane	3.221	62	130063	9.736	ppbv	100
38) Trichloroethene	4.551	130	65940	9.976	ppbv	96
39) 1,2-Dichloropropane	4.315	63	60010	9.248	ppbv	95
40) 1,4-Dioxane	4.583	88	28306m	9.316	ppbv	
41) Tetrahydrofuran	3.059	42	59604	9.440	ppbv	98
42) Bromodichloromethane	4.489	83	163210	9.999	ppbv	95
43) Methyl Methacrylate	4.881	69	67516	10.438	ppbv	99
44) 2,2,4-Trimethylpentane	4.642	57	277217	9.867	ppbv	100
45) t-1,3-Dichloropropene	6.415	75	89638m	10.563	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL050125

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 03:48:56 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.609	75	104665	10.378	ppbv	97
47) 1,1,2-Trichloroethane	6.584	97	67821	9.530	ppbv	95
48) Dibromochloromethane	7.393	129	121381	10.103	ppbv	97
49) Bromoform	9.487	173	102150	10.138	ppbv	100
50) 4-Methyl-2-Pentanone	5.752	43	159280	9.478	ppbv	99
51) 2-Hexanone	7.438	43	126396	9.407	ppbv #	98
52) Tetrachloroethene	8.228	164	58272	8.832	ppbv	96
53) Toluene	6.940	91	191922	10.461	ppbv	99
54) 1,2-Dibromoethane	7.658	107	105149	9.664	ppbv	95
56) 1,1,1,2-Tetrachloroethane	8.927	131	92337	9.959	ppbv	97
57) Chlorobenzene	8.927	112	148759	9.431	ppbv	92
58) Ethyl Benzene	9.357	91	280660	10.698	ppbv	99
59) m/p-Xylene	9.555	91	468984m	21.324	ppbv	
60) o-Xylene	9.966	91	238347	10.824	ppbv	97
61) Styrene	9.872	104	91128	11.090	ppbv	97
62) Isopropylbenzene	10.542	105	337992	10.641	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.966	83	140306	9.242	ppbv	100
64) n-propylbenzene	10.992	120	79983	10.303	ppbv	99
65) tert-Butylbenzene	11.536	119	302506	10.672	ppbv	99
66) Benzyl Chloride	11.617	91	34933	10.663	ppbv	97
67) sec-Butylbenzene	11.769	105	425444	10.562	ppbv	99
69) p-Isopropyltoluene	11.927	119	363563	10.712	ppbv	98
70) n-Butylbenzene	12.283	91	371369	10.585	ppbv	99
71) 2-Chlorotoluene	10.905	91	264063	10.599	ppbv	99
72) 4-Ethyltoluene	11.128	105	287375	11.069	ppbv	99
73) 1,3,5-Trimethylbenzene	11.206	105	247128	10.823	ppbv	93
74) 1,2,4-Trimethylbenzene	11.539	105	278306	10.074	ppbv	98
75) 1,3-Dichlorobenzene	11.610	146	150026	9.754	ppbv	99
76) 1,4-Dichlorobenzene	11.675	146	151740	10.425	ppbv	97
77) 1,2-Dichlorobenzene	11.950	146	146314	9.719	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	133688	8.893	ppbv	99
79) Naphthalene	13.516	128	251318	10.806	ppbv	98
80) Naphthalene,2-methyl-	14.458	142	198604	11.489	ppbv	99
81) 1,2,4-Trichlorobenzene	13.442	180	129406	10.433	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL050125

Quant Time: May 02 03:48:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	1.655	1.712	-3.4	104	0.00
3	Chlorodifluoromethane	1.750	1.736	0.8	98	0.00
4	Chloromethane	0.692	0.692	0.0	99	0.00
5 T	Vinyl Chloride	0.696	0.648	6.9	102	0.00
6 T	Bromomethane	0.329	0.331	-0.6	103	0.00
7	Chloroethane	0.257	0.262	-1.9	101	0.00
8 T	Dichlorotetrafluoroethane	1.364	1.406	-3.1	101	0.00
9 T	Propene	0.572	0.581	-1.6	98	0.00
10 T	Heptane	1.484	1.489	-0.3	92	0.00
11 T	Trichlorofluoromethane	1.527	1.585	-3.8	103	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.219	1.277	-4.8	103	0.00
13	Ethanol	0.116	0.090	22.4	101	0.00
14 T	Bromoethene	0.450	0.474	-5.3	106	0.00
15 T	Acetone	1.826	1.502	17.7	101	0.00
16 T	1,3-Butadiene	0.805	0.780	3.1	102	0.00
17	tert-Butyl alcohol	1.608	1.606	0.1	96	0.00
18 T	1,1-Dichloroethene	0.569	0.563	1.1	101	0.00
19 T	Isopropyl Alcohol	1.010	0.960	5.0	98	0.00
20 T	Methylene Chloride	0.549	0.491	10.6	97	0.00
21 T	Allyl Chloride	1.174	1.155	1.6	95	0.00
22 T	trans-1,2-Dichloroethene	0.665	0.676	-1.7	109	0.00
23 T	Vinyl Acetate	1.742	1.780	-2.2	92	0.00
24 T	1,1-Dichloroethane	1.357	1.403	-3.4	104	0.00
25 T	Ethyl Acetate	2.923	2.930	-0.2	92	0.00
26 T	Hexane	1.273	1.308	-2.7	96	0.00
27 T	Carbon Disulfide	1.486	1.577	-6.1	99	0.00
28 T	Methyl tert-Butyl Ether	0.765	0.761	0.5	101	0.00
29 T	Chloroform	2.386	2.398	-0.5	97	0.00
30 T	Cyclohexane	1.065	1.111	-4.3	95	0.00
31 T	cis-1,2-Dichloroethene	1.486	1.508	-1.5	100	0.00
32 T	1,1,1-Trichloroethane	2.589	2.541	1.9	99	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
34 T	2-Butanone	0.723	0.685	5.3	95	0.00
35 T	Carbon Tetrachloride	0.908	0.908	0.0	100	0.00
36 T	Benzene	0.944	0.933	1.2	95	0.00
37 T	1,2-Dichloroethane	0.727	0.708	2.6	97	0.00
38 T	Trichloroethene	0.360	0.359	0.3	94	0.00
39 T	1,2-Dichloropropane	0.353	0.327	7.4	94	0.00
40 T	1,4-Dioxane	0.165	0.154	6.7	93	0.00
41 T	Tetrahydrofuran	0.344	0.325	5.5	91	0.00
42 T	Bromodichloromethane	0.889	0.889	0.0	97	0.00
43	Methyl Methacrylate	0.352	0.368	-4.5	92	0.00
44 T	2,2,4-Trimethylpentane	1.530	1.509	1.4	96	0.00
45 T	t-1,3-Dichloropropene	0.462	0.488	-5.6	93	0.00
46 T	cis-1,3-Dichloropropene	0.549	0.570	-3.8	94	0.00
47 T	1,1,2-Trichloroethane	0.387	0.369	4.7	94	0.00
48 T	Dibromochloromethane	0.654	0.661	-1.1	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL050125

Quant Time: May 02 03:48:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.549	0.556	-1.3	96	0.00
50 T	4-Methyl-2-Pentanone	0.915	0.867	5.2	91	0.00
51 T	2-Hexanone	0.731	0.688	5.9	87	0.00
52 T	Tetrachloroethene	0.359	0.317	11.7	95	0.00
53 T	Toluene	0.999	1.045	-4.6	94	0.00
54 T	1,2-Dibromoethane	0.592	0.572	3.4	94	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
56	1,1,1,2-Tetrachloroethane	0.591	0.588	0.5	98	0.00
57 T	Chlorobenzene	1.005	0.948	5.7	95	0.00
58 T	Ethyl Benzene	1.671	1.788	-7.0	93	0.00
59 T	m/p-Xylene	1.401	1.494	-6.6	94	0.03
60 T	o-Xylene	1.403	1.518	-8.2	96	0.00
61 T	Styrene	0.523	0.580	-10.9	90	0.00
62	Isopropylbenzene	2.023	2.153	-6.4	95	0.00
63 T	1,1,2,2-Tetrachloroethane	0.967	0.894	7.5	95	0.00
64	n-propylbenzene	0.495	0.510	-3.0	95	0.00
65	tert-Butylbenzene	1.806	1.927	-6.7	96	0.00
66 T	Benzyl Chloride	0.209	0.223	-6.7	87	0.00
67	sec-Butylbenzene	2.566	2.710	-5.6	96	0.00
68 S	1-Bromo-4-Fluorobenzene	0.927	0.959	-3.5	91	0.00
69	p-Isopropyltoluene	2.162	2.316	-7.1	97	0.00
70	n-Butylbenzene	2.235	2.366	-5.9	97	0.00
71	2-Chlorotoluene	1.587	1.682	-6.0	95	0.00
72 T	4-Ethyltoluene	1.654	1.831	-10.7	96	0.00
73 T	1,3,5-Trimethylbenzene	1.455	1.574	-8.2	96	0.00
74 T	1,2,4-Trimethylbenzene	1.760	1.773	-0.7	97	0.00
75 T	1,3-Dichlorobenzene	0.980	0.956	2.4	96	0.00
76 T	1,4-Dichlorobenzene	0.927	0.967	-4.3	96	0.00
77 T	1,2-Dichlorobenzene	0.959	0.932	2.8	97	0.00
78 T	Hexachloro-1,3-Butadiene	0.958	0.852	11.1	99	0.00
79 T	Naphthalene	1.482	1.601	-8.0	94	0.00
80 T	Naphthalene,2-methyl-	1.101	1.265	-14.9	97	0.00
81 T	1,2,4-Trichlorobenzene	0.790	0.824	-4.3	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICSVVL050125

Quant Time: May 02 03:48:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	10.000	10.346	-3.5	104	0.00
3	Chlorodifluoromethane	10.000	9.921	0.8	98	0.00
4	Chloromethane	10.000	10.001	-0.0	99	0.00
5 T	Vinyl Chloride	10.000	9.307	6.9	102	0.00
6 T	Bromomethane	10.000	10.055	-0.5	103	0.00
7	Chloroethane	10.000	10.179	-1.8	101	0.00
8 T	Dichlorotetrafluoroethane	10.000	10.307	-3.1	101	0.00
9 T	Propene	10.000	10.153	-1.5	98	0.00
10 T	Heptane	10.000	10.033	-0.3	92	0.00
11 T	Trichlorofluoromethane	10.000	10.380	-3.8	103	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	10.474	-4.7	103	0.00
13	Ethanol	10.000	7.773	22.3	101	0.00
14 T	Bromoethene	10.000	10.522	-5.2	106	0.00
15 T	Acetone	10.000	8.230	17.7	101	0.00
16 T	1,3-Butadiene	10.000	9.689	3.1	102	0.00
17	tert-Butyl alcohol	10.000	9.986	0.1	96	0.00
18 T	1,1-Dichloroethene	10.000	9.892	1.1	101	0.00
19 T	Isopropyl Alcohol	10.000	9.499	5.0	98	0.00
20 T	Methylene Chloride	10.000	8.951	10.5	97	0.00
21 T	Allyl Chloride	10.000	9.835	1.6	95	0.00
22 T	trans-1,2-Dichloroethene	10.000	10.160	-1.6	109	0.00
23 T	Vinyl Acetate	10.000	10.220	-2.2	92	0.00
24 T	1,1-Dichloroethane	10.000	10.340	-3.4	104	0.00
25 T	Ethyl Acetate	10.000	10.025	-0.3	92	0.00
26 T	Hexane	10.000	10.275	-2.8	96	0.00
27 T	Carbon Disulfide	10.000	10.612	-6.1	99	0.00
28 T	Methyl tert-Butyl Ether	10.000	9.942	0.6	101	0.00
29 T	Chloroform	10.000	10.050	-0.5	97	0.00
30 T	Cyclohexane	10.000	10.434	-4.3	95	0.00
31 T	cis-1,2-Dichloroethene	10.000	10.148	-1.5	100	0.00
32 T	1,1,1-Trichloroethane	10.000	9.816	1.8	99	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	91	0.00
34 T	2-Butanone	10.000	9.475	5.3	95	0.00
35 T	Carbon Tetrachloride	10.000	9.998	0.0	100	0.00
36 T	Benzene	10.000	9.888	1.1	95	0.00
37 T	1,2-Dichloroethane	10.000	9.736	2.6	97	0.00
38 T	Trichloroethene	10.000	9.976	0.2	94	0.00
39 T	1,2-Dichloropropane	10.000	9.248	7.5	94	0.00
40 T	1,4-Dioxane	10.000	9.316	6.8	93	0.00
41 T	Tetrahydrofuran	10.000	9.440	5.6	91	0.00
42 T	Bromodichloromethane	10.000	9.999	0.0	97	0.00
43	Methyl Methacrylate	10.000	10.438	-4.4	92	0.00
44 T	2,2,4-Trimethylpentane	10.000	9.867	1.3	96	0.00
45 T	t-1,3-Dichloropropene	10.000	10.563	-5.6	93	0.00
46 T	cis-1,3-Dichloropropene	10.000	10.378	-3.8	94	0.00
47 T	1,1,2-Trichloroethane	10.000	9.530	4.7	94	0.00
48 T	Dibromochloromethane	10.000	10.103	-1.0	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL050125

Quant Time: May 02 03:48:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	Bromoform	10.000	10.138	-1.4	96	0.00
50 T	4-Methyl-2-Pentanone	10.000	9.478	5.2	91	0.00
51 T	2-Hexanone	10.000	9.407	5.9	87	0.00
52 T	Tetrachloroethene	10.000	8.832	11.7	95	0.00
53 T	Toluene	10.000	10.461	-4.6	94	0.00
54 T	1,2-Dibromoethane	10.000	9.664	3.4	94	0.00

55 I	Chlorobenzene-d5	10.000	10.000	0.0	90	0.00
56	1,1,1,2-Tetrachloroethane	10.000	9.959	0.4	98	0.00
57 T	Chlorobenzene	10.000	9.431	5.7	95	0.00
58 T	Ethyl Benzene	10.000	10.698	-7.0	93	0.00
59 T	m/p-Xylene	20.000	21.324	-6.6	94	0.03
60 T	o-Xylene	10.000	10.824	-8.2	96	0.00
61 T	Styrene	10.000	11.090	-10.9	90	0.00
62	Isopropylbenzene	10.000	10.641	-6.4	95	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	9.242	7.6	95	0.00
64	n-propylbenzene	10.000	10.303	-3.0	95	0.00
65	tert-Butylbenzene	10.000	10.672	-6.7	96	0.00
66 T	Benzyl Chloride	10.000	10.663	-6.6	87	0.00
67	sec-Butylbenzene	10.000	10.562	-5.6	96	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	10.342	-3.4	91	0.00
69	p-Isopropyltoluene	10.000	10.712	-7.1	97	0.00
70	n-Butylbenzene	10.000	10.585	-5.9	97	0.00
71	2-Chlorotoluene	10.000	10.599	-6.0	95	0.00
72 T	4-Ethyltoluene	10.000	11.069	-10.7	96	0.00
73 T	1,3,5-Trimethylbenzene	10.000	10.823	-8.2	96	0.00
74 T	1,2,4-Trimethylbenzene	10.000	10.074	-0.7	97	0.00
75 T	1,3-Dichlorobenzene	10.000	9.754	2.5	96	0.00
76 T	1,4-Dichlorobenzene	10.000	10.425	-4.3	96	0.00
77 T	1,2-Dichlorobenzene	10.000	9.719	2.8	97	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	8.893	11.1	99	0.00
79 T	Naphthalene	10.000	10.806	-8.1	94	0.00
80 T	Naphthalene,2-methyl-	10.000	11.489	-14.9	97	0.00
81 T	1,2,4-Trichlorobenzene	10.000	10.433	-4.3	96	0.00

(#) = Out of Range



SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
Data File : VL042486.D
Acq On : 06 May 2025 08:16
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri May 02 03:16:31 2025



AutoFind: Scans 3180, 3181, 3182; Background Corrected with Scan 3165

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	24.9	28768	PASS
75	95	30	66	63.6	73411	PASS
95	95	100	100	100.0	115408	PASS
96	95	5	9	6.2	7151	PASS
173	174	0.00	2	0.6	392	PASS
174	95	50	120	56.1	64757	PASS
175	174	4	9	7.6	4952	PASS
176	174	93	101	94.3	61040	PASS
177	176	5	9	6.1	3717	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042487.D
 Acq On : 06 May 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/07/2025
 Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 07 01:35:37 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.791	49	69320	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	188157	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	161366	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.381 95 165339 11.051 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 110.500%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	131525	11.466	ppbv	96
3) Chlorodifluoromethane	1.477	51	119859	9.883	ppbv	98
4) Chloromethane	1.535	50	51848	10.814	ppbv	91
5) Vinyl Chloride	1.580	62	47536	9.847	ppbv	95
6) Bromomethane	1.667	94	25689	11.263	ppbv	99
7) Chloroethane	1.706	64	18985	10.651	ppbv	97
8) Dichlorotetrafluoroethane	1.554	85	105839	11.192	ppbv	97
9) Propene	1.486	41	34471	8.689	ppbv	96
10) Heptane	4.982	43	103210	10.030	ppbv	96
11) Trichlorofluoromethane	1.878	101	123855	11.700	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.140	101	93934	11.114	ppbv	98
13) Ethanol	1.722	45	6351	7.889	ppbv	98
14) Bromoethene	1.784	108	36335	11.645	ppbv	92
15) Acetone	1.836	43	111680	8.825	ppbv	99
16) 1,3-Butadiene	1.609	39	55608	9.963	ppbv	95
17) tert-Butyl alcohol	2.043	59	118491	10.628	ppbv	97
18) 1,1-Dichloroethene	2.036	96	41887	10.614	ppbv	98
19) Isopropyl Alcohol	1.888	45	69030	9.856	ppbv	100
20) Methylene Chloride	2.062	84	37393	9.829	ppbv	97
21) Allyl Chloride	2.098	41	84167	10.342	ppbv	97
22) trans-1,2-Dichloroethene	2.341	96	47256	10.253	ppbv	95
23) Vinyl Acetate	2.464	43	118463	9.809	ppbv	97
24) 1,1-Dichloroethane	2.409	63	100156	10.651	ppbv	99
25) Ethyl Acetate	2.839	43	201213	9.931	ppbv	99
26) Hexane	2.826	57	88211	9.993	ppbv	97
27) Carbon Disulfide	2.150	76	120690	11.716	ppbv	93
28) Methyl tert-Butyl Ether	2.441	73	51653	9.740	ppbv	98
29) Chloroform	2.852	83	172606	10.435	ppbv	93
30) Cyclohexane	3.859	84	74929	10.153	ppbv	98
31) cis-1,2-Dichloroethene	2.723	61	97864	9.503	ppbv	93
32) 1,1,1-Trichloroethane	3.370	97	182668	10.179	ppbv	99
34) 2-Butanone	2.561	43	131262	9.645	ppbv	97
35) Carbon Tetrachloride	3.765	117	190178	11.127	ppbv	100
36) Benzene	3.661	78	181690	10.233	ppbv	99
37) 1,2-Dichloroethane	3.221	62	146010	10.669	ppbv	98
38) Trichloroethene	4.555	130	71976	10.630	ppbv	90
39) 1,2-Dichloropropane	4.322	63	63612	9.570	ppbv #	87
40) 1,4-Dioxane	4.590	88	29247	9.396	ppbv #	100
41) Tetrahydrofuran	3.059	42	64241	9.932	ppbv	98
42) Bromodichloromethane	4.496	83	185825	11.113	ppbv	91
43) Methyl Methacrylate	4.891	69	72618	10.960	ppbv	96
44) 2,2,4-Trimethylpentane	4.645	57	311522	10.824	ppbv	99
45) t-1,3-Dichloropropene	6.422	75	93178	10.719	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042487.D
 Acq On : 06 May 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/07/2025
 Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 07 01:35:37 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.606	75	108203	10.473	ppbv	95
47) 1,1,2-Trichloroethane	6.587	97	75237	10.320	ppbv	93
48) Dibromochloromethane	7.393	129	134135	10.899	ppbv	99
49) Bromoform	9.490	173	119371	11.565	ppbv	98
50) 4-Methyl-2-Pentanone	5.759	43	177017	10.283	ppbv	96
51) 2-Hexanone	7.442	43	141871	10.308	ppbv	99
52) Tetrachloroethene	8.228	164	63220	9.354	ppbv	86
53) Toluene	6.940	91	205094	10.913	ppbv	99
54) 1,2-Dibromoethane	7.659	107	117565	10.548	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.934	131	104985	11.016	ppbv	98
57) Chlorobenzene	8.927	112	169084	10.428	ppbv	92
58) Ethyl Benzene	9.361	91	308243	11.431	ppbv	100
59) m/p-Xylene	9.555	91	540767m	23.920	ppbv	
60) o-Xylene	9.969	91	271925	12.013	ppbv	99
61) Styrene	9.876	104	99907	11.828	ppbv	97
62) Isopropylbenzene	10.546	105	384743	11.784	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.969	83	167724	10.748	ppbv	99
64) n-propylbenzene	10.992	120	91898	11.517	ppbv	95
65) tert-Butylbenzene	11.536	119	349935	12.010	ppbv	98
66) Benzyl Chloride	11.620	91	43100	12.798	ppbv	99
67) sec-Butylbenzene	11.772	105	496533	11.992	ppbv	99
69) p-Isopropyltoluene	11.931	119	416243	11.931	ppbv	99
70) n-Butylbenzene	12.287	91	437958	12.144	ppbv	100
71) 2-Chlorotoluene	10.905	91	303775	11.862	ppbv	100
72) 4-Ethyltoluene	11.131	105	331569	12.424	ppbv	99
73) 1,3,5-Trimethylbenzene	11.209	105	287473	12.248	ppbv	98
74) 1,2,4-Trimethylbenzene	11.542	105	324889	11.441	ppbv	94
75) 1,3-Dichlorobenzene	11.610	146	175086	11.074	ppbv	99
76) 1,4-Dichlorobenzene	11.675	146	178127	11.906	ppbv	98
77) 1,2-Dichlorobenzene	11.950	146	170747	11.034	ppbv	100
78) Hexachloro-1,3-Butadiene	13.886	225	152349	9.859	ppbv	99
79) Naphthalene	13.517	128	292585	12.239	ppbv	97
80) Naphthalene,2-methyl-	14.462	142	225180	12.673	ppbv	99
81) 1,2,4-Trichlorobenzene	13.446	180	142121	11.147	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042487.D
 Acq On : 06 May 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 07 01:35:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	1.655	1.897	-14.6	125	0.00
3	Chlorodifluoromethane	1.750	1.729	1.2	107	0.00
4	Chloromethane	0.692	0.748	-8.1	116	0.00
5 T	Vinyl Chloride	0.696	0.686	1.4	117	0.00
6 T	Bromomethane	0.329	0.371	-12.8	126	0.00
7	Chloroethane	0.257	0.274	-6.6	115	0.00
8 T	Dichlorotetrafluoroethane	1.364	1.527	-12.0	120	0.00
9 T	Propene	0.572	0.497	13.1	92	0.00
10 T	Heptane	1.484	1.489	-0.3	100	0.00
11 T	Trichlorofluoromethane	1.527	1.787	-17.0	126	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.219	1.355	-11.2	120	0.00
13	Ethanol	0.116	0.092	20.7	112	0.00
14 T	Bromoethene	0.450	0.524	-16.4	128	0.00
15 T	Acetone	1.826	1.611	11.8	119	0.00
16 T	1,3-Butadiene	0.805	0.802	0.4	114	0.00
17	tert-Butyl alcohol	1.608	1.709	-6.3	111	0.00
18 T	1,1-Dichloroethene	0.569	0.604	-6.2	119	0.00
19 T	Isopropyl Alcohol	1.010	0.996	1.4	110	0.00
20 T	Methylene Chloride	0.549	0.539	1.8	117	0.00
21 T	Allyl Chloride	1.174	1.214	-3.4	110	0.00
22 T	trans-1,2-Dichloroethene	0.665	0.682	-2.6	120	0.00
23 T	Vinyl Acetate	1.742	1.709	1.9	96	0.00
24 T	1,1-Dichloroethane	1.357	1.445	-6.5	117	0.00
25 T	Ethyl Acetate	2.923	2.903	0.7	100	0.00
26 T	Hexane	1.273	1.273	0.0	102	0.00
27 T	Carbon Disulfide	1.486	1.741	-17.2	119	0.00
28 T	Methyl tert-Butyl Ether	0.765	0.745	2.6	108	0.00
29 T	Chloroform	2.386	2.490	-4.4	110	0.00
30 T	Cyclohexane	1.065	1.081	-1.5	100	0.00
31 T	cis-1,2-Dichloroethene	1.486	1.412	5.0	102	0.00
32 T	1,1,1-Trichloroethane	2.589	2.635	-1.8	112	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
34 T	2-Butanone	0.723	0.698	3.5	99	0.00
35 T	Carbon Tetrachloride	0.908	1.011	-11.3	114	0.00
36 T	Benzene	0.944	0.966	-2.3	101	0.00
37 T	1,2-Dichloroethane	0.727	0.776	-6.7	109	0.00
38 T	Trichloroethene	0.360	0.383	-6.4	103	0.01
39 T	1,2-Dichloropropane	0.353	0.338	4.2	100	0.01
40 T	1,4-Dioxane	0.165	0.155	6.1	97	0.00
41 T	Tetrahydrofuran	0.344	0.341	0.9	98	0.00
42 T	Bromodichloromethane	0.889	0.988	-11.1	110	0.01
43	Methyl Methacrylate	0.352	0.386	-9.7	99	0.01
44 T	2,2,4-Trimethylpentane	1.530	1.656	-8.2	108	0.01
45 T	t-1,3-Dichloropropene	0.462	0.495	-7.1	96	0.01
46 T	cis-1,3-Dichloropropene	0.549	0.575	-4.7	97	0.00
47 T	1,1,2-Trichloroethane	0.387	0.400	-3.4	105	0.00
48 T	Dibromochloromethane	0.654	0.713	-9.0	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042487.D
 Acq On : 06 May 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 07 01:35:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.549	0.634	-15.5	112	0.01
50 T	4-Methyl-2-Pentanone	0.915	0.941	-2.8	101	0.01
51 T	2-Hexanone	0.731	0.754	-3.1	98	0.00
52 T	Tetrachloroethene	0.359	0.336	6.4	103	0.00
53 T	Toluene	0.999	1.090	-9.1	100	0.01
54 T	1,2-Dibromoethane	0.592	0.625	-5.6	105	0.01
55 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
56	1,1,1,2-Tetrachloroethane	0.591	0.651	-10.2	111	0.01
57 T	Chlorobenzene	1.005	1.048	-4.3	107	0.00
58 T	Ethyl Benzene	1.671	1.910	-14.3	103	0.00
59 T	m/p-Xylene	1.401	1.676	-19.6	109	0.03
60 T	o-Xylene	1.403	1.685	-20.1	110	0.00
61 T	Styrene	0.523	0.619	-18.4	98	0.00
62	Isopropylbenzene	2.023	2.384	-17.8	109	0.00
63 T	1,1,2,2-Tetrachloroethane	0.967	1.039	-7.4	114	0.00
64	n-propylbenzene	0.495	0.570	-15.2	109	0.00
65	tert-Butylbenzene	1.806	2.169	-20.1	111	0.00
66 T	Benzyl Chloride	0.209	0.267	-27.8	107	0.00
67	sec-Butylbenzene	2.566	3.077	-19.9	112	0.00
68 S	1-Bromo-4-Fluorobenzene	0.927	1.025	-10.6	100	0.00
69	p-Isopropyltoluene	2.162	2.579	-19.3	111	0.00
70	n-Butylbenzene	2.235	2.714	-21.4	114	0.00
71	2-Chlorotoluene	1.587	1.883	-18.7	109	0.00
72 T	4-Ethyltoluene	1.654	2.055	-24.2	111	0.00
73 T	1,3,5-Trimethylbenzene	1.455	1.781	-22.4	112	0.00
74 T	1,2,4-Trimethylbenzene	1.760	2.013	-14.4	113	0.00
75 T	1,3-Dichlorobenzene	0.980	1.085	-10.7	113	0.00
76 T	1,4-Dichlorobenzene	0.927	1.104	-19.1	113	0.00
77 T	1,2-Dichlorobenzene	0.959	1.058	-10.3	114	0.00
78 T	Hexachloro-1,3-Butadiene	0.958	0.944	1.5	113	0.00
79 T	Naphthalene	1.482	1.813	-22.3	110	0.00
80 T	Naphthalene,2-methyl-	1.101	1.395	-26.7	110	0.00
81 T	1,2,4-Trichlorobenzene	0.790	0.881	-11.5	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042487.D
 Acq On : 06 May 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 07 01:35:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	10.000	11.466	-14.7	125	0.00
3	Chlorodifluoromethane	10.000	9.883	1.2	107	0.00
4	Chloromethane	10.000	10.814	-8.1	116	0.00
5 T	Vinyl Chloride	10.000	9.847	1.5	117	0.00
6 T	Bromomethane	10.000	11.263	-12.6	126	0.00
7	Chloroethane	10.000	10.651	-6.5	115	0.00
8 T	Dichlorotetrafluoroethane	10.000	11.192	-11.9	120	0.00
9 T	Propene	10.000	8.689	13.1	92	0.00
10 T	Heptane	10.000	10.030	-0.3	100	0.00
11 T	Trichlorofluoromethane	10.000	11.700	-17.0	126	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	11.114	-11.1	120	0.00
13	Ethanol	10.000	7.889	21.1	112	0.00
14 T	Bromoethene	10.000	11.645	-16.4	128	0.00
15 T	Acetone	10.000	8.825	11.8	119	0.00
16 T	1,3-Butadiene	10.000	9.963	0.4	114	0.00
17	tert-Butyl alcohol	10.000	10.628	-6.3	111	0.00
18 T	1,1-Dichloroethene	10.000	10.614	-6.1	119	0.00
19 T	Isopropyl Alcohol	10.000	9.856	1.4	110	0.00
20 T	Methylene Chloride	10.000	9.829	1.7	117	0.00
21 T	Allyl Chloride	10.000	10.342	-3.4	110	0.00
22 T	trans-1,2-Dichloroethene	10.000	10.253	-2.5	120	0.00
23 T	Vinyl Acetate	10.000	9.809	1.9	96	0.00
24 T	1,1-Dichloroethane	10.000	10.651	-6.5	117	0.00
25 T	Ethyl Acetate	10.000	9.931	0.7	100	0.00
26 T	Hexane	10.000	9.993	0.1	102	0.00
27 T	Carbon Disulfide	10.000	11.716	-17.2	119	0.00
28 T	Methyl tert-Butyl Ether	10.000	9.740	2.6	108	0.00
29 T	Chloroform	10.000	10.435	-4.4	110	0.00
30 T	Cyclohexane	10.000	10.153	-1.5	100	0.00
31 T	cis-1,2-Dichloroethene	10.000	9.503	5.0	102	0.00
32 T	1,1,1-Trichloroethane	10.000	10.179	-1.8	112	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	93	0.00
34 T	2-Butanone	10.000	9.645	3.6	99	0.00
35 T	Carbon Tetrachloride	10.000	11.127	-11.3	114	0.00
36 T	Benzene	10.000	10.233	-2.3	101	0.00
37 T	1,2-Dichloroethane	10.000	10.669	-6.7	109	0.00
38 T	Trichloroethene	10.000	10.630	-6.3	103	0.01
39 T	1,2-Dichloropropane	10.000	9.570	4.3	100	0.01
40 T	1,4-Dioxane	10.000	9.396	6.0	97	0.00
41 T	Tetrahydrofuran	10.000	9.932	0.7	98	0.00
42 T	Bromodichloromethane	10.000	11.113	-11.1	110	0.01
43	Methyl Methacrylate	10.000	10.960	-9.6	99	0.01
44 T	2,2,4-Trimethylpentane	10.000	10.824	-8.2	108	0.01
45 T	t-1,3-Dichloropropene	10.000	10.719	-7.2	96	0.01
46 T	cis-1,3-Dichloropropene	10.000	10.473	-4.7	97	0.00
47 T	1,1,2-Trichloroethane	10.000	10.320	-3.2	105	0.00
48 T	Dibromochloromethane	10.000	10.899	-9.0	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042487.D
 Acq On : 06 May 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 07 01:35:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	Bromoform	10.000	11.565	-15.6	112	0.01
50 T	4-Methyl-2-Pentanone	10.000	10.283	-2.8	101	0.01
51 T	2-Hexanone	10.000	10.308	-3.1	98	0.00
52 T	Tetrachloroethene	10.000	9.354	6.5	103	0.00
53 T	Toluene	10.000	10.913	-9.1	100	0.01
54 T	1,2-Dibromoethane	10.000	10.548	-5.5	105	0.01
55 I	Chlorobenzene-d5	10.000	10.000	0.0	93	0.00
56	1,1,1,2-Tetrachloroethane	10.000	11.016	-10.2	111	0.01
57 T	Chlorobenzene	10.000	10.428	-4.3	107	0.00
58 T	Ethyl Benzene	10.000	11.431	-14.3	103	0.00
59 T	m/p-Xylene	20.000	23.920	-19.6	109	0.03
60 T	o-Xylene	10.000	12.013	-20.1	110	0.00
61 T	Styrene	10.000	11.828	-18.3	98	0.00
62	Isopropylbenzene	10.000	11.784	-17.8	109	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	10.748	-7.5	114	0.00
64	n-propylbenzene	10.000	11.517	-15.2	109	0.00
65	tert-Butylbenzene	10.000	12.010	-20.1	111	0.00
66 T	Benzyl Chloride	10.000	12.798	-28.0	107	0.00
67	sec-Butylbenzene	10.000	11.992	-19.9	112	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	11.051	-10.5	100	0.00
69	p-Isopropyltoluene	10.000	11.931	-19.3	111	0.00
70	n-Butylbenzene	10.000	12.144	-21.4	114	0.00
71	2-Chlorotoluene	10.000	11.862	-18.6	109	0.00
72 T	4-Ethyltoluene	10.000	12.424	-24.2	111	0.00
73 T	1,3,5-Trimethylbenzene	10.000	12.248	-22.5	112	0.00
74 T	1,2,4-Trimethylbenzene	10.000	11.441	-14.4	113	0.00
75 T	1,3-Dichlorobenzene	10.000	11.074	-10.7	113	0.00
76 T	1,4-Dichlorobenzene	10.000	11.906	-19.1	113	0.00
77 T	1,2-Dichlorobenzene	10.000	11.034	-10.3	114	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	9.859	1.4	113	0.00
79 T	Naphthalene	10.000	12.239	-22.4	110	0.00
80 T	Naphthalene,2-methyl-	10.000	12.673	-26.7	110	0.00
81 T	1,2,4-Trichlorobenzene	10.000	11.147	-11.5	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042488.D
 Acq On : 06 May 2025 09:49
 Operator : SY/MD
 Sample : VL0506ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0506ABL01

Quant Time: May 07 01:36:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	71766	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	194806	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	158449	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	146316	9.960	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.600%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
Data File : VL042488.D
Acq On : 06 May 2025 09:49
Operator : SY/MD
Sample : VL0506ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0506ABL01

Quant Time: May 07 01:36:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 02 03:16:31 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042489.D
 Acq On : 06 May 2025 10:44
 Operator : SY/MD
 Sample : VL0506ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0506ABS01

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/07/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 07 01:36:51 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	68512	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	189766	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	162176	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.380 95 162347 10.797 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 108.000%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	128425	11.327	ppbv	99
3) Chlorodifluoromethane	1.476	51	120834	10.080	ppbv	98
4) Chloromethane	1.531	50	51363	10.839	ppbv	94
5) Vinyl Chloride	1.580	62	50552	10.596	ppbv	95
6) Bromomethane	1.667	94	24844	11.021	ppbv #	81
7) Chloroethane	1.706	64	19253	10.929	ppbv	98
8) Dichlorotetrafluoroethane	1.554	85	108256	11.583	ppbv	99
9) Propene	1.483	41	36687	9.357	ppbv	100
10) Heptane	4.985	43	108748	10.693	ppbv	99
11) Trichlorofluoromethane	1.878	101	118964	11.371	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.140	101	92491	11.072	ppbv	97
13) Ethanol	1.722	45	5717	7.185	ppbv	96
14) Bromoethene	1.781	108	35036	11.361	ppbv	99
15) Acetone	1.836	43	110911	8.868	ppbv	98
16) 1,3-Butadiene	1.609	39	57769	10.473	ppbv	100
17) tert-Butyl alcohol	2.043	59	121629	11.038	ppbv	98
18) 1,1-Dichloroethene	2.033	96	41573	10.658	ppbv	91
19) Isopropyl Alcohol	1.887	45	70143	10.133	ppbv	100
20) Methylene Chloride	2.062	84	38025	10.113	ppbv	97
21) Allyl Chloride	2.095	41	87030	10.820	ppbv	98
22) trans-1,2-Dichloroethene	2.341	96	46194	10.140	ppbv	92
23) Vinyl Acetate	2.463	43	122507	10.263	ppbv	99
24) 1,1-Dichloroethane	2.408	63	99878	10.747	ppbv	98
25) Ethyl Acetate	2.839	43	207809	10.377	ppbv	99
26) Hexane	2.826	57	94005	10.775	ppbv	95
27) Carbon Disulfide	2.150	76	118732	11.662	ppbv #	89
28) Methyl tert-Butyl Ether	2.441	73	52295	9.977	ppbv	97
29) Chloroform	2.849	83	178140	10.896	ppbv	98
30) Cyclohexane	3.858	84	78419	10.751	ppbv	98
31) cis-1,2-Dichloroethene	2.722	61	102999	10.119	ppbv	97
32) 1,1,1-Trichloroethane	3.370	97	185042	10.433	ppbv	99
34) 2-Butanone	2.561	43	135534	9.874	ppbv	95
35) Carbon Tetrachloride	3.765	117	192452	11.164	ppbv	96
36) Benzene	3.658	78	188082	10.503	ppbv	96
37) 1,2-Dichloroethane	3.221	62	150741	10.922	ppbv	99
38) Trichloroethene	4.554	130	75153	11.005	ppbv	97
39) 1,2-Dichloropropane	4.315	63	67616	10.086	ppbv	92
40) 1,4-Dioxane	4.583	88	30720	9.786	ppbv #	100
41) Tetrahydrofuran	3.056	42	66397	10.179	ppbv	98
42) Bromodichloromethane	4.496	83	188247	11.163	ppbv	96
43) Methyl Methacrylate	4.888	69	73710	11.030	ppbv	95
44) 2,2,4-Trimethylpentane	4.642	57	317017	10.921	ppbv	99
45) t-1,3-Dichloropropene	6.419	75	95245	10.864	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042489.D
 Acq On : 06 May 2025 10:44
 Operator : SY/MD
 Sample : VL0506ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0506ABS01

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/07/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 07 01:36:51 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.603	75	112024	10.751	ppbv	96
47) 1,1,2-Trichloroethane	6.590	97	80000	10.880	ppbv	99
48) Dibromochloromethane	7.390	129	137019	11.039	ppbv	99
49) Bromoform	9.487	173	115649	11.110	ppbv	99
50) 4-Methyl-2-Pentanone	5.755	43	179961	10.365	ppbv	97
51) 2-Hexanone	7.441	43	143801	10.359	ppbv #	99
52) Tetrachloroethene	8.228	164	61852	9.074	ppbv	94
53) Toluene	6.940	91	211514	11.159	ppbv	99
54) 1,2-Dibromoethane	7.655	107	119785	10.656	ppbv	90
56) 1,1,1,2-Tetrachloroethane	8.927	131	107810	11.256	ppbv	99
57) Chlorobenzene	8.927	112	170186	10.444	ppbv	98
58) Ethyl Benzene	9.361	91	316187	11.667	ppbv	100
59) m/p-Xylene	9.555	91	546871m	24.069	ppbv	
60) o-Xylene	9.966	91	275574	12.114	ppbv	98
61) Styrene	9.875	104	100944	11.891	ppbv	95
62) Isopropylbenzene	10.542	105	386743	11.786	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.966	83	168155	10.722	ppbv	99
64) n-propylbenzene	10.992	120	95485	11.906	ppbv	98
65) tert-Butylbenzene	11.536	119	347665	11.872	ppbv	97
66) Benzyl Chloride	11.617	91	44409	13.121	ppbv	96
67) sec-Butylbenzene	11.772	105	500648	12.031	ppbv	99
69) p-Isopropyltoluene	11.931	119	418184	11.927	ppbv	99
70) n-Butylbenzene	12.283	91	439757	12.133	ppbv	99
71) 2-Chlorotoluene	10.905	91	304635	11.836	ppbv	99
72) 4-Ethyltoluene	11.128	105	332254	12.388	ppbv	99
73) 1,3,5-Trimethylbenzene	11.209	105	285001	12.082	ppbv	94
74) 1,2,4-Trimethylbenzene	11.539	105	326996	11.457	ppbv	95
75) 1,3-Dichlorobenzene	11.610	146	176191	11.089	ppbv	99
76) 1,4-Dichlorobenzene	11.675	146	173576	11.543	ppbv	98
77) 1,2-Dichlorobenzene	11.950	146	171125	11.003	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	153031	9.854	ppbv	99
79) Naphthalene	13.517	128	289175	12.036	ppbv	98
80) Naphthalene,2-methyl-	14.458	142	236741	13.257	ppbv	99
81) 1,2,4-Trichlorobenzene	13.442	180	142190	11.096	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042489.D
 Acq On : 06 May 2025 10:44
 Operator : SY/MD
 Sample : VL0506ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0506ABS01

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 05/07/2025
 Supervised By : Mahesh Dadoda 05/09/2025


Quant Time: May 07 01:36:51 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042492.D
 Acq On : 06 May 2025 12:19
 Operator : SY/MD
 Sample : QC042225 CAN 10669
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC042225 CAN 10669

Quant Time: May 07 01:38:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	63187	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	172595	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	143665	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.384	95	131884	9.901	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.000%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
Data File : VL042492.D
Acq On : 06 May 2025 12:19
Operator : SY/MD
Sample : QC042225 CAN 10669
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC042225 CAN 10669

Quant Time: May 07 01:38:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 02 03:16:31 2025
Response via : Initial Calibration





Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
Data File : VL043030.D
Acq On : 02 Oct 2025 09:11
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri Oct 03 08:23:34 2025



AutoFind: Scans 3173, 3174, 3175; Background Corrected with Scan 3150

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	22.6	54675	PASS
75	95	30	66	56.0	135808	PASS
95	95	100	100	100.0	242325	PASS
96	95	5	9	6.8	16496	PASS
173	174	0.00	2	0.3	425	PASS
174	95	50	120	65.8	159467	PASS
175	174	4	9	6.9	10992	PASS
176	174	93	101	97.8	156032	PASS
177	176	5	9	6.3	9819	PASS

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL100225AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Fri Oct 03 08:23:34 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL043037.D 0.1 =VL043036.D 0.5 =VL043031.D 1 =VL043034.D 2 =VL043033.D 10 =VL043032.D 15 =VL043038.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.748	1.803	1.698	1.587	1.591	1.685	5.66
3) Chlorodifluoro...			1.726	1.701	1.684	1.527	1.513	1.630	6.23
4) Chloromethane			0.644	0.663	0.614	0.577	0.580	0.616	6.17
5) T Vinyl Chloride	0.573	0.687	0.605	0.573	0.564	0.539	0.548	0.584	8.59
6) T Bromomethane			0.260	0.283	0.270	0.248	0.252	0.263	5.35
7) Chloroethane			0.259	0.257	0.256	0.223	0.224	0.244	7.62
8) T Dichlorotetra...			1.429	1.430	1.438	1.301	1.307	1.381	5.09
9) T Propene			0.657	0.691	0.697	0.633	0.609	0.657	5.69
10) T Heptane			1.331	1.565	1.650	1.760	1.776	1.616	11.19
11) T Trichlorofluor...			1.770	1.866	1.761	1.621	1.642	1.732	5.83
12) T 1,1,2-Trichlor...			1.449	1.410	1.392	1.284	1.286	1.364	5.50
13) Ethanol			0.085	0.049	0.041	0.027	0.026	0.045#	52.72
14) T Bromoethene			0.507	0.499	0.501	0.469	0.469	0.489	3.80
15) T Acetone			1.669	1.561	1.472	1.143	1.155	1.400	17.12
16) T 1,3-Butadiene			0.609	0.587	0.611	0.536	0.543	0.577	6.17
17) tert-Butyl alc...			1.663	1.586	1.437	1.377	1.344	1.481	9.28
18) T 1,1-Dichloroet...			0.607	0.626	0.610	0.575	0.572	0.598	3.90
19) T Isopropyl Alcohol			0.768	0.727	0.667	0.603	0.589	0.671	11.56
20) T Methylene Chlo...			0.809	0.684	0.577	0.490	0.496	0.611	22.19
21) T Allyl Chloride			0.982	0.888	0.920	0.843	0.852	0.897	6.32
22) T trans-1,2-Dich...			0.723	0.713	0.676	0.635	0.648	0.679	5.74
23) T Vinyl Acetate			0.388	0.316	0.308	0.244	0.231	0.297	21.19
24) T 1,1-Dichloroet...			1.347	1.365	1.295	1.206	1.221	1.287	5.57
25) T Ethyl Acetate			3.265	3.663	3.644	3.520	3.471	3.513	4.56
26) T Hexane			1.261	1.271	1.413	1.401	1.411	1.351	5.80
27) T Carbon Disulfide			1.681	1.655	1.576	1.521	1.543	1.595	4.37
28) T Methyl tert-Bu...			0.853	0.872	0.868	0.823	0.813	0.846	3.14
29) T Chloroform			2.393	2.395	2.326	2.206	2.199	2.304	4.19
30) T Cyclohexane			0.947	1.068	1.145	1.191	1.200	1.110	9.46
31) T cis-1,2-Dichlo...			1.329	1.390	1.436	1.403	1.388	1.389	2.79
32) T 1,1,1-Trichlor...	2.543	2.098	2.204	2.216	2.255	2.170	2.186	2.239	6.38
33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.704	0.685	0.695	0.655	0.654	0.679	3.42
35) T Carbon Tetrach...	1.006	0.828	0.816	0.805	0.770	0.761	0.773	0.823	10.30
36) T Benzene			0.901	0.984	0.971	0.964	0.957	0.955	3.32
37) T 1,2-Dichloroet...			0.584	0.577	0.570	0.547	0.551	0.566	2.79
38) T Trichloroethene	0.355	0.386	0.404	0.424	0.402	0.409	0.407	0.398	5.58
39) T 1,2-Dichloropr...			0.360	0.362	0.354	0.350	0.345	0.354	2.02

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\ Method File : VL100225AIR.M									
40) T	1,4-Dioxane		0.146	0.151	0.152	0.157	0.154	0.152	2.71
41) T	Tetrahydrofuran		0.297	0.338	0.365	0.376	0.378	0.351	9.74
42) T	Bromodichlorom...		0.818	0.805	0.791	0.788	0.793	0.799	1.55
43) T	Methyl Methacr...		0.281	0.286	0.308	0.353	0.358	0.317	11.50
44) T	2,2,4-Trimethy...		1.388	1.494	1.589	1.593	1.601	1.533	6.01
45) T	t-1,3-Dichloro...		0.268	0.268	0.279	0.314	0.322	0.290	8.91
46) T	cis-1,3-Dichlo...		0.374	0.386	0.396	0.440	0.449	0.409	8.23
47) T	1,1,2-Trichlor...		0.414	0.405	0.395	0.376	0.379	0.394	4.18
48) T	Dibromochlorom...		0.610	0.644	0.644	0.661	0.655	0.643	3.10
49) T	Bromoform		0.492	0.516	0.542	0.560	0.575	0.537	6.24
50) T	4-Methyl-2-Pen...		0.719	0.763	0.820	0.910	0.926	0.827	10.90
51) T	2-Hexanone		0.438	0.515	0.578	0.693	0.720	0.589	20.15
52) T	Tetrachloroethene	0.424 0.328	0.339	0.343	0.352	0.345	0.346	0.354	9.02
53) T	Toluene		0.790	0.881	0.966	1.087	1.099	0.965	13.74
54) T	1,2-Dibromoethane	0.445	0.448	0.472	0.471	0.481	0.484	0.467	3.56
-----ISTD-----									
55) I	Chlorobenzene-d5								
56) T	1,1,1,2-Tetrac...		0.631	0.623	0.614	0.591	0.590	0.610	3.04
57) T	Chlorobenzene		1.088	1.096	1.086	1.015	1.002	1.057	4.24
58) T	Ethyl Benzene		1.232	1.379	1.598	1.759	1.778	1.549	15.41
59) T	m/p-Xylene		1.012	1.231	1.401	1.443	1.446	1.307	14.29
60) T	o-Xylene		1.020	1.216	1.385	1.439	1.448	1.302	14.07
61) T	Styrene		0.398	0.449	0.523	0.666	0.691	0.545	23.81
62) T	Isopropylbenzene		1.675	1.919	2.082	2.141	2.174	1.998	10.30
63) T	1,1,2,2-Tetrac...	1.080 0.884	0.919	0.949	0.949	0.870	0.886	0.934	7.66
64) T	n-propylbenzene		0.384	0.491	0.540	0.562	0.579	0.511	15.38
65) T	tert-Butylbenzene		1.384	1.730	1.932	1.928	1.918	1.778	13.28
66) T	Benzyl Chloride		0.064	0.066	0.066	0.069	0.075	0.068	6.61
67) T	sec-Butylbenzene		1.912	2.476	2.742	2.652	2.626	2.482	13.41
68) S	1-Bromo-4-Fluo...	0.688 0.692	0.727	0.726	0.732	0.735	0.743	0.720	3.00
69) T	p-Isopropyltol...		1.499	1.879	2.234	2.250	2.231	2.019	16.33
70) T	n-Butylbenzene		1.361	1.776	2.148	2.222	2.212	1.944	19.22
71) T	2-Chlorotoluene		1.156	1.348	1.522	1.583	1.621	1.446	13.34
72) T	4-Ethyltoluene		1.145	1.426	1.647	1.751	1.778	1.549	17.12
73) T	1,3,5-Trimethy...		1.041	1.271	1.423	1.470	1.496	1.340	14.06
74) T	1,2,4-Trimethy...		1.133	1.492	1.676	1.633	1.630	1.513	14.77
75) T	1,3-Dichlorobe...		0.932	0.986	1.031	0.976	0.994	0.984	3.60
76) T	1,4-Dichlorobe...		0.825	0.929	0.990	0.967	0.980	0.938	7.20
77) T	1,2-Dichlorobe...		0.951	1.037	1.036	0.930	0.945	0.980	5.35
78) T	Hexachloro-1,3...		0.683	0.705	0.731	0.645	0.650	0.683	5.29
79) T	Naphthalene	0.333	0.611	0.746	1.115	1.288	1.326	0.903	44.50
80) T	Naphthalene,2-...		0.645	0.167	0.278	0.567	0.629	0.457	48.04
81) T	1,2,4-Trichlor...		0.383	0.476	0.629	0.696	0.731	0.583	25.49

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043031.D
 Acq On : 02 Oct 2025 09:50
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/03/2025
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:00:22 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Oct 03 07:59:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.774	49	172062	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.936	114	504004	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.859	117	415115	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.357 95 301605 10.086 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.900%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.492	85	15037	0.519	ppbv	97
3) Chlorodifluoromethane	1.470	51	14848	0.529	ppbv	96
4) Chloromethane	1.528	50	5542	0.523	ppbv	91
5) Vinyl Chloride	1.573	62	5205	0.518	ppbv	99
6) Bromomethane	1.660	94	2235	0.495	ppbv	87
7) Chloroethane	1.696	64	2228	0.531	ppbv	94
8) Dichlorotetrafluoroethane	1.547	85	12294	0.517	ppbv	99
9) Propene	1.479	41	5652	0.500	ppbv	95
10) Heptane	4.949	43	11455m	0.411	ppbv	
11) Trichlorofluoromethane	1.871	101	15230	0.511	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.130	101	12463	0.531	ppbv	99
13) Ethanol	1.716	45	729	0.509	ppbv #	31
14) Bromoethene	1.774	108	4364	0.519	ppbv #	91
15) Acetone	1.829	43	14360	0.596	ppbv	100
16) 1,3-Butadiene	1.602	39	5235	0.527	ppbv	97
17) tert-Butyl alcohol	2.033	59	14304	0.561	ppbv	96
18) 1,1-Dichloroethene	2.026	96	5223	0.508	ppbv	93
19) Isopropyl Alcohol	1.877	45	6611	0.573	ppbv #	32
20) Methylene Chloride	2.052	84	6961	0.662	ppbv	92
21) Allyl Chloride	2.088	41	8450	0.547	ppbv	97
22) trans-1,2-Dichloroethene	2.327	96	6223	0.533	ppbv	89
23) Vinyl Acetate	2.447	43	3334	0.652	ppbv #	82
24) 1,1-Dichloroethane	2.398	63	11586	0.523	ppbv	99
25) Ethyl Acetate	2.822	43	28090	0.465	ppbv	98
26) Hexane	2.813	57	10845	0.466	ppbv	92
27) Carbon Disulfide	2.140	76	14458	0.527	ppbv #	36
28) Methyl tert-Butyl Ether	2.431	73	7340	0.505	ppbv	94
29) Chloroform	2.835	83	20586	0.519	ppbv	97
30) Cyclohexane	3.839	84	8149	0.427	ppbv	93
31) cis-1,2-Dichloroethene	2.706	61	11430	0.478	ppbv	95
32) 1,1,1-Trichloroethane	3.350	97	18962	0.492	ppbv	97
34) 2-Butanone	2.547	43	17736	0.519	ppbv	100
35) Carbon Tetrachloride	3.742	117	20564	0.496	ppbv	93
36) Benzene	3.638	78	22717	0.472	ppbv	93
37) 1,2-Dichloroethane	3.204	62	14711	0.516	ppbv	98
38) Trichloroethene	4.515	130	10181m	0.507	ppbv	
39) 1,2-Dichloropropane	4.289	63	9070	0.508	ppbv #	86
40) 1,4-Dioxane	4.570	88	3676m	0.480	ppbv	
41) Tetrahydrofuran	3.046	42	7477	0.423	ppbv	95
42) Bromodichloromethane	4.463	83	20614	0.512	ppbv #	97
43) Methyl Methacrylate	4.852	69	7092	0.444	ppbv	91
44) 2,2,4-Trimethylpentane	4.616	57	34977	0.453	ppbv	96
45) t-1,3-Dichloropropene	6.376	75	6760m	0.467	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043031.D
 Acq On : 02 Oct 2025 09:50
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/03/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:00:22 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Oct 03 07:59:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.561	75	9423m	0.458	ppbv	
47) 1,1,2-Trichloroethane	6.551	97	10438	0.526	ppbv	91
48) Dibromochloromethane	7.357	129	15360	0.474	ppbv	94
49) Bromoform	9.461	173	12398	0.458	ppbv	95
50) 4-Methyl-2-Pentanone	5.726	43	18115	0.434	ppbv	100
51) 2-Hexanone	7.418	43	11043m	0.373	ppbv	
52) Tetrachloroethene	8.205	164	8537	0.476	ppbv	94
53) Toluene	6.904	91	19917m	0.410	ppbv	
54) 1,2-Dibromoethane	7.626	107	11279	0.479	ppbv	89
56) 1,1,1,2-Tetrachloroethane	8.901	131	13091	0.517	ppbv #	1
57) Chlorobenzene	8.898	112	22573	0.514	ppbv	95
58) Ethyl Benzene	9.335	91	25575	0.398	ppbv	96
59) m/p-Xylene	9.529	91	42005m	0.774	ppbv	
60) o-Xylene	9.943	91	21166	0.392	ppbv	97
61) Styrene	9.849	104	8258	0.365	ppbv	98
62) Isopropylbenzene	10.522	105	34761	0.419	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.943	83	19066	0.490	ppbv	99
64) n-propylbenzene	10.972	120	7963	0.375	ppbv	87
65) tert-Butylbenzene	11.516	119	28729	0.389	ppbv	97
66) Benzyl Chloride	11.600	91	1321	0.462	ppbv #	92
67) sec-Butylbenzene	11.752	105	39676	0.385	ppbv	98
69) p-Isopropyltoluene	11.911	119	31118	0.371	ppbv	98
70) n-Butylbenzene	12.267	91	28253	0.350	ppbv	95
71) 2-Chlorotoluene	10.885	91	24004	0.400	ppbv	98
72) 4-Ethyltoluene	11.108	105	23762	0.369	ppbv #	96
73) 1,3,5-Trimethylbenzene	11.186	105	21617	0.389	ppbv	98
74) 1,2,4-Trimethylbenzene	11.523	105	23511	0.374	ppbv	88
75) 1,3-Dichlorobenzene	11.590	146	19346	0.474	ppbv	99
76) 1,4-Dichlorobenzene	11.655	146	17115	0.440	ppbv	98
77) 1,2-Dichlorobenzene	11.934	146	19742	0.485	ppbv	99
78) Hexachloro-1,3-Butadiene	13.866	225	14174	0.500	ppbv	99
79) Naphthalene	13.500	128	12682	0.479	ppbv #	96
80) Naphthalene,2-methyl-	14.442	142	13384	1.165	ppbv	95
81) 1,2,4-Trichlorobenzene	13.426	180	7952	0.329	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043032.D
 Acq On : 02 Oct 2025 10:23
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/03/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:01:20 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Fri Oct 03 07:59:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.774	49	175694	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.939	114	523534	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.862	117	448396	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.361 95 329578 10.203 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.000%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	278856	9.417	ppbv	100
3) Chlorodifluoromethane	1.473	51	268344	9.368	ppbv	100
4) Chloromethane	1.528	50	101430	9.378	ppbv	100
5) Vinyl Chloride	1.576	62	94623	9.218	ppbv	100
6) Bromomethane	1.664	94	43581	9.448	ppbv	100
7) Chloroethane	1.699	64	39123	9.134	ppbv	100
8) Dichlorotetrafluoroethane	1.551	85	228586	9.421	ppbv	100
9) Propene	1.479	41	111264	9.633	ppbv	100
10) Heptane	4.956	43	309134	10.867	ppbv	100
11) Trichlorofluoromethane	1.871	101	284734	9.357	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.130	101	225532	9.410	ppbv	100
13) Ethanol	1.716	45	4785	10.163	ppbv	100
14) Bromoethene	1.777	108	82329	9.582	ppbv	100
15) Acetone	1.829	43	200896	8.166	ppbv	100
16) 1,3-Butadiene	1.606	39	94239	9.292	ppbv	100
17) tert-Butyl alcohol	2.033	59	241909	9.297	ppbv	100
18) 1,1-Dichloroethene	2.026	96	101035	9.617	ppbv	100
19) Isopropyl Alcohol	1.877	45	105917	8.986	ppbv	100
20) Methylene Chloride	2.052	84	86085	8.017	ppbv	100
21) Allyl Chloride	2.088	41	148094	9.397	ppbv	100
22) trans-1,2-Dichloroethene	2.331	96	111481	9.346	ppbv	100
23) Vinyl Acetate	2.450	43	42908	8.216	ppbv	100
24) 1,1-Dichloroethane	2.399	63	211973	9.375	ppbv	100
25) Ethyl Acetate	2.819	43	618469	10.022	ppbv	100
26) Hexane	2.813	57	246127	10.368	ppbv	100
27) Carbon Disulfide	2.143	76	267294	9.537	ppbv	100
28) Methyl tert-Butyl Ether	2.428	73	144534	9.729	ppbv	100
29) Chloroform	2.836	83	387615	9.577	ppbv	100
30) Cyclohexane	3.839	84	209266	10.728	ppbv	100
31) cis-1,2-Dichloroethene	2.709	61	246505	10.101	ppbv	100
32) 1,1,1-Trichloroethane	3.353	97	381212	9.691	ppbv	100
34) 2-Butanone	2.544	43	342678	9.647	ppbv	100
35) Carbon Tetrachloride	3.745	117	398481	9.251	ppbv	100
36) Benzene	3.638	78	504680	10.090	ppbv	100
37) 1,2-Dichloroethane	3.201	62	286590	9.674	ppbv	100
38) Trichloroethene	4.525	130	214186	10.277	ppbv	100
39) 1,2-Dichloropropane	4.292	63	182987	9.870	ppbv	100
40) 1,4-Dioxane	4.554	88	82195	10.324	ppbv #	98
41) Tetrahydrofuran	3.036	42	196992	10.730	ppbv	100
42) Bromodichloromethane	4.467	83	412493	9.861	ppbv	100
43) Methyl Methacrylate	4.852	69	184960	11.139	ppbv	100
44) 2,2,4-Trimethylpentane	4.616	57	834185	10.393	ppbv	100
45) t-1,3-Dichloropropene	6.386	75	164433	10.931	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043032.D
 Acq On : 02 Oct 2025 10:23
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/03/2025
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:01:20 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Oct 03 07:59:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.574	75	230231	10.772	ppbv	100
47) 1,1,2-Trichloroethane	6.551	97	196990	9.552	ppbv	100
48) Dibromochloromethane	7.364	129	345945	10.278	ppbv	100
49) Bromoform	9.464	173	292987	10.423	ppbv	100
50) 4-Methyl-2-Pentanone	5.713	43	476242	10.994	ppbv	100
51) 2-Hexanone	7.412	43	362983	11.817	ppbv	100
52) Tetrachloroethene	8.205	164	180664	9.693	ppbv	100
53) Toluene	6.907	91	568843	11.265	ppbv	100
54) 1,2-Dibromoethane	7.629	107	251974	10.309	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.907	131	264925	9.693	ppbv	100
57) Chlorobenzene	8.904	112	455049	9.600	ppbv	100
58) Ethyl Benzene	9.338	91	788510	11.351	ppbv	100
59) m/p-Xylene	9.532	91	1293839m	22.074	ppbv	
60) o-Xylene	9.946	91	645403	11.058	ppbv	100
61) Styrene	9.853	104	298597	12.212	ppbv	100
62) Isopropylbenzene	10.526	105	960158	10.716	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.946	83	390289	9.278	ppbv	100
64) n-propylbenzene	10.976	120	252073	10.998	ppbv	99
65) tert-Butylbenzene	11.516	119	864475	10.841	ppbv	100
66) Benzyl Chloride	11.604	91	30823	9.975	ppbv	100
67) sec-Butylbenzene	11.752	105	1189296	10.688	ppbv	100
69) p-Isopropyltoluene	11.914	119	1008951	11.146	ppbv	100
70) n-Butylbenzene	12.267	91	996449	11.433	ppbv	100
71) 2-Chlorotoluene	10.885	91	709807	10.946	ppbv	100
72) 4-Ethyltoluene	11.112	105	785118	11.302	ppbv	100
73) 1,3,5-Trimethylbenzene	11.189	105	659038	10.966	ppbv	100
74) 1,2,4-Trimethylbenzene	11.523	105	732252	10.796	ppbv	100
75) 1,3-Dichlorobenzene	11.594	146	437848	9.925	ppbv	100
76) 1,4-Dichlorobenzene	11.659	146	433712	10.311	ppbv	100
77) 1,2-Dichlorobenzene	11.934	146	417019	9.491	ppbv	100
78) Hexachloro-1,3-Butadiene	13.869	225	289398	9.452	ppbv	100
79) Naphthalene	13.500	128	577659	10.038	ppbv	100
80) Naphthalene,2-methyl-	14.445	142	254437	9.510	ppbv	100
81) 1,2,4-Trichlorobenzene	13.429	180	312011	11.936	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043033.D
 Acq On : 02 Oct 2025 10:57
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/03/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:02:17 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Oct 03 07:59:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.771	49	179740	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.936	114	541054	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.859	117	447624	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.357 95 327830 10.167 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 101.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.492	85	61057	2.016	ppbv	97
3) Chlorodifluoromethane	1.470	51	60544	2.066	ppbv	97
4) Chloromethane	1.528	50	22071	1.995	ppbv	98
5) Vinyl Chloride	1.573	62	20286	1.932	ppbv	98
6) Bromomethane	1.661	94	9696	2.055	ppbv	97
7) Chloroethane	1.696	64	9205	2.101	ppbv	96
8) Dichlorotetrafluoroethane	1.547	85	51709	2.083	ppbv	98
9) Propene	1.479	41	25044	2.119	ppbv	93
10) Heptane	4.952	43	59306	2.038	ppbv	98
11) Trichlorofluoromethane	1.871	101	63315	2.034	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.130	101	50042	2.041	ppbv	100
13) Ethanol	1.716	45	1472	2.168	ppbv #	58
14) Bromoethene	1.774	108	18006	2.048	ppbv	98
15) Acetone	1.829	43	52918	2.103	ppbv	95
16) 1,3-Butadiene	1.602	39	21968	2.117	ppbv	96
17) tert-Butyl alcohol	2.030	59	51640	1.940	ppbv	98
18) 1,1-Dichloroethene	2.023	96	21932	2.040	ppbv	98
19) Isopropyl Alcohol	1.877	45	23979	1.989	ppbv #	80
20) Methylene Chloride	2.052	84	20728	1.887	ppbv	95
21) Allyl Chloride	2.085	41	33088	2.052	ppbv	98
22) trans-1,2-Dichloroethene	2.331	96	24306	1.992	ppbv	95
23) Vinyl Acetate	2.450	43	11081	2.074	ppbv #	92
24) 1,1-Dichloroethane	2.395	63	46548	2.012	ppbv	99
25) Ethyl Acetate	2.819	43	130995	2.075	ppbv	99
26) Hexane	2.813	57	50777	2.091	ppbv	96
27) Carbon Disulfide	2.140	76	56650	1.976	ppbv #	70
28) Methyl tert-Butyl Ether	2.428	73	31189	2.052	ppbv	99
29) Chloroform	2.832	83	83628	2.020	ppbv	99
30) Cyclohexane	3.836	84	41150	2.062	ppbv	96
31) cis-1,2-Dichloroethene	2.706	61	51609	2.067	ppbv	95
32) 1,1,1-Trichloroethane	3.350	97	81078	2.015	ppbv	99
34) 2-Butanone	2.544	43	75222	2.049	ppbv	100
35) Carbon Tetrachloride	3.739	117	83331	1.872	ppbv	98
36) Benzene	3.635	78	105052	2.032	ppbv	98
37) 1,2-Dichloroethane	3.201	62	61665	2.014	ppbv	100
38) Trichloroethene	4.519	130	43500	2.020	ppbv	98
39) 1,2-Dichloropropane	4.289	63	38286	1.998	ppbv	97
40) 1,4-Dioxane	4.564	88	16471m	2.002	ppbv	
41) Tetrahydrofuran	3.043	42	39501	2.082	ppbv	98
42) Bromodichloromethane	4.464	83	85622	1.981	ppbv	94
43) Methyl Methacrylate	4.852	69	33288	1.940	ppbv	96
44) 2,2,4-Trimethylpentane	4.609	57	171998	2.074	ppbv	99
45) t-1,3-Dichloropropene	6.383	75	30227	1.944	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043033.D
 Acq On : 02 Oct 2025 10:57
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/03/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:02:17 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Oct 03 07:59:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.567	75	42818	1.938	ppbv	90
47) 1,1,2-Trichloroethane	6.548	97	42719	2.004	ppbv	93
48) Dibromochloromethane	7.357	129	69735	2.005	ppbv	100
49) Bromoform	9.461	173	58655	2.019	ppbv	99
50) 4-Methyl-2-Pentanone	5.719	43	88751	1.982	ppbv	100
51) 2-Hexanone	7.412	43	62520	1.969	ppbv #	96
52) Tetrachloroethene	8.202	164	38059	1.976	ppbv	97
53) Toluene	6.907	91	104582	2.004	ppbv	100
54) 1,2-Dibromoethane	7.629	107	50973	2.018	ppbv	95
56) 1,1,1,2-Tetrachloroethane	8.901	131	54924	2.013	ppbv	95
57) Chlorobenzene	8.901	112	97190	2.054	ppbv	99
58) Ethyl Benzene	9.335	91	143037	2.063	ppbv	97
59) m/p-Xylene	9.529	91	250936m	4.289	ppbv	
60) o-Xylene	9.946	91	124006	2.128	ppbv	98
61) Styrene	9.849	104	46782	1.917	ppbv	98
62) Isopropylbenzene	10.523	105	186358	2.084	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.943	83	84916	2.022	ppbv	99
64) n-propylbenzene	10.972	120	48346	2.113	ppbv	90
65) tert-Butylbenzene	11.516	119	172968	2.173	ppbv	99
66) Benzyl Chloride	11.597	91	5886	1.908	ppbv	99
67) sec-Butylbenzene	11.752	105	245463	2.210	ppbv	99
69) p-Isopropyltoluene	11.911	119	200038	2.214	ppbv	99
70) n-Butylbenzene	12.267	91	192286	2.210	ppbv	98
71) 2-Chlorotoluene	10.885	91	136276	2.105	ppbv	99
72) 4-Ethyltoluene	11.108	105	147421	2.126	ppbv	99
73) 1,3,5-Trimethylbenzene	11.186	105	127419	2.124	ppbv	94
74) 1,2,4-Trimethylbenzene	11.519	105	150033	2.216	ppbv	98
75) 1,3-Dichlorobenzene	11.591	146	92272	2.095	ppbv	99
76) 1,4-Dichlorobenzene	11.655	146	88602	2.110	ppbv	99
77) 1,2-Dichlorobenzene	11.930	146	92779	2.115	ppbv	99
78) Hexachloro-1,3-Butadiene	13.866	225	65408	2.140	ppbv	99
79) Naphthalene	13.500	128	99793	2.006	ppbv	97
80) Naphthalene,2-methyl-	14.445	142	24913	1.530	ppbv	92
81) 1,2,4-Trichlorobenzene	13.426	180	56312	2.158	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043034.D
 Acq On : 02 Oct 2025 11:29
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/03/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:03:15 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Oct 03 07:59:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.774	49	174129	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.936	114	522004	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.859	117	431704	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.357 95 313212 10.071 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.492	85	31391	1.070	ppbv	98
3) Chlorodifluoromethane	1.470	51	29622	1.043	ppbv	97
4) Chloromethane	1.528	50	11539	1.076	ppbv	91
5) Vinyl Chloride	1.573	62	9984	0.981	ppbv	98
6) Bromomethane	1.661	94	4926	1.078	ppbv	97
7) Chloroethane	1.699	64	4474	1.054	ppbv	90
8) Dichlorotetrafluoroethane	1.547	85	24892	1.035	ppbv	98
9) Propene	1.479	41	12033	1.051	ppbv	97
10) Heptane	4.952	43	27247	0.966	ppbv	98
11) Trichlorofluoromethane	1.871	101	32491	1.077	ppbv	95
12) 1,1,2-Trichlorotrifluo...	2.130	101	24551	1.034	ppbv	99
13) Ethanol	1.716	45	849	0.777	ppbv #	48
14) Bromoethene	1.774	108	8693	1.021	ppbv #	94
15) Acetone	1.829	43	27190	1.115	ppbv	97
16) 1,3-Butadiene	1.602	39	10229	1.018	ppbv	100
17) tert-Butyl alcohol	2.033	59	27609	1.071	ppbv	96
18) 1,1-Dichloroethene	2.026	96	10892	1.046	ppbv	94
19) Isopropyl Alcohol	1.877	45	12661	1.084	ppbv	86
20) Methylene Chloride	2.052	84	11906	1.119	ppbv #	92
21) Allyl Chloride	2.088	41	15455	0.989	ppbv	96
22) trans-1,2-Dichloroethene	2.331	96	12411	1.050	ppbv	99
23) Vinyl Acetate	2.450	43	5498	1.062	ppbv #	77
24) 1,1-Dichloroethane	2.399	63	23773	1.061	ppbv	97
25) Ethyl Acetate	2.823	43	63776	1.043	ppbv	100
26) Hexane	2.813	57	22126	0.940	ppbv	80
27) Carbon Disulfide	2.140	76	28818	1.037	ppbv #	62
28) Methyl tert-Butyl Ether	2.428	73	15176	1.031	ppbv	99
29) Chloroform	2.832	83	41697	1.039	ppbv	99
30) Cyclohexane	3.839	84	18605	0.962	ppbv	97
31) cis-1,2-Dichloroethene	2.709	61	24206	1.001	ppbv	97
32) 1,1,1-Trichloroethane	3.350	97	38581	0.990	ppbv	95
34) 2-Butanone	2.544	43	35775	1.010	ppbv	100
35) Carbon Tetrachloride	3.742	117	41999	0.978	ppbv	94
36) Benzene	3.638	78	51343	1.029	ppbv	95
37) 1,2-Dichloroethane	3.201	62	30102	1.019	ppbv	99
38) Trichloroethene	4.519	130	22132	1.065	ppbv	90
39) 1,2-Dichloropropane	4.289	63	18915	1.023	ppbv	90
40) 1,4-Dioxane	4.564	88	7860m	0.990	ppbv	
41) Tetrahydrofuran	3.043	42	17630	0.963	ppbv	98
42) Bromodichloromethane	4.460	83	42012	1.007	ppbv	89
43) Methyl Methacrylate	4.849	69	14911	0.901	ppbv	95
44) 2,2,4-Trimethylpentane	4.609	57	77988	0.975	ppbv	96
45) t-1,3-Dichloropropene	6.380	75	13974	0.932	ppbv	92

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043034.D
 Acq On : 02 Oct 2025 11:29
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/03/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:03:15 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Oct 03 07:59:23 2025


Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.570	75	20141m	0.945	ppbv	
47) 1,1,2-Trichloroethane	6.548	97	21164	1.029	ppbv #	85
48) Dibromochloromethane	7.357	129	33636	1.002	ppbv	100
49) Bromoform	9.464	173	26918	0.960	ppbv	98
50) 4-Methyl-2-Pentanone	5.729	43	39803	0.922	ppbv	98
51) 2-Hexanone	7.415	43	26904	0.878	ppbv #	95
52) Tetrachloroethene	8.202	164	17908	0.964	ppbv	97
53) Toluene	6.904	91	45974	0.913	ppbv	98
54) 1,2-Dibromoethane	7.626	107	24648	1.011	ppbv	85
56) 1,1,1,2-Tetrachloroethane	8.904	131	26884	1.022	ppbv	95
57) Chlorobenzene	8.898	112	47296	1.036	ppbv	97
58) Ethyl Benzene	9.335	91	59544	0.890	ppbv	97
59) m/p-Xylene	9.532	91	106322m	1.884	ppbv	
60) o-Xylene	9.943	91	52489	0.934	ppbv	100
61) Styrene	9.849	104	19376	0.823	ppbv	97
62) Isopropylbenzene	10.523	105	82825	0.960	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.943	83	40981	1.012	ppbv	99
64) n-propylbenzene	10.972	120	21190	0.960	ppbv	96
65) tert-Butylbenzene	11.516	119	74677	0.973	ppbv	99
66) Benzyl Chloride	11.597	91	2866	0.963	ppbv	98
67) sec-Butylbenzene	11.752	105	106876	0.998	ppbv	100
69) p-Isopropyltoluene	11.911	119	81108	0.931	ppbv	99
70) n-Butylbenzene	12.267	91	76664	0.914	ppbv	97
71) 2-Chlorotoluene	10.885	91	58174	0.932	ppbv	100
72) 4-Ethyltoluene	11.108	105	61540	0.920	ppbv	98
73) 1,3,5-Trimethylbenzene	11.189	105	54851	0.948	ppbv	94
74) 1,2,4-Trimethylbenzene	11.519	105	64405	0.986	ppbv	98
75) 1,3-Dichlorobenzene	11.591	146	42580	1.003	ppbv	98
76) 1,4-Dichlorobenzene	11.655	146	40098	0.990	ppbv	100
77) 1,2-Dichlorobenzene	11.930	146	44771	1.058	ppbv	99
78) Hexachloro-1,3-Butadiene	13.866	225	30422	1.032	ppbv	99
79) Naphthalene	13.497	128	32191	0.831	ppbv	98
80) Naphthalene,2-methyl-	14.442	142	7210	0.923	ppbv	99
81) 1,2,4-Trichlorobenzene	13.426	180	20534	0.816	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations

Quant Time: Oct 03 08:03:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Oct 03 07:59:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043036.D
 Acq On : 02 Oct 2025 12:35
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/03/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:04:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Oct 03 07:59:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.771	49	175730	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.936	114	510906	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.859	117	418551	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.358	95	289539	9.603	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.000%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.573	62	1208	0.118	ppbv	96
32) 1,1,1-Trichloroethane	3.350	97	3686	0.094	ppbv	96
35) Carbon Tetrachloride	3.739	117	4230	0.101	ppbv #	91
38) Trichloroethene	4.522	130	1971m	0.097	ppbv	
52) Tetrachloroethene	8.199	164	1674	0.092	ppbv #	75
54) 1,2-Dibromoethane	7.626	107	2275m	0.095	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.946	83	3701	0.094	ppbv	93
79) Naphthalene	13.500	128	1393	0.261	ppbv #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
Data File : VL043036.D
Acq On : 02 Oct 2025 12:35
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/03/2025
Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:04:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Oct 03 07:59:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043037.D
 Acq On : 02 Oct 2025 13:07
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/03/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:05:09 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Oct 03 07:59:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.774	49	172214	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.939	114	493591	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.862	117	408184	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.361	95	280877	9.552	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.500%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.573	62	296	0.029	ppbv #	78
32) 1,1,1-Trichloroethane	3.353	97	1314m	0.034	ppbv	
35) Carbon Tetrachloride	3.748	117	1490m	0.037	ppbv	
38) Trichloroethene	4.525	130	525m	0.027	ppbv	
52) Tetrachloroethene	8.205	164	628m	0.036	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.947	83	1322	0.035	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
Data File : VL043037.D
Acq On : 02 Oct 2025 13:07
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Oct 03 08:05:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Oct 03 07:59:23 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 10/03/2025
Supervised By : Mahesh Dadoda 10/10/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043038.D
 Acq On : 02 Oct 2025 13:41
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/03/2025
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:06:14 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Fri Oct 03 07:59:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.774	49	174781	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.939	114	521075	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.862	117	446323	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.361 95 331763 10.318 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 103.200%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.493	85	417041	14.157	ppbv	100
3) Chlorodifluoromethane	1.470	51	396783	13.924	ppbv	100
4) Chloromethane	1.528	50	152053	14.131	ppbv	100
5) Vinyl Chloride	1.573	62	143710	14.073	ppbv	98
6) Bromomethane	1.661	94	66123	14.410	ppbv	97
7) Chloroethane	1.700	64	58804	13.801	ppbv	97
8) Dichlorotetrafluoroethane	1.551	85	342721	14.198	ppbv	100
9) Propene	1.480	41	159689	13.898	ppbv	97
10) Heptane	4.956	43	465543	16.450	ppbv	99
11) Trichlorofluoromethane	1.871	101	430473	14.220	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.130	101	337263	14.145	ppbv	97
13) Ethanol	1.713	45	6726	14.884	ppbv	94
14) Bromoethene	1.774	108	123040	14.394	ppbv	99
15) Acetone	1.826	43	302724	12.370	ppbv	100
16) 1,3-Butadiene	1.603	39	142349	14.108	ppbv	99
17) tert-Butyl alcohol	2.030	59	352251	13.608	ppbv	100
18) 1,1-Dichloroethene	2.027	96	150012	14.353	ppbv	98
19) Isopropyl Alcohol	1.878	45	154373	13.166	ppbv	98
20) Methylene Chloride	2.053	84	130116	12.181	ppbv	96
21) Allyl Chloride	2.088	41	223371	14.247	ppbv	99
22) trans-1,2-Dichloroethene	2.331	96	169834	14.313	ppbv	96
23) Vinyl Acetate	2.454	43	60446	11.634	ppbv #	96
24) 1,1-Dichloroethane	2.399	63	320234	14.237	ppbv	100
25) Ethyl Acetate	2.820	43	909988	14.822	ppbv	100
26) Hexane	2.813	57	370003	15.667	ppbv	98
27) Carbon Disulfide	2.143	76	404552	14.510	ppbv	99
28) Methyl tert-Butyl Ether	2.428	73	213111	14.420	ppbv	99
29) Chloroform	2.836	83	576414	14.316	ppbv	99
30) Cyclohexane	3.836	84	314603	16.212	ppbv	99
31) cis-1,2-Dichloroethene	2.710	61	363768	14.984	ppbv	99
32) 1,1,1-Trichloroethane	3.354	97	573214	14.648	ppbv	100
34) 2-Butanone	2.545	43	511000	14.453	ppbv	100
35) Carbon Tetrachloride	3.745	117	604277	14.095	ppbv	97
36) Benzene	3.638	78	748168	15.028	ppbv	98
37) 1,2-Dichloroethane	3.205	62	431043	14.619	ppbv	100
38) Trichloroethene	4.525	130	317772	15.319	ppbv	99
39) 1,2-Dichloropropane	4.289	63	269729	14.617	ppbv	98
40) 1,4-Dioxane	4.554	88	120237	15.173	ppbv #	95
41) Tetrahydrofuran	3.036	42	295205	16.155	ppbv	100
42) Bromodichloromethane	4.464	83	619944	14.890	ppbv	96
43) Methyl Methacrylate	4.852	69	279674	16.923	ppbv	99
44) 2,2,4-Trimethylpentane	4.616	57	1251093	15.661	ppbv	99
45) t-1,3-Dichloropropene	6.386	75	251549	16.802	ppbv	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043038.D
 Acq On : 02 Oct 2025 13:41
 Operator : SY/MD
 Sample : VSTDICC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/03/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:06:14 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Oct 03 07:59:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.571	75	351325	16.515	ppbv	100
47) 1,1,2-Trichloroethane	6.555	97	296142	14.428	ppbv	98
48) Dibromochloromethane	7.364	129	512219	15.290	ppbv	100
49) Bromoform	9.468	173	449641	16.072	ppbv	100
50) 4-Methyl-2-Pentanone	5.713	43	723710	16.786	ppbv	100
51) 2-Hexanone	7.409	43	562773	18.407	ppbv	100
52) Tetrachloroethene	8.205	164	270386	14.575	ppbv	98
53) Toluene	6.911	91	858713	17.085	ppbv	100
54) 1,2-Dibromoethane	7.629	107	378193	15.546	ppbv	97
56) 1,1,1,2-Tetrachloroethane	8.908	131	394995	14.519	ppbv	100
57) Chlorobenzene	8.904	112	670949	14.220	ppbv	99
58) Ethyl Benzene	9.338	91	1190670	17.220	ppbv	99
59) m/p-Xylene	9.536	91	1936045m	33.184	ppbv	
60) o-Xylene	9.947	91	969480	16.688	ppbv	99
61) Styrene	9.853	104	462847	19.017	ppbv	100
62) Isopropylbenzene	10.526	105	1455778	16.324	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.947	83	593157	14.167	ppbv	100
64) n-propylbenzene	10.976	120	387627	16.991	ppbv	95
65) tert-Butylbenzene	11.520	119	1283802	16.175	ppbv	99
66) Benzyl Chloride	11.604	91	50421	16.394	ppbv	99
67) sec-Butylbenzene	11.756	105	1758282	15.875	ppbv	98
69) p-Isopropyltoluene	11.915	119	1493674	16.578	ppbv	99
70) n-Butylbenzene	12.271	91	1480662	17.067	ppbv	98
71) 2-Chlorotoluene	10.889	91	1085558	16.819	ppbv	100
72) 4-Ethyltoluene	11.112	105	1190551	17.218	ppbv	99
73) 1,3,5-Trimethylbenzene	11.190	105	1001678	16.745	ppbv	100
74) 1,2,4-Trimethylbenzene	11.523	105	1091026	16.160	ppbv	95
75) 1,3-Dichlorobenzene	11.594	146	665205	15.149	ppbv	100
76) 1,4-Dichlorobenzene	11.659	146	656125	15.671	ppbv	99
77) 1,2-Dichlorobenzene	11.934	146	632662	14.465	ppbv	99
78) Hexachloro-1,3-Butadiene	13.870	225	435408	14.288	ppbv	100
79) Naphthalene	13.501	128	887779	14.984	ppbv	99
80) Naphthalene,2-methyl-	14.446	142	421038	15.372	ppbv	99
81) 1,2,4-Trichlorobenzene	13.429	180	489642	18.818	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043039.D
 Acq On : 02 Oct 2025 14:14
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL100225

Quant Time: Oct 03 08:24:32 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Oct 03 08:22:04 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 10/03/2025
 Supervised By : Mahesh Dadoda 10/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.777	49	178544	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.939	114	537412	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.866	117	454012	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.364 95 333843 10.207 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.100%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	287085	9.540	ppbv	100
3) Chlorodifluoromethane	1.473	51	275618	9.468	ppbv	99
4) Chloromethane	1.528	50	104427	9.501	ppbv	98
5) Vinyl Chloride	1.577	62	98850	9.476	ppbv	97
6) Bromomethane	1.664	94	46326	9.883	ppbv	97
7) Chloroethane	1.700	64	40636	9.336	ppbv	98
8) Dichlorotetrafluoroethane	1.551	85	235901	9.567	ppbv	99
9) Propene	1.480	41	115956	9.879	ppbv	99
10) Heptane	4.956	43	311038	10.779	ppbv	100
11) Trichlorofluoromethane	1.871	101	291376	9.422	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.133	101	228314	9.374	ppbv	97
13) Ethanol	1.716	45	4643	9.646	ppbv	98
14) Bromoethene	1.777	108	85349	9.774	ppbv	99
15) Acetone	1.829	43	204381	8.176	ppbv	100
16) 1,3-Butadiene	1.606	39	95179	9.234	ppbv	99
17) tert-Butyl alcohol	2.033	59	249385	9.431	ppbv	99
18) 1,1-Dichloroethene	2.026	96	101929	9.547	ppbv	95
19) Isopropyl Alcohol	1.878	45	109403	9.134	ppbv	99
20) Methylene Chloride	2.056	84	89214	8.176	ppbv	97
21) Allyl Chloride	2.088	41	151234	9.443	ppbv	99
22) trans-1,2-Dichloroethene	2.331	96	115000	9.487	ppbv	97
23) Vinyl Acetate	2.454	43	45886	8.646	ppbv	97
24) 1,1-Dichloroethane	2.399	63	217319	9.458	ppbv	98
25) Ethyl Acetate	2.823	43	628433	10.021	ppbv	100
26) Hexane	2.816	57	253815	10.521	ppbv	98
27) Carbon Disulfide	2.143	76	273348	9.598	ppbv	99
28) Methyl tert-Butyl Ether	2.431	73	147450	9.767	ppbv	99
29) Chloroform	2.836	83	388247	9.439	ppbv	100
30) Cyclohexane	3.839	84	216007	10.896	ppbv	98
31) cis-1,2-Dichloroethene	2.709	61	249555	10.063	ppbv	97
32) 1,1,1-Trichloroethane	3.354	97	387043	9.682	ppbv	100
34) 2-Butanone	2.544	43	346463	9.501	ppbv	100
35) Carbon Tetrachloride	3.745	117	399311	9.031	ppbv	96
36) Benzene	3.642	78	514339	10.017	ppbv	98
37) 1,2-Dichloroethane	3.205	62	289339	9.515	ppbv	100
38) Trichloroethene	4.528	130	213727	9.992	ppbv	98
39) 1,2-Dichloropropane	4.292	63	183886	9.662	ppbv	97
40) 1,4-Dioxane	4.558	88	80249	9.831	ppbv #	98
41) Tetrahydrofuran	3.040	42	201471	10.690	ppbv	100
42) Bromodichloromethane	4.467	83	413457	9.629	ppbv	98
43) Methyl Methacrylate	4.855	69	189993	11.147	ppbv	99
44) 2,2,4-Trimethylpentane	4.619	57	847766	10.290	ppbv	99
45) t-1,3-Dichloropropene	6.389	75	164370	10.538	ppbv	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043039.D
 Acq On : 02 Oct 2025 14:14
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL100225

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/03/2025
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:24:32 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Oct 03 08:22:04 2025


Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.574	75	232895	10.597	ppbv	98
47) 1,1,2-Trichloroethane	6.558	97	196496	9.282	ppbv	97
48) Dibromochloromethane	7.367	129	336813	9.749	ppbv	99
49) Bromoform	9.468	173	275751	9.557	ppbv	100
50) 4-Methyl-2-Pentanone	5.720	43	484498	10.896	ppbv	100
51) 2-Hexanone	7.412	43	355624	11.236	ppbv	99
52) Tetrachloroethene	8.208	164	182062	9.576	ppbv	99
53) Toluene	6.911	91	572257	11.040	ppbv	100
54) 1,2-Dibromoethane	7.632	107	247276	9.855	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.908	131	259863	9.390	ppbv	99
57) Chlorobenzene	8.904	112	447850	9.331	ppbv	100
58) Ethyl Benzene	9.338	91	775232	11.022	ppbv	98
59) m/p-Xylene	9.535	91	1265386m	21.330	ppbv	
60) o-Xylene	9.950	91	628486	10.635	ppbv	99
61) Styrene	9.856	104	284955	11.510	ppbv	100
62) Isopropylbenzene	10.526	105	939236	10.353	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.950	83	360837	8.511	ppbv	99
64) n-propylbenzene	10.976	120	240769	10.375	ppbv	98
65) tert-Butylbenzene	11.520	119	814322	10.086	ppbv	100
66) Benzyl Chloride	11.604	91	26124	8.466	ppbv	99
67) sec-Butylbenzene	11.756	105	1122743	9.965	ppbv	99
69) p-Isopropyltoluene	11.914	119	937600	10.230	ppbv	100
70) n-Butylbenzene	12.270	91	913299	10.349	ppbv	100
71) 2-Chlorotoluene	10.888	91	673074	10.251	ppbv	100
72) 4-Ethyltoluene	11.112	105	744145	10.579	ppbv	100
73) 1,3,5-Trimethylbenzene	11.193	105	621377	10.212	ppbv	100
74) 1,2,4-Trimethylbenzene	11.526	105	684740	9.971	ppbv	93
75) 1,3-Dichlorobenzene	11.594	146	409217	9.161	ppbv	98
76) 1,4-Dichlorobenzene	11.659	146	398965	9.368	ppbv	100
77) 1,2-Dichlorobenzene	11.934	146	384233	8.636	ppbv	99
78) Hexachloro-1,3-Butadiene	13.869	225	289545	9.340	ppbv	99
79) Naphthalene	13.500	128	552627	9.523	ppbv	99
80) Naphthalene,2-methyl-	14.445	142	223063	8.323	ppbv	98
81) 1,2,4-Trichlorobenzene	13.429	180	317372	11.990	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations APPROVED

Quant Time: Oct 03 08:24:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Oct 03 08:22:04 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043039.D
 Acq On : 02 Oct 2025 14:14
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL100225

Quant Time: Oct 03 08:24:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Oct 03 08:22:04 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	1.685	1.608	4.6	103	0.00
3	Chlorodifluoromethane	1.630	1.544	5.3	103	0.00
4	Chloromethane	0.616	0.585	5.0	103	0.00
5 T	Vinyl Chloride	0.584	0.554	5.1	104	0.00
6 T	Bromomethane	0.263	0.259	1.5	106	0.00
7	Chloroethane	0.244	0.228	6.6	104	0.00
8 T	Dichlorotetrafluoroethane	1.381	1.321	4.3	103	0.00
9 T	Propene	0.657	0.649	1.2	104	0.00
10 T	Heptane	1.616	1.742	-7.8	101	0.00
11 T	Trichlorofluoromethane	1.732	1.632	5.8	102	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.364	1.279	6.2	101	0.00
13	Ethanol	0.045	0.026#	42.2#	97	0.00
14 T	Bromoethene	0.489	0.478	2.2	104	0.00
15 T	Acetone	1.400	1.145	18.2	102	0.00
16 T	1,3-Butadiene	0.577	0.533	7.6	101	0.00
17	tert-Butyl alcohol	1.481	1.397	5.7	103	0.00
18 T	1,1-Dichloroethene	0.598	0.571	4.5	101	0.00
19 T	Isopropyl Alcohol	0.671	0.613	8.6	103	0.00
20 T	Methylene Chloride	0.611	0.500	18.2	104	0.00
21 T	Allyl Chloride	0.897	0.847	5.6	102	0.00
22 T	trans-1,2-Dichloroethene	0.679	0.644	5.2	103	0.00
23 T	Vinyl Acetate	0.297	0.257	13.5	107	0.00
24 T	1,1-Dichloroethane	1.287	1.217	5.4	103	0.00
25 T	Ethyl Acetate	3.513	3.520	-0.2	102	0.00
26 T	Hexane	1.351	1.422	-5.3	103	0.00
27 T	Carbon Disulfide	1.595	1.531	4.0	102	0.00
28 T	Methyl tert-Butyl Ether	0.846	0.826	2.4	102	0.00
29 T	Chloroform	2.304	2.175	5.6	100	0.00
30 T	Cyclohexane	1.110	1.210	-9.0	103	0.00
31 T	cis-1,2-Dichloroethene	1.389	1.398	-0.6	101	0.00
32 T	1,1,1-Trichloroethane	2.239	2.168	3.2	102	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
34 T	2-Butanone	0.679	0.645	5.0	101	0.00
35 T	Carbon Tetrachloride	0.823	0.743	9.7	100	0.00
36 T	Benzene	0.955	0.957	-0.2	102	0.00
37 T	1,2-Dichloroethane	0.566	0.538	4.9	101	0.00
38 T	Trichloroethene	0.398	0.398	0.0	100	0.00
39 T	1,2-Dichloropropane	0.354	0.342	3.4	100	0.00
40 T	1,4-Dioxane	0.152	0.149	2.0	98	0.00
41 T	Tetrahydrofuran	0.351	0.375	-6.8	102	0.00
42 T	Bromodichloromethane	0.799	0.769	3.8	100	0.00
43	Methyl Methacrylate	0.317	0.354	-11.7	103	0.00
44 T	2,2,4-Trimethylpentane	1.533	1.577	-2.9	102	0.00
45 T	t-1,3-Dichloropropene	0.290	0.306	-5.5	100	0.00
46 T	cis-1,3-Dichloropropene	0.409	0.433	-5.9	101	0.00
47 T	1,1,2-Trichloroethane	0.394	0.366	7.1	100	0.00
48 T	Dibromochloromethane	0.643	0.627	2.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043039.D
 Acq On : 02 Oct 2025 14:14
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL100225

Quant Time: Oct 03 08:24:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Oct 03 08:22:04 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.537	0.513	4.5	94	0.00
50 T	4-Methyl-2-Pentanone	0.827	0.902	-9.1	102	0.00
51 T	2-Hexanone	0.589	0.662	-12.4	98	0.00
52 T	Tetrachloroethene	0.354	0.339	4.2	101	0.00
53 T	Toluene	0.965	1.065	-10.4	101	0.00
54 T	1,2-Dibromoethane	0.467	0.460	1.5	98	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
56	1,1,1,2-Tetrachloroethane	0.610	0.572	6.2	98	0.00
57 T	Chlorobenzene	1.057	0.986	6.7	98	0.00
58 T	Ethyl Benzene	1.549	1.708	-10.3	98	0.00
59 T	m/p-Xylene	1.307	1.394	-6.7	98	0.00
60 T	o-Xylene	1.302	1.384	-6.3	97	0.00
61 T	Styrene	0.545	0.628	-15.2	95	0.00
62	Isopropylbenzene	1.998	2.069	-3.6	98	0.00
63 T	1,1,2,2-Tetrachloroethane	0.934	0.795	14.9	92	0.00
64	n-propylbenzene	0.511	0.530	-3.7	96	0.00
65	tert-Butylbenzene	1.778	1.794	-0.9	94	0.00
66 T	Benzyl Chloride	0.068	0.058	14.7	85	0.00
67	sec-Butylbenzene	2.482	2.473	0.4	94	0.00
68 S	1-Bromo-4-Fluorobenzene	0.720	0.735	-2.1	101	0.00
69	p-Isopropyltoluene	2.019	2.065	-2.3	93	0.00
70	n-Butylbenzene	1.944	2.012	-3.5	92	0.00
71	2-Chlorotoluene	1.446	1.483	-2.6	95	0.00
72 T	4-Ethyltoluene	1.549	1.639	-5.8	95	0.00
73 T	1,3,5-Trimethylbenzene	1.340	1.369	-2.2	94	0.00
74 T	1,2,4-Trimethylbenzene	1.513	1.508	0.3	94	0.00
75 T	1,3-Dichlorobenzene	0.984	0.901	8.4	93	0.00
76 T	1,4-Dichlorobenzene	0.938	0.879	6.3	92	0.00
77 T	1,2-Dichlorobenzene	0.980	0.846	13.7	92	0.00
78 T	Hexachloro-1,3-Butadiene	0.683	0.638	6.6	100	0.00
79 T	Naphthalene	0.903	1.217	-34.8#	96	0.00
80 T	Naphthalene,2-methyl-	0.457	0.491	-7.4	88	0.00
81 T	1,2,4-Trichlorobenzene	0.583	0.699	-19.9	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043039.D
 Acq On : 02 Oct 2025 14:14
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL100225

Quant Time: Oct 03 08:24:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Oct 03 08:22:04 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	10.000	9.540	4.6	103	0.00
3	Chlorodifluoromethane	10.000	9.468	5.3	103	0.00
4	Chloromethane	10.000	9.501	5.0	103	0.00
5 T	Vinyl Chloride	10.000	9.476	5.2	104	0.00
6 T	Bromomethane	10.000	9.883	1.2	106	0.00
7	Chloroethane	10.000	9.336	6.6	104	0.00
8 T	Dichlorotetrafluoroethane	10.000	9.567	4.3	103	0.00
9 T	Propene	10.000	9.879	1.2	104	0.00
10 T	Heptane	10.000	10.779	-7.8	101	0.00
11 T	Trichlorofluoromethane	10.000	9.422	5.8	102	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	9.374	6.3	101	0.00
13	Ethanol	10.000	9.646	3.5	97	0.00
14 T	Bromoethene	10.000	9.774	2.3	104	0.00
15 T	Acetone	10.000	8.176	18.2	102	0.00
16 T	1,3-Butadiene	10.000	9.234	7.7	101	0.00
17	tert-Butyl alcohol	10.000	9.431	5.7	103	0.00
18 T	1,1-Dichloroethene	10.000	9.547	4.5	101	0.00
19 T	Isopropyl Alcohol	10.000	9.134	8.7	103	0.00
20 T	Methylene Chloride	10.000	8.176	18.2	104	0.00
21 T	Allyl Chloride	10.000	9.443	5.6	102	0.00
22 T	trans-1,2-Dichloroethene	10.000	9.487	5.1	103	0.00
23 T	Vinyl Acetate	10.000	8.646	13.5	107	0.00
24 T	1,1-Dichloroethane	10.000	9.458	5.4	103	0.00
25 T	Ethyl Acetate	10.000	10.021	-0.2	102	0.00
26 T	Hexane	10.000	10.521	-5.2	103	0.00
27 T	Carbon Disulfide	10.000	9.598	4.0	102	0.00
28 T	Methyl tert-Butyl Ether	10.000	9.767	2.3	102	0.00
29 T	Chloroform	10.000	9.439	5.6	100	0.00
30 T	Cyclohexane	10.000	10.896	-9.0	103	0.00
31 T	cis-1,2-Dichloroethene	10.000	10.063	-0.6	101	0.00
32 T	1,1,1-Trichloroethane	10.000	9.682	3.2	102	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	103	0.00
34 T	2-Butanone	10.000	9.501	5.0	101	0.00
35 T	Carbon Tetrachloride	10.000	9.031	9.7	100	0.00
36 T	Benzene	10.000	10.017	-0.2	102	0.00
37 T	1,2-Dichloroethane	10.000	9.515	4.8	101	0.00
38 T	Trichloroethene	10.000	9.992	0.1	100	0.00
39 T	1,2-Dichloropropane	10.000	9.662	3.4	100	0.00
40 T	1,4-Dioxane	10.000	9.831	1.7	98	0.00
41 T	Tetrahydrofuran	10.000	10.690	-6.9	102	0.00
42 T	Bromodichloromethane	10.000	9.629	3.7	100	0.00
43	Methyl Methacrylate	10.000	11.147	-11.5	103	0.00
44 T	2,2,4-Trimethylpentane	10.000	10.290	-2.9	102	0.00
45 T	t-1,3-Dichloropropene	10.000	10.538	-5.4	100	0.00
46 T	cis-1,3-Dichloropropene	10.000	10.597	-6.0	101	0.00
47 T	1,1,2-Trichloroethane	10.000	9.282	7.2	100	0.00
48 T	Dibromochloromethane	10.000	9.749	2.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043039.D
 Acq On : 02 Oct 2025 14:14
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL100225

Quant Time: Oct 03 08:24:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Oct 03 08:22:04 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	Bromoform	10.000	9.557	4.4	94	0.00
50 T	4-Methyl-2-Pentanone	10.000	10.896	-9.0	102	0.00
51 T	2-Hexanone	10.000	11.236	-12.4	98	0.00
52 T	Tetrachloroethene	10.000	9.576	4.2	101	0.00
53 T	Toluene	10.000	11.040	-10.4	101	0.00
54 T	1,2-Dibromoethane	10.000	9.855	1.4	98	0.00
55 I	Chlorobenzene-d5	10.000	10.000	0.0	101	0.00
56	1,1,1,2-Tetrachloroethane	10.000	9.390	6.1	98	0.00
57 T	Chlorobenzene	10.000	9.331	6.7	98	0.00
58 T	Ethyl Benzene	10.000	11.022	-10.2	98	0.00
59 T	m/p-Xylene	20.000	21.330	-6.6	98	0.00
60 T	o-Xylene	10.000	10.635	-6.3	97	0.00
61 T	Styrene	10.000	11.510	-15.1	95	0.00
62	Isopropylbenzene	10.000	10.353	-3.5	98	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	8.511	14.9	92	0.00
64	n-propylbenzene	10.000	10.375	-3.8	96	0.00
65	tert-Butylbenzene	10.000	10.086	-0.9	94	0.00
66 T	Benzyl Chloride	10.000	8.466	15.3	85	0.00
67	sec-Butylbenzene	10.000	9.965	0.4	94	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	10.207	-2.1	101	0.00
69	p-Isopropyltoluene	10.000	10.230	-2.3	93	0.00
70	n-Butylbenzene	10.000	10.349	-3.5	92	0.00
71	2-Chlorotoluene	10.000	10.251	-2.5	95	0.00
72 T	4-Ethyltoluene	10.000	10.579	-5.8	95	0.00
73 T	1,3,5-Trimethylbenzene	10.000	10.212	-2.1	94	0.00
74 T	1,2,4-Trimethylbenzene	10.000	9.971	0.3	94	0.00
75 T	1,3-Dichlorobenzene	10.000	9.161	8.4	93	0.00
76 T	1,4-Dichlorobenzene	10.000	9.368	6.3	92	0.00
77 T	1,2-Dichlorobenzene	10.000	8.636	13.6	92	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	9.340	6.6	100	0.00
79 T	Naphthalene	10.000	9.523	4.8	96	0.00
80 T	Naphthalene,2-methyl-	10.000	8.323	16.8	88	0.00
81 T	1,2,4-Trichlorobenzene	10.000	11.990	-19.9	102	0.00

(#) = Out of Range



SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
Data File : VL043043.D
Acq On : 06 Oct 2025 07:57
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri Oct 03 08:23:34 2025



AutoFind: Scans 3174, 3175, 3176; Background Corrected with Scan 3159

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	23.7	56904	PASS
75	95	30	66	56.6	136077	PASS
95	95	100	100	100.0	240213	PASS
96	95	5	9	6.7	16210	PASS
173	174	0.00	2	0.6	956	PASS
174	95	50	120	62.9	150976	PASS
175	174	4	9	7.4	11194	PASS
176	174	93	101	95.7	144528	PASS
177	176	5	9	6.5	9360	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
 Data File : VL043044.D
 Acq On : 06 Oct 2025 08:38
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/08/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 07 01:32:06 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Fri Oct 03 08:23:34 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.780	49	181342	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	515782	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	434070	10.000	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.370 95 324488 10.377 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 103.800%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	287157	9.395	ppbv	99
3) Chlorodifluoromethane	1.476	51	281863	9.533	ppbv	99
4) Chloromethane	1.531	50	107064	9.590	ppbv	98
5) Vinyl Chloride	1.580	62	97626	9.214	ppbv	96
6) Bromomethane	1.667	94	45638	9.586	ppbv	95
7) Chloroethane	1.703	64	40916	9.255	ppbv	98
8) Dichlorotetrafluoroethane	1.554	85	235363	9.398	ppbv	99
9) Propene	1.483	41	108909	9.135	ppbv	99
10) Heptane	4.968	43	319007	10.884	ppbv	100
11) Trichlorofluoromethane	1.874	101	292298	9.306	ppbv	99
12) 1,1,2-Trichlorotrifluoro...	2.136	101	228489	9.236	ppbv	99
13) Ethanol	1.719	45	4994	10.290	ppbv	97
14) Bromoethene	1.780	108	83347	9.398	ppbv	98
15) Acetone	1.832	43	211361	8.324	ppbv	99
16) 1,3-Butadiene	1.609	39	98478	9.407	ppbv	98
17) tert-Butyl alcohol	2.036	59	252229	9.391	ppbv	100
18) 1,1-Dichloroethene	2.033	96	100506	9.268	ppbv	96
19) Isopropyl Alcohol	1.884	45	113044	9.292	ppbv	99
20) Methylene Chloride	2.059	84	89172	8.046	ppbv	97
21) Allyl Chloride	2.094	41	154207	9.480	ppbv	100
22) trans-1,2-Dichloroethene	2.337	96	112836	9.165	ppbv	98
23) Vinyl Acetate	2.457	43	48730	9.040	ppbv	98
24) 1,1-Dichloroethane	2.405	63	217854	9.335	ppbv	99
25) Ethyl Acetate	2.829	43	638620	10.026	ppbv	100
26) Hexane	2.819	57	250238	10.213	ppbv	98
27) Carbon Disulfide	2.146	76	277548	9.595	ppbv	100
28) Methyl tert-Butyl Ether	2.434	73	145380	9.481	ppbv	100
29) Chloroform	2.842	83	394464	9.442	ppbv	99
30) Cyclohexane	3.849	84	208637	10.362	ppbv	98
31) cis-1,2-Dichloroethene	2.716	61	244012	9.688	ppbv	99
32) 1,1,1-Trichloroethane	3.360	97	384289	9.465	ppbv	99
34) 2-Butanone	2.551	43	347993	9.944	ppbv	99
35) Carbon Tetrachloride	3.755	117	408168	9.619	ppbv	97
36) Benzene	3.648	78	502165	10.190	ppbv	98
37) 1,2-Dichloroethane	3.211	62	293779	10.066	ppbv	99
38) Trichloroethene	4.535	130	210056	10.233	ppbv	99
39) 1,2-Dichloropropane	4.302	63	183913	10.069	ppbv	99
40) 1,4-Dioxane	4.567	88	82396	10.517	ppbv #	95
41) Tetrahydrofuran	3.046	42	198764	10.989	ppbv	99
42) Bromodichloromethane	4.480	83	420134	10.194	ppbv	97
43) Methyl Methacrylate	4.865	69	181458	11.093	ppbv	98
44) 2,2,4-Trimethylpentane	4.629	57	857112	10.839	ppbv	99
45) t-1,3-Dichloropropene	6.402	75	158655	10.598	ppbv	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
 Data File : VL043044.D
 Acq On : 06 Oct 2025 08:38
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/08/2025
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 07 01:32:06 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Fri Oct 03 08:23:34 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.587	75	224260	10.632	ppbv	100
47) 1,1,2-Trichloroethane	6.571	97	201486	9.917	ppbv	99
48) Dibromochloromethane	7.376	129	343727	10.366	ppbv	99
49) Bromoform	9.474	173	294450	10.633	ppbv	99
50) 4-Methyl-2-Pentanone	5.732	43	485069	11.366	ppbv	99
51) 2-Hexanone	7.422	43	362864	11.945	ppbv #	99
52) Tetrachloroethene	8.218	164	178049	9.758	ppbv	94
53) Toluene	6.920	91	558213	11.220	ppbv	100
54) 1,2-Dibromoethane	7.642	107	251743	10.454	ppbv	97
56) 1,1,1,2-Tetrachloroethane	8.917	131	266983	10.090	ppbv	99
57) Chlorobenzene	8.914	112	457632	9.973	ppbv	98
58) Ethyl Benzene	9.348	91	773986	11.509	ppbv	100
59) m/p-Xylene	9.545	91	1291856m	22.776	ppbv	
60) o-Xylene	9.956	91	650260	11.509	ppbv	98
61) Styrene	9.865	104	285984	12.082	ppbv	99
62) Isopropylbenzene	10.532	105	966428	11.142	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.956	83	409295	10.098	ppbv	100
64) n-propylbenzene	10.982	120	256334	11.553	ppbv	99
65) tert-Butylbenzene	11.526	119	880541	11.407	ppbv	99
66) Benzyl Chloride	11.610	91	33291	11.284	ppbv	99
67) sec-Butylbenzene	11.762	105	1214749	11.277	ppbv	100
69) p-Isopropyltoluene	11.921	119	1022374	11.667	ppbv	100
70) n-Butylbenzene	12.277	91	1013351	12.010	ppbv	100
71) 2-Chlorotoluene	10.895	91	715947	11.405	ppbv	99
72) 4-Ethyltoluene	11.118	105	797033	11.852	ppbv	100
73) 1,3,5-Trimethylbenzene	11.199	105	666778	11.461	ppbv	100
74) 1,2,4-Trimethylbenzene	11.529	105	751232	11.441	ppbv	93
75) 1,3-Dichlorobenzene	11.600	146	448438	10.501	ppbv	100
76) 1,4-Dichlorobenzene	11.665	146	445649	10.944	ppbv	100
77) 1,2-Dichlorobenzene	11.940	146	430137	10.112	ppbv	100
78) Hexachloro-1,3-Butadiene	13.876	225	285716	9.640	ppbv	100
79) Naphthalene	13.507	128	563992	10.118	ppbv	99
80) Naphthalene,2-methyl-	14.452	142	228905	8.885	ppbv	99
81) 1,2,4-Trichlorobenzene	13.435	180	304505	12.033	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
 Data File : VL043044.D
 Acq On : 06 Oct 2025 08:38
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 07 01:32:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Oct 03 08:23:34 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	1.685	1.584	6.0	103	0.00
3	Chlorodifluoromethane	1.630	1.554	4.7	105	0.00
4	Chloromethane	0.616	0.590	4.2	106	0.00
5 T	Vinyl Chloride	0.584	0.538	7.9	103	0.00
6 T	Bromomethane	0.263	0.252	4.2	105	0.00
7	Chloroethane	0.244	0.226	7.4	105	0.00
8 T	Dichlorotetrafluoroethane	1.381	1.298	6.0	103	0.00
9 T	Propene	0.657	0.601	8.5	98	0.00
10 T	Heptane	1.616	1.759	-8.8	103	0.01
11 T	Trichlorofluoromethane	1.732	1.612	6.9	103	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.364	1.260	7.6	101	0.00
13	Ethanol	0.045	0.028#	37.8#	104	0.00
14 T	Bromoethene	0.489	0.460	5.9	101	0.00
15 T	Acetone	1.400	1.166	16.7	105	0.00
16 T	1,3-Butadiene	0.577	0.543	5.9	104	0.00
17	tert-Butyl alcohol	1.481	1.391	6.1	104	0.00
18 T	1,1-Dichloroethene	0.598	0.554	7.4	99	0.00
19 T	Isopropyl Alcohol	0.671	0.623	7.2	107	0.00
20 T	Methylene Chloride	0.611	0.492	19.5	104	0.00
21 T	Allyl Chloride	0.897	0.850	5.2	104	0.00
22 T	trans-1,2-Dichloroethene	0.679	0.622	8.4	101	0.00
23 T	Vinyl Acetate	0.297	0.269	9.4	114	0.00
24 T	1,1-Dichloroethane	1.287	1.201	6.7	103	0.00
25 T	Ethyl Acetate	3.513	3.522	-0.3	103	0.00
26 T	Hexane	1.351	1.380	-2.1	102	0.00
27 T	Carbon Disulfide	1.595	1.531	4.0	104	0.00
28 T	Methyl tert-Butyl Ether	0.846	0.802	5.2	101	0.00
29 T	Chloroform	2.304	2.175	5.6	102	0.00
30 T	Cyclohexane	1.110	1.151	-3.7	100	0.00
31 T	cis-1,2-Dichloroethene	1.389	1.346	3.1	99	0.00
32 T	1,1,1-Trichloroethane	2.239	2.119	5.4	101	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
34 T	2-Butanone	0.679	0.675	0.6	102	0.00
35 T	Carbon Tetrachloride	0.823	0.791	3.9	102	0.00
36 T	Benzene	0.955	0.974	-2.0	100	0.00
37 T	1,2-Dichloroethane	0.566	0.570	-0.7	103	0.00
38 T	Trichloroethene	0.398	0.407	-2.3	98	0.00
39 T	1,2-Dichloropropane	0.354	0.357	-0.8	101	0.00
40 T	1,4-Dioxane	0.152	0.160	-5.3	100	0.01
41 T	Tetrahydrofuran	0.351	0.385	-9.7	101	0.00
42 T	Bromodichloromethane	0.799	0.815	-2.0	102	0.01
43	Methyl Methacrylate	0.317	0.352	-11.0	98	0.01
44 T	2,2,4-Trimethylpentane	1.533	1.662	-8.4	103	0.01
45 T	t-1,3-Dichloropropene	0.290	0.308	-6.2	96	0.02
46 T	cis-1,3-Dichloropropene	0.409	0.435	-6.4	97	0.01
47 T	1,1,2-Trichloroethane	0.394	0.391	0.8	102	0.02
48 T	Dibromochloromethane	0.643	0.666	-3.6	99	0.01

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
 Data File : VL043044.D
 Acq On : 06 Oct 2025 08:38
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 07 01:32:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Oct 03 08:23:34 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.537	0.571	-6.3	100	0.00
50 T	4-Methyl-2-Pentanone	0.827	0.940	-13.7	102	0.02
51 T	2-Hexanone	0.589	0.704	-19.5	100	0.00
52 T	Tetrachloroethene	0.354	0.345	2.5	99	0.01
53 T	Toluene	0.965	1.082	-12.1	98	0.01
54 T	1,2-Dibromoethane	0.467	0.488	-4.5	100	0.01
55 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.01
56	1,1,1,2-Tetrachloroethane	0.610	0.615	-0.8	101	0.00
57 T	Chlorobenzene	1.057	1.054	0.3	101	0.00
58 T	Ethyl Benzene	1.549	1.783	-15.1	98	0.00
59 T	m/p-Xylene	1.307	1.488	-13.8	100	0.01
60 T	o-Xylene	1.302	1.498	-15.1	101	0.00
61 T	Styrene	0.545	0.659	-20.9	96	0.01
62	Isopropylbenzene	1.998	2.226	-11.4	101	0.00
63 T	1,1,2,2-Tetrachloroethane	0.934	0.943	-1.0	105	0.00
64	n-propylbenzene	0.511	0.591	-15.7	102	0.00
65	tert-Butylbenzene	1.778	2.029	-14.1	102	0.00
66 T	Benzyl Chloride	0.068	0.077	-13.2	108	0.00
67	sec-Butylbenzene	2.482	2.799	-12.8	102	0.00
68 S	1-Bromo-4-Fluorobenzene	0.720	0.748	-3.9	98	0.00
69	p-Isopropyltoluene	2.019	2.355	-16.6	101	0.00
70	n-Butylbenzene	1.944	2.335	-20.1	102	0.00
71	2-Chlorotoluene	1.446	1.649	-14.0	101	0.00
72 T	4-Ethyltoluene	1.549	1.836	-18.5	102	0.00
73 T	1,3,5-Trimethylbenzene	1.340	1.536	-14.6	101	0.00
74 T	1,2,4-Trimethylbenzene	1.513	1.731	-14.4	103	0.00
75 T	1,3-Dichlorobenzene	0.984	1.033	-5.0	102	0.00
76 T	1,4-Dichlorobenzene	0.938	1.027	-9.5	103	0.00
77 T	1,2-Dichlorobenzene	0.980	0.991	-1.1	103	0.00
78 T	Hexachloro-1,3-Butadiene	0.683	0.658	3.7	99	0.00
79 T	Naphthalene	0.903	1.299	-43.9#	98	0.00
80 T	Naphthalene,2-methyl-	0.457	0.527	-15.3	90	0.00
81 T	1,2,4-Trichlorobenzene	0.583	0.702	-20.4	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
 Data File : VL043044.D
 Acq On : 06 Oct 2025 08:38
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 07 01:32:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Oct 03 08:23:34 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	10.000	9.395	6.1	103	0.00
3	Chlorodifluoromethane	10.000	9.533	4.7	105	0.00
4	Chloromethane	10.000	9.590	4.1	106	0.00
5 T	Vinyl Chloride	10.000	9.214	7.9	103	0.00
6 T	Bromomethane	10.000	9.586	4.1	105	0.00
7	Chloroethane	10.000	9.255	7.4	105	0.00
8 T	Dichlorotetrafluoroethane	10.000	9.398	6.0	103	0.00
9 T	Propene	10.000	9.135	8.7	98	0.00
10 T	Heptane	10.000	10.884	-8.8	103	0.01
11 T	Trichlorofluoromethane	10.000	9.306	6.9	103	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	9.236	7.6	101	0.00
13	Ethanol	10.000	10.290	-2.9	104	0.00
14 T	Bromoethene	10.000	9.398	6.0	101	0.00
15 T	Acetone	10.000	8.324	16.8	105	0.00
16 T	1,3-Butadiene	10.000	9.407	5.9	104	0.00
17	tert-Butyl alcohol	10.000	9.391	6.1	104	0.00
18 T	1,1-Dichloroethene	10.000	9.268	7.3	99	0.00
19 T	Isopropyl Alcohol	10.000	9.292	7.1	107	0.00
20 T	Methylene Chloride	10.000	8.046	19.5	104	0.00
21 T	Allyl Chloride	10.000	9.480	5.2	104	0.00
22 T	trans-1,2-Dichloroethene	10.000	9.165	8.4	101	0.00
23 T	Vinyl Acetate	10.000	9.040	9.6	114	0.00
24 T	1,1-Dichloroethane	10.000	9.335	6.6	103	0.00
25 T	Ethyl Acetate	10.000	10.026	-0.3	103	0.00
26 T	Hexane	10.000	10.213	-2.1	102	0.00
27 T	Carbon Disulfide	10.000	9.595	4.0	104	0.00
28 T	Methyl tert-Butyl Ether	10.000	9.481	5.2	101	0.00
29 T	Chloroform	10.000	9.442	5.6	102	0.00
30 T	Cyclohexane	10.000	10.362	-3.6	100	0.00
31 T	cis-1,2-Dichloroethene	10.000	9.688	3.1	99	0.00
32 T	1,1,1-Trichloroethane	10.000	9.465	5.4	101	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	99	0.00
34 T	2-Butanone	10.000	9.944	0.6	102	0.00
35 T	Carbon Tetrachloride	10.000	9.619	3.8	102	0.00
36 T	Benzene	10.000	10.190	-1.9	100	0.00
37 T	1,2-Dichloroethane	10.000	10.066	-0.7	103	0.00
38 T	Trichloroethene	10.000	10.233	-2.3	98	0.00
39 T	1,2-Dichloropropane	10.000	10.069	-0.7	101	0.00
40 T	1,4-Dioxane	10.000	10.517	-5.2	100	0.01
41 T	Tetrahydrofuran	10.000	10.989	-9.9	101	0.00
42 T	Bromodichloromethane	10.000	10.194	-1.9	102	0.01
43	Methyl Methacrylate	10.000	11.093	-10.9	98	0.01
44 T	2,2,4-Trimethylpentane	10.000	10.839	-8.4	103	0.01
45 T	t-1,3-Dichloropropene	10.000	10.598	-6.0	96	0.02
46 T	cis-1,3-Dichloropropene	10.000	10.632	-6.3	97	0.01
47 T	1,1,2-Trichloroethane	10.000	9.917	0.8	102	0.02
48 T	Dibromochloromethane	10.000	10.366	-3.7	99	0.01

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
 Data File : VL043044.D
 Acq On : 06 Oct 2025 08:38
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 07 01:32:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Oct 03 08:23:34 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	Bromoform	10.000	10.633	-6.3	100	0.00
50 T	4-Methyl-2-Pentanone	10.000	11.366	-13.7	102	0.02
51 T	2-Hexanone	10.000	11.945	-19.5	100	0.00
52 T	Tetrachloroethene	10.000	9.758	2.4	99	0.01
53 T	Toluene	10.000	11.220	-12.2	98	0.01
54 T	1,2-Dibromoethane	10.000	10.454	-4.5	100	0.01
55 I	Chlorobenzene-d5	10.000	10.000	0.0	97	0.01
56	1,1,1,2-Tetrachloroethane	10.000	10.090	-0.9	101	0.00
57 T	Chlorobenzene	10.000	9.973	0.3	101	0.00
58 T	Ethyl Benzene	10.000	11.509	-15.1	98	0.00
59 T	m/p-Xylene	20.000	22.776	-13.9	100	0.01
60 T	o-Xylene	10.000	11.509	-15.1	101	0.00
61 T	Styrene	10.000	12.082	-20.8	96	0.01
62	Isopropylbenzene	10.000	11.142	-11.4	101	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	10.098	-1.0	105	0.00
64	n-propylbenzene	10.000	11.553	-15.5	102	0.00
65	tert-Butylbenzene	10.000	11.407	-14.1	102	0.00
66 T	Benzyl Chloride	10.000	11.284	-12.8	108	0.00
67	sec-Butylbenzene	10.000	11.277	-12.8	102	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	10.377	-3.8	98	0.00
69	p-Isopropyltoluene	10.000	11.667	-16.7	101	0.00
70	n-Butylbenzene	10.000	12.010	-20.1	102	0.00
71	2-Chlorotoluene	10.000	11.405	-14.0	101	0.00
72 T	4-Ethyltoluene	10.000	11.852	-18.5	102	0.00
73 T	1,3,5-Trimethylbenzene	10.000	11.461	-14.6	101	0.00
74 T	1,2,4-Trimethylbenzene	10.000	11.441	-14.4	103	0.00
75 T	1,3-Dichlorobenzene	10.000	10.501	-5.0	102	0.00
76 T	1,4-Dichlorobenzene	10.000	10.944	-9.4	103	0.00
77 T	1,2-Dichlorobenzene	10.000	10.112	-1.1	103	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	9.640	3.6	99	0.00
79 T	Naphthalene	10.000	10.118	-1.2	98	0.00
80 T	Naphthalene,2-methyl-	10.000	8.885	11.2	90	0.00
81 T	1,2,4-Trichlorobenzene	10.000	12.033	-20.3	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
 Data File : VL043045.D
 Acq On : 06 Oct 2025 09:19
 Operator : SY/MD
 Sample : VL1006ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1006ABL01

Quant Time: Oct 07 01:32:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Oct 03 08:23:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.781	49	178856	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	516182	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	413895	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	290223	9.734	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.300%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
Data File : VL043045.D
Acq On : 06 Oct 2025 09:19
Operator : SY/MD
Sample : VL1006ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1006ABL01

Quant Time: Oct 07 01:32:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Oct 03 08:23:34 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
 Data File : VL043046.D
 Acq On : 06 Oct 2025 10:12
 Operator : SY/MD
 Sample : VL1006ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1006ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/08/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 07 01:33:20 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2025

QLast Update : Fri Oct 03 08:23:34 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.780	49	188357	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.946	114	537601	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	447189	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.367 95 338181 10.498 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 105.000%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	295209	9.299	ppbv	99
3) Chlorodifluoromethane	1.473	51	282563	9.201	ppbv	100
4) Chloromethane	1.531	50	106576	9.191	ppbv	97
5) Vinyl Chloride	1.576	62	101063	9.184	ppbv	99
6) Bromomethane	1.664	94	47460	9.598	ppbv	94
7) Chloroethane	1.703	64	42252	9.201	ppbv	96
8) Dichlorotetrafluoroethane	1.551	85	247037	9.497	ppbv	97
9) Propene	1.483	41	113498	9.166	ppbv	99
10) Heptane	4.965	43	320280	10.521	ppbv	100
11) Trichlorofluoromethane	1.874	101	302747	9.280	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.133	101	239211	9.310	ppbv	99
13) Ethanol	1.719	45	4927	9.710	ppbv	99
14) Bromoethene	1.777	108	87214	9.468	ppbv	99
15) Acetone	1.832	43	213544	8.097	ppbv	99
16) 1,3-Butadiene	1.606	39	99231	9.126	ppbv	99
17) tert-Butyl alcohol	2.033	59	260280	9.330	ppbv	100
18) 1,1-Dichloroethene	2.030	96	102725	9.120	ppbv	96
19) Isopropyl Alcohol	1.881	45	111933	8.858	ppbv	99
20) Methylene Chloride	2.059	84	90931	7.899	ppbv	98
21) Allyl Chloride	2.091	41	157109	9.299	ppbv	99
22) trans-1,2-Dichloroethene	2.337	96	117166	9.162	ppbv	98
23) Vinyl Acetate	2.457	43	42991	7.678	ppbv	99
24) 1,1-Dichloroethane	2.402	63	225170	9.289	ppbv	99
25) Ethyl Acetate	2.826	43	647632	9.789	ppbv	100
26) Hexane	2.819	57	258251	10.147	ppbv	100
27) Carbon Disulfide	2.146	76	283466	9.434	ppbv	99
28) Methyl tert-Butyl Ether	2.434	73	147013	9.230	ppbv	99
29) Chloroform	2.842	83	402053	9.266	ppbv	100
30) Cyclohexane	3.845	84	213947	10.230	ppbv	98
31) cis-1,2-Dichloroethene	2.712	61	247932	9.477	ppbv	95
32) 1,1,1-Trichloroethane	3.360	97	392565	9.309	ppbv	99
34) 2-Butanone	2.547	43	347648	9.530	ppbv	100
35) Carbon Tetrachloride	3.751	117	414732	9.377	ppbv	96
36) Benzene	3.648	78	513451	9.997	ppbv	99
37) 1,2-Dichloroethane	3.208	62	296234	9.738	ppbv	100
38) Trichloroethene	4.535	130	213507	9.979	ppbv	99
39) 1,2-Dichloropropane	4.302	63	185321	9.734	ppbv	96
40) 1,4-Dioxane	4.564	88	80157	9.816	ppbv #	96
41) Tetrahydrofuran	3.043	42	199341	10.573	ppbv	100
42) Bromodichloromethane	4.473	83	422079	9.826	ppbv	99
43) Methyl Methacrylate	4.862	69	185044	10.853	ppbv	99
44) 2,2,4-Trimethylpentane	4.625	57	871720	10.577	ppbv	99
45) t-1,3-Dichloropropene	6.399	75	157655	10.104	ppbv	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
 Data File : VL043046.D
 Acq On : 06 Oct 2025 10:12
 Operator : SY/MD
 Sample : VL1006ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1006ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/08/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 07 01:33:20 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2025

QLast Update : Fri Oct 03 08:23:34 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.583	75	228082	10.375	ppbv	97
47) 1,1,2-Trichloroethane	6.561	97	201822	9.530	ppbv	96
48) Dibromochloromethane	7.373	129	341429	9.879	ppbv	100
49) Bromoform	9.474	173	279106	9.669	ppbv	100
50) 4-Methyl-2-Pentanone	5.726	43	489303	11.000	ppbv	98
51) 2-Hexanone	7.419	43	361924	11.431	ppbv #	97
52) Tetrachloroethene	8.215	164	178891	9.406	ppbv	100
53) Toluene	6.917	91	563494	10.867	ppbv	100
54) 1,2-Dibromoethane	7.639	107	248606	9.905	ppbv	92
56) 1,1,1,2-Tetrachloroethane	8.914	131	264054	9.687	ppbv	99
57) Chlorobenzene	8.911	112	456710	9.661	ppbv	99
58) Ethyl Benzene	9.344	91	767600	11.080	ppbv	98
59) m/p-Xylene	9.542	91	1275234m	21.824	ppbv	
60) o-Xylene	9.956	91	632243	10.862	ppbv	98
61) Styrene	9.862	104	275280	11.289	ppbv	99
62) Isopropylbenzene	10.532	105	945764	10.584	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.953	83	373448	8.943	ppbv	99
64) n-propylbenzene	10.982	120	246252	10.773	ppbv	99
65) tert-Butylbenzene	11.523	119	829195	10.427	ppbv	100
66) Benzyl Chloride	11.607	91	28281	9.305	ppbv	99
67) sec-Butylbenzene	11.759	105	1139275	10.266	ppbv	100
69) p-Isopropyltoluene	11.917	119	953493	10.562	ppbv	99
70) n-Butylbenzene	12.274	91	920477	10.590	ppbv	100
71) 2-Chlorotoluene	10.891	91	679459	10.507	ppbv	99
72) 4-Ethyltoluene	11.118	105	752627	10.863	ppbv	100
73) 1,3,5-Trimethylbenzene	11.196	105	630941	10.527	ppbv	100
74) 1,2,4-Trimethylbenzene	11.529	105	702152	10.380	ppbv	97
75) 1,3-Dichlorobenzene	11.600	146	409449	9.307	ppbv	99
76) 1,4-Dichlorobenzene	11.665	146	401258	9.565	ppbv	100
77) 1,2-Dichlorobenzene	11.940	146	389675	8.892	ppbv	99
78) Hexachloro-1,3-Butadiene	13.872	225	292520	9.580	ppbv	99
79) Naphthalene	13.503	128	548561	9.592	ppbv	99
80) Naphthalene,2-methyl-	14.449	142	207324	7.891	ppbv	99
81) 1,2,4-Trichlorobenzene	13.432	180	312207	11.975	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
 Data File : VL043048.D
 Acq On : 06 Oct 2025 11:23
 Operator : SY/MD
 Sample : QC092425 CAN 10590
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC092425 CAN 10590

Quant Time: Oct 07 01:34:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Oct 03 08:23:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.781	49	170545	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.946	114	476003	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	387537	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	268829	9.629	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.300%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100625\
Data File : VL043048.D
Acq On : 06 Oct 2025 11:23
Operator : SY/MD
Sample : QC092425 CAN 10590
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC092425 CAN 10590

Quant Time: Oct 07 01:34:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Oct 03 08:23:34 2025
Response via : Initial Calibration



Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL050125

Review By	Semsettin Yesilyurt	Review On	5/2/2025 7:59:23 AM		
Supervise By	Mahesh Dadoda	Supervise On	5/5/2025 2:55:39 PM		
SubDirectory	VL050125	HP Acquire Method	TO15AIR.M	HP Processing Method	VL050125AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2603			
Initial Calibration Stds		AP2606,AP2607,AP2608			
CCC		AP2606			
Internal Standard/PEM		AP2603			
ICV/I.BLK		AP2609			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042457.D	01 May 2025 08:19	SY/MD	Ok
2	VSTDICCC010	VL042458.D	01 May 2025 09:03	SY/MD	Ok,M
3	VSTDICCC002	VL042459.D	01 May 2025 09:37	SY/MD	Ok,M
4	VSTDICCC001	VL042460.D	01 May 2025 10:08	SY/MD	Ok,M
5	VSTDICCC0.5	VL042461.D	01 May 2025 10:40	SY/MD	Ok,M
6	VSTDICCC0.1	VL042462.D	01 May 2025 11:12	SY/MD	Ok,M
7	VSTDICCC0.03	VL042463.D	01 May 2025 11:44	SY/MD	Not Ok
8	VSTDICCC0.03	VL042464.D	01 May 2025 12:15	SY/MD	Not Ok
9	VSTDICCC015	VL042465.D	01 May 2025 12:49	SY/MD	Ok,M
10	VSTDICCC0.03	VL042466.D	01 May 2025 14:10	SY/MD	Ok,M
11	VSTDICV010	VL042467.D	01 May 2025 15:46	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL050625

Review By	Semsettin Yesilyurt	Review On	5/9/2025 2:37:46 PM
Supervise By	Maresh Dadoda	Supervise On	5/9/2025 2:51:31 PM
SubDirectory	VL050625	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL050125AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2603		
Initial Calibration Stds	AP2606,AP2607,AP2608		
CCC	AP2606		
Internal Standard/PEM	AP2603		
ICV/I.BLK	AP2609		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042486.D	06 May 2025 08:16	SY/MD	Ok
2	VSTDCCC010	VL042487.D	06 May 2025 09:00	SY/MD	Ok,M
3	VL0506ABL01	VL042488.D	06 May 2025 09:49	SY/MD	Ok
4	VL0506ABS01	VL042489.D	06 May 2025 10:44	SY/MD	Ok,M
5	QC043025 CAN 10605	VL042490.D	06 May 2025 11:15	SY/MD	Ok
6	QC042925 CAN 10590	VL042491.D	06 May 2025 11:47	SY/MD	Ok
7	QC042225 CAN 10669	VL042492.D	06 May 2025 12:19	SY/MD	Ok
8	QC042325 CAN 10683	VL042493.D	06 May 2025 12:51	SY/MD	Ok
9	QC0429325 CAN 10458	VL042494.D	06 May 2025 13:23	SY/MD	Ok
10	MDL1 0.45	VL042495.D	06 May 2025 14:26	SY/MD	Ok,M
11	MDL2 0.45	VL042496.D	06 May 2025 14:57	SY/MD	Ok,M
12	MDL3 0.45	VL042497.D	06 May 2025 15:29	SY/MD	Not Ok
13	MDL1 0.03	VL042498.D	06 May 2025 16:01	SY/MD	Ok,M
14	MDL2 0.03	VL042499.D	06 May 2025 16:32	SY/MD	Not Ok
15	MDL3 0.03	VL042500.D	06 May 2025 17:04	SY/MD	Ok,M
16	MDL1 0.40	VL042501.D	06 May 2025 17:36	SY/MD	Ok,M
17	MDL2 0.40	VL042502.D	06 May 2025 18:07	SY/MD	Ok,M
18	MDL3 0.40	VL042503.D	06 May 2025 18:39	SY/MD	Ok,M
19	MDL1 0.1	VL042504.D	06 May 2025 19:10	SY/MD	Ok,M
20	MDL2 0.1	VL042505.D	06 May 2025 19:42	SY/MD	Ok,M
21	MDL3 0.1	VL042506.D	06 May 2025 20:13	SY/MD	Ok,M

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL050625

Review By	Semsettin Yesilyurt	Review On	5/9/2025 2:37:46 PM		
Supervise By	Mahesh Dadoda	Supervise On	5/9/2025 2:51:31 PM		
SubDirectory	VL050625	HP Acquire Method	TO15AIR.M	HP Processing Method	VL050125AIR.M
STD. NAME	STD REF.#				
Tune/Reschk	AP2603				
Initial Calibration Stds	AP2606,AP2607,AP2608				
CCC	AP2606				
Internal Standard/PEM	AP2603				
ICV/I.BLK	AP2609				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	VSTDCCC010	VL042507.D	06 May 2025 20:45	SY/MD	Ok,M
----	------------	------------	-------------------	-------	------

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL100225

Review By	Semsettin Yesilyurt	Review On	10/3/2025 9:49:03 AM		
Supervise By	Maresh Dadoda	Supervise On	10/10/2025 2:54:57 AM		
SubDirectory	VL100225	HP Acquire Method	TO15AIR	HP Processing Method	VL100225AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2682			
Initial Calibration Stds		AP2678,AP2679,AP2680			
CCC		AP2678			
Internal Standard/PEM		AP2682			
ICV/I.BLK		AP2681			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL043030.D	02 Oct 2025 09:11	SY/MD	Ok
2	VSTDICC0.5	VL043031.D	02 Oct 2025 09:50	SY/MD	Ok,M
3	VSTDICCC010	VL043032.D	02 Oct 2025 10:23	SY/MD	Ok,M
4	VSTDICC002	VL043033.D	02 Oct 2025 10:57	SY/MD	Ok,M
5	VSTDICC001	VL043034.D	02 Oct 2025 11:29	SY/MD	Ok,M
6	VSTDICC0.5	VL043035.D	02 Oct 2025 12:02	SY/MD	Not Ok
7	VSTDICC0.1	VL043036.D	02 Oct 2025 12:35	SY/MD	Ok,M
8	VSTDICC0.03	VL043037.D	02 Oct 2025 13:07	SY/MD	Ok,M
9	VSTDICC015	VL043038.D	02 Oct 2025 13:41	SY/MD	Ok,M
10	VSTDICV010	VL043039.D	02 Oct 2025 14:14	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL100625

Review By	Semsettin Yesilyurt	Review On	10/8/2025 10:36:14 AM		
Supervise By	Mahesh Dadoda	Supervise On	10/10/2025 2:51:29 AM		
SubDirectory	VL100625	HP Acquire Method	TO15AIR	HP Processing Method	VL100225AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2682			
Initial Calibration Stds		AP2678,AP2679,AP2680			
CCC		AP2678			
Internal Standard/PEM		AP2682			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL043043.D	06 Oct 2025 07:57	SY/MD	Ok
2	VSTDCCC010	VL043044.D	06 Oct 2025 08:38	SY/MD	Ok,M
3	VL1006ABL01	VL043045.D	06 Oct 2025 09:19	SY/MD	Ok
4	VL1006ABS01	VL043046.D	06 Oct 2025 10:12	SY/MD	Ok,M
5	QC092925 CAN 10198	VL043047.D	06 Oct 2025 10:50	SY/MD	Ok
6	QC092425 CAN 10590	VL043048.D	06 Oct 2025 11:23	SY/MD	Ok
7	CAN 10296	VL043049.D	06 Oct 2025 11:56	SY/MD	Ok
8	CAN 10312	VL043050.D	06 Oct 2025 12:29	SY/MD	Ok
9	Q2893-14 0.5 LOQ	VL043051.D	06 Oct 2025 13:38	SY/MD	ReRun
10	Q2893-14 0.03 LOQ	VL043052.D	06 Oct 2025 14:11	SY/MD	Ok,M

M : Manual Integration