

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: RTPE01
Lab Code: ACE SDG NO.: Q3627
Lab File ID: VL043191.D Date Analyzed: 11/13/2025
Instrument ID: MSVOA_L Time Analyzed: 09:01
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	141757	2.79	379323	3.97	324654	8.89
UPPER LIMIT	198460	3.12	531052	4.30	454516	9.22
LOWER LIMIT	85054.2	2.46	227594	3.64	194792	8.56
EPA SAMPLE NO.						
UNIT-8-IA-1	143373	2.79	385279	3.97	320935	8.89
UNIT-7-IA-1	144607	2.80	396555	3.97	342186	8.89
UNIT-7-IA-1DUP	145038	2.79	398990	3.97	346200	8.89
AA-East	141257	2.79	362075	3.97	303679	8.89
UNIT-8-SS-1	136843	2.79	359287	3.97	305253	8.89
UNIT-8-SS-1DL	138276	2.80	360106	3.97	298832	8.89
UNIT-7-SS-1	147747	2.80	400343	3.97	381448	8.89
UNIT-7-SS-1DL	140657	2.79	374439	3.97	343535	8.89
VL1113ABL01	136642	2.79	369240	3.97	306160	8.89
VL1113ABS01	140848	2.79	381641	3.97	326347	8.89

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.