

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3629

Project: Linden Generating Station - Towers

Run No.: LB137885

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137885BS							
TDS	mg/L	100	98.0		98	1	90-110	11/13/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3629

Project: Linden Generating Station - Towers

Run No.: LB137913

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137913BS							
TSS	mg/L	550	538		98	1	90-110	11/17/2025