

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Project: Linden Generating Station - LRSA

RunNo.: LB137888

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.510	0.5	102	90-110	11/13/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.491	0.5	98	90-110	11/13/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.508	0.5	102	90-110	11/13/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Project: Linden Generating Station - LRSA

RunNo.: LB137904

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 COD	mg/L	51.184	50	102	95-105	11/14/2025
Sample ID: CCV2 COD	mg/L	48.150	50	96	95-105	11/14/2025
Sample ID: CCV3 COD	mg/L	48.150	50	96	95-105	11/14/2025
Sample ID: CCV4 COD	mg/L	49.161	50	98	95-105	11/14/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Project: Linden Generating Station - LRSA

RunNo.: LB137905

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.094	0.099	95	85-115	11/14/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	11/14/2025
Sample ID: CCV2 Cyanide	mg/L	0.25	0.25	100	90-110	11/14/2025
Sample ID: CCV3 Cyanide	mg/L	0.24	0.25	96	90-110	11/14/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Project: Linden Generating Station - LRSA

RunNo.: LB137908

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	11/14/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.503	0.5	101	90-110	11/14/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.506	0.5	101	90-110	11/14/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Project: Linden Generating Station - LRSA

RunNo.: LB137922

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	11/17/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.95	1	95	90-110	11/17/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.95	1	95	90-110	11/17/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	11/17/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.97	1	97	90-110	11/17/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Project: Linden Generating Station - LRSA

RunNo.: LB137926

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Phenolics	mg/L	1	1	100	90-110	11/17/2025
Sample ID: CCV1 Phenolics	mg/L	0.97	1	97	90-110	11/17/2025
Sample ID: CCV2 Phenolics	mg/L	0.98	1	98	90-110	11/17/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Project: Linden Generating Station - LRSA

RunNo.: LB137990

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Field pH	pH	7.05	7	101	90-110	11/13/2025
Sample ID: CCV1 Field pH	pH	7.04	7	101	90-110	11/13/2025
Sample ID: CCV2 Field pH	pH	7.02	7	100	90-110	11/13/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Project: Linden Generating Station - LRSA

RunNo.: LB137990

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------