

Preparation Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Project: Linden Generating Station - LRSA

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB137888BL Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.003	0.01	11/13/2025
Sample ID: LB137904BL COD	mg/L	< 5.0000	5.0000	U	1.5	10.0	11/14/2025
Sample ID: LB137906BL BOD5	mg/L	< 0.2000	0.2000	U	0.20	2.0	11/14/2025
Sample ID: LB137908BL Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.003	0.01	11/14/2025
Sample ID: LB137913BL TSS	mg/L	< 2.0000	2.0000	U	1	4	11/17/2025
Sample ID: PB170548BL Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/14/2025
Sample ID: PB170566BL Phenolics	mg/L	< 0.0250	0.0250	U	0.015	0.05	11/17/2025
Sample ID: PB170582BL Ammonia as N	mg/L	< 0.0500	0.0500	U	0.03	0.1	11/17/2025