



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Contract: PSEG04

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	20.5	20.0	102	65 - 135	P	11/17/2025	13:35	LB137929
	Cadmium	6.08	6.0	101	65 - 135	P	11/17/2025	13:35	LB137929
	Chromium	10.4	10.0	104	65 - 135	P	11/17/2025	13:35	LB137929
	Copper	22.4	20.0	112	65 - 135	P	11/17/2025	13:35	LB137929
	Iron	107	100	107	65 - 135	P	11/17/2025	13:35	LB137929
	Lead	12.4	12.0	103	65 - 135	P	11/17/2025	13:35	LB137929
	Nickel	42.9	40.0	107	65 - 135	P	11/17/2025	13:35	LB137929
	Zinc	43.8	40.0	110	65 - 135	P	11/17/2025	13:35	LB137929