

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170558BL	WATER			Batch Number:	PB170558		Prep Date:	11/14/2025	
	Arsenic	2.56	<5	U	10.0	P	11/17/2025	14:30	LB137929
	Cadmium	0.25	<1.5	U	3.00	P	11/17/2025	14:30	LB137929
	Chromium	1.06	<2.5	U	5.00	P	11/17/2025	14:30	LB137929
	Copper	2.30	<5	U	10.0	P	11/17/2025	14:30	LB137929
	Iron	11.7	<25	U	50.0	P	11/17/2025	14:30	LB137929
	Lead	1.15	<3	U	6.00	P	11/17/2025	14:30	LB137929
	Nickel	1.53	<10	U	20.0	P	11/17/2025	14:30	LB137929
	Zinc	8.33	<10	U	20.0	P	11/17/2025	14:30	LB137929