

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Contract: PSEG04

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	11.0			-20	20	11/17/2025	13:40	LB137929
	Cadmium	5.85	1.0	585	-5	7	11/17/2025	13:40	LB137929
	Chromium	51.1	52.0	98	42	62	11/17/2025	13:40	LB137929
	Copper	-17.2	2.0	860	-18	22	11/17/2025	13:40	LB137929
	Iron	96200	101000	95	85600	116500	11/17/2025	13:40	LB137929
	Lead	9.64			-12	12	11/17/2025	13:40	LB137929
	Nickel	5.91	2.0	296	-38	42	11/17/2025	13:40	LB137929
	Zinc	6.79			-40	40	11/17/2025	13:40	LB137929
ICSAB01	Arsenic	116	104	112	88.4	120	11/17/2025	13:44	LB137929
	Cadmium	994	972	102	826	1120	11/17/2025	13:44	LB137929
	Chromium	565	542	104	460	624	11/17/2025	13:44	LB137929
	Copper	460	511	90	434	588	11/17/2025	13:44	LB137929
	Iron	98300	99300	99	84400	114500	11/17/2025	13:44	LB137929
	Lead	58.0	49.0	118	37	61	11/17/2025	13:44	LB137929
	Nickel	1000	954	105	810	1100	11/17/2025	13:44	LB137929
	Zinc	1050	952	110	809	1095	11/17/2025	13:44	LB137929