

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q3630

Contract: PSEG04

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170558BS							
Arsenic	ug/L	400	382		96	80 - 120	P
Cadmium	ug/L	100	95.3		95	80 - 120	P
Chromium	ug/L	200	199		100	80 - 120	P
Copper	ug/L	150	153		102	80 - 120	P
Iron	ug/L	1500	1420		95	80 - 120	P
Lead	ug/L	500	469		94	80 - 120	P
Nickel	ug/L	250	244		98	80 - 120	P
Zinc	ug/L	100	100		100	80 - 120	P