

### Initial and Continuing Calibration Verification

**Client:** Vartan Manufacturing Corp.

**SDG No.:** Q3642

**Project:** Semi Annual 2025

**RunNo.:** LB138054

| Analyte               |      | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|-----------------------|------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID:<br>Cyanide | ICV1 | mg/L  | 0.098  | 0.099      | 99            | 85-115                    | 11/26/2025       |
| Sample ID:<br>Cyanide | CCV1 | mg/L  | 0.24   | 0.25       | 96            | 90-110                    | 11/26/2025       |
| Sample ID:<br>Cyanide | CCV2 | mg/L  | 0.24   | 0.25       | 96            | 90-110                    | 11/26/2025       |