

Laboratory Control Sample Summary

Client: Vartan Manufacturing Corp.

SDG No.: Q3642

Project: Semi Annual 2025

Run No.: LB138054

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170751BS							
Cyanide	mg/L	0.1	0.098		98	1	85-115	11/26/2025