



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Vartan Manufacturing Corp.

SDG No.: Q3642

Contract: F&GE01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	6.21	6.0	104	65 - 135	P	12/01/2025	13:50	LB138072
	Chromium	9.87	10.0	99	65 - 135	P	12/01/2025	13:50	LB138072
	Copper	22.0	20.0	110	65 - 135	P	12/01/2025	13:50	LB138072
	Lead	13.2	12.0	110	65 - 135	P	12/01/2025	13:50	LB138072
	Molybdenum	207	200	103	65 - 135	P	12/01/2025	13:50	LB138072
	Nickel	41.6	40.0	104	65 - 135	P	12/01/2025	13:50	LB138072
	Silver	10.2	10.0	102	65 - 135	P	12/01/2025	13:50	LB138072
	Zinc	44.3	40.0	111	65 - 135	P	12/01/2025	13:50	LB138072
CRA	Mercury	0.054	0.05	108	60 - 140	CV	12/02/2025	11:10	LB138077