



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vartan Manufacturing Corp.

SDG No.: Q3642

Contract: F&GE01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB133	Mercury	0.027	+/-0.2	U	0.20	CV	12/02/2025	10:58	LB138077

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vartan Manufacturing Corp.

SDG No.: Q3642

Contract: F&GE01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB47	Mercury	0.027	+/-0.2	U	0.20	CV	12/02/2025	11:03	LB138077
CCB48	Mercury	0.027	+/-0.2	U	0.20	CV	12/02/2025	11:44	LB138077
CCB49	Mercury	0.027	+/-0.2	U	0.20	CV	12/02/2025	11:53	LB138077

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vartan Manufacturing Corp.

SDG No.: Q3642

Contract: F&GE01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	13:46	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	13:46	LB138072
	Copper	4.60	+/-10	U	20.0	P	12/01/2025	13:46	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	13:46	LB138072
	Molybdenum	18.7	+/-100	U	200	P	12/01/2025	13:46	LB138072
	Nickel	3.06	+/-20	U	40.0	P	12/01/2025	13:46	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	13:46	LB138072
	Zinc	16.7	+/-20	U	40.0	P	12/01/2025	13:46	LB138072

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vartan Manufacturing Corp.

SDG No.: Q3642

Contract: F&GE01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	14:34	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	14:34	LB138072
	Copper	4.60	+/-10	U	20.0	P	12/01/2025	14:34	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	14:34	LB138072
	Molybdenum	18.7	+/-100	U	200	P	12/01/2025	14:34	LB138072
	Nickel	3.06	+/-20	U	40.0	P	12/01/2025	14:34	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	14:34	LB138072
	Zinc	16.7	+/-20	U	40.0	P	12/01/2025	14:34	LB138072
CCB02	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	15:31	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	15:31	LB138072
	Copper	4.60	+/-10	U	20.0	P	12/01/2025	15:31	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	15:31	LB138072
	Molybdenum	18.7	+/-100	U	200	P	12/01/2025	15:31	LB138072
	Nickel	3.06	+/-20	U	40.0	P	12/01/2025	15:31	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	15:31	LB138072
	Zinc	16.7	+/-20	U	40.0	P	12/01/2025	15:31	LB138072
CCB03	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	16:38	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	16:38	LB138072
	Copper	4.60	+/-10	U	20.0	P	12/01/2025	16:38	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	16:38	LB138072
	Molybdenum	18.7	+/-100	U	200	P	12/01/2025	16:38	LB138072
	Nickel	3.06	+/-20	U	40.0	P	12/01/2025	16:38	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	16:38	LB138072
	Zinc	16.7	+/-20	U	40.0	P	12/01/2025	16:38	LB138072
CCB04	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	17:29	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	17:29	LB138072
	Copper	4.60	+/-10	U	20.0	P	12/01/2025	17:29	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	17:29	LB138072
	Molybdenum	18.7	+/-100	U	200	P	12/01/2025	17:29	LB138072
	Nickel	3.06	+/-20	U	40.0	P	12/01/2025	17:29	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	17:29	LB138072
	Zinc	16.7	+/-20	U	40.0	P	12/01/2025	17:29	LB138072
CCB05	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	18:20	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	18:20	LB138072
	Copper	4.60	+/-10	U	20.0	P	12/01/2025	18:20	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	18:20	LB138072
	Molybdenum	18.7	+/-100	U	200	P	12/01/2025	18:20	LB138072
	Nickel	3.06	+/-20	U	40.0	P	12/01/2025	18:20	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	18:20	LB138072

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vartan Manufacturing Corp.

SDG No.: Q3642

Contract: F&GE01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Zinc	16.7	+/-20	U	40.0	P	12/01/2025	18:20	LB138072
CCB06	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	19:11	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	19:11	LB138072
	Copper	4.60	+/-10	U	20.0	P	12/01/2025	19:11	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	19:11	LB138072
	Molybdenum	18.7	+/-100	U	200	P	12/01/2025	19:11	LB138072
	Nickel	3.06	+/-20	U	40.0	P	12/01/2025	19:11	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	19:11	LB138072
	Zinc	16.7	+/-20	U	40.0	P	12/01/2025	19:11	LB138072
CCB07	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	19:28	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	19:28	LB138072
	Copper	4.60	+/-10	U	20.0	P	12/01/2025	19:28	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	19:28	LB138072
	Molybdenum	18.7	+/-100	U	200	P	12/01/2025	19:28	LB138072
	Nickel	3.06	+/-20	U	40.0	P	12/01/2025	19:28	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	19:28	LB138072
	Zinc	16.7	+/-20	U	40.0	P	12/01/2025	19:28	LB138072

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vartan Manufacturing Corp.

SDG No.: Q3642

Contract: F&GE01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------