

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Vartan Manufacturing Corp.

SDG No.: Q3642

Contract: F&GE01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170769BS							
Cadmium	ug/L	100	96.6		97	85 - 115	P
Chromium	ug/L	200	197		98	85 - 115	P
Copper	ug/L	150	151		101	85 - 115	P
Lead	ug/L	500	473		95	85 - 115	P
Molybdenum	ug/L	200	203		102	85 - 115	P
Nickel	ug/L	250	240		96	85 - 115	P
Silver	ug/L	37.5	38.0		101	85 - 115	P
Zinc	ug/L	100	106		106	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Vartan Manufacturing Corp.

SDG No.: Q3642

Contract: F&GE01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170773BS Mercury	ug/L	4.0	3.81		95	85 - 115	CV