

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN102325

Review By	John Carlone	Review On	10/24/2025 8:16:57 AM
Supervise By	Maresh Dadoda	Supervise On	10/27/2025 2:53:14 PM
SubDirectory	VN102325	HP Acquire Method	HP Processing Method 82N102325W.M

STD. NAME	STD REF.#
Tune/Reschk	VP135909
Initial Calibration Stds	VP135921,VP135922,VP135923,VP135924,VP135925,VP135926
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP135927
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088077.D	23 Oct 2025 09:09		JC\MD	Ok
2	VSTDICC001	VSTDICC001	VN088078.D	23 Oct 2025 09:42	COM.#74 Peak not detected in 1PPB	JC\MD	Ok,M
3	VSTDICC005	VSTDICC005	VN088079.D	23 Oct 2025 10:38		JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN088080.D	23 Oct 2025 10:59		JC\MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VN088081.D	23 Oct 2025 11:20	Method failed for Comp. #74	JC\MD	Ok,M
6	VSTDICC100	VSTDICC100	VN088082.D	23 Oct 2025 11:41		JC\MD	Ok,M
7	VSTDICC150	VSTDICC150	VN088083.D	23 Oct 2025 12:02		JC\MD	Ok,M
8	IBLK	IBLK	VN088084.D	23 Oct 2025 12:23		JC\MD	Ok
9	VSTDICV050	ICVVN102325	VN088085.D	23 Oct 2025 14:54		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN111725

Review By	John Carlone	Review On	11/18/2025 1:22:10 PM
Supervise By	Amit Patel	Supervise On	11/19/2025 10:33:40 AM
SubDirectory	VN111725	HP Acquire Method	HP Processing Method 82N102325W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP136238		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136239,VP136240		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088291.D	17 Nov 2025 08:32		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN088292.D	17 Nov 2025 10:26		JC\MD	Ok,M
3	VN1117MBL01	VN1117MBL01	VN088293.D	17 Nov 2025 11:00		JC\MD	Ok
4	VN1117WBL01	VN1117WBL01	VN088294.D	17 Nov 2025 11:21		JC\MD	Ok
5	VN1117WBS01	VN1117WBS01	VN088295.D	17 Nov 2025 11:41		JC\MD	Ok,M
6	VN1117WBSD01	VN1117WBSD01	VN088296.D	17 Nov 2025 12:16		JC\MD	Ok,M
7	IBLK	IBLK	VN088297.D	17 Nov 2025 12:37		JC\MD	Ok
8	Q3643-01	STORAGE-BLANK-SO	VN088298.D	17 Nov 2025 12:57	vial B pH<2	JC\MD	Ok
9	Q3643-03	STORAGE-BLANK-WA	VN088299.D	17 Nov 2025 13:18	vial B pH<2	JC\MD	Ok
10	PB170545TB	PB170545TB	VN088300.D	17 Nov 2025 13:39		JC\MD	Ok
11	Q3628-04	TP-5	VN088301.D	17 Nov 2025 14:00	vial A pH#5.0	JC\MD	Ok
12	Q3647-02	VNJ-212	VN088302.D	17 Nov 2025 14:21	vial A pH#5.0	JC\MD	Ok
13	VSTDCCC050	VSTDCCC050EC	VN088303.D	17 Nov 2025 15:07		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX110725

Review By	John Carlone	Review On	11/10/2025 8:28:24 AM
Supervise By	Amit Patel	Supervise On	11/10/2025 9:54:56 AM
SubDirectory	VX110725	HP Acquire Method	MSVOA_X HP Processing Method 82X110725W.M

STD. NAME	STD REF.#
Tune/Reschk	VP136104
Initial Calibration Stds	VP136097,VP136098,VP135099,VP136100,VP136101,VP135102
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP136103
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048515.D	07 Nov 2025 13:06		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX048516.D	07 Nov 2025 13:35		JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX048517.D	07 Nov 2025 13:56		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX048518.D	07 Nov 2025 14:16		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX048519.D	07 Nov 2025 14:37	spiked low	JC/MD	Not Ok
6	VSTDIC100	VSTDIC100	VX048520.D	07 Nov 2025 14:58		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX048521.D	07 Nov 2025 15:18		JC/MD	Ok,M
8	VIBLK	VIBLK	VX048522.D	07 Nov 2025 15:39		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX048523.D	07 Nov 2025 16:08	spike error	JC/MD	Not Ok
10	VSTDICCC050	VSTDICCC050	VX048524.D	07 Nov 2025 16:29		JC/MD	Ok,M
11	VSTDICV050	ICVVX110725	VX048525.D	07 Nov 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX111425

Review By	John Carlone	Review On	11/17/2025 8:03:24 AM
Supervise By	Maresh Dadoda	Supervise On	11/17/2025 9:25:03 AM
SubDirectory	VX111425	HP Acquire Method	HP Processing Method 82X110

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136205
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136206,VP136207

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048601.D	14 Nov 2025 08:15		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048602.D	14 Nov 2025 08:47	pH#Lot#V12668	JC/MD	Ok,M
3	VX1114MBL01	VX1114MBL01	VX048603.D	14 Nov 2025 11:19	Surrogate fail	JC/MD	Not Ok
4	VX1114WBL01	VX1114WBL01	VX048604.D	14 Nov 2025 11:39		JC/MD	Ok
5	VX1114WBS01	VX1114WBS01	VX048605.D	14 Nov 2025 13:36		JC/MD	Ok,M
6	VX1114WBSD01	VX1114WBSD01	VX048606.D	14 Nov 2025 13:56		JC/MD	Ok,M
7	IBLK	IBLK	VX048607.D	14 Nov 2025 14:18		JC/MD	Ok
8	Q3643-01	STORAGE-BLANK-SO	VX048608.D	14 Nov 2025 14:39	vial A pH<2 Surrogate fail	JC/MD	ReRun
9	Q3643-02	STORAGE-BLANK-WA	VX048609.D	14 Nov 2025 15:00	vial A pH<2	JC/MD	Ok
10	Q3643-03	STORAGE-BLANK-WA	VX048610.D	14 Nov 2025 15:21	vial A pH<2 Surrogate fail	JC/MD	ReRun
11	Q3643-04	STORAGE-BLANK-SAI	VX048611.D	14 Nov 2025 15:41	vial A pH<2	JC/MD	Ok

M : Manual Integration