

Surrogate Summary

SDG No.: **Q3643**

Client: **Chemtech Consulting Group**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3643-01	STORAGE-BLANK-SOIL-REF-6	1,2-Dichloroethane-d4	50	55.2	110		74	125
		Dibromofluoromethane	50	50.1	100		75	124
		Toluene-d8	50	45.5	91		86	113
		4-Bromofluorobenzene	50	46.5	93		77	121
Q3643-02	STORAGE-BLANK-WATER-REF-5	1,2-Dichloroethane-d4	50	51.9	104		74	125
		Dibromofluoromethane	50	57.4	115		75	124
		Toluene-d8	50	49.8	100		86	113
		4-Bromofluorobenzene	50	57.0	114		77	121
Q3643-03	STORAGE-BLANK-WATER-REF-4	1,2-Dichloroethane-d4	50	55.3	111		74	125
		Dibromofluoromethane	50	49.5	99		75	124
		Toluene-d8	50	45.5	91		86	113
		4-Bromofluorobenzene	50	47.3	95		77	121
Q3643-04	STORAGE-BLANK-SAMPLE-MANAC	1,2-Dichloroethane-d4	50	58.6	117		74	125
		Dibromofluoromethane	50	47.6	95		75	124
		Toluene-d8	50	51.0	102		86	113
		4-Bromofluorobenzene	50	57.2	114		77	121
VN1117WBL01	VN1117WBL01	1,2-Dichloroethane-d4	50	61.1	122		74	125
		Dibromofluoromethane	50	50.3	101		75	124
		Toluene-d8	50	45.7	91		86	113
		4-Bromofluorobenzene	50	48.5	97		77	121
VN1117WBS01	VN1117WBS01	1,2-Dichloroethane-d4	50	54.4	109		74	125
		Dibromofluoromethane	50	51.7	103		75	124
		Toluene-d8	50	48.6	97		86	113
		4-Bromofluorobenzene	50	49.3	99		77	121
VN1117WBSD0	VN1117WBSD01	1,2-Dichloroethane-d4	50	53.5	107		74	125
		Dibromofluoromethane	50	51.2	102		75	124
		Toluene-d8	50	47.9	96		86	113
		4-Bromofluorobenzene	50	47.6	95		77	121
VX1114WBL01	VX1114WBL01	1,2-Dichloroethane-d4	50	58.2	116		74	125
		Dibromofluoromethane	50	43.8	88		75	124
		Toluene-d8	50	49.1	98		86	113
		4-Bromofluorobenzene	50	55.9	112		77	121
VX1114WBS01	VX1114WBS01	1,2-Dichloroethane-d4	50	47.0	94		74	125
		Dibromofluoromethane	50	41.3	83		75	124
		Toluene-d8	50	44.5	89		86	113
		4-Bromofluorobenzene	50	44.6	89		77	121