

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3643
 Lab File ID: VN088292.D Date Analyzed: 11/17/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:26
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	261515	8.21	487635	9.08	426259	11.84
UPPER LIMIT	523030	8.706	975270	9.576	852518	12.341
LOWER LIMIT	130758	7.706	243818	8.576	213130	11.341
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	180499	8.21	408220	9.08	392968	11.85
STORAGE-BLANK-WATER-REF-4	180533	8.21	408977	9.08	396641	11.84
VN1117WBL01	219462	8.21	508968	9.08	502457	11.84
VN1117WBS01	221239	8.21	421404	9.08	388771	11.84
VN1117WBSD01	234156	8.20	431101	9.08	386674	11.84

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.