

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3643
 Lab File ID: VX048602.D Date Analyzed: 11/14/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:47
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	140447	5.53	235074	6.74	205635	10.04
UPPER LIMIT	280894	6.032	470148	7.239	411270	10.537
LOWER LIMIT	70223.5	5.032	117537	6.239	102818	9.537
EPA SAMPLE NO.						
STORAGE-BLANK-WATER-REF-5	170242	5.54	262793	6.75	276084	10.04
STORAGE-BLANK-SAMPLE-MANAGEMENT	152243	5.54	286864	6.75	308578	10.04
VX1114WBL01	149161	5.54	302183	6.75	367753	10.04
VX1114WBS01	140425	5.53	234812	6.75	222283	10.04

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.