

Laboratory Control Sample Summary

Client:	Dal Tile - Quartz Plant	SDG No.:	Q3644
Project:	Quartz Plant - Stormwater Bench Monitoring	Run No.:	LB137913

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137913BS							
TSS	mg/L	550	538		98	1	90-110	11/14/2025