

Laboratory Control Sample Summary

Client: HDR, Inc.

SDG No.: Q3649

Project: PVWC Linear Construction

Run No.: LB137917

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170570BS							
Hexavalent Chromium	mg/Kg	20	19.8		99	1	84-110	11/17/2025

Laboratory Control Sample Summary

Client: HDR, Inc.

SDG No.: Q3649

Project: PVWC Linear Construction

Run No.: LB137941

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170589BS							
Cyanide	mg/Kg	5	4.80		96	1	85-115	11/18/2025