

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV119	Mercury	4.22	4.0	106	90 - 110	CV	11/18/2025	09:54	LB137932

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Mercury	4.90	5.0	98	90 - 110	CV	11/18/2025	10:01	LB137932
CCV08	Mercury	4.69	5.0	94	90 - 110	CV	11/18/2025	10:32	LB137932
CCV09	Mercury	4.93	5.0	98	90 - 110	CV	11/18/2025	11:02	LB137932
CCV10	Mercury	4.65	5.0	93	90 - 110	CV	11/18/2025	11:16	LB137932

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7920	8000	99	90 - 110	P	11/17/2025	13:23	LB137929
	Antimony	4010	4000	100	90 - 110	P	11/17/2025	13:23	LB137929
	Arsenic	3890	4000	97	90 - 110	P	11/17/2025	13:23	LB137929
	Barium	7960	8000	99	90 - 110	P	11/17/2025	13:23	LB137929
	Beryllium	200	200	100	90 - 110	P	11/17/2025	13:23	LB137929
	Cadmium	1940	2000	97	90 - 110	P	11/17/2025	13:23	LB137929
	Calcium	19700	20000	99	90 - 110	P	11/17/2025	13:23	LB137929
	Chromium	814	800	102	90 - 110	P	11/17/2025	13:23	LB137929
	Cobalt	1990	2000	100	90 - 110	P	11/17/2025	13:23	LB137929
	Copper	1020	1000	102	90 - 110	P	11/17/2025	13:23	LB137929
	Iron	4000	4000	100	90 - 110	P	11/17/2025	13:23	LB137929
	Lead	3830	4000	96	90 - 110	P	11/17/2025	13:23	LB137929
	Magnesium	19800	20000	99	90 - 110	P	11/17/2025	13:23	LB137929
	Manganese	2010	2000	101	90 - 110	P	11/17/2025	13:23	LB137929
	Nickel	1990	2000	100	90 - 110	P	11/17/2025	13:23	LB137929
	Potassium	19500	20000	98	90 - 110	P	11/17/2025	13:23	LB137929
	Selenium	3930	4000	98	90 - 110	P	11/17/2025	13:23	LB137929
	Silver	957	1000	96	90 - 110	P	11/17/2025	13:23	LB137929
	Sodium	19800	20000	99	90 - 110	P	11/17/2025	13:23	LB137929
	Thallium	3800	4000	95	90 - 110	P	11/17/2025	13:23	LB137929
	Vanadium	1990	2000	99	90 - 110	P	11/17/2025	13:23	LB137929
	Zinc	2000	2000	100	90 - 110	P	11/17/2025	13:23	LB137929

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10100	10000	101	90 - 110	P	11/17/2025	13:57	LB137929
	Antimony	5040	5000	101	90 - 110	P	11/17/2025	13:57	LB137929
	Arsenic	5010	5000	100	90 - 110	P	11/17/2025	13:57	LB137929
	Barium	10100	10000	101	90 - 110	P	11/17/2025	13:57	LB137929
	Beryllium	251	250	100	90 - 110	P	11/17/2025	13:57	LB137929
	Cadmium	2510	2500	100	90 - 110	P	11/17/2025	13:57	LB137929
	Calcium	25200	25000	101	90 - 110	P	11/17/2025	13:57	LB137929
	Chromium	1030	1000	103	90 - 110	P	11/17/2025	13:57	LB137929
	Cobalt	2510	2500	100	90 - 110	P	11/17/2025	13:57	LB137929
	Copper	1270	1250	102	90 - 110	P	11/17/2025	13:57	LB137929
	Iron	5100	5000	102	90 - 110	P	11/17/2025	13:57	LB137929
	Lead	5020	5000	100	90 - 110	P	11/17/2025	13:57	LB137929
	Magnesium	25100	25000	100	90 - 110	P	11/17/2025	13:57	LB137929
	Manganese	2560	2500	102	90 - 110	P	11/17/2025	13:57	LB137929
	Nickel	2510	2500	100	90 - 110	P	11/17/2025	13:57	LB137929
	Potassium	24800	25000	99	90 - 110	P	11/17/2025	13:57	LB137929
	Selenium	5010	5000	100	90 - 110	P	11/17/2025	13:57	LB137929
	Silver	1260	1250	101	90 - 110	P	11/17/2025	13:57	LB137929
	Sodium	25000	25000	100	90 - 110	P	11/17/2025	13:57	LB137929
	Thallium	4900	5000	98	90 - 110	P	11/17/2025	13:57	LB137929
	Vanadium	2510	2500	101	90 - 110	P	11/17/2025	13:57	LB137929
	Zinc	2540	2500	102	90 - 110	P	11/17/2025	13:57	LB137929
CCV02	Aluminum	10200	10000	102	90 - 110	P	11/17/2025	14:47	LB137929
	Antimony	5140	5000	103	90 - 110	P	11/17/2025	14:47	LB137929
	Arsenic	5120	5000	102	90 - 110	P	11/17/2025	14:47	LB137929
	Barium	10100	10000	101	90 - 110	P	11/17/2025	14:47	LB137929
	Beryllium	253	250	101	90 - 110	P	11/17/2025	14:47	LB137929
	Cadmium	2540	2500	102	90 - 110	P	11/17/2025	14:47	LB137929
	Calcium	25200	25000	101	90 - 110	P	11/17/2025	14:47	LB137929
	Chromium	1030	1000	103	90 - 110	P	11/17/2025	14:47	LB137929
	Cobalt	2540	2500	101	90 - 110	P	11/17/2025	14:47	LB137929
	Copper	1290	1250	103	90 - 110	P	11/17/2025	14:47	LB137929
	Iron	5180	5000	104	90 - 110	P	11/17/2025	14:47	LB137929
	Lead	5080	5000	102	90 - 110	P	11/17/2025	14:47	LB137929

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	25100	25000	100	90 - 110	P	11/17/2025	14:47	LB137929
	Manganese	2560	2500	102	90 - 110	P	11/17/2025	14:47	LB137929
	Nickel	2540	2500	102	90 - 110	P	11/17/2025	14:47	LB137929
	Potassium	25400	25000	102	90 - 110	P	11/17/2025	14:47	LB137929
	Selenium	5170	5000	103	90 - 110	P	11/17/2025	14:47	LB137929
	Silver	1280	1250	102	90 - 110	P	11/17/2025	14:47	LB137929
	Sodium	25600	25000	102	90 - 110	P	11/17/2025	14:47	LB137929
	Thallium	4970	5000	99	90 - 110	P	11/17/2025	14:47	LB137929
	Vanadium	2520	2500	101	90 - 110	P	11/17/2025	14:47	LB137929
	Zinc	2580	2500	103	90 - 110	P	11/17/2025	14:47	LB137929
CCV03	Aluminum	9640	10000	96	90 - 110	P	11/17/2025	15:44	LB137929
	Antimony	4840	5000	97	90 - 110	P	11/17/2025	15:44	LB137929
	Arsenic	4790	5000	96	90 - 110	P	11/17/2025	15:44	LB137929
	Barium	9720	10000	97	90 - 110	P	11/17/2025	15:44	LB137929
	Beryllium	242	250	97	90 - 110	P	11/17/2025	15:44	LB137929
	Cadmium	2370	2500	95	90 - 110	P	11/17/2025	15:44	LB137929
	Calcium	24000	25000	96	90 - 110	P	11/17/2025	15:44	LB137929
	Chromium	948	1000	95	90 - 110	P	11/17/2025	15:44	LB137929
	Cobalt	2390	2500	96	90 - 110	P	11/17/2025	15:44	LB137929
	Copper	1210	1250	97	90 - 110	P	11/17/2025	15:44	LB137929
	Iron	4740	5000	95	90 - 110	P	11/17/2025	15:44	LB137929
	Lead	4750	5000	95	90 - 110	P	11/17/2025	15:44	LB137929
	Magnesium	23800	25000	95	90 - 110	P	11/17/2025	15:44	LB137929
	Manganese	2400	2500	96	90 - 110	P	11/17/2025	15:44	LB137929
	Nickel	2380	2500	95	90 - 110	P	11/17/2025	15:44	LB137929
	Potassium	23200	25000	93	90 - 110	P	11/17/2025	15:44	LB137929
	Selenium	4800	5000	96	90 - 110	P	11/17/2025	15:44	LB137929
	Silver	1190	1250	95	90 - 110	P	11/17/2025	15:44	LB137929
	Sodium	23200	25000	93	90 - 110	P	11/17/2025	15:44	LB137929
	Thallium	4520	5000	90	90 - 110	P	11/17/2025	15:44	LB137929
	Vanadium	2400	2500	96	90 - 110	P	11/17/2025	15:44	LB137929
	Zinc	2390	2500	95	90 - 110	P	11/17/2025	15:44	LB137929
CCV04	Aluminum	10200	10000	102	90 - 110	P	11/17/2025	16:34	LB137929
	Antimony	5160	5000	103	90 - 110	P	11/17/2025	16:34	LB137929

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5130	5000	103	90 - 110	P	11/17/2025	16:34	LB137929
	Barium	10000	10000	100	90 - 110	P	11/17/2025	16:34	LB137929
	Beryllium	254	250	102	90 - 110	P	11/17/2025	16:34	LB137929
	Cadmium	2510	2500	101	90 - 110	P	11/17/2025	16:34	LB137929
	Calcium	25200	25000	101	90 - 110	P	11/17/2025	16:34	LB137929
	Chromium	1020	1000	102	90 - 110	P	11/17/2025	16:34	LB137929
	Cobalt	2530	2500	101	90 - 110	P	11/17/2025	16:34	LB137929
	Copper	1290	1250	103	90 - 110	P	11/17/2025	16:34	LB137929
	Iron	5130	5000	103	90 - 110	P	11/17/2025	16:34	LB137929
	Lead	5030	5000	101	90 - 110	P	11/17/2025	16:34	LB137929
	Magnesium	25200	25000	101	90 - 110	P	11/17/2025	16:34	LB137929
	Manganese	2560	2500	102	90 - 110	P	11/17/2025	16:34	LB137929
	Nickel	2520	2500	101	90 - 110	P	11/17/2025	16:34	LB137929
	Potassium	25200	25000	101	90 - 110	P	11/17/2025	16:34	LB137929
	Selenium	5160	5000	103	90 - 110	P	11/17/2025	16:34	LB137929
	Silver	1280	1250	103	90 - 110	P	11/17/2025	16:34	LB137929
	Sodium	25600	25000	102	90 - 110	P	11/17/2025	16:34	LB137929
	Thallium	5180	5000	104	90 - 110	P	11/17/2025	16:34	LB137929
	Vanadium	2530	2500	101	90 - 110	P	11/17/2025	16:34	LB137929
	Zinc	2610	2500	104	90 - 110	P	11/17/2025	16:34	LB137929
CCV05	Aluminum	10200	10000	102	90 - 110	P	11/17/2025	17:24	LB137929
	Antimony	5150	5000	103	90 - 110	P	11/17/2025	17:24	LB137929
	Arsenic	5130	5000	102	90 - 110	P	11/17/2025	17:24	LB137929
	Barium	9840	10000	98	90 - 110	P	11/17/2025	17:24	LB137929
	Beryllium	250	250	100	90 - 110	P	11/17/2025	17:24	LB137929
	Cadmium	2530	2500	101	90 - 110	P	11/17/2025	17:24	LB137929
	Calcium	25000	25000	100	90 - 110	P	11/17/2025	17:24	LB137929
	Chromium	1030	1000	103	90 - 110	P	11/17/2025	17:24	LB137929
	Cobalt	2540	2500	102	90 - 110	P	11/17/2025	17:24	LB137929
	Copper	1290	1250	103	90 - 110	P	11/17/2025	17:24	LB137929
	Iron	4960	5000	99	90 - 110	P	11/17/2025	17:24	LB137929
	Lead	5020	5000	100	90 - 110	P	11/17/2025	17:24	LB137929
	Magnesium	25000	25000	100	90 - 110	P	11/17/2025	17:24	LB137929
	Manganese	2580	2500	103	90 - 110	P	11/17/2025	17:24	LB137929

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2530	2500	101	90 - 110	P	11/17/2025	17:24	LB137929
	Potassium	24300	25000	97	90 - 110	P	11/17/2025	17:24	LB137929
	Selenium	5110	5000	102	90 - 110	P	11/17/2025	17:24	LB137929
	Silver	1260	1250	101	90 - 110	P	11/17/2025	17:24	LB137929
	Sodium	24600	25000	99	90 - 110	P	11/17/2025	17:24	LB137929
	Thallium	5340	5000	107	90 - 110	P	11/17/2025	17:24	LB137929
	Vanadium	2520	2500	101	90 - 110	P	11/17/2025	17:24	LB137929
	Zinc	2580	2500	103	90 - 110	P	11/17/2025	17:24	LB137929
CCV06	Aluminum	9870	10000	99	90 - 110	P	11/17/2025	18:15	LB137929
	Antimony	5040	5000	101	90 - 110	P	11/17/2025	18:15	LB137929
	Arsenic	5010	5000	100	90 - 110	P	11/17/2025	18:15	LB137929
	Barium	9690	10000	97	90 - 110	P	11/17/2025	18:15	LB137929
	Beryllium	244	250	98	90 - 110	P	11/17/2025	18:15	LB137929
	Cadmium	2450	2500	98	90 - 110	P	11/17/2025	18:15	LB137929
	Calcium	24400	25000	98	90 - 110	P	11/17/2025	18:15	LB137929
	Chromium	1000	1000	100	90 - 110	P	11/17/2025	18:15	LB137929
	Cobalt	2470	2500	99	90 - 110	P	11/17/2025	18:15	LB137929
	Copper	1260	1250	101	90 - 110	P	11/17/2025	18:15	LB137929
	Iron	4940	5000	99	90 - 110	P	11/17/2025	18:15	LB137929
	Lead	4890	5000	98	90 - 110	P	11/17/2025	18:15	LB137929
	Magnesium	24300	25000	97	90 - 110	P	11/17/2025	18:15	LB137929
	Manganese	2500	2500	100	90 - 110	P	11/17/2025	18:15	LB137929
	Nickel	2460	2500	98	90 - 110	P	11/17/2025	18:15	LB137929
	Potassium	24100	25000	97	90 - 110	P	11/17/2025	18:15	LB137929
	Selenium	5020	5000	100	90 - 110	P	11/17/2025	18:15	LB137929
	Silver	1250	1250	100	90 - 110	P	11/17/2025	18:15	LB137929
	Sodium	24500	25000	98	90 - 110	P	11/17/2025	18:15	LB137929
	Thallium	4960	5000	99	90 - 110	P	11/17/2025	18:15	LB137929
	Vanadium	2450	2500	98	90 - 110	P	11/17/2025	18:15	LB137929
	Zinc	2570	2500	103	90 - 110	P	11/17/2025	18:15	LB137929
CCV07	Aluminum	9850	10000	98	90 - 110	P	11/17/2025	19:07	LB137929
	Antimony	4980	5000	100	90 - 110	P	11/17/2025	19:07	LB137929
	Arsenic	4980	5000	100	90 - 110	P	11/17/2025	19:07	LB137929
	Barium	9940	10000	99	90 - 110	P	11/17/2025	19:07	LB137929

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	246	250	99	90 - 110	P	11/17/2025	19:07	LB137929
	Cadmium	2410	2500	96	90 - 110	P	11/17/2025	19:07	LB137929
	Calcium	24400	25000	98	90 - 110	P	11/17/2025	19:07	LB137929
	Chromium	987	1000	99	90 - 110	P	11/17/2025	19:07	LB137929
	Cobalt	2430	2500	97	90 - 110	P	11/17/2025	19:07	LB137929
	Copper	1260	1250	100	90 - 110	P	11/17/2025	19:07	LB137929
	Iron	4960	5000	99	90 - 110	P	11/17/2025	19:07	LB137929
	Lead	4790	5000	96	90 - 110	P	11/17/2025	19:07	LB137929
	Magnesium	24100	25000	97	90 - 110	P	11/17/2025	19:07	LB137929
	Manganese	2450	2500	98	90 - 110	P	11/17/2025	19:07	LB137929
	Nickel	2420	2500	97	90 - 110	P	11/17/2025	19:07	LB137929
	Potassium	24500	25000	98	90 - 110	P	11/17/2025	19:07	LB137929
	Selenium	4970	5000	99	90 - 110	P	11/17/2025	19:07	LB137929
	Silver	1250	1250	100	90 - 110	P	11/17/2025	19:07	LB137929
	Sodium	24600	25000	98	90 - 110	P	11/17/2025	19:07	LB137929
	Thallium	4840	5000	97	90 - 110	P	11/17/2025	19:07	LB137929
	Vanadium	2440	2500	98	90 - 110	P	11/17/2025	19:07	LB137929
	Zinc	2580	2500	103	90 - 110	P	11/17/2025	19:07	LB137929
CCV08	Aluminum	10300	10000	102	90 - 110	P	11/17/2025	19:24	LB137929
	Antimony	5180	5000	104	90 - 110	P	11/17/2025	19:24	LB137929
	Arsenic	5170	5000	104	90 - 110	P	11/17/2025	19:24	LB137929
	Barium	10200	10000	102	90 - 110	P	11/17/2025	19:24	LB137929
	Beryllium	254	250	102	90 - 110	P	11/17/2025	19:24	LB137929
	Cadmium	2510	2500	100	90 - 110	P	11/17/2025	19:24	LB137929
	Calcium	25400	25000	102	90 - 110	P	11/17/2025	19:24	LB137929
	Chromium	1030	1000	103	90 - 110	P	11/17/2025	19:24	LB137929
	Cobalt	2530	2500	101	90 - 110	P	11/17/2025	19:24	LB137929
	Copper	1300	1250	104	90 - 110	P	11/17/2025	19:24	LB137929
	Iron	5130	5000	103	90 - 110	P	11/17/2025	19:24	LB137929
	Lead	4990	5000	100	90 - 110	P	11/17/2025	19:24	LB137929
	Magnesium	25300	25000	101	90 - 110	P	11/17/2025	19:24	LB137929
	Manganese	2570	2500	103	90 - 110	P	11/17/2025	19:24	LB137929
	Nickel	2520	2500	101	90 - 110	P	11/17/2025	19:24	LB137929
	Potassium	25300	25000	101	90 - 110	P	11/17/2025	19:24	LB137929

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5180	5000	104	90 - 110	P	11/17/2025	19:24	LB137929
	Silver	1290	1250	103	90 - 110	P	11/17/2025	19:24	LB137929
	Sodium	25600	25000	102	90 - 110	P	11/17/2025	19:24	LB137929
	Thallium	5140	5000	103	90 - 110	P	11/17/2025	19:24	LB137929
	Vanadium	2550	2500	102	90 - 110	P	11/17/2025	19:24	LB137929
	Zinc	2660	2500	106	90 - 110	P	11/17/2025	19:24	LB137929
CCV09	Aluminum	9890	10000	99	90 - 110	P	11/17/2025	19:47	LB137929
	Antimony	5200	5000	104	90 - 110	P	11/17/2025	19:47	LB137929
	Arsenic	5220	5000	104	90 - 110	P	11/17/2025	19:47	LB137929
	Barium	9950	10000	100	90 - 110	P	11/17/2025	19:47	LB137929
	Beryllium	243	250	97	90 - 110	P	11/17/2025	19:47	LB137929
	Cadmium	2430	2500	97	90 - 110	P	11/17/2025	19:47	LB137929
	Calcium	24700	25000	99	90 - 110	P	11/17/2025	19:47	LB137929
	Chromium	990	1000	99	90 - 110	P	11/17/2025	19:47	LB137929
	Cobalt	2450	2500	98	90 - 110	P	11/17/2025	19:47	LB137929
	Copper	1280	1250	103	90 - 110	P	11/17/2025	19:47	LB137929
	Iron	5060	5000	101	90 - 110	P	11/17/2025	19:47	LB137929
	Lead	4900	5000	98	90 - 110	P	11/17/2025	19:47	LB137929
	Magnesium	24200	25000	97	90 - 110	P	11/17/2025	19:47	LB137929
	Manganese	2650	2500	106	90 - 110	P	11/17/2025	19:47	LB137929
	Nickel	2440	2500	98	90 - 110	P	11/17/2025	19:47	LB137929
	Potassium	24800	25000	99	90 - 110	P	11/17/2025	19:47	LB137929
	Selenium	5220	5000	104	90 - 110	P	11/17/2025	19:47	LB137929
	Silver	1250	1250	100	90 - 110	P	11/17/2025	19:47	LB137929
	Sodium	24800	25000	99	90 - 110	P	11/17/2025	19:47	LB137929
	Thallium	5020	5000	100	90 - 110	P	11/17/2025	19:47	LB137929
	Vanadium	2470	2500	99	90 - 110	P	11/17/2025	19:47	LB137929
	Zinc	2620	2500	105	90 - 110	P	11/17/2025	19:47	LB137929

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLCCV	Aluminum	120	100	120	80 - 120	P	11/17/2025	19:57	LB137929
	Antimony	55.6	50.0	111	80 - 120	P	11/17/2025	19:57	LB137929
	Arsenic	20.5	20.0	103	80 - 120	P	11/17/2025	19:57	LB137929
	Barium	102	100	102	80 - 120	P	11/17/2025	19:57	LB137929
	Beryllium	6.51	6.0	109	80 - 120	P	11/17/2025	19:57	LB137929
	Cadmium	6.08	6.0	101	80 - 120	P	11/17/2025	19:57	LB137929
	Calcium	2090	2000	104	80 - 120	P	11/17/2025	19:57	LB137929
	Chromium	10.4	10.0	104	80 - 120	P	11/17/2025	19:57	LB137929
	Cobalt	30.7	30.0	102	80 - 120	P	11/17/2025	19:57	LB137929
	Copper	22.6	20.0	113	80 - 120	P	11/17/2025	19:57	LB137929
	Iron	117	100	117	80 - 120	P	11/17/2025	19:57	LB137929
	Lead	14.3	12.0	119	80 - 120	P	11/17/2025	19:57	LB137929
	Magnesium	2190	2000	110	80 - 120	P	11/17/2025	19:57	LB137929
	Manganese	23.0	20.0	115	80 - 120	P	11/17/2025	19:57	LB137929
	Nickel	42.5	40.0	106	80 - 120	P	11/17/2025	19:57	LB137929
	Potassium	2060	2000	103	80 - 120	P	11/17/2025	19:57	LB137929
	Selenium	19.2	20.0	96	80 - 120	P	11/17/2025	19:57	LB137929
	Silver	11.9	10.0	119	80 - 120	P	11/17/2025	19:57	LB137929
	Sodium	1890	2000	94	80 - 120	P	11/17/2025	19:57	LB137929
	Thallium	39.0	40.0	97	80 - 120	P	11/17/2025	19:57	LB137929
	Vanadium	42.1	40.0	105	80 - 120	P	11/17/2025	19:57	LB137929
	Zinc	39.3	40.0	98	80 - 120	P	11/17/2025	19:57	LB137929