



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	20.5	20.0	102	65 - 135	P	11/17/2025	13:35	LB137929
	Barium	104	100	104	65 - 135	P	11/17/2025	13:35	LB137929
	Cadmium	6.08	6.0	101	65 - 135	P	11/17/2025	13:35	LB137929
	Chromium	10.4	10.0	104	65 - 135	P	11/17/2025	13:35	LB137929
	Lead	12.4	12.0	103	65 - 135	P	11/17/2025	13:35	LB137929
	Selenium	22.9	20.0	115	65 - 135	P	11/17/2025	13:35	LB137929
	Silver	10.6	10.0	106	65 - 135	P	11/17/2025	13:35	LB137929
CRA	Mercury	0.16	0.2	82	70 - 130	CV	11/17/2025	15:26	LB137920
CRI01	Arsenic	20.8	20.0	104	65 - 135	P	11/20/2025	13:26	LB137997
	Barium	98.7	100	99	65 - 135	P	11/20/2025	13:26	LB137997
	Cadmium	5.95	6.0	99	65 - 135	P	11/20/2025	13:26	LB137997
	Chromium	10.1	10.0	101	65 - 135	P	11/20/2025	13:26	LB137997
	Lead	11.7	12.0	98	65 - 135	P	11/20/2025	13:26	LB137997
	Selenium	21.8	20.0	109	65 - 135	P	11/20/2025	13:26	LB137997
	Silver	11.0	10.0	110	65 - 135	P	11/20/2025	13:26	LB137997