

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170583BS							
Arsenic	ug/L	4000	3900		98	80 - 120	P
Barium	ug/L	1000	1010		101	80 - 120	P
Cadmium	ug/L	1000	961		96	80 - 120	P
Chromium	ug/L	2000	2000		100	80 - 120	P
Lead	ug/L	5000	4680		94	80 - 120	P
Selenium	ug/L	10000	9790		98	80 - 120	P
Silver	ug/L	380	381		100	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: HDR, Inc.

SDG No.: Q3649

Contract: HDRI08

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170585BS Mercury	ug/L	4.0	3.76		94	80 - 120	CV