

Continuing Calibration Report for SequenceID : FE111725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE056867.D

Aliphatic C9-C12	10349661.000	60.000	3.213	6.836	172494.350	171162.641	-0.778
Aliphatic C12-C16	7040992.000	40.000	6.837	10.281	176024.800	178754.319	1.527
Aliphatic C16-C21	10352194.000	60.000	10.282	13.655	172536.567	186505.496	7.490
Aliphatic C21-C28	12779054.000	80.000	13.656	17.323	159738.175	192766.144	17.134
Aliphatic C28-C40	17024804.000	120.000	17.324	22.257	141873.367	159314.647	10.948
Aliphatic EPH	57546705.000	360.000	3.213	22.257	159851.958	175414.751	8.872