

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Continuing Calib Date: 11/17/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 15:46

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.49	5.49	5.39	5.59	0.00
Aroclor-1242-2 (2)	5.51	5.51	5.41	5.61	0.00
Aroclor-1242-3 (3)	5.57	5.57	5.47	5.67	0.00
Aroclor-1242-4 (4)	5.67	5.67	5.57	5.77	0.01
Aroclor-1242-5 (5)	6.39	6.39	6.29	6.49	0.00
Tetrachloro-m-xylene	4.35	4.35	4.25	4.45	0.00
Decachlorobiphenyl	9.93	9.93	9.83	10.03	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Continuing Calib Date: 11/17/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 15:46

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.72	4.73	4.63	4.83	0.01
Aroclor-1242-2 (2)	4.74	4.75	4.65	4.85	0.01
Aroclor-1242-3 (3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1242-4 (4)	5.00	5.01	4.91	5.11	0.01
Aroclor-1242-5 (5)	5.52	5.53	5.43	5.63	0.01
Tetrachloro-m-xylene	3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl	8.64	8.64	8.54	8.74	0.00