

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3649
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115139.D **Time Analyzed:** 15:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.485	5.387	5.587	505.110	500.000	1.0
Aroclor-1016-2	5.507	5.408	5.608	500.960	500.000	0.2
Aroclor-1016-3	5.568	5.470	5.670	523.560	500.000	4.7
Aroclor-1016-4	5.665	5.566	5.766	509.240	500.000	1.8
Aroclor-1016-5	5.956	5.858	6.058	482.070	500.000	-3.6
Aroclor-1260-1	7.070	6.971	7.171	474.310	500.000	-5.1
Aroclor-1260-2	7.323	7.225	7.425	509.100	500.000	1.8
Aroclor-1260-3	7.681	7.582	7.782	514.220	500.000	2.8
Aroclor-1260-4	7.903	7.803	8.003	490.890	500.000	-1.8
Aroclor-1260-5	8.210	8.110	8.310	487.200	500.000	-2.6
Decachlorobiphenyl	9.930	9.830	10.030	53.660	50.000	7.3
Tetrachloro-m-xylene	4.346	4.248	4.448	54.000	50.000	8.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: HDRI08
Lab Code: ACE
SDG NO.: Q3649
GC Column: ZB-MR2
ID: 0.32 (mm)
 Initi. Calib. Date(s): 11/04/2025 11/04/2025

Client Sample No.: CCAL01
Date Analyzed: 11/17/2025
Lab Sample No.: AR1660CCC500
Data File : PO115139.D
Time Analyzed: 15:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.724	4.628	4.828	507.050	500.000	1.4
Aroclor-1016-2	4.744	4.647	4.847	502.940	500.000	0.6
Aroclor-1016-3	4.918	4.821	5.021	519.670	500.000	3.9
Aroclor-1016-4	4.960	4.864	5.064	518.090	500.000	3.6
Aroclor-1016-5	5.172	5.076	5.276	514.460	500.000	2.9
Aroclor-1260-1	6.203	6.106	6.306	474.520	500.000	-5.1
Aroclor-1260-2	6.390	6.294	6.494	457.240	500.000	-8.6
Aroclor-1260-3	6.543	6.446	6.646	470.100	500.000	-6.0
Aroclor-1260-4	7.014	6.917	7.117	476.400	500.000	-4.7
Aroclor-1260-5	7.255	7.158	7.358	473.150	500.000	-5.4
Decachlorobiphenyl	8.639	8.541	8.741	52.460	50.000	4.9
Tetrachloro-m-xylene	3.647	3.548	3.748	52.080	50.000	4.2